



NSAI
Standards

Irish Standard
I.S. EN 16450:2017

Ambient air - Automated measuring systems for the measurement of the concentration of particulate matter (PM₁₀; PM_{2,5})

I.S. EN 16450:2017

Incorporating amendments/corrigenda/National Annexes issued since publication:

The National Standards Authority of Ireland (NSAI) produces the following categories of formal documents:

I.S. xxx: Irish Standard — national specification based on the consensus of an expert panel and subject to public consultation.

S.R. xxx: Standard Recommendation — recommendation based on the consensus of an expert panel and subject to public consultation.

SWiFT xxx: A rapidly developed recommendatory document based on the consensus of the participants of an NSAI workshop.

This document replaces/revises/consolidates the NSAI adoption of the document(s) indicated on the CEN/CENELEC cover/Foreword and the following National document(s):

NOTE: The date of any NSAI previous adoption may not match the date of its original CEN/CENELEC document.

This document is based on:

EN 16450:2017

Published:

2017-03-15

This document was published under the authority of the NSAI and comes into effect on:

2017-04-02

ICS number:

13.040.20

NOTE: If blank see CEN/CENELEC cover page

NSAI
1 Swift Square,
Northwood, Santry
Dublin 9

T +353 1 807 3800
F +353 1 807 3838
E standards@nsai.ie
W NSAI.ie

Sales:
T +353 1 857 6730
F +353 1 857 6729
W standards.ie

Údarás um Chaighdeáin Náisiúnta na hÉireann

National Foreword

I.S. EN 16450:2017 is the adopted Irish version of the European Document EN 16450:2017, Ambient air - Automated measuring systems for the measurement of the concentration of particulate matter (PM10; PM2,5)

This document does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

For relationships with other publications refer to the NSAI web store.

Compliance with this document does not of itself confer immunity from legal obligations.

In line with international standards practice the decimal point is shown as a comma (,) throughout this document.

This page is intentionally left blank

EUROPEAN STANDARD

EN 16450

NORME EUROPÉENNE

EUROPÄISCHE NORM

March 2017

ICS 13.040.20

Supersedes CEN/TS 16450:2013

English Version

**Ambient air - Automated measuring systems for the
measurement of the concentration of particulate matter
(PM10; PM2,5)**

Air ambiant - Systèmes automatisés de mesurage de la
concentration de matière particulaire (PM10; PM2,5)

Außenluft - Automatische Messeinrichtungen zur
Bestimmung der Staubkonzentration (PM10; PM2,5)

This European Standard was approved by CEN on 16 January 2017.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

EN 16450:2017 (E)

Contents		Page
European foreword		3
1	Scope	4
2	Normative references	5
3	Terms and definitions	5
4	Symbols and abbreviated terms	9
5	Principle	10
6	Sampling	12
7	Type testing	13
8	Field operation and ongoing quality control	29
9	Data handling, validation and data reports	38
10	Reporting of AMS data	39
11	Test reports and documentation – Field operation	39
Annex A (informative) Examples of principles of AMS for monitoring particulate matter		40
A.1	Oscillating microbalance	40
A.2	β-ray attenuation	41
A.3	Light scattering	42
A.4	System consisting of a central instrument and an array of regional instruments	42
Annex B (normative) Orthogonal regression algorithms		43
Annex C (normative) Performing calibrations of the AMS		45
Annex D (normative) Elements of type testing report		47
Annex E (informative) Elements of suitability evaluation report		49
Bibliography		50

European foreword

This document (EN 16450:2017) has been prepared by Technical Committee CEN/TC 264 "Air quality", the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by September 2017, and conflicting national standards shall be withdrawn at the latest by September 2017.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes CEN/TS 16450:2013.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association.

According to the CEN-CENELEC Internal Regulations, the national standards organisations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

EN 16450:2017 (E)

1 Scope

In order to be in compliance with EU Air Quality Directive requirements, the reference methods given in the Directive 2008/50/EC [1] for the measurement of mass concentrations of particulate matter are not commonly used for operation in routine monitoring networks. These networks usually apply automated continuous measurement systems (AMS), such as those based on the use of oscillating microbalances, β -ray attenuation, or *in situ* optical methods. Such AMS are typically capable of producing 24-h average measurement values over a measurement range up to 1 000 $\mu\text{g}/\text{m}^3$ and 1-h average measurement values up to 10 000 $\mu\text{g}/\text{m}^3$, if applicable, where the volume of air is the volume at ambient conditions near the inlet.

The 1-h average values may be used for:

- a) direct information of the public;
- b) aggregation to produce daily or yearly average concentration values for regulatory reporting purposes.

NOTE National regulatory reporting purposes could require other time basis for averages (i.e. monthly).

Directive 2008/50/EC allows the use of such systems after demonstration of equivalence with the reference method, i.e. after demonstration that these systems meet the Data Quality Objectives for continuous measurements. Guidelines for the demonstration of equivalence are given in Reference [2].

This European Standard lays down the minimum performance requirements and test procedures for the type testing of appropriate AMS for particulate matter. The standard includes the evaluation of its equivalence with the reference method as laid down in Directive 2008/50/EC.

Further, this European Standard describes minimum requirements for ongoing quality assurance – quality control (QA/QC) of AMS deployed in the field. These requirements are necessary to ensure that uncertainties of measured concentrations are kept within the required limits during extended periods of continuous monitoring in the field, and include procedures for maintenance, calibration and control checks.

Additional procedures are described that determine whether an instrument's equivalence to the reference method is maintained through possible pollution climate changes, over periods longer than five years.

Lastly, this European Standard describes harmonized requirements and procedures for the treatment and validation of raw measurement data that are used for the assembly of daily or yearly average concentration values. Experience with existing methods for data treatment and validation – for similar AMS – has shown that the different ways of data treatment and validation applied may lead to significant differences in reported results for similar data sets [3].

When the European Standard is used for purposes other than measurements required by Directive 2008/50/EC, the range and uncertainty requirements may not apply.

This European Standard contains information for different groups of users.

Clauses 5 and 6 and Annex A contain general information about the principles of automated continuous measurement systems for particulate matter, and relevant equipment.

Clause 7 and Annexes B and C are specifically directed towards test houses and laboratories that perform type testing of automated continuous measurement systems for particulate matter. These clauses contain information about:

- c) type testing conditions, test procedures and test requirements;
- d) system performance requirements;
- e) evaluation of the type testing results;
- f) evaluation of the uncertainty of the measurement results of the automated continuous measurement systems for particulate matter based on the type testing results.

Clauses 8 to 11 are aimed at monitoring networks performing the practical measurements of particulate matter in ambient air. These clauses contain information about:

- g) initial installation of the system in the monitoring network and acceptance testing;
- h) ongoing quality assurance/quality control;
- i) on-going verification of suitability;
- j) treatment, validation and reporting of measurement results.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 12341, *Ambient air - Standard gravimetric measurement method for the determination of the PM10 or PM2,5 mass concentration of suspended particulate matter*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1

ambient air

outdoor air in the troposphere, excluding workplaces as defined by Directive 89/654/EEC [4] where provisions concerning health and safety at work apply and to which members of the public do not have regular access

[SOURCE: Directive 2008/50/EC [1]]

3.2

automated measuring system

AMS

entirety of all measuring instruments and additional devices necessary for obtaining a measurement result

3.3

availability of the AMS

fraction of the time period for which valid measuring data of the ambient air concentration is available from an AMS

[SOURCE: EN 14211 [5]]

This is a free preview. Purchase the entire publication at the link below:

[Product Page](#)

-
- [Looking for additional Standards? Visit Intertek Inform Infostore](#)
 - [Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation](#)
-