



NSAI
Standards

Irish Standard
I.S. EN ISO 19238:2017

Radiological protection - Performance criteria for service laboratories performing biological dosimetry by cytogenetics (ISO 19238:2014)

I.S. EN ISO 19238:2017

Incorporating amendments/corrigenda/National Annexes issued since publication:

The National Standards Authority of Ireland (NSAI) produces the following categories of formal documents:

I.S. xxx: Irish Standard — national specification based on the consensus of an expert panel and subject to public consultation.

S.R. xxx: Standard Recommendation — recommendation based on the consensus of an expert panel and subject to public consultation.

SWiFT xxx: A rapidly developed recommendatory document based on the consensus of the participants of an NSAI workshop.

This document replaces/revises/consolidates the NSAI adoption of the document(s) indicated on the CEN/CENELEC cover/Foreword and the following National document(s):

NOTE: The date of any NSAI previous adoption may not match the date of its original CEN/CENELEC document.

This document is based on:

EN ISO 19238:2017

Published:

2017-10-18

This document was published under the authority of the NSAI and comes into effect on:

2017-11-05

ICS number:

13.280

17.240

NOTE: If blank see CEN/CENELEC cover page

NSAI
1 Swift Square,
Northwood, Santry
Dublin 9

T +353 1 807 3800
F +353 1 807 3838
E standards@nsai.ie
W NSAI.ie

Sales:
T +353 1 857 6730
F +353 1 857 6729
W standards.ie

Údarás um Chaighdeáin Náisiúnta na hÉireann

National Foreword

I.S. EN ISO 19238:2017 is the adopted Irish version of the European Document EN ISO 19238:2017, Radiological protection - Performance criteria for service laboratories performing biological dosimetry by cytogenetics (ISO 19238:2014)

This document does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

For relationships with other publications refer to the NSAI web store.

Compliance with this document does not of itself confer immunity from legal obligations.

In line with international standards practice the decimal point is shown as a comma (,) throughout this document.

This page is intentionally left blank

EUROPEAN STANDARD

EN ISO 19238

NORME EUROPÉENNE

EUROPÄISCHE NORM

October 2017

ICS 13.280; 17.240

English Version

Radiological protection - Performance criteria for service laboratories performing biological dosimetry by cytogenetics (ISO 19238:2014)

Radioprotection - Critères de performance pour les laboratoires de service pratiquant la dosimétrie biologique par cytogénétique (ISO 19238:2014)

This European Standard was approved by CEN on 13 September 2017.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents	Page
European foreword.....	3

European foreword

The text of ISO 19238:2014 has been prepared by Technical Committee ISO/TC 85 “Nuclear energy, nuclear technologies, and radiological protection” of the International Organization for Standardization (ISO) and has been taken over as EN ISO 19238:2017 by Technical Committee CEN/TC 430 “Nuclear energy, nuclear technologies, and radiological protection” the secretariat of which is held by AFNOR.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by April 2018, and conflicting national standards shall be withdrawn at the latest by April 2018.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 19238:2014 has been approved by CEN as EN ISO 19238:2017 without any modification.

This page is intentionally left blank

INTERNATIONAL STANDARD

ISO
19238

Second edition
2014-02-01

Radiological protection — Performance criteria for service laboratories performing biological dosimetry by cytogenetics

*Radioprotection — Critères de performance pour les laboratoires de
service pratiquant la dosimétrie biologique par cytogénétique*



Reference number
ISO 19238:2014(E)

© ISO 2014

ISO 19238:2014(E)



COPYRIGHT PROTECTED DOCUMENT

© ISO 2014

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

	Page
Foreword	v
Introduction	vi
1 Scope	1
2 Terms and definitions	1
3 Dicentric assay	3
4 Responsibility of the customer	3
5 Responsibility of the service laboratory	4
5.1 Setup and sustainment of the QA program.....	4
5.2 Responsibility during service.....	4
6 Confidentiality of personal information	5
6.1 Overview.....	5
6.2 Applications of the principle of confidentiality.....	5
7 Laboratory safety requirements	6
7.1 Overview.....	6
7.2 Microbiological safety requirements.....	6
7.3 Chemical safety.....	6
7.4 Optical safety requirements.....	8
7.5 Safety plan.....	8
8 Calibration curve(s)	9
8.1 Culturing.....	9
8.2 Calibration source(s).....	10
8.3 Establishment of calibration curve(s).....	10
8.4 Minimum resolvable dose measurement.....	11
9 Scoring unstable chromosome aberrations	11
9.1 Procedure for scoring first-division metaphases.....	11
9.2 Criteria for scoring.....	11
10 Criteria for converting a measured aberration frequency into an estimate of absorbed dose	12
10.1 Overview.....	12
10.2 Comparison with controls.....	12
10.3 Testing the distribution of aberrations per cell.....	12
10.4 Determination of estimated whole-body dose and confidence limits.....	12
10.5 Acute and non-acute exposure cases.....	13
10.6 Partial-body and prior-exposure cases.....	13
11 Reporting of results	15
11.1 General.....	15
11.2 Content of the report (see Annex C for a standard form).....	15
11.3 Interpretation of the results.....	16
12 Quality assurance and quality control	16
12.1 Overview.....	16
12.2 Specific requirements.....	17
Annex A (informative) Sample instructions for customer	19
Annex B (informative) Sample questionnaire	20
Annex C (informative) Sample of report	22
Annex D (informative) Fitting of the low-LET dose-response curve by the method of maximum likelihood and calculating the error of dose estimate	23
Annex E (informative) Odds ratio method for cases of suspected exposure to a low dose	26

ISO 19238:2014(E)

Annex F (informative) Sample data sheet for recording aberrations	27
Bibliography	28

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2. www.iso.org/directives

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received. www.iso.org/patents

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

The committee responsible for this document is ISO/TC 85, *Nuclear energy, nuclear technologies, and radiological protection*, Subcommittee SC 2, *Radiological protection*.

This second edition cancels and replaces the first edition (ISO 19238:2004), of which it constitutes a minor revision.

ISO 19238:2014(E)

Introduction

The wide use of ionising radiations for medical, industrial, agricultural, research, and military purposes increases the risk of overexposure of radiation workers and individuals of the general population. Biological dosimetry, based on the study of chromosomal aberrations, mainly the dicentric assay, has become a routine component of accidental dose assessment. Experience with its application in hundreds of cases of suspected or verified overexposures has proved the value of this method and also defined its limitations. It should be emphasized that cytogenetic analysis is used as a dosimeter and provides one input into the compendium of information needed for assessment of a radiological accident.

Many studies in animals and man have shown that one can establish a good correlation between the results obtained *in vivo* and *in vitro*, so that *in vitro* established dose-effect relationships from irradiated blood samples can be used as calibration curves. The dicentric yield is dependent on radiation quality and dose rate so that information about these variables needs to be established for each investigation. If known, these exposure characteristics are important for refining the dose estimates. The specificity of this technique is enhanced by the fact that generally 1 dicentric is observed per 1 000 metaphase spreads in the normal population, and that this frequency is approximately independent of age and sex. The precision of the technique thus depends on the number of cells observed, the background level, and the calibration curve used. Theoretically, it is possible to detect exposure as low as 0,01 Gy. However, for these very low doses, it is necessary to analyse tens of thousands of metaphase spreads. In practice, this level of detection is neither feasible nor necessary. The upper limits to dose detection extend well into the range of doses that are lethal to humans.

The primary purpose of this International Standard is to provide a guideline to all laboratories in order to perform the dicentric assay using documented and validated procedures. Secondly, it can facilitate the comparison of results obtained in different laboratories, particularly for international collaborations or intercomparisons. Finally, laboratories newly commissioned to carry out the dicentric assay should conform to this International Standard in order to perform it reproducibly and accurately.

This International Standard is written in the form of procedures to be adopted for biological dosimetry for overexposures involving, at most, a few casualties. The criteria required for such measurements will usually depend upon the application of the results: radiation protection management, medical management when appropriate, record keeping, and legal requirements. In the special situation of a mass radiation casualty and limited resources, the technique can be applied for emergency triage analysis. The standard recommended scoring criteria would then be relaxed as appropriate to the situation.

A part of the information in this International Standard is contained in other international guidelines and scientific publications, primarily in the International Atomic Energy Agency's (IAEA) Technical Reports Series on Biological Dosimetry. However, this International Standard expands and standardizes the quality assurance and quality control, the criteria of accreditation, and the evaluation of performance. This International Standard is generally compliant with ISO/IEC 17025, with particular consideration given to the specific needs of biological dosimetry. The expression of uncertainties in dose estimations given in this International Standard comply with the ISO guide to the expression of uncertainty in measurement (ISO/IEC Guide 98-1) and the ISO 5725 on accuracy (trueness and precision) of measurement methods and results.

Radiological protection — Performance criteria for service laboratories performing biological dosimetry by cytogenetics

1 Scope

This International Standard provides criteria for quality assurance and quality control, evaluation of the performance, and the accreditation of biological dosimetry by cytogenetic service laboratories.

This International Standard addresses

- a) the confidentiality of personal information, for the customer and the service laboratory,
- b) the laboratory safety requirements,
- c) the calibration sources and calibration dose ranges useful for establishing the reference dose-effect curves that contribute to the dose estimation from chromosome aberration frequency and the minimum resolvable doses,
- d) the scoring procedure for unstable chromosome aberrations used for biological dosimetry,
- e) the criteria for converting a measured aberration frequency into an estimate of absorbed dose,
- f) the reporting of results,
- g) the quality assurance and quality control,
- h) informative annexes containing sample instructions for customer, sample questionnaire, sample of report, fitting of the low dose-response curve by the method of maximum likelihood and calculating the error of dose estimate, odds ratio method for cases of suspected exposure to a low dose, and sample data sheet for recording aberrations.

2 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

2.1

acentric

terminal or interstitial chromosome fragment of varying size, referred to as an excess acentric fragment when it is formed independently of a dicentric or centric ring chromosome aberration

2.2

background level

spontaneous frequency (or number) of chromosome aberrations recorded in control samples or individuals

2.3

bias

statistical sampling or testing error caused by systematically favouring some outcomes over others

2.4

centric ring

aberrant circular chromosome resulting from the joining of two breaks on separate arms of the same chromosome

Note 1 to entry: It is generally accompanied by an acentric fragment.

This is a free preview. Purchase the entire publication at the link below:

[Product Page](#)

-
- [Looking for additional Standards? Visit Intertek Inform Infostore](#)
 - [Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation](#)
-