



NSAI
Standards

Irish Standard
I.S. EN ISO 20785-1:2017

Dosimetry for exposures to cosmic radiation in civilian aircraft - Part 1: Conceptual basis for measurements (ISO 20785-1:2012)

© CEN 2017 No copying without NSAI permission except as permitted by copyright law.

I.S. EN ISO 20785-1:2017

Incorporating amendments/corrigenda/National Annexes issued since publication:

The National Standards Authority of Ireland (NSAI) produces the following categories of formal documents:

I.S. xxx: Irish Standard — national specification based on the consensus of an expert panel and subject to public consultation.

S.R. xxx: Standard Recommendation — recommendation based on the consensus of an expert panel and subject to public consultation.

SWiFT xxx: A rapidly developed recommendatory document based on the consensus of the participants of an NSAI workshop.

This document replaces/revises/consolidates the NSAI adoption of the document(s) indicated on the CEN/CENELEC cover/Foreword and the following National document(s):

NOTE: The date of any NSAI previous adoption may not match the date of its original CEN/CENELEC document.

This document is based on:

EN ISO 20785-1:2017

Published:

2017-10-11

*This document was published
under the authority of the NSAI
and comes into effect on:*

2017-10-30

ICS number:

13.280

49.020

NOTE: If blank see CEN/CENELEC cover page

NSAI
1 Swift Square,
Northwood, Santry
Dublin 9

T +353 1 807 3800
F +353 1 807 3838
E standards@nsai.ie
W NSAI.ie

Sales:
T +353 1 857 6730
F +353 1 857 6729
W standards.ie

Údarás um Chaighdeáin Náisiúnta na hÉireann

National Foreword

I.S. EN ISO 20785-1:2017 is the adopted Irish version of the European Document EN ISO 20785-1:2017, Dosimetry for exposures to cosmic radiation in civilian aircraft - Part 1: Conceptual basis for measurements (ISO 20785-1:2012)

This document does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

For relationships with other publications refer to the NSAI web store.

Compliance with this document does not of itself confer immunity from legal obligations.

In line with international standards practice the decimal point is shown as a comma (,) throughout this document.

This page is intentionally left blank

EUROPEAN STANDARD

EN ISO 20785-1

NORME EUROPÉENNE

EUROPÄISCHE NORM

October 2017

ICS 13.280; 49.020

English Version

Dosimetry for exposures to cosmic radiation in civilian aircraft - Part 1: Conceptual basis for measurements (ISO 20785-1:2012)

Dosimétrie pour l'exposition au rayonnement
cosmique à bord d'un avion civil - Partie 1: Fondement
théorique des mesurages (ISO 20785-1:2012)

This European Standard was approved by CEN on 13 September 2017.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

EN ISO 20785-1:2017 (E)

Contents	Page
European foreword.....	3

European foreword

The text of ISO 20785-1:2012 has been prepared by Technical Committee ISO/TC 85 “Nuclear energy, nuclear technologies, and radiological protection” of the International Organization for Standardization (ISO) and has been taken over as EN ISO 20785-1:2017 by Technical Committee CEN/TC 430 “Nuclear energy, nuclear technologies, and radiological protection” the secretariat of which is held by AFNOR.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by April 2018, and conflicting national standards shall be withdrawn at the latest by April 2018.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 20785-1:2012 has been approved by CEN as EN ISO 20785-1:2017 without any modification.

This page is intentionally left blank

INTERNATIONAL STANDARD

**ISO
20785-1**

Second edition
2012-12-15

Dosimetry for exposures to cosmic radiation in civilian aircraft —

Part 1: Conceptual basis for measurements

*Dosimétrie pour l'exposition au rayonnement cosmique à bord d'un
avion civil —*

Partie 1: Fondement théorique des mesurages



Reference number
ISO 20785-1:2012(E)

© ISO 2012

ISO 20785-1:2012(E)



COPYRIGHT PROTECTED DOCUMENT

© ISO 2012

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

	Page
Foreword	iv
Introduction	v
1 Scope	1
2 Terms and definitions	1
2.1 General.....	1
2.2 Quantities and units	2
2.3 Atmospheric radiation field.....	8
3 General considerations	10
3.1 General description of the cosmic radiation field in the atmosphere	10
3.2 General calibration considerations for the dosimetry of cosmic radiation fields in aircraft.....	11
3.3 Conversion coefficients.....	13
4 Dosimetric devices	13
4.1 Introduction	13
4.2 Active devices.....	14
4.3 Passive devices.....	17
Annex A (informative) Representative particle fluence rate energy distributions for the cosmic radiation field at flight altitudes for solar minimum and maximum conditions and for minimum and maximum vertical cut-off rigidity [80]	20
Bibliography	24

ISO 20785-1:2012(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 20785-1 was prepared by Technical Committee ISO/TC 85, *Nuclear energy, nuclear technologies, and radiological protection*, Subcommittee SC 2, *Radiological protection*.

This second edition cancels and replaces the first edition (ISO 20785-1:2006), which has been technically revised.

ISO 20785 consists of the following parts, under the general title *Dosimetry for exposures to cosmic radiation in civilian aircraft*:

- *Part 1: Conceptual basis for measurements*
- *Part 2: Characterization of instrument response*

Measurements at aviation altitudes is to form the subject of a future Part 3.

Introduction

Aircraft crews are exposed to elevated levels of cosmic radiation of galactic and solar origin and secondary radiation produced in the atmosphere, the aircraft structure and its contents. Following recommendations of the International Commission on Radiological Protection in Publication 60,^[1] confirmed by Publication 103,^[2] the European Union (EU) introduced a revised Basic Safety Standards Directive^[3] which included exposure to natural sources of ionizing radiation, including cosmic radiation, as occupational exposure. The Directive requires account to be taken of the exposure of aircraft crews liable to receive more than 1 mSv per year. It then identifies the following four protection measures: (i) to assess the exposure of the crew concerned; (ii) to take into account the assessed exposure when organizing working schedules with a view to reducing the doses of highly exposed crews; (iii) to inform the workers concerned of the health risks their work involves; and (iv) to apply the same special protection during pregnancy to female crews in respect of the “child to be born” as to other female workers. The EU Council Directive has already been incorporated into laws and regulations of EU Member States and is being included in the aviation safety standards and procedures of the Joint Aviation Authorities and the European Air Safety Agency. Other countries such as Canada and Japan have issued advisories to their airline industries to manage aircraft crew exposure.

For regulatory and legislative purposes, the radiation protection quantities of interest are the equivalent dose (to the foetus) and the effective dose. The cosmic radiation exposure of the body is essentially uniform and the maternal abdomen provides no effective shielding to the foetus. As a result, the magnitude of equivalent dose to the foetus can be put equal to that of the effective dose received by the mother. Doses on board aircraft are generally predictable, and events comparable to unplanned exposure in other radiological workplaces cannot normally occur (with the rare exceptions of extremely intense and energetic solar particle events). Personal dosimeters for routine use are not considered necessary. The preferred approach for the assessment of doses of aircraft crews, where necessary, is to calculate directly the effective dose per unit time, as a function of geographic location, altitude and solar cycle phase, and to fold these values with flight and staff roster information to obtain estimates of effective doses for individuals. This approach is supported by guidance from the European Commission and the ICRP in Publication 75.^[4]

The role of calculations in this procedure is unique in routine radiation protection and it is widely accepted that the calculated doses should be validated by measurement. The effective dose is not directly measurable. The operational quantity of interest is ambient dose equivalent, $H^*(10)$. In order to validate the assessed doses obtained in terms of effective dose, calculations can be made of ambient dose equivalent rates or route doses in terms of ambient dose equivalent, and values of this quantity determined by measurements traceable to national standards. The validation of calculations of ambient dose equivalent for a particular calculation method may be taken as a validation of the calculation of the effective dose by the same computer code, but this step in the process may need to be confirmed. The alternative is to establish *a priori* that the operational quantity ambient dose equivalent is a good estimator of effective dose and equivalent dose to the foetus for the radiation fields being considered, in the same way that the use of the operational quantity personal dose equivalent is justified for the estimation of effective dose for radiation workers. Ambient dose equivalent rate as a function of geographic location, altitude and solar cycle phase is then calculated and folded with flight and staff roster information.

The radiation field in aircraft at altitude is complex, with many types of ionizing radiation present, with energies ranging up to many GeV. The determination of ambient dose equivalent for such a complex radiation field is difficult. In many cases, the methods used for the determination of ambient dose equivalent in aircraft are similar to those used at high-energy accelerators in research laboratories. Therefore, it is possible to recommend dosimetric methods and methods for the calibration of dosimetric devices, as well as the techniques for maintaining the traceability of dosimetric measurements to national standards. Dosimetric measurements made to evaluate ambient dose equivalent must be performed using accurate and reliable methods that ensure the quality of readings provided to workers and regulatory authorities. This part of ISO 20785 gives a conceptual basis for the characterization of the response of instruments for the determination of ambient dose equivalent in aircraft.

ISO 20785-1:2012(E)

Requirements for the determination and recording of the cosmic radiation exposure of aircraft crews have been introduced into the national legislation of EU Member States and other countries. Harmonization of methods used for determining ambient dose equivalent and for calibrating instruments is desirable to ensure the compatibility of measurements performed with such instruments.

This part of ISO 20785 is intended for the use of primary and secondary calibration laboratories for ionizing radiation, by radiation protection personnel employed by governmental agencies, and by industrial corporations concerned with the determination of ambient dose equivalent for aircraft crews.

Dosimetry for exposures to cosmic radiation in civilian aircraft —

Part 1: Conceptual basis for measurements

1 Scope

This part of ISO 20785 gives the conceptual basis for the determination of ambient dose equivalent for the evaluation of exposure to cosmic radiation in civilian aircraft and for the calibration of instruments used for that purpose.

2 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

2.1 General

2.1.1

calibration

operation that, under specified conditions, establishes a relation between the conventional quantity, H_0 , and the indication, G

Note 1 to entry: A calibration may be expressed by a statement, calibration function, calibration diagram, calibration curve, or calibration table. In some cases, it may consist of an additive or multiplicative correction of the indication with associated measurement uncertainty.

Note 2 to entry: Calibration should not be confused with adjustment of a measuring system, often mistakenly called “self-calibration”, or with verification of calibration.

Note 3 to entry: Often, the first step alone in the above definition is perceived as being calibration.

2.1.2

calibration coefficient

N_{coeff}

quotient of the conventional quantity value to be measured and the corrected indication of the instrument

Note 1 to entry: The calibration coefficient is equivalent to the calibration factor multiplied by the instrument constant.

Note 2 to entry: The reciprocal of the calibration coefficient, N_{coeff} , is the response.

Note 3 to entry: For the calibration of some instruments, e.g. ionization chambers, the instrument constant and the calibration factor are not identified separately but are applied together as the calibration coefficient.

Note 4 to entry: It is necessary, in order to avoid confusion, to state the quantity to be measured, for example: the calibration coefficient with respect to fluence, N_ϕ , the calibration coefficient with respect to kerma, N_K , the calibration coefficient with respect to absorbed dose, N_D .

This is a free preview. Purchase the entire publication at the link below:

[Product Page](#)

-
- Looking for additional Standards? Visit Intertek Inform Infostore
 - Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation
-