



NSAI
Standards

Irish Standard
I.S. EN ISO 10704:2019

Water quality - Gross alpha and gross beta activity - Test method using thin source deposit (ISO 10704:2019)

I.S. EN ISO 10704:2019

Incorporating amendments/corrigenda/National Annexes issued since publication:

The National Standards Authority of Ireland (NSAI) produces the following categories of formal documents:

I.S. xxx: Irish Standard — national specification based on the consensus of an expert panel and subject to public consultation.

S.R. xxx: Standard Recommendation — recommendation based on the consensus of an expert panel and subject to public consultation.

SWiFT xxx: A rapidly developed recommendatory document based on the consensus of the participants of an NSAI workshop.

This document replaces/revises/consolidates the NSAI adoption of the document(s) indicated on the CEN/CENELEC cover/Foreword and the following National document(s):

NOTE: The date of any NSAI previous adoption may not match the date of its original CEN/CENELEC document.

This document is based on:

EN ISO 10704:2019

Published:

2019-03-27

*This document was published
under the authority of the NSAI
and comes into effect on:*

2019-04-14

ICS number:

13.060.60

13.280

NOTE: If blank see CEN/CENELEC cover page

NSAI
1 Swift Square,
Northwood, Santry
Dublin 9

T +353 1 807 3800
F +353 1 807 3838
E standards@nsai.ie
W NSAI.ie

Sales:
T +353 1 857 6730
F +353 1 857 6729
W standards.ie

Údarás um Chaighdeáin Náisiúnta na hÉireann

National Foreword

I.S. EN ISO 10704:2019 is the adopted Irish version of the European Document EN ISO 10704:2019, Water quality - Gross alpha and gross beta activity - Test method using thin source deposit (ISO 10704:2019)

This document does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

For relationships with other publications refer to the NSAI web store.

Compliance with this document does not of itself confer immunity from legal obligations.

In line with international standards practice the decimal point is shown as a comma (,) throughout this document.

This page is intentionally left blank

EUROPEAN STANDARD

EN ISO 10704

NORME EUROPÉENNE

EUROPÄISCHE NORM

March 2019

ICS 13.060.60; 13.280

Supersedes EN ISO 10704:2015

English Version

Water quality - Gross alpha and gross beta activity - Test method using thin source deposit (ISO 10704:2019)

Qualité de l'eau - Activités alpha globale et bêta globale
- Méthode d'essai par dépôt d'une source fine (ISO
10704:2019)

Wasserbeschaffenheit - Bestimmung der Gesamt-
Alpha- und der Gesamt-Beta-Aktivität in nicht-
salzhaltigem Wasser - Dünnschichtverfahren (ISO
10704:2019)

This European Standard was approved by CEN on 16 February 2019.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents	Page
European foreword.....	3

European foreword

This document (EN ISO 10704:2019) has been prepared by Technical Committee ISO/TC 147 "Water quality" in collaboration with Technical Committee CEN/TC 230 "Water analysis" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by September 2019, and conflicting national standards shall be withdrawn at the latest by September 2019.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 10704:2015.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 10704:2019 has been approved by CEN as EN ISO 10704:2019 without any modification.

This page is intentionally left blank

INTERNATIONAL STANDARD

**ISO
10704**

Second edition
2019-02

Water quality — Gross alpha and gross beta activity — Test method using thin source deposit

Qualité de l'eau — Activités alpha globale et bêta globale — Méthode d'essai par dépôt d'une source fine



Reference number
ISO 10704:2019(E)

© ISO 2019

ISO 10704:2019(E)



COPYRIGHT PROTECTED DOCUMENT

© ISO 2019

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

	Page
Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms, definitions and symbols	2
4 Principle	3
5 Chemical reagents and equipment	3
5.1 Reagents.....	3
5.1.1 General.....	3
5.1.2 Standard solutions.....	3
5.1.3 Wetting or surfactant agents.....	4
5.1.4 Volatile organic solvents.....	4
5.1.5 Water.....	4
5.1.6 Specific reagents for alpha-emitting radionuclides co-precipitation.....	4
5.2 Equipment.....	4
5.2.1 Laboratory equipment for direct evaporation.....	4
5.2.2 General equipment.....	4
5.2.3 Special equipment for alpha-emitting radionuclide co-precipitation.....	5
5.2.4 Measurement equipment.....	5
6 Sampling	5
7 Procedure	5
7.1 Preliminary.....	5
7.2 Source preparation.....	5
7.2.1 Preparation of planchet.....	5
7.2.2 Evaporation.....	6
7.2.3 Co-precipitation.....	6
7.3 Counting stage.....	7
7.4 Background and blank determination.....	7
7.5 Preparation of counting standard for calibration.....	7
7.6 Preparation of calibration source for self-absorption determination.....	8
7.6.1 General.....	8
7.6.2 Spiking one of two test portions.....	8
7.6.3 Self-absorption curve.....	8
8 Expression of results	9
8.1 General.....	9
8.2 Alpha activity concentration.....	9
8.3 Beta activity concentration.....	9
8.4 Standard uncertainty of the alpha activity concentration.....	10
8.5 Standard uncertainty of the beta activity concentration.....	10
8.6 Decision threshold.....	12
8.6.1 Decision threshold of the alpha activity concentration.....	12
8.6.2 Decision threshold of the beta activity concentration.....	12
8.7 Detection limit.....	12
8.7.1 Detection limit of the alpha activity concentration.....	12
8.7.2 Detection limit of the beta activity concentration.....	13
8.8 Confidence limits.....	13
9 Control of interferences	13
9.1 General.....	13
9.2 Relative humidity.....	14
9.3 Geometry of the deposit.....	14
9.4 Crosstalk.....	14

ISO 10704:2019(E)

9.5	Gamma emitters.....	15
9.6	Low beta energy.....	15
9.7	Chlorides	15
9.8	Organic matter	15
9.9	Contamination.....	15
9.10	Losses of activity	15
9.11	Contribution of the natural radionuclides	15
9.12	Losses of activity	16
10	Test report.....	16
Annex A (informative) Numerical applications.....		18
Bibliography		19

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 147, *Water quality*, Subcommittee SC 3, *Radioactivity measurements*.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

This second edition cancels and replaces the first edition (ISO 10704:2009), which has been technically revised. The main changes compared to the previous edition are as follows:

- Introduction: an introduction has been added;
- [Clause 1](#): the scope has been modified to specify applicability to emergency situations and applicability of waste water as a test sample; information about the exclusion of low energy beta emitters has also been added;
- [Clause 4](#): the filtration has been specified to be carried out at 0,45 µ;
- [5.1.2.2](#): ¹³⁷Cs has been introduced as a standard that can be used;
- [5.2.4](#): the recommended thickness has been increased to up to 400 µg/cm²;
- [7.6.3.1](#): in order to evaluate self-absorption phenomena, spiking method has been recommended to mimic the nature of the salt;
- [Clause 8](#):
 - a new [Formula \(9\)](#) has been introduced to obtain the beta activity concentration when systematic correction is not required;
 - the subsequent Formulae have been renumbered;
- [Clause 9](#): several limitations and interferences have been given;
- [9.1](#): the natural radionuclides contributions have been given.

ISO 10704:2019(E)

Introduction

Radioactivity from several naturally-occurring and anthropogenic sources is present throughout the environment. Thus, water bodies (e.g. surface waters, ground waters, sea waters) can contain radionuclides of natural, human-made, or both origins:

- natural radionuclides, including ^{40}K , ^3H , ^{14}C , and those originating from the thorium and uranium decay series, in particular ^{226}Ra , ^{228}Ra , ^{234}U , ^{238}U , ^{210}Po and ^{210}Pb can be found in water for natural reasons (e.g. desorption from the soil and washoff by rain water) or can be released from technological processes involving naturally occurring radioactive materials (e.g. the mining and processing of mineral sands or phosphate fertilizers production and use);
- human-made radionuclides such as transuranium elements (americium, plutonium, neptunium, curium), ^3H , ^{14}C , ^{90}Sr , and gamma emitting radionuclides can also be found in natural waters. Small quantities of these radionuclides are discharged from nuclear fuel cycle facilities into the environment as a result of authorized routine releases. Some of these radionuclides used for medical and industrial applications are also released into the environment after use. Anthropogenic radionuclides are also found in waters as a result of past fallout contaminations resulting from the explosion in the atmosphere of nuclear devices and accidents such as those that occurred in Chernobyl and Fukushima.

Radionuclide activity concentration in water bodies can vary according to local geological characteristics and climatic conditions and can be locally and temporally enhanced by releases from nuclear installation during planned, existing, and emergency exposure situations[1]. Drinking-water can thus contain radionuclides at activity concentrations which could present a risk to human health.

The radionuclides present in liquid effluents are usually controlled before being discharged into the environment[2] and water bodies. Drinking waters are monitored for their radioactivity as recommended by the World Health Organization (WHO)[3] so that proper actions can be taken to ensure that there is no adverse health effect to the public. Following these international recommendations, national regulations usually specify radionuclide authorized concentration limits for liquid effluent discharged to the environment and radionuclide guidance levels for waterbodies and drinking waters for planned, existing, and emergency exposure situations. Compliance with these limits can be assessed using measurement results with their associated uncertainties as specified by ISO/IEC Guide 98-3 and ISO 5667-20[4].

Depending on the exposure situation, there are different limits and guidance levels that would result in an action to reduce health risk. As an example, during a planned or existing situation, the WHO guidelines for guidance level in drinking water is 0,5 Bq/l for gross alpha activity and 1 Bq/l for gross beta activity.

NOTE The guidance level is the activity concentration with an intake of 2 l/d of drinking water for one year that results in an effective dose of 0,1 mSv/a for members of the public. This is an effective dose that represents a very low level of risk and which is not expected to give rise to any detectable adverse health effects[3].

Thus, the test method can be adapted so that the characteristic limits, decision threshold, detection limit and uncertainties ensure that the radionuclide activity concentrations test results can be verified to be below the guidance levels required by a national authority for either planned/existing situations or for an emergency situation[5][6][7].

Usually, the test methods can be adjusted to measure the activity concentration of the radionuclide(s) in either wastewaters before storage or in liquid effluents before being discharged to the environment. The test results will enable the plant/installation operator to verify that, before their discharge, wastewaters/liquid effluent radioactive activity concentrations do not exceed authorized limits.

The test method(s) described in this document can be used during planned, existing and emergency exposure situations as well as for wastewaters and liquid effluents with specific modifications that could increase the overall uncertainty, detection limit, and threshold.

The test method(s) can be used for water samples after proper sampling, sample handling, and test sample preparation (see the relevant part of the ISO 5667 series).

An International Standard on a test method of gross alpha and gross beta activity concentrations in water samples is justified for test laboratories carrying out these measurements, required sometimes by national authorities, as laboratories might have to obtain a specific accreditation for radionuclide measurement in drinking water samples.

This document is one of a set of International Standards on test methods dealing with the measurement of the activity concentration of radionuclides in water samples.

Water quality — Gross alpha and gross beta activity — Test method using thin source deposit

WARNING — Persons using this document should be familiar with normal laboratory practice. This document does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of the user to establish appropriate safety and health practices.

IMPORTANT — It is absolutely essential that tests conducted according to this document be carried out by suitably trained staff.

1 Scope

This document specifies a method for the determination of gross alpha and gross beta activity concentration for alpha- and beta-emitting radionuclides. Gross alpha and gross beta activity measurement is not intended to give an absolute determination of the activity concentration of all alpha and beta emitting radionuclides in a test sample, but is a screening analysis to ensure particular reference levels of specific alpha and beta emitters have not been exceeded. This type of determination is also known as gross alpha and gross beta index. Gross alpha and gross beta analysis is not expected to be as accurate nor as precise as specific radionuclide analysis after radiochemical separations.

Maximum beta energies of approximately 0,1 MeV or higher are well measured. It is possible that low energy beta emitters can not be detected (e.g. ^3H , ^{55}Fe , ^{241}Pu) or can only be partially detected (e.g. ^{14}C , ^{35}S , ^{63}Ni , ^{210}Pb , ^{228}Ra).

The method covers non-volatile radionuclides, since some gaseous or volatile radionuclides (e.g. radon and radioiodine) can be lost during the source preparation.

The method is applicable to test samples of drinking water, rainwater, surface and ground water as well as cooling water, industrial water, domestic and industrial wastewater after proper sampling, sample handling, and test sample preparation (filtration when necessary and taking into account the amount of dissolved material in the water).

The method described in this document is applicable in the event of an emergency situation, because the results can be obtained in less than 1 h. Detection limits reached for gross alpha and gross beta are less than 10 Bq/l and 20 Bq/l respectively. The evaporation of 10 ml sample is carried out in 20 min followed by 10 min counting with window-proportional counters.

It is the laboratory's responsibility to ensure the suitability of this test method for the water samples tested.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 3696, *Water for analytical laboratory use — Specification and test methods*

ISO 5667-1, *Water quality — Sampling — Part 1: Guidance on the design of sampling programmes and sampling techniques*

ISO 5667-3, *Water quality — Sampling — Part 3: Preservation and handling of water samples*

ISO 11929, *Determination of the characteristic limits (decision threshold, detection limit and limits of the confidence interval) for measurements of ionizing radiation — Fundamentals and application*

This is a free preview. Purchase the entire publication at the link below:

[Product Page](#)

-
- Looking for additional Standards? Visit Intertek Inform Infostore
 - Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation
-