



NSAI
Standards

Irish Standard
I.S. EN 50676:2019

Electrical equipment used for detection and concentration measurement of refrigerant gases - Performance requirements and test methods

I.S. EN 50676:2019

Incorporating amendments/corrigenda/National Annexes issued since publication:

The National Standards Authority of Ireland (NSAI) produces the following categories of formal documents:

I.S. xxx: Irish Standard — national specification based on the consensus of an expert panel and subject to public consultation.

S.R. xxx: Standard Recommendation — recommendation based on the consensus of an expert panel and subject to public consultation.

SWiFT xxx: A rapidly developed recommendatory document based on the consensus of the participants of an NSAI workshop.

This document replaces/revises/consolidates the NSAI adoption of the document(s) indicated on the CEN/CENELEC cover/Foreword and the following National document(s):

NOTE: The date of any NSAI previous adoption may not match the date of its original CEN/CENELEC document.

This document is based on:

EN 50676:2019

Published:

2019-12-20

*This document was published
under the authority of the NSAI
and comes into effect on:*

2020-01-08

ICS number:

NOTE: If blank see CEN/CENELEC cover page

NSAI
1 Swift Square,
Northwood, Santry
Dublin 9

T +353 1 807 3800
F +353 1 807 3838
E standards@nsai.ie
W NSAI.ie

Sales:
T +353 1 857 6730
F +353 1 857 6729
W standards.ie

Údarás um Chaighdeáin Náisiúnta na hÉireann

National Foreword

I.S. EN 50676:2019 is the adopted Irish version of the European Document EN 50676:2019, Electrical equipment used for detection and concentration measurement of refrigerant gases - Performance requirements and test methods

This document does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

For relationships with other publications refer to the NSAI web store.

Compliance with this document does not of itself confer immunity from legal obligations.

In line with international standards practice the decimal point is shown as a comma (,) throughout this document.

This page is intentionally left blank

EUROPEAN STANDARD

EN 50676

NORME EUROPÉENNE

EUROPÄISCHE NORM

December 2019

ICS 13.320; 27.200

English Version

**Electrical equipment used for detection and concentration
measurement of refrigerant gases - Performance requirements
and test methods**

Appareils électriques utilisés pour la détection et la mesure
de la concentration de gaz frigorigènes - Exigences de
performance et méthodes d'essai

Elektrische Geräte zur Detektion und
Konzentrationsmessung von Kältemittelgasen -
Anforderungen an das Betriebsverhalten und Prüfverfahren

This European Standard was approved by CENELEC on 2019-11-04. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.



European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents	Page
European foreword	3
Introduction	3
1 Scope	5
2 Normative references	5
3 Terms and definitions	6
Figure 1 — Warm-up time in clean air (typical)	8
4 General requirements	8
4.1 Introduction	8
4.2 Construction	9
Table 1 — Measuring ranges, reference values and maximum thresholds	9
4.3 Adjustments	10
4.4 Gas detection transmitter for use with separate gas detection control units	10
4.5 Separate gas detection control units for use with gas detection transmitter(s)	10
4.6 Equipment using software and/or digital technologies	10
4.7 Labelling and marking	10
4.8 Instruction manual	11
5 Test methods	11
5.1 Introduction	11
5.2 General requirements for tests	11
5.3 Samples and sequence of tests	11
5.4 Normal conditions for test	12
5.5 Tests	13
Table 2 — Test gas for poison gas test	15
Annex A (normative) Gas specific performance requirements (EN 45544-1)	17
Table A.1 — Gas specific performance requirements	17
Annex B (normative) Performance requirements under standard test conditions	18
Table B.1 — Performance requirements under standard test conditions	18
Bibliography	20

European foreword

This document (EN 50676:2019) has been prepared by CLC/TC 216 “Gas detectors”.

The following dates are fixed:

- latest date by which this document has (dop) 2020-11-04
to be implemented at national level by
publication of an identical national
standard or by endorsement
- latest date by which the national (dow) 2022-11-04
standards conflicting with this document
have to be withdrawn

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC shall not be held responsible for identifying any or all such patent rights.

EN 50676:2019 (E)

Introduction

This document specifies test methods and performance requirements for all electrical equipment used for gas detection in refrigerant applications as defined in EN 378-1:2016 by means of concentration measurement.

This document is addressed to the manufacturers of such equipment and test laboratories which validate it.

The tendency to use low-GWP refrigerant gases in the refrigeration and HVAC market (F-Gas Regulation) has intensified the considerations of safety measures for low-GWP gases as some are flammable, toxic and can cause lack of oxygen.

This document covers all refrigerant gases and defines performance requirements for the detection equipment, mentioned in EN 378-3:2016 as gas detectors or sensors, used in refrigerant applications. The level of safety is expected to be the same as in the already existing performance standards for general-purpose equipment, i.e. EN 60079-29-1 concerning refrigerant flammable gases and EN 45544 series concerning refrigerant toxic gases in atmospheres. Refrigerant gases not mentioned by EN 378-1:2016 are also covered by this standard following the categorization scheme of EN 378-1:2016.

1 Scope

This document specifies general requirements for the construction, testing and performance of electrically operated refrigerant fixed gas detection equipment in safety applications. This document does not specify requirements for portable locating leak detectors for refrigerant application as already covered by EN 14624:2012.

This document is applicable to equipment whose primary purpose is to provide an indication, alarm and/or other output function to warn of the presence of refrigerant gases in an industrial or commercial environment and, in some cases, to initiate automatic or manual protective actions. It is applicable to equipment in which the sensor automatically generates an electrical signal when gas is present.

This standard does not apply to gas detection equipment:

- for non-refrigerant application;
- used for air pollution monitoring;
- sampling systems, which are not integral part of the gas detection equipment;
- open path gas detection;
- residential applications;
- process control;
- for applications in mines;
- portable locating leak detectors for refrigerant application.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 378-1:2016, *Refrigerating systems and heat pumps - Safety and environmental requirements - Part 1: Basic requirements, definitions, classification and selection criteria*

EN 45544-1, *Workplace atmospheres - Electrical apparatus used for the direct detection and direct concentration measurement of toxic gases and vapours - Part 1: General requirements and test methods*

EN 45544-2, *Workplace atmospheres - Electrical apparatus used for the direct detection and direct concentration measurement of toxic gases and vapours - Part 2: Performance requirements for apparatus used for exposure measurement*

EN 45544-3, *Workplace atmospheres - Electrical apparatus used for the direct detection and direct concentration measurement of toxic gases and vapours - Part 3: Performance requirements for apparatus used for general gas detection*

EN 45544-4, *Workplace atmospheres - Electrical apparatus used for the direct detection and direct concentration measurement of toxic gases and vapours - Part 4: Guide for selection, installation, use and maintenance*

EN 60079-29-1:2016, *Explosive atmospheres - Part 29-1: Gas detectors - Performance requirements of detectors for flammable gases*

IEC 60335-2-40, *Household and similar electrical appliances - Safety - Part 2-40: Particular requirements for electrical heat pumps, air-conditioners and dehumidifiers*

This is a free preview. Purchase the entire publication at the link below:

[Product Page](#)

-
- Looking for additional Standards? Visit Intertek Inform Infostore
 - Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation
-