



**NSAI**  
Standards

Irish Standard  
I.S. EN ISO 11607-1;2020

Packaging for terminally sterilized medical devices - Part 1: Requirements for materials, sterile barrier systems and packaging systems (ISO 11607-1:2019)

**I.S. EN ISO 11607-1;2020**

*Incorporating amendments/corrigenda/National Annexes issued since publication:*

The National Standards Authority of Ireland (NSAI) produces the following categories of formal documents:

I.S. xxx: Irish Standard — national specification based on the consensus of an expert panel and subject to public consultation.

S.R. xxx: Standard Recommendation — recommendation based on the consensus of an expert panel and subject to public consultation.

SWiFT xxx: A rapidly developed recommendatory document based on the consensus of the participants of an NSAI workshop.

*This document replaces/revises/consolidates the NSAI adoption of the document(s) indicated on the CEN/CENELEC cover/Foreword and the following National document(s):*

*NOTE: The date of any NSAI previous adoption may not match the date of its original CEN/CENELEC document.*

*This document is based on:*

EN ISO 11607-1:2020

*Published:*

2020-01-15

*This document was published under the authority of the NSAI and comes into effect on:*

2020-01-23

ICS number:

11.080.30

NOTE: If blank see CEN/CENELEC cover page

NSAI  
1 Swift Square,  
Northwood, Santry  
Dublin 9

T +353 1 807 3800  
F +353 1 807 3838  
E standards@nsai.ie  
W NSAI.ie

Sales:  
T +353 1 857 6730  
F +353 1 857 6729  
W standards.ie

Údarás um Chaighdeáin Náisiúnta na hÉireann

## National Foreword

I.S. EN ISO 11607-1;2020 is the adopted Irish version of the European Document EN ISO 11607-1;2020, Packaging for terminally sterilized medical devices - Part 1: Requirements for materials, sterile barrier systems and packaging systems (ISO 11607-1:2019)

This document does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

For relationships with other publications refer to the NSAI web store.

**Compliance with this document does not of itself confer immunity from legal obligations.**

*In line with international standards practice the decimal point is shown as a comma (,) throughout this document.*

This page is intentionally left blank

EUROPEAN STANDARD

EN ISO 11607-1

NORME EUROPÉENNE

EUROPÄISCHE NORM

January 2020

ICS 11.080.30

Supersedes EN ISO 11607-1:2009

English Version

## Packaging for terminally sterilized medical devices - Part 1: Requirements for materials, sterile barrier systems and packaging systems (ISO 11607-1:2019)

Emballages des dispositifs médicaux stérilisés au stade terminal - Partie 1: Exigences relatives aux matériaux, aux systèmes de barrière stérile et aux systèmes d'emballage (ISO 11607-1:2019)

Verpackungen für in der Endverpackung zu sterilisierende Medizinprodukte - Teil 1: Anforderungen an Materialien, Sterilbarrieresysteme und Verpackungssysteme (ISO 11607-1:2019)

This European Standard was approved by CEN on 3 November 2018.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

**EN ISO 11607-1:2020 (E)**

<b>Contents</b>	<b>Page</b>
<b>European foreword.....</b>	<b>3</b>

## **European foreword**

This document (EN ISO 11607-1:2020) has been prepared by Technical Committee ISO/TC 198 "Sterilization of health care products" in collaboration with Technical Committee CEN/TC 102 "Sterilizers and associated equipment for processing of medical devices" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by July 2020, and conflicting national standards shall be withdrawn at the latest by July 2020.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 11607-1:2009.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

## **Endorsement notice**

The text of ISO 11607-1:2019 has been approved by CEN as EN ISO 11607-1:2020 without any modification.

This page is intentionally left blank



# INTERNATIONAL STANDARD

# ISO 11607-1

Second edition  
2019-02

---

---

## **Packaging for terminally sterilized medical devices —**

### **Part 1: Requirements for materials, sterile barrier systems and packaging systems**

*Emballages des dispositifs médicaux stérilisés au stade terminal —*

*Partie 1: Exigences relatives aux matériaux, aux systèmes de barrière  
stérile et aux systèmes d'emballage*



Reference number  
ISO 11607-1:2019(E)

© ISO 2019

**ISO 11607-1:2019(E)**



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2019

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva  
Phone: +41 22 749 01 11  
Fax: +41 22 749 09 47  
Email: [copyright@iso.org](mailto:copyright@iso.org)  
Website: [www.iso.org](http://www.iso.org)

Published in Switzerland

# Contents

	Page
<b>Foreword</b> .....	<b>iv</b>
<b>Introduction</b> .....	<b>v</b>
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>1</b>
<b>3 Terms and definitions</b> .....	<b>1</b>
<b>4 General requirements</b> .....	<b>6</b>
4.1 Quality systems.....	6
4.2 Risk management.....	6
4.3 Sampling.....	6
4.4 Test methods.....	6
4.5 Documentation.....	7
<b>5 Materials, preformed sterile barrier systems and sterile barrier systems</b> .....	<b>7</b>
5.1 General requirements.....	7
5.2 Microbial barrier properties.....	10
5.3 Compatibility with the sterilization process.....	11
5.4 Labelling system.....	11
5.5 Storage and transport of materials and preformed sterile barrier systems.....	11
<b>6 Design and development for packaging systems</b> .....	<b>12</b>
6.1 General.....	12
6.2 Design.....	12
<b>7 Usability evaluation for aseptic presentation</b> .....	<b>13</b>
<b>8 Packaging system performance and stability</b> .....	<b>14</b>
8.1 General.....	14
8.2 Packaging system performance testing.....	14
8.3 Stability testing.....	15
<b>9 Packaging system validation and changes</b> .....	<b>15</b>
<b>10 Inspection immediately prior to aseptic presentation</b> .....	<b>16</b>
<b>11 Information to be provided</b> .....	<b>16</b>
<b>Annex A (informative) Guidance on medical packaging</b> .....	<b>17</b>
<b>Annex B (informative) Standardized test methods, guides and procedures that can be used to demonstrate conformity with the requirements of this document</b> .....	<b>20</b>
<b>Annex C (normative) Test method for resistance of impermeable materials to the passage of air</b>	<b>31</b>
<b>Annex D (informative) Environmental aspects</b> .....	<b>32</b>
<b>Annex E (informative) Draft guidance on ways to differentiate a sterile barrier system from protective packaging</b> .....	<b>33</b>
<b>Bibliography</b> .....	<b>38</b>

## ISO 11607-1:2019(E)

### Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by Technical Committee ISO/TC 198, *Sterilization of health care products*.

This second edition cancels and replaces the first edition (ISO 11607-1:2006), which has been technically revised. It also incorporates the amendment ISO 11607-1:2006/Amd.1:2014.

The main changes compared to the previous edition are as follows:

- the definitions have been aligned with the latest version of ISO 11139;
- new requirements for the evaluation of usability for aseptic presentation have been added;
- new requirements for the inspection of sterile barrier system integrity prior to use have been added;
- a new subclause with requirements for revalidation in accordance with ISO 11607-2 has been added;
- [Annex B](#) has been updated and various national, international and European test methods have been added or deleted;
- a new [Annex D](#) has been added with environmental considerations;
- a new [Annex E](#) has been added with draft guidance on ways to differentiate a sterile barrier system from protective packaging.

A list of all parts in the ISO 11607 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).

## Introduction

The process of designing and developing a packaging system for terminally sterilized medical devices is a complicated and critical endeavour. The device components and the packaging system should be combined to create a sterile medical device that performs efficiently, safely and effectively in the hands of the user.

This document specifies requirements for the design of sterile barrier systems and packaging systems for terminally sterilized medical devices, the basic attributes required of materials and preformed sterile barrier systems, and design validation requirements. This document is written as a general (horizontal) standard considering a wide range of potential materials, medical devices, packaging system designs and sterilization methods. It can be applied by suppliers of materials or of preformed sterile barrier systems, by medical device manufacturers or health care facilities. ISO 11607-2 describes the process development and validation requirements for forming, sealing and assembly processes and addresses controls during normal operations.

Guidance for ISO 11607 series can be found in ISO/TS 16775.

European standards that provide requirements for particular materials and preformed sterile barrier systems are available and known as the EN 868 series. Conformity with the EN 868 series can be used to demonstrate conformity with one or more of the requirements of this document.

The goal of a terminally sterilized medical device packaging system is to allow sterilization, provide physical protection, maintain sterility up to the point of use and allow aseptic presentation. The specific nature of the medical device, the intended sterilization methods(s), the intended use, expiry date, transport and storage all influence the packaging system design and choice of materials.

The term “sterile barrier system” was introduced in ISO 11607-1:2006 to describe the minimum packaging required to perform the unique functions required of medical packaging: to allow sterilization, to provide an acceptable microbial barrier, and to allow for aseptic presentation. “Protective packaging” protects the sterile barrier system, and together they form the packaging system. “Preformed sterile barrier systems” would include any partially assembled sterile barrier systems such as pouches, header bags or hospital packaging reels. An overview of sterile barrier systems is given in [Annex A](#).

The sterile barrier system is essential to ensure the safety of terminally sterilized medical devices. Regulatory authorities recognize the critical nature of sterile barrier systems by considering them as an accessory or a component of a medical device. Preformed sterile barrier systems sold to health care facilities for use in internal sterilization are considered medical devices in many parts of the world.



# Packaging for terminally sterilized medical devices —

## Part 1: Requirements for materials, sterile barrier systems and packaging systems

### 1 Scope

This document specifies requirements and test methods for materials, preformed sterile barrier systems, sterile barrier systems and packaging systems that are intended to maintain sterility of terminally sterilized medical devices until the point of use.

It is applicable to industry, to health care facilities, and to wherever medical devices are placed in sterile barrier systems and sterilized.

It does not cover all requirements for sterile barrier systems and packaging systems for medical devices that are manufactured aseptically. Additional requirements can be necessary for drug/device combinations.

It does not describe a quality assurance system for control of all stages of manufacture.

It does not apply to packaging materials and/or systems used to contain a contaminated medical device during transportation of the item to the site of reprocessing or disposal.

### 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 5636-5, *Paper and board — Determination of air permeance (medium range) — Part 5: Gurley method*

ISO 11607-2, *Packaging for terminally sterilized medical devices — Part 2: Validation requirements for forming, sealing and assembly processes*

### 3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

#### 3.1

##### **aseptic presentation**

transfer of sterile contents from its sterile barrier system using conditions and procedures that minimize the risk of microbial contamination

[SOURCE: ISO 11139:2018, 3.13]

This is a free preview. Purchase the entire publication at the link below:

[Product Page](#)

- 
- [Looking for additional Standards? Visit Intertek Inform Infostore](#)
  - [Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation](#)
-