



NSAI
Standards

Irish Standard
I.S. EN ISO 8637-1:2020

Extracorporeal systems for blood purification - Part 1: Haemodialysers, haemodiafilters, haemofilters and haemoconcentrators (ISO 8637-1:2017)

I.S. EN ISO 8637-1:2020

Incorporating amendments/corrigenda/National Annexes issued since publication:

The National Standards Authority of Ireland (NSAI) produces the following categories of formal documents:

I.S. xxx: Irish Standard — national specification based on the consensus of an expert panel and subject to public consultation.

S.R. xxx: Standard Recommendation — recommendation based on the consensus of an expert panel and subject to public consultation.

SWiFT xxx: A rapidly developed recommendatory document based on the consensus of the participants of an NSAI workshop.

This document replaces/revises/consolidates the NSAI adoption of the document(s) indicated on the CEN/CENELEC cover/Foreword and the following National document(s):

NOTE: The date of any NSAI previous adoption may not match the date of its original CEN/CENELEC document.

This document is based on:

EN ISO 8637-1:2020

Published:

2020-04-15

This document was published under the authority of the NSAI and comes into effect on:

2020-05-03

ICS number:

11.040.40

NOTE: If blank see CEN/CENELEC cover page

NSAI
1 Swift Square,
Northwood, Santry
Dublin 9

T +353 1 807 3800
F +353 1 807 3838
E standards@nsai.ie
W NSAI.ie

Sales:
T +353 1 857 6730
F +353 1 857 6729
W standards.ie

Údarás um Chaighdeáin Náisiúnta na hÉireann

National Foreword

I.S. EN ISO 8637-1:2020 is the adopted Irish version of the European Document EN ISO 8637-1:2020, Extracorporeal systems for blood purification - Part 1: Haemodialysers, haemodiafilters, haemofilters and haemoconcentrators (ISO 8637-1:2017)

This document does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

For relationships with other publications refer to the NSAI web store.

Compliance with this document does not of itself confer immunity from legal obligations.

In line with international standards practice the decimal point is shown as a comma (,) throughout this document.

This page is intentionally left blank

EUROPEAN STANDARD

EN ISO 8637-1

NORME EUROPÉENNE

EUROPÄISCHE NORM

April 2020

ICS 11.040.40

Supersedes EN ISO 8637:2014

English Version

**Extracorporeal systems for blood purification - Part 1:
Haemodialysers, haemodiafilters, haemofilters and
haemoconcentrators (ISO 8637-1:2017)**

Systèmes extracorporels pour la purification du sang -
Partie 1: Hémodialyseurs, hémodiafiltres, hémo-filtres
et hémoconcentrateurs (ISO 8637-1:2017)

Kardiovaskuläre Implantate und extrakorporale
Systeme - Teil 1: Hämodialysatoren, Hämodiafilter,
Hämo-filter und Hämo-konzentratoren (ISO 8637-
1:2017)

This European Standard was approved by CEN on 8 April 2020.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

EN ISO 8637-1:2020 (E)

Contents	Page
European foreword.....	3

European foreword

This document (EN ISO 8637-1:2020) has been prepared by Technical Committee ISO/TC 150 "Implants for surgery" in collaboration with Technical Committee CEN/TC 205 "Non-active medical devices" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by October 2020, and conflicting national standards shall be withdrawn at the latest by April 2023.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 8637:2014.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 8637-1:2017 has been approved by CEN as EN ISO 8637-1:2020 without any modification.

This page is intentionally left blank

INTERNATIONAL STANDARD

ISO
8637-1

First edition
2017-11

Extracorporeal systems for blood purification —

Part 1: Haemodialysers, haemodiafilters, haemofilters and haemoconcentrators

Systèmes extracorporels pour la purification du sang —

*Partie 1: Hémodialyseurs, hémodiafiltres, hémofiltres et
hémococoncentrateurs*



Reference number
ISO 8637-1:2017(E)

© ISO 2017

ISO 8637-1:2017(E)



COPYRIGHT PROTECTED DOCUMENT

© ISO 2017, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Contents

	Page
Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 Requirements	4
4.1 Biological safety.....	4
4.2 Sterility.....	4
4.3 Non-pyrogenicity.....	4
4.4 Mechanical characteristics.....	4
4.4.1 Structural integrity.....	4
4.4.2 Blood compartment integrity.....	5
4.4.3 Haemodialyser, haemodiafilter and haemofilter blood compartment ports.....	5
4.4.4 Haemodialyser and haemodiafilter dialysis fluid compartment ports.....	8
4.4.5 Haemofilter filtrate ports.....	8
4.4.6 Haemoconcentrator blood and filtrate ports.....	8
4.5 Performance characteristics.....	8
4.5.1 Solute clearance of haemodialysers and haemodiafilters.....	8
4.5.2 Sieving coefficient of haemodiafilters, haemofilters and haemoconcentrators.....	9
4.5.3 Ultrafiltration coefficient.....	9
4.5.4 Volume of the blood compartment.....	9
4.5.5 Pressure drop of the blood compartment.....	9
4.6 Expiry date.....	9
5 Test methods	9
5.1 General.....	9
5.2 Biological safety.....	10
5.3 Sterility.....	10
5.4 Non-pyrogenicity.....	10
5.5 Mechanical characteristics.....	10
5.5.1 Structural integrity.....	10
5.5.2 Blood compartment integrity.....	11
5.5.3 Haemodialyser, haemodiafilter and haemofilter blood compartment ports.....	11
5.5.4 Haemodialyser and haemodiafilter dialysis fluid compartment ports.....	11
5.5.5 Haemofilter filtrate ports.....	11
5.5.6 Haemoconcentrator blood and filtrate ports.....	11
5.6 Performance characteristics.....	13
5.6.1 Clearance.....	13
5.6.2 Sieving coefficient of haemodiafilters, haemofilters and haemoconcentrators.....	15
5.6.3 Ultrafiltration coefficient.....	16
5.6.4 Volume of the blood compartment.....	17
5.6.5 Pressure drop of the blood compartment.....	17
5.7 Expiry date.....	17
6 Labelling	17
6.1 Labelling on the device.....	17
6.2 Labelling on unit containers.....	18
6.3 Labelling on the outer containers.....	18
6.4 Information to be given in the accompanying documentation.....	19
Bibliography	22

ISO 8637-1:2017(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical committee ISO/TC 150, *Implants for surgery*, Subcommittee SC 2, *Cardiovascular implants and extracorporeal systems*.

This first edition of ISO 8637-1:2017 cancels and replaces the third edition of ISO 8637:2010 and ISO 8637:2010/Amd1:2013, which has been technically revised. The following changes have been done:

— [Figure 1](#), [Figure 2](#), and [Figure 3](#) have been revised.

A list of all the parts in the ISO 8637 series can be found on the ISO website.

Introduction

This document is concerned with devices intended for haemodialysis, haemodiafiltration, haemofiltration and haemoconcentration in humans. The requirements specified in this document will help to ensure safety and satisfactory function.

It was not found practicable to specify materials of construction. This document therefore requires only that materials which have been used have been tested and that the methods and results are made available upon request. There is no intention to specify, or to set limits on, the performance characteristics of the devices because such restrictions are unnecessary for the qualified user and would limit the alternatives available when choosing a device for a specific application.

The dimensions of the blood ports and the dialysis fluid or filtrate ports have been specified to ensure compatibility of the device with the extracorporeal blood circuit specified in ISO 8637-2. The design and dimensions have been selected in order to minimize the risk of leakage of blood and the ingress of air.

This document reflects the consensus of physicians, manufacturers and other interested parties for devices that are approved for clinical use. Conformance with this document is voluntary and it does not supersede any national regulation.

Extracorporeal systems for blood purification —

Part 1:

Haemodialysers, haemodiafilters, haemofilters and haemoconcentrators

1 Scope

WARNING — Other requirements and other EU Directives may be applicable to the product(s) falling within the scope of this standard.

This document specifies requirements for haemodialysers, haemodiafilters, haemofilters and haemoconcentrators, hereinafter collectively referred to as “the device”, for use in humans.

This document does not apply to:

- extracorporeal blood circuits;
- plasmafilters;
- haemoperfusion devices;
- vascular access devices;
- blood pumps;
- pressure monitors for the extracorporeal blood circuit;
- air detection devices;
- systems to prepare, maintain or monitor dialysis fluid;
- systems or equipment intended to perform haemodialysis, haemodiafiltration, haemofiltration or haemoconcentration;
- reprocessing procedures and equipment.

NOTE Requirements for the extracorporeal blood circuit for haemodialysers, haemodiafilters and haemofilters are specified in ISO 8637-2.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 10993-1, *Biological evaluation of medical devices — Part 1: Evaluation and testing within a risk management process*

ISO 10993-4, *Biological evaluation of medical devices — Part 4: Selection of tests for interactions with blood*

ISO 10993-7, *Biological evaluation of medical devices — Part 7: Ethylene oxide sterilization residuals*

ISO 10993-11, *Biological evaluation of medical devices — Part 11: Tests for systemic toxicity*

This is a free preview. Purchase the entire publication at the link below:

[Product Page](#)

-
- [Looking for additional Standards? Visit Intertek Inform Infostore](#)
 - [Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation](#)
-