



NSAI
Standards

Irish Standard
I.S. EN 17389:2020

Stationary source emissions - Quality assurance and quality control procedures for automated dust arrestment plant monitors

I.S. EN 17389:2020

Incorporating amendments/corrigenda/National Annexes issued since publication:

The National Standards Authority of Ireland (NSAI) produces the following categories of formal documents:

I.S. xxx: Irish Standard — national specification based on the consensus of an expert panel and subject to public consultation.

S.R. xxx: Standard Recommendation — recommendation based on the consensus of an expert panel and subject to public consultation.

SWiFT xxx: A rapidly developed recommendatory document based on the consensus of the participants of an NSAI workshop.

This document replaces/revises/consolidates the NSAI adoption of the document(s) indicated on the CEN/CENELEC cover/Foreword and the following National document(s):

NOTE: The date of any NSAI previous adoption may not match the date of its original CEN/CENELEC document.

This document is based on:

EN 17389:2020

Published:

2020-05-06

This document was published under the authority of the NSAI and comes into effect on:

2020-06-10

ICS number:

13.040.40

NOTE: If blank see CEN/CENELEC cover page

NSAI
1 Swift Square,
Northwood, Santry
Dublin 9

T +353 1 807 3800
F +353 1 807 3838
E standards@nsai.ie
W NSAI.ie

Sales:
T +353 1 857 6730
F +353 1 857 6729
W standards.ie

Údarás um Chaighdeáin Náisiúnta na hÉireann

National Foreword

I.S. EN 17389:2020 is the adopted Irish version of the European Document EN 17389:2020, Stationary source emissions - Quality assurance and quality control procedures for automated dust arrestment plant monitors

This document does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

For relationships with other publications refer to the NSAI web store.

Compliance with this document does not of itself confer immunity from legal obligations.

In line with international standards practice the decimal point is shown as a comma (,) throughout this document.

This page is intentionally left blank

EUROPEAN STANDARD

EN 17389

NORME EUROPÉENNE

EUROPÄISCHE NORM

May 2020

ICS 13.040.40

English Version

Stationary source emissions - Quality assurance and quality control procedures for automated dust arrestment plant monitors

Émissions de sources fixes - Procédures d'assurance
qualité et de contrôle qualité applicables aux
analyseurs automatiques pour la surveillance des
systèmes de dépoussiérage

Emissionen aus stationären Quellen - Verfahren zur
Qualitätssicherung und Qualitätslenkung von
automatischen Geräten zur Überwachung von
Staubabscheidern

This European Standard was approved by CEN on 6 April 2020.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents	Page
European foreword.....	3
Introduction	4
1 Scope.....	5
2 Normative references.....	5
3 Terms and definitions	5
4 Symbols and abbreviations	10
4.1 Symbols.....	10
4.2 Abbreviations.....	10
5 Principle	10
6 Selection and installation of a suitable instrument.....	10
6.1 Selection	10
6.2 Installation	11
7 Configuration of instruments	11
7.1 General.....	11
7.2 Configuration of filter dust monitors.....	11
7.2.1 General.....	11
7.2.2 Service test.....	11
7.2.3 Parallel measurements with the SRM.....	12
7.2.4 Data evaluation and calibration function.....	13
7.2.5 Setting of dust alarm limits.....	14
7.3 Configuration of a filter leakage monitor	14
7.3.1 General.....	14
7.3.2 Service test.....	15
7.3.3 Setting of range and dust alarm limits	15
8 Ongoing quality assurance during operation	15
9 Annual surveillance test.....	16
Annex A (informative) Extension of the calibration range by changing the dust concentration	17
A.1 General.....	17
A.2 Changing plant load.....	17
A.3 Changing operation of dust arrestment plant	17
A.4 Injection of dust.....	17
Annex B (informative) Explanation of bag filter operation, bag cleaning and filter leakage monitoring.....	19
Figure B.1 — Dust pulses associated with bag cleaning superimposed on the baseline emission – mass concentration c as a function of time t (5 min intervals)	19
Figure B.2 — Changes in dust emissions associated with filter leak – mass concentration c as a function of time in days d of a calendar month.....	20
Bibliography.....	21

European foreword

This document (EN 17389:2020) has been prepared by Technical Committee CEN/TC 264 “Air quality”, the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by November 2020, and conflicting national standards shall be withdrawn at the latest by November 2020.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

According to the CEN-CENELEC Internal Regulations, the national standards organisations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

EN 17389:2020 (E)

Introduction

The type of monitoring of dust emissions at industrial plants depends on the monitoring objectives. In general, three monitoring objectives can be distinguished:

- Monitoring of plants with emission limit values (ELV), which require quantitative measurements with permanently installed automated measuring systems (AMS) providing measured values with a maximum permissible measurement uncertainty specified by legislation, in order to determine exceedances of the ELV and number of exceedances for specified time periods forms the first level of monitoring. Large combustion plants and waste incineration plants are examples of plants under this monitoring objective. The measurements are performed with AMS, which are performance tested and certified according to EN 15267-1, EN 15267-2 and EN 15267-3. The quality assurance measures for these AMS are laid down in EN 14181 and EN 13284-2.
- Monitoring of plants with ELV, which require qualitative measurements in order to demonstrate that the dust emissions are below the specified ELV and the dust arrestment plant works properly form the second level of monitoring. Foundries are typical examples of plants under this monitoring objective. The measurements are performed with filter dust monitors, which can be calibrated in mass concentration units (e.g. mg/m³) but have a larger measurement uncertainty than quantitative measurements. Filter dust monitors are performance tested and certified according to EN 15859. The quality assurance measures for these instruments are laid down in this document.
- Monitoring of dust arrestment plants with indicative measurements in order to indicate a possible problem with the dust arrestment plant by monitoring a change in the emissions level or a change in the magnitude of the dust pulses created by the cleaning process forms the third level of monitoring. The measurements are performed with filter leakage monitors which are performance tested and certified according to EN 15859. The quality assurance measures for these instruments are laid down in this document.

This document provides supporting information on the quality assurance and quality control procedures related to automated dust arrestment plant monitors, which cover both *filter dust monitors* and *filter leakage monitors*.

For the purposes of this document, the term *instrument* is used to encompass both types of automated dust arrestment plant monitors. The terms *filter dust monitor* and *filter leakage monitor* are only used where it is necessary to distinguish between the two types.

This document includes provisions for the selection, installation, configuration, ongoing quality assurance and annual surveillance test of automated dust arrestment plant monitors.

This document supports requirements for filter leakage monitors and filter dust monitors specified e.g. in national legislation or in a number of industrial sector BREF Documents.

1 Scope

This document specifies the quality assurance and quality control procedures related to automated dust arrestment plant monitors.

This document applies to two types of instruments commonly used for dust arrestment plant control purposes:

- filter dust monitors that are configured in mass concentration units (e.g. mg/m³) and are used for dust arrestment control purposes;
- filter leakage monitors that indicate a change in the emission levels or a change in the magnitude of the dust pulses created by the cleaning process of the dust arrestment plant.

This document applies to instruments certified according to the requirements of EN 15859.

This document provides information on the configuration, ongoing quality assurance (with automatic internal zero point and reference point checks) and annual surveillance tests of instruments. This ensures that the instrument is providing information to demonstrate that the dust arrestment plant is working correctly and controlling dust pollution to the required levels.

The configuration of the alarm levels of filter dust monitors is performed by parallel measurements with the standard reference method in EN 13284-1.

This document specifies the set-up of filter leakage monitors used to monitor a change in response caused by deterioration in the operation of the dust arrestment plant.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 13284-1, *Stationary source emissions - Determination of low range mass concentration of dust - Part 1: Manual gravimetric method*

EN 15259, *Air quality - Measurement of stationary source emissions - Requirements for measurement sections and sites and for the measurement objective, plan and report*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <https://www.iso.org/obp>

3.1

dust

particles, of any shape, structure or density, dispersed in the gas phase at the sampling point conditions which may be collected by filtration under specified conditions after representative sampling of the gas to be analysed

Note 1 to entry: Adapted from EN 13284-1:2017, 3.1.

This is a free preview. Purchase the entire publication at the link below:

[Product Page](#)

-
- [Looking for additional Standards? Visit Intertek Inform Infostore](#)
 - [Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation](#)
-