



**NSAI**  
Standards

Irish Standard  
I.S. EN ISO 7886-3:2020

# Sterile hypodermic syringes for single use - Part 3: Auto-disabled syringes for fixed-dose immunization (ISO 7886-3:2020)

**I.S. EN ISO 7886-3:2020**

*Incorporating amendments/corrigenda/National Annexes issued since publication:*

The National Standards Authority of Ireland (NSAI) produces the following categories of formal documents:

I.S. xxx: Irish Standard — national specification based on the consensus of an expert panel and subject to public consultation.

S.R. xxx: Standard Recommendation — recommendation based on the consensus of an expert panel and subject to public consultation.

SWiFT xxx: A rapidly developed recommendatory document based on the consensus of the participants of an NSAI workshop.

*This document replaces/revises/consolidates the NSAI adoption of the document(s) indicated on the CEN/CENELEC cover/Foreword and the following National document(s):*

*NOTE: The date of any NSAI previous adoption may not match the date of its original CEN/CENELEC document.*

*This document is based on:*

EN ISO 7886-3:2020

*Published:*

2020-05-20

*This document was published  
under the authority of the NSAI  
and comes into effect on:*

2020-06-19

ICS number:

11.040.25

NOTE: If blank see CEN/CENELEC cover page

NSAI  
1 Swift Square,  
Northwood, Santry  
Dublin 9

T +353 1 807 3800  
F +353 1 807 3838  
E standards@nsai.ie  
W NSAI.ie

Sales:  
T +353 1 857 6730  
F +353 1 857 6729  
W standards.ie

Údarás um Chaighdeáin Náisiúnta na hÉireann

## National Foreword

I.S. EN ISO 7886-3:2020 is the adopted Irish version of the European Document EN ISO 7886-3:2020, Sterile hypodermic syringes for single use - Part 3: Auto-disabled syringes for fixed-dose immunization (ISO 7886-3:2020)

This document does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

For relationships with other publications refer to the NSAI web store.

**Compliance with this document does not of itself confer immunity from legal obligations.**

*In line with international standards practice the decimal point is shown as a comma (,) throughout this document.*

This page is intentionally left blank

EUROPEAN STANDARD  
NORME EUROPÉENNE  
EUROPÄISCHE NORM

**EN ISO 7886-3**

May 2020

ICS 11.040.25

Supersedes EN ISO 7886-3:2009

English Version

**Sterile hypodermic syringes for single use - Part 3: Auto-disabled syringes for fixed-dose immunization (ISO 7886-3:2020)**

Seringues hypodermiques stériles, non réutilisables -  
Partie 3: Seringues autobloquantes pour vaccination à  
dose fixe (ISO 7886-3:2020)

Sterile Einmalspritzen für medizinische Zwecke - Teil  
3: Selbstblockierende Spritzen für die Injektion mit  
fixer Impfstoffdosis (ISO 7886-3:2020)

This European Standard was approved by CEN on 24 April 2020.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels**

**EN ISO 7886-3:2020 (E)**

**Contents**

Page

<b>European foreword.....</b>	<b>3</b>
-------------------------------	----------

## **European foreword**

This document (EN ISO 7886-3:2020) has been prepared by Technical Committee ISO/TC 84 "Devices for administration of medicinal products and catheters" in collaboration with Technical Committee CEN/TC 205 "Non-active medical devices" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by November 2020, and conflicting national standards shall be withdrawn at the latest by November 2020.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 7886-3:2009.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

## **Endorsement notice**

The text of ISO 7886-3:2020 has been approved by CEN as EN ISO 7886-3:2020 without any modification.

This page is intentionally left blank



# INTERNATIONAL STANDARD

**ISO  
7886-3**

Second edition  
2020-05

---

---

## **Sterile hypodermic syringes for single use —**

### **Part 3: Auto-disabled syringes for fixed-dose immunization**

*Seringues hypodermiques stériles, non réutilisables —*

*Partie 3: Seringues autobloquantes pour vaccination à dose fixe*



Reference number  
ISO 7886-3:2020(E)

© ISO 2020

**ISO 7886-3:2020(E)**



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2020

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva  
Phone: +41 22 749 01 11  
Fax: +41 22 749 09 47  
Email: [copyright@iso.org](mailto:copyright@iso.org)  
Website: [www.iso.org](http://www.iso.org)

Published in Switzerland

# Contents

	Page
<b>Foreword</b> .....	<b>v</b>
<b>Introduction</b> .....	<b>vi</b>
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>1</b>
<b>3 Terms and definitions</b> .....	<b>1</b>
<b>4 Nomenclature</b> .....	<b>2</b>
<b>5 General requirements</b> .....	<b>3</b>
<b>6 Extraneous matter</b> .....	<b>4</b>
6.1 General.....	4
6.2 Limits for acidity or alkalinity.....	4
6.3 Limits for extractable metals.....	4
<b>7 Lubricant</b> .....	<b>4</b>
<b>8 Tolerance on nominal capacity</b> .....	<b>4</b>
<b>9 Graduated scale</b> .....	<b>5</b>
9.1 Scale.....	5
9.2 Position of scale.....	5
<b>10 Barrel</b> .....	<b>5</b>
10.1 Dimensions.....	5
10.2 Barrel flanges.....	5
<b>11 Plunger stopper/plunger assembly</b> .....	<b>5</b>
11.1 Design.....	5
11.2 Fit of the plunger stopper/plunger in the barrel.....	6
11.3 Fiducial line.....	6
<b>12 Needle</b> .....	<b>6</b>
12.1 General.....	6
12.2 Integrated needle.....	6
12.3 Non-integrated needle.....	6
12.4 Sharps protection features.....	6
<b>13 Performance</b> .....	<b>7</b>
13.1 General.....	7
13.2 Dead space.....	7
13.3 Freedom from air and liquid leakage.....	7
13.4 Auto-disable syringe feature.....	7
13.5 Performance after shipping.....	7
<b>14 Packaging</b> .....	<b>8</b>
14.1 Unit packaging providing sterile barrier.....	8
14.2 Multiple unit pack.....	8
14.3 User packaging.....	8
<b>15 Information supplied by the manufacturer</b> .....	<b>8</b>
15.1 General.....	8
15.2 Syringes.....	8
15.3 Unit packaging providing sterile barrier.....	8
15.4 User packaging.....	9
15.5 Storage containers.....	9
15.6 Transport wrapping.....	10
<b>Annex A (normative) Method for preparation of extracts</b> .....	<b>11</b>
<b>Annex B (informative) Test method for forces required to operate piston</b> .....	<b>12</b>

**ISO 7886-3:2020(E)**

<b>Annex C (normative) Test method for testing auto-disable syringe feature .....</b>	<b>14</b>
<b>Bibliography .....</b>	<b>15</b>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by Technical Committee ISO/TC 84, *Devices for administration of medicinal products and catheters*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 205, *Non-active medical devices*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This second edition cancels and replaces the first edition (ISO 7886-3:2005), which has been technically revised. The main changes compared to the previous edition are as follows:

— update of the references, mainly ISO 7886-1:2017.

A list of all parts in the ISO 7886 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).

## **ISO 7886-3:2020(E)**

### **Introduction**

ISO 7886 was first published in 1984. It was subsequently decided to divide it into two parts: ISO 7886-1 retaining essentially the scope of ISO 7886:1984 and ISO 7886-2 being applicable to sterile, single-use syringes for use with power-driven pumps.

The preparation of this document was recognized as a high priority to prevent the reuse of fixed dose immunization syringes. Reuse of injection equipment in the absence of sterilization has increasingly led to transmission of blood-borne pathogens.

The World Health Organization (WHO) had produced a specification for syringes that are rendered inactive after one use (commonly referred to as “auto-disabled” syringes). It was agreed that an additional part of the ISO 7886 series would be needed to cover “auto-disabled” syringes, while leaving in place ISO 7886-1 and ISO 7886-2 without modification, as a large number of devices in common use would not be intended to conform with the auto-disable properties suggested.

It has been discussed to limit the syringe types to only comprise the type having an auto-disable syringe feature that is automatically activated and remains effective from the time that the injection is commenced. An assessment of potential hazards based only on hypothetical use indicates that the type having an auto-disable syringe feature that is automatically activated and remains effective from the time of the injection being initiated is potentially safer than the other types. However, no consensus could be reached on either deleting types or retaining them, as no reliable risk data from field use exists at present. It was therefore agreed to retain all types and restrict this revision to alignment with ISO 7886-1:2017 and initiate a new revision if new field studies or incident reports indicate a need for a revision.

It is recognized that syringes designed to reduce the risk of needle stick injuries can also conform with this document.

In some countries national regulations might take precedence over the requirements in this document.

Guidance on transition periods for implementing the requirements of this document is given in ISO/TR 19244.

# Sterile hypodermic syringes for single use —

## Part 3:

## Auto-disabled syringes for fixed-dose immunization

### 1 Scope

This document specifies the properties and performance of sterile single-use hypodermic syringes with an auto-disable syringe feature intended to deliver a fixed dose of vaccine immediately after filling. The syringes can be made of plastic, rubber or other materials and can be with or without needle and needle protection feature.

This document does not specify the design of the auto-disable syringe feature.

This document is not applicable to syringes for use with insulin (covered by ISO 8537), syringes for use with power-driven syringe pumps (covered by ISO 7886-2), reuse prevention syringes (covered by ISO 7886-4) or syringes designed to be prefilled (covered by the ISO 11040 series). It does not address compatibility with injection fluids/vaccines.

### 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 7864:2016, *Sterile hypodermic needles for single use — Requirements and test methods*

ISO 7886-1:2017, *Sterile hypodermic syringes for single use — Part 1: Syringes for manual use*

ISO 8537:2016, *Sterile single-use syringes, with or without needle, for insulin*

ISO 9626, *Stainless steel needle tubing for the manufacture of medical devices — Requirements and test methods*

ISO 10993-1, *Biological evaluation of medical devices — Part 1: Evaluation and testing within a risk management process*

ISO 14971, *Medical devices — Application of risk management to medical devices*

ISO 23908, *Sharps injury protection — Requirements and test methods — Sharps protection features for single-use hypodermic needles, introducers for catheters and needles used for blood sampling*

ISO 80369-7, *Small-bore connectors for liquids and gases in healthcare applications — Part 7: Connectors for intravascular or hypodermic applications*

ASTM D999-01, *Standard methods for vibration testing of shipping containers*

ASTM D5276-98, *Standard test method for drop test of loaded containers by free fall*

### 3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 7886-1 and ISO 8537 and the following apply.

This is a free preview. Purchase the entire publication at the link below:

[Product Page](#)

- 
- Looking for additional Standards? Visit Intertek Inform Infostore
  - Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation
-