



NSAI
Standards

Irish Standard
I.S. EN 60601-1-3:2008

Medical electrical equipment -- Part 1
-3: General requirements for basic
safety and essential performance -
Collateral Standard: Radiation
protection in diagnostic X-ray
equipment (IEC 60601-1-3:2008 (EQV))

I.S. EN 60601-1-3:2008

Incorporating amendments/corrigenda issued since publication:

EN 60601-1-3:2008/AC:2010

This document replaces:
EN 60601-1-3:1994

This document is based on:
EN 60601-1-3:2008
EN 60601-1-3:1994

Published:
24 April, 2008
14 September, 1994

This document was published
under the authority of the NSAI and
comes into effect on:

28 July, 2008

ICS number:
11.040.50
13.280

NSAI
1 Swift Square,
Northwood, Santry
Dublin 9

T +353 1 807 3800
F +353 1 807 3838
E standards@nsai.ie
W NSAI.ie

Sales:
T +353 1 857 6730
F +353 1 857 6729
W standards.ie

Údarás um Chaighdeáin Náisiúnta na hÉireann

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN 60601-1-3

April 2008

ICS 11.040.50; 13.280

Supersedes EN 60601-1-3:1994
Incorporates corrigendum March 2010

English version

**Medical electrical equipment -
Part 1-3: General requirements for basic safety
and essential performance -
Collateral Standard: Radiation protection in diagnostic X-ray equipment
(IEC 60601-1-3:2008)**

Appareils électromédicaux -
Partie 1-3: Exigences générales
pour la sécurité de base
et les performances essentielles -
Norme collatérale: Radioprotection
dans les appareils à rayonnement X
de diagnostic
(CEI 60601-1-3:2008)

Medizinische elektrische Geräte -
Teil 1-3: Allgemeine Festlegungen
für die Sicherheit einschließlich
der wesentlichen Leistungsmerkmale -
Ergänzungsnorm: Strahlenschutz
von diagnostischen Röntgengeräten
(IEC 60601-1-3:2008)

This European Standard was approved by CENELEC on 2008-03-01. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the Central Secretariat has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

CENELEC

European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

Central Secretariat: rue de Stassart 35, B - 1050 Brussels

Foreword

The text of document 62B/673/FDIS, future edition 2 of IEC 60601-1-3, prepared by SC 62B, Diagnostic imaging equipment, of IEC TC 62, Electrical equipment in medical practice, was submitted to the IEC-CENELEC parallel vote and was approved by CENELEC as EN 60601-1-3 on 2008-03-01.

The following date was fixed:

- latest date by which the EN has to be implemented
at national level by publication of an identical
national standard or by endorsement (dop) 2008-12-01
- latest date by which the national standards conflicting
with the EN have to be withdrawn (dow) 2012-06-01

This European Standard supersedes EN 60601-1-3:1994.

This EN 60601-1-3 has been restructured and aligned to EN 60601-1:2006 and focussed on general requirements for RADIATION PROTECTION that apply to all diagnostic X-RAY EQUIPMENT. Requirements particular to specific equipment have been removed and will be covered in particular standards. For a description of the changes, see the mapping in Annex C.

This European Standard has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association and covers essential requirements of EC Directive 93/42/EEC. See Annex ZZ.

This European Standard constitutes a collateral standard to EN 60601-1:2006, hereafter referred to as the general standard.

In the 60601 series of publications, collateral standards specify general requirements for safety applicable to:

- a subgroup of MEDICAL ELECTRICAL EQUIPMENT (e.g. RADIOLOGICAL equipment); or
- a specific characteristic of all MEDICAL ELECTRICAL EQUIPMENT, not fully addressed in the general standard (e.g. alarm systems).

In this collateral standard, the following print types are used:

- requirements and definitions: in roman type;
- *test specifications: in italic type;*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type;
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS COLLATERAL STANDARD OR AS NOTED: IN SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the thirteen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes Subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

Clauses, subclauses and definitions for which a rationale is provided in informative Annex A are marked with an asterisk (*).

Annexes ZA and ZZ have been added by CENELEC.

Endorsement notice

The text of the International Standard IEC 60601-1-3:2008 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following notes have to be added for the standards indicated:

IEC 60601-2-7	NOTE	Harmonized as EN 60601-2-7:1998 (not modified).
IEC 60601-2-28	NOTE	Harmonized as EN 60601-2-28:1993 (not modified).
IEC 60601-2-29	NOTE	Harmonized as EN 60601-2-29:1999 (not modified).
IEC 60601-2-32	NOTE	Harmonized as EN 60601-2-32:1994 (not modified).
IEC 60601-2-43	NOTE	Harmonized as EN 60601-2-43:2000 (not modified).
IEC 60601-2-44	NOTE	Harmonized as EN 60601-2-44:2001 (not modified).
IEC 60601-2-45	NOTE	Harmonized as EN 60601-2-45:2001 (not modified).
IEC 60580	NOTE	Harmonized as EN 60580:2000 (not modified).
IEC 60627	NOTE	Harmonized as EN 60627:2001 (not modified).
IEC 61262	NOTE	Harmonized in EN 61262 series (not modified).
IEC 62220	NOTE	Harmonized in EN 62220 series (not modified).
IEC 62220-1	NOTE	Harmonized as EN 62220-1:2003 (not modified).

The contents of the corrigendum of March 2010 have been included in this copy.

Annex ZA (normative)

Normative references to international publications with their corresponding European publications

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE When an international publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 60336	- ¹⁾	Medical electrical equipment - X-ray tube assemblies for medical diagnosis - Characteristics of focal spots	EN 60336	2005 ²⁾
IEC 60522	1999	Determination of the permanent filtration of X-ray tube assemblies	EN 60522	1999
IEC 60601-1	2005	Medical electrical equipment - Part 1: General requirements for basic safety and essential performance	EN 60601-1	2006
IEC/TR 60788	2004	Medical electrical equipment - Glossary of defined terms	-	-
ISO 497	- ¹⁾	Guide to the choice of series of preferred numbers and series containing more rounded values of preferred numbers	-	-

¹⁾ Undated reference.

²⁾ Valid edition at date of issue.

This is a free preview. Purchase the entire publication at the link below:

[Product Page](#)

-
- Looking for additional Standards? Visit Intertek Inform Infostore
 - Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation
-