



NSAI
Standards

Irish Standard
I.S. EN 61010-2-201:2013

Safety requirements for electrical
equipment for measurement, control
and laboratory use -- Part 2-201:
Particular requirements for control
equipment (IEC 61010-2-201:2013
(EQV))

I.S. EN 61010-2-201:2013

Incorporating amendments/corrigenda issued since publication:

EN 61010-2-201:2013/AC:2013

The National Standards Authority of Ireland (NSAI) produces the following categories of formal documents:

I.S. xxx: Irish Standard – national specification based on the consensus of an expert panel and subject to public consultation.

S.R. xxx: Standard Recommendation - recommendation based on the consensus of an expert panel and subject to public consultation.

SWiFT xxx: A rapidly developed recommendatory document based on the consensus of the participants of an NSAI workshop.

<i>This document replaces:</i>	<i>This document is based on:</i> EN 61010-2-201:2013	<i>Published:</i> 10 May, 2013
This document was published under the authority of the NSAI and comes into effect on: 27 May, 2013		ICS number: 13.110; 17.020 19.020 25.040.40
NSAI 1 Swift Square, Northwood, Santry Dublin 9	T +353 1 807 3800 F +353 1 807 3838 E standards@nsai.ie W NSAI.ie	Sales: T +353 1 857 6730 F +353 1 857 6729 W standards.ie
Údarás um Chaighdeáin Náisiúnta na hÉireann		



Corrigendum to EN 61010-2-201:2013

English version

Title page

In the header of the title page, **delete** "Supersedes EN 61131-2:2007 (partially)".

Foreword

In the foreword, **delete** the sentence "This document partially supersedes EN 61131-2:2007."

September 2013

This page is intentionally left BLANK.

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN 61010-2-201

May 2013

ICS 13.110; 17.020; 19.020; 25.040.40

Supersedes EN 61131-2:2007 (partially)

English version

**Safety requirements for electrical equipment for measurement, control
and laboratory use -**

**Part 2-201: Particular requirements for control equipment
(IEC 61010-2-201:2013)**

Règles de sécurité pour appareils
électriques de mesure, de régulation et
de laboratoire -
Partie 2-201: Exigences particulières pour
les équipements de commande
(CEI 61010-2-201:2013)

Sicherheitsbestimmungen für elektrische
Mess-, Steuer-, Regel- und Laborgeräte -
Teil 2-201: Besondere Anforderungen für
Steuer- und Regelgeräte
(IEC 61010-2-201:2013)

This European Standard was approved by CENELEC on 2013-04-01. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

CENELEC

European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

Management Centre: Avenue Marnix 17, B - 1000 Brussels

Foreword

The text of document 65/515/FDIS, future edition 1 of IEC 61010-2-201, prepared by IEC TC 65 "Industrial-process measurement, control and automation" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN 61010-2-201:2013.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2014-01-01
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2016-04-01

This document partially supersedes EN 61131-2:2007.

This Part 2-201 is intended to be used in conjunction with EN 61010-1:2010. Consideration may be given to future editions of, or amendments to, EN 61010-1.

This Part 2-201 supplements or modifies the corresponding clauses in EN 61010-1 so as to convert that publication into the European standard: *Particular requirements for control equipment*.

Where a particular subclause of Part 1 is not mentioned in this part 2, that subclause applies as far as is reasonable. Where this part states "addition", "modification", "replacement", or "deletion", the relevant requirement, test specification or note in Part 1 should be adapted accordingly.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC [and/or CEN] shall not be held responsible for identifying any or all such patent rights.

This standard covers the Principle Elements of the Safety Objectives for Electrical Equipment Designed for Use within Certain Voltage Limits (LVD - 2006/95/EC).

Endorsement notice

The text of the International Standard IEC 61010-2-201:2013 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following notes have to be added for the standards indicated:

IEC 60079 series	NOTE	Harmonised in EN 60079 series.
IEC 60364 series	NOTE	Harmonised in HD 60364 series.
IEC 60364-4-41	NOTE	Harmonised as HD 60364-4-41.
IEC 60664-5:2007	NOTE	Harmonised as EN 60664-5:2007 (not modified).
IEC 60715:1981 + A1:1995	NOTE	Harmonised as EN 60715:2001 (not modified).
IEC 60721-2-3:1987	NOTE	Harmonised as HD 478.2.3 S1:1990 (not modified).
IEC 61131-2:2007	NOTE	Harmonised as EN 61131-2:2007 (not modified).
IEC 61131-6:2012	NOTE	Harmonised as EN 61131-6:2012 (not modified).
IEC 61140:2001	NOTE	Harmonised as EN 61140:2002 (not modified).
IEC 61326 series	NOTE	Harmonised in EN 61326 series.

IEC 61508 series	NOTE	Harmonised in EN 61508 series.
IEC 61643 series	NOTE	Harmonised in EN 61643 series.
IEC 61643-21	NOTE	Harmonised as EN 61643-21.
IEC 61643-311	NOTE	Harmonised as EN 61643-311.
IEC 61643-321	NOTE	Harmonised as EN 61643-321.
IEC 61643-331	NOTE	Harmonised as EN 61643-331.
IEC 61800 series	NOTE	Harmonised in EN 61800 series.
IEC 62133:2002	NOTE	Harmonised as EN 62133:2003 (not modified).
IEC 62368 series	NOTE	Harmonised in EN 62368 series.

Annex ZA (normative)

Normative references to international publications with their corresponding European publications

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE When an international publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

Addition to Annex ZA of EN 61010-1:2010:

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 60068-2-31	2008	Environmental testing - Part 2-31: Tests - Test Ec: Rough handling shocks, primarily for equipment-type specimens	EN 60068-2-31	2008
IEC 60384-14	2005	Fixed capacitors for use in electronic equipment - Part 14: Sectional specification - Fixed capacitors for electromagnetic interference suppression and connection to the supply mains	EN 60384-14	2005
IEC 60664-1	2007	Insulation coordination for equipment within low-voltage systems - Part 1: Principles, requirements and tests	EN 60664-1	2007
IEC 60695-2-11 + corr. January	2000 2001	Fire hazard testing - Part 2-11: Glowing/hot-wire based test methods - Glow-wire flammability test method for end-products	EN 60695-2-11	2001
IEC 60947-5-1	2003	Low-voltage switchgear and controlgear - Part 5-1: Control circuit devices and switching elements - Electromechanical control circuit devices	EN 60947-5-1 + corr. July	2004 2005
IEC 60947-7-1	2009	Low-voltage switchgear and controlgear - Part 7-1: Ancillary equipment - Terminal blocks for copper conductors	EN 60947-7-1	2009
IEC 61010-1 + corr. May	2010 2011	Safety requirements for electrical equipment for measurement, control and laboratory use - Part 1: General requirements	EN 61010-1	2010
IEC 61010-2-030	-	Safety requirements for electrical equipment for measurement, control and laboratory use - Part 2-030: Particular requirements for testing and measuring circuits	EN 61010-2-030	-
IEC 61051-2	1991	Varistors for use in electronic equipment - Part 2: Sectional specification for surge suppression varistors	-	-

CONTENTS

FOREWORD.....	5
INTRODUCTION.....	7
1 Scope and object.....	8
1.1.1 Equipment included in scope	8
1.1.2 Equipment excluded from scope	9
1.2.1 Aspects included in scope	9
1.2.2 Aspects excluded from scope	9
2 Normative references	10
3 Terms and definitions	10
4 Tests	12
4.1 General	12
4.3.2 State of equipment	12
4.4 Testing in single fault condition	12
5 Marking and documentation.....	14
5.4.3 Equipment installation	14
6 Protection against electric shock	14
6.1.2 Exceptions.....	14
6.2.1 General	14
6.2.2 Examination	14
6.2.3 Openings above parts that are hazardous live	14
6.2.4 Openings for pre-set controls	15
6.2.101 Accessibility of interfaces/ports/terminals	15
6.2.102 Control equipment	17
6.6.1 General	20
6.6.2 Terminals for external circuits.....	20
6.6.3 Circuits with terminals which are hazardous live	20
6.6.4 Terminals for stranded conductors.....	20
6.7.2 Insulation for mains circuits of overvoltage category II with a nominal supply voltage up to 300 V	22
6.7.3 Insulation for secondary circuits derived from mains circuits of overvoltage category II up to 300 V	24
6.7.101 Insulation for field wiring terminals of overvoltage category II with a nominal voltage up to 1 000 V	26
6.8.3 Test procedures	26
6.10 Connection to the mains supply source and connections between parts of equipment	26
6.11 Disconnection from supply source	27
7 Protection against mechanical hazards.....	27
7.1.101 Open and panel mounted equipment	27
7.2 Sharp edges.....	27
7.3.3 Risk assessment for mechanical hazards to body parts	27
7.3.4 Limitation of force and pressure.....	28
7.3.5 Gap limitations between moving parts	28
7.7 Expelled parts	28
8 Resistance to mechanical stresses	28
8.1.101 Open equipment	28

I.S. EN 61010-2-201:2013

61010-2-201 © IEC:2013

– 3 –

8.1.102	Panel mounted equipment	28
8.2.2	Impact test	28
8.3	Drop test	28
8.3.1	Equipment other than hand-held equipment and direct plug-in equipment	29
8.3.2	Hand-held equipment and direct plug-in equipment	29
9	Protection against the spread of fire	29
9.2	Eliminating or reducing the sources of ignition within the equipment.....	29
9.3.2	Constructional requirements	29
10	Equipment temperature limits and resistance to heat.....	30
10.1	Surface temperature limits for protection against burns	30
10.3	Other temperature measurements	31
10.4.1	General	31
10.4.2	Temperature measurement of heating equipment	32
10.4.3	Equipment intended for installation in a cabinet or a wall.....	32
10.5.2	Non-metallic enclosures	33
11	Protection against hazards from fluids	33
11.6	Specially protected equipment.....	33
12	Protection against radiation, including laser sources, and against sonic and ultrasonic pressure	33
13	Protection against liberated gases and substances, explosion and implosion	34
13.1	Poisonous and injurious gases and substances	34
13.2.1	Components	34
13.2.2	Batteries and battery charging	34
14	Components and subassemblies	34
14.101	Components bridging insulation	34
14.101.1	Capacitors	34
14.101.2	Surge suppressors	34
14.102	Switching devices	35
15	Protection by interlocks	35
16	Hazards resulting from application.....	35
17	Risk assessment	35
Annexes	35
Annex F (normative)	Routine tests	36
Annex L (informative)	Index of defined terms	38
Annex AA (informative)	General approach to safety for control equipment	39
Annex BB (informative)	System drawing of isolation boundaries	41
Annex CC (informative)	Historical techniques for secondary circuits	49
Annex DD (informative)	Cross references between IEC 61010-2-201 and IEC 61010-1:2010 or IEC 61131-2:2007	53
Bibliography	54
Figure 101	– Typical interface/port diagram of control equipment.....	16
Figure 102	– Requirements for insulation between separate circuits and between circuits and accessible conductive parts	21
Figure 103	– Mechanical hazards requirements for panel mounted equipment	27
Figure 104	– Safety enclosure with HMI installed through a wall	30

Figure 105 – Panel mounted HMI device extending through the wall of a cabinet.....	33
Figure AA.1 – Control equipment access and safety concerns	39
Figure BB.1 – Typical system enclosure layout	42
Figure BB.2 – Simplified system schematic.....	43
Figure BB.3 – Hazard situation of the control equipment.....	44
Figure BB.4 – Application of the standard to the control equipment safety drawing	45
Figure BB.5 – Reinforced insulation	46
Figure BB.6 – Basic insulation	47
Figure BB.7 – Reinforced insulation, basic insulation and limiting impedance	48
 Table 101 – Overload test circuit values	 13
Table 102 – Endurance test circuit values.....	13
Table 103 – Operator accessibility for open and enclosed equipment	16
Table 4 – Clearance and creepage distances for mains circuits of overvoltage category II up to 300 V	23
Table 5 – Test voltages for solid insulation between mains and between mains and secondary circuits overvoltage category II up to 300 V ^d	24
Table 6 – Clearances and test voltages for secondary circuits derived from mains circuits of overvoltage category II up to 300 V.....	25
Table 104 – Minimum creepages and clearances in air of overvoltage category II up to 1 000 V at field-wiring terminals.....	26
Table 105 – Drop tests.....	29
Table 19 – Surface temperature limits, under normal conditions	31
Table CC.1 – Limits of output current and output power for inherently limited power sources.....	51
Table CC.2 – Limits of output current, output power and ratings for over-current protective devices for non-inherently limited power sources.....	52
Table DD.1 – Cross-references between IEC 61010-2-201 and IEC 61010-1 or IEC 61131-2	53

INTERNATIONAL ELECTROTECHNICAL COMMISSION

**SAFETY REQUIREMENTS FOR ELECTRICAL EQUIPMENT
FOR MEASUREMENT, CONTROL, AND LABORATORY USE –****Part 2-201: Particular requirements for control equipment**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 61010-2-201 has been prepared by IEC technical committee 65: Industrial-process measurement, control and automation.

The text of this standard is based on the following documents:

FDIS	Report on voting
65/515/FDIS	65/521/RVD

Full information on the voting for the approval of this standard can be found in the report on voting indicated in the above table.

This Part 2-201 is intended to be used in conjunction with IEC 61010-1. It was established on the basis of the third edition (2010). Consideration may be given to future editions of, or amendments to, IEC 61010-1.

This Part 2-201 supplements or modifies the corresponding clauses in IEC 61010-1 so as to convert that publication into the IEC standard: *Particular requirements for control equipment*.

Where a particular subclause of Part 1 is not mentioned in this part 2, that subclause applies as far as is reasonable. Where this part states “addition”, “modification”, “replacement”, or “deletion”, the relevant requirement, test specification or note in Part 1 should be adapted accordingly.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

It has the status of a basic safety publication in accordance with IEC Guide 104.

A list of all parts in the IEC 61010 series, published under the general title *Safety requirements for electrical equipment for measurement, control and laboratory use*, can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

IMPORTANT – The 'colour inside' logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

INTRODUCTION

This IEC 61010-2-201 document constitutes Part 2-201 of a planned series of standards on industrial-process measurement, control and automation equipment.

This part specifies the complete safety requirements for control equipment (e.g. programmable controller (PLC)), the components of Distributed Control Systems, I/O devices, Human Machine Interface (HMI)).

Safety terms of general use are defined in IEC 61010-1. More specific terms are defined in each part.

This part incorporates the safety related requirements of Programmable Controllers.

Annex DD provides a cross reference between clauses of this standard and those of IEC 61010-1 or IEC 61131-2:2007.

SAFETY REQUIREMENTS FOR ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL, AND LABORATORY USE –

Part 2-201: Particular requirements for control equipment

1 Scope and object

This clause of Part 1 is applicable, except as follows.

1.1.1 Equipment included in scope

Replacement:

This part of IEC 61010 specifies safety requirements and related verification tests for control equipment of the following types:

- Programmable controllers (PLC and PAC);
- the components of Distributed Control Systems (DCS);
- the components of remote I/O – systems;
- industrial PC (computers) and Programming and Debugging Tools (PADTs);
- Human-Machine Interfaces (HMI);
- any product performing the function of control equipment and/or their associated peripherals,

which have as their intended use the control and command of machines, automated manufacturing and industrial processes, e.g. discrete and continuous control.

Components of the above named equipment and in the scope of this standard are:

- (auxiliary) stand-alone power supplies;
- peripherals such as digital and analogue I/O, remote-I/O;
- industrial network equipment.

Control equipment and their associated peripherals are intended to be used in an industrial environment and may be provided as open or enclosed equipment.

NOTE 1 Control equipment intended also for use in other environments or for other purposes (example; for use in building installations to control light or other electrical installations, or for use on cars, trains or ships) can have additional conformity requirements defined by the safety standard(s) for these applications. These requirements can involve as example: insulation, spacings and power restrictions.

NOTE 2 Computing devices and similar equipment within the scope of IEC 60950 (planned to be replaced by IEC 62368) and conforming to its requirements are considered to be suitable for use with control equipment within the scope of this standard. However, some of the requirements of IEC 60950 for resistance to moisture and liquids are less stringent than those in IEC 61010-1:2010, 5.4.4 second paragraph.

Control equipment covered in this standard is intended for use in overvoltage category II (IEC 60664-1) in low-voltage installations, where the rated equipment supply voltage does not exceed a.c. 1 000 V r.m.s. (50/60 Hz), or d.c. 1 500 V.

NOTE 3 If equipment in the scope of this part is applied to overvoltage category III and IV installations, then the requirements of Annex K of Part 1 apply.

The requirements of ISO/IEC Guide 51 and IEC Guide 104, as they relate to this Part, are incorporated herein.

1.1.2 Equipment excluded from scope

Replacement:

This standard does not deal with aspects of the overall automated system, e.g. a complete assembly line. Control equipment (e.g. DCS and PLC), their application program and their associated peripherals are considered as components (components in this context are items which perform no useful function by themselves) of an overall automated system.

Since control equipment (e.g. DCS and PLC) are component devices, safety considerations for the overall automated system including installation and application are beyond the scope of this standard. Refer to IEC 60364 series of standards or applicable national/local regulations for electrical installation and guidelines.

1.2.1 Aspects included in scope

Replacement:

The purpose of the requirements of this standard is to ensure that all hazards to the operator, service personnel and the surrounding area are reduced to a tolerable level.

NOTE By using the terms "operator" and "service personnel" this standard considers the perception of hazards depending on training and skills. Annex AA gives a general approach in this regard.

Requirements for protection against particular types of hazard are given in Clauses 6 to 13, as follows:

- a) electric shock or burn (see Clause 6);
- b) mechanical hazards (see Clauses 7 and 8);
- c) spread of fire from the control equipment (see Clause 9);
- d) excessive temperature (see Clause 10);
- e) effects of fluids and fluid pressure (see Clause 11);
- f) effects of radiation, including lasers sources, and sonic and ultrasonic pressure (see Clause 12);
- g) liberated gases, explosion and implosion (see Clause 13);

Requirements for protection against hazards arising from reasonably foreseeable misuse and ergonomic factors are specified in Clause 16.

Risk assessment for hazards or environments not fully covered above is specified in Clause 17.

NOTE Attention is drawn to the existence of additional requirements regarding the health and safety of labour forces.

1.2.2 Aspects excluded from scope

Replacement:

This standard does not cover:

- a) reliability, functionality, performance, or other properties of the control equipment not related to safety;
- b) mechanical or climatic requirements for operation, transport or storage;
- c) EMC requirements (See e.g. IEC 61326 or IEC 61131-2);
- d) protective measures for explosive atmospheres (See e.g. IEC 60079 series);
- e) functional safety (See e.g. IEC 61508 or IEC 61131-6).

This is a free preview. Purchase the entire publication at the link below:

[Product Page](#)

-
- Looking for additional Standards? Visit Intertek Inform Infostore
 - Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation
-