



**NSAI**  
Standards

Irish Standard  
I.S. EN 15199-2:2020

Petroleum products - Determination of boiling range distribution by gas chromatography method - Part 2: Heavy distillates and residual fuels

**I.S. EN 15199-2:2020**

*Incorporating amendments/corrigenda/National Annexes issued since publication:*

The National Standards Authority of Ireland (NSAI) produces the following categories of formal documents:

I.S. xxx: Irish Standard — national specification based on the consensus of an expert panel and subject to public consultation.

S.R. xxx: Standard Recommendation — recommendation based on the consensus of an expert panel and subject to public consultation.

SWiFT xxx: A rapidly developed recommendatory document based on the consensus of the participants of an NSAI workshop.

*This document replaces/revises/consolidates the NSAI adoption of the document(s) indicated on the CEN/CENELEC cover/Foreword and the following National document(s):*

*NOTE: The date of any NSAI previous adoption may not match the date of its original CEN/CENELEC document.*

*This document is based on:*

EN 15199-2:2020

*Published:*

2020-12-23

*This document was published under the authority of the NSAI and comes into effect on:*

2021-01-27

ICS number:

75.080

NOTE: If blank see CEN/CENELEC cover page

NSAI  
1 Swift Square,  
Northwood, Santry  
Dublin 9

T +353 1 807 3800  
F +353 1 807 3838  
E standards@nsai.ie  
W NSAI.ie

Sales:  
T +353 1 857 6730  
F +353 1 857 6729  
W standards.ie

Údarás um Chaighdeáin Náisiúnta na hÉireann

## National Foreword

I.S. EN 15199-2:2020 is the adopted Irish version of the European Document EN 15199-2:2020, Petroleum products - Determination of boiling range distribution by gas chromatography method - Part 2: Heavy distillates and residual fuels

This document does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

For relationships with other publications refer to the NSAI web store.

**Compliance with this document does not of itself confer immunity from legal obligations.**

*In line with international standards practice the decimal point is shown as a comma (,) throughout this document.*

This page is intentionally left blank

EUROPEAN STANDARD

EN 15199-2

NORME EUROPÉENNE

EUROPÄISCHE NORM

December 2020

ICS 75.080

Supersedes EN 15199-2:2006

English Version

## Petroleum products - Determination of boiling range distribution by gas chromatography method - Part 2: Heavy distillates and residual fuels

Produits pétroliers - Détermination de la répartition dans l'intervalle de distillation par méthode de chromatographie en phase gazeuse - Partie 2 : Fiouls lourds et fiouls résiduels

Mineralölerzeugnisse - Gaschromatographische Bestimmung des Siedeverlaufes - Teil 2: Schweröle und Rückstandsöle

This European Standard was approved by CEN on 23 November 2020.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels**

## Contents

Page

European foreword .....	3
1 Scope.....	4
2 Normative references.....	4
3 Terms and definitions.....	4
4 Principle.....	6
5 Reagents and materials .....	6
6 Apparatus .....	9
7 Sampling.....	11
8 Preparation of the apparatus .....	11
8.1 Gas chromatograph preparation.....	11
8.2 System performance check .....	11
9 Sample and reference material preparation .....	11
10 Calibration.....	12
11 Procedure .....	14
12 Visual inspection of the chromatograms.....	15
13 Calculation.....	15
14 Expression of results .....	15
15 Precision .....	15
15.1 General .....	15
15.2 Repeatability .....	15
15.3 Reproducibility.....	15
16 Test report.....	17
Annex A (normative) Calculation procedure .....	18
Annex B (normative) System performance check.....	21
Annex C (normative) Boiling points of n-alkanes.....	23
Annex D (informative) Additional guidance for the calculation algorithm.....	25
Bibliography .....	29

## European foreword

This document (EN 15199-2:2020) has been prepared by Technical Committee CEN/TC 19 “Gaseous and liquid fuels, lubricants and related products of petroleum, synthetic and biological origin”, the secretariat of which is held by NEN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by June 2021, and conflicting national standards shall be withdrawn at the latest by June 2021.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN 15199-2:2006.

The main changes in this edition include:

- updated text to give better guidance to operators executing the test;
- additional clarification in the sample preparation section.

EN 15199 consists of the following parts, under the general title *Petroleum products — Determination of boiling range distribution by gas chromatography method*:

- *Part 1: Middle distillates and lubricating base oils;*
- *Part 2: Heavy distillates and residual fuels;*
- *Part 3: Crude oil;*
- *Part 4: Light fractions of crude oil.*

This document specifies the determination of boiling range distribution of materials with initial boiling points (IBP) above 100 °C and final boiling points (FBP) above 750 °C. For testing materials with initial boiling points (IBP) above 100 °C and final boiling point (FBP) below 750 °C, Part 1 of the standard can be used. For testing materials with initial boiling points (IBP) below 100 °C and final boiling points (FBP) above 750 °C, such as crude oils, Part 3 is applicable. Part 4 describes the determination of boiling range distribution of hydrocarbons up to n-nonane in crude oil.

This document is a joint development between the EI [5], ASTM [4] and CEN.

According to the CEN-CENELEC Internal Regulations, the national standards organisations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom

## EN 15199-2:2020 (E)

### 1 Scope

This document specifies a method for the determination of the boiling range distribution of petroleum products by capillary gas chromatography using flame ionization detection. The standard is applicable to materials having a vapour pressure low enough to permit sampling at ambient temperature, and which have a boiling range of at least 100 °C. The standard is applicable to materials with initial boiling points (IBP) above 100 °C and final boiling points (FBP) above 750 °C, for example, heavy distillate fuels and residuals. The method is not applicable to bituminous samples.

The test method is not applicable for the analysis of petroleum or petroleum products containing low molecular weight components (for example naphthas, reformates, gasolines) or middle distillates like Diesel and Jet fuel.

Petroleum or petroleum products containing blending components, which contain hetero atoms (for example alcohols, ethers, acids, or esters) or residue, are not to be analysed by this test method.

**NOTE** For the purposes of this document, the terms “% (m/m)” and “% (V/V)” are used to represent respectively the mass fraction and the volume fraction.

**WARNING** — The use of this document may involve hazardous materials, operations and equipment. This document does not purport to address all of the safety problems associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and to determine the applicability of regulatory limitations prior to use.

### 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN ISO 3170, *Petroleum liquids — Manual sampling (ISO 3170)*

EN ISO 3171, *Petroleum liquids — Automatic pipeline sampling (ISO 3171)*

### 3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

#### 3.1

##### **initial boiling point**

##### **IBP**

temperature corresponding to the retention time at which a net area (3.7) counts equal to 0,5 % of the total sample area (3.6) under the chromatogram is obtained (see Figure 1)

#### 3.2

##### **final boiling point**

##### **FBP**

temperature corresponding to the retention time at which a net area (3.7) counts equal to 99,5 % of the total sample area (3.6) under the chromatogram is obtained (see Figure 1)



This is a free preview. Purchase the entire publication at the link below:

[Product Page](#)

- 
- [Looking for additional Standards? Visit Intertek Inform Infostore](#)
  - [Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation](#)
-