

Irish Standard I.S. EN ISO 2719:2016&A1:2021

Determination of flash point - Pensky-Martens closed cup method (ISO 2719:2016)

 $\ensuremath{\mathbb C}$ CEN 2021 $\hfill No copying without NSAI permission except as permitted by copyright law.$

I.S. EN ISO 2719:2016&A1:2021

Incorporating amendments/corrigenda/National Annexes issued since publication:

EN ISO 2719:2016/A1:2021

The National Standards Authority of Ireland (NSAI) produces the following categories of formal documents:

I.S. xxx: Irish Standard – national specification based on the consensus of an expert panel and subject to public consultation.

S.R. xxx: Standard Recommendation — recommendation based on the consensus of an expert panel and subject to public consultation.

SWiFT xxx: A rapidly developed recommendatory document based on the consensus of the participants of an NSAI workshop.

This document replaces/revises/consolidates the NSAI adoption of the document(s) indicated on the CEN/CENELEC cover/Foreword and the following National document(s):

NOTE: The date of any NSAI previous adoption may not match the date of its original CEN/CENELEC document.

This document is based on: EN ISO 2719:2016 *Published:* 2016-07-20

This document was published		ICS number:			
and comes into effect on:		75.080			
		75.100			
2021-04-18					
		NOTE: If blank see CEN/CENELEC cover page			
NSAI	T +353 1	. 807 3800 Sales:			
1 Swift Square,	F +353 1	. 807 3838 T +353 1 857 6730			
Northwood, Santry	E standa	ards@nsai.ie F +353 1 857 6729			
Dublin 9	W NSAI.i	ie W standards.ie			
Údarás um Chaighdeáin Náisiúnta na hÉireann					

National Foreword

I.S. EN ISO 2719:2016&A1:2021 is the adopted Irish version of the European Document EN ISO 2719:2016, Determination of flash point - Pensky-Martens closed cup method (ISO 2719:2016)

This document does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

Compliance with this document does not of itself confer immunity from legal obligations.

In line with international standards practice the decimal point is shown as a comma (,) throughout this document.

This is a free page sample. Access the full version online.

This page is intentionally left blank

EUROPEAN STANDARD NORME EUROPÉENNE

EN ISO 2719:2016/A1

EUROPÄISCHE NORM

March 2021

ICS 75.080

English Version

Determination of flash point - Pensky-Martens closed cup method - Amendment 1: Thermometers correction (ISO 2719:2016/Amd 1:2021)

Détermination du point d'éclair - Méthode Pensky-Martens en vase clos - Amendement 1: Correction concernant les thermomètres (ISO 2719:2016/Amd 1:2021) Bestimmung des Flammpunktes - Verfahren nach Pensky-Martens mit geschlossenem Tiegel - Änderung 1: Korrektur des Thermometers (ISO 2719:2016/Amd 1:2021)

This amendment A1 modifies the European Standard EN ISO 2719:2016; it was approved by CEN on 12 February 2021.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for inclusion of this amendment into the relevant national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This amendment exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION COMITÉ EUROPÉEN DE NORMALISATION EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

© 2021 CEN All rights of exploitation in any form and by any means reserved worldwide for CEN national Members.

Ref. No. EN ISO 2719:2016/A1:2021 E

This is a free page sample. Access the full version online. I.S. EN ISO 2719:2016&A1:2021

EN ISO 2719:2016/A1:2021 (E)

Contents	Page
European foreword	

This is a free page sample. Access the full version online. I.S. EN ISO 2719:2016&A1:2021

European foreword

This document (EN ISO 2719:2016/A1:2021) has been prepared by Technical Committee ISO/TC 28 "Petroleum and related products, fuels and lubricants from natural or synthetic sources" in collaboration with Technical Committee CEN/TC 19 "Gaseous and liquid fuels, lubricants and related products of petroleum, synthetic and biological origin." the secretariat of which is held by NEN.

This Amendment to the European Standard EN ISO 2719:2016 shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by September 2021, and conflicting national standards shall be withdrawn at the latest by September 2021.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 2719:2016/Amd 1:2021 has been approved by CEN as EN ISO 2719:2016/A1:2021 without any modification.

This is a free page sample. Access the full version online.

This page is intentionally left blank

This is a free page sample. Access the full version online. I.S. EN ISO 2719:2016&A1:2021

INTERNATIONAL STANDARD

ISO 2719

Fourth edition 2016-06-15

Determination of flash point — Pensky-Martens closed cup method

Détermination du point d'éclair — Méthode Pensky-Martens en vase clos



Reference number ISO 2719:2016(E)



© ISO 2016, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office Ch. de Blandonnet 8 • CP 401 CH-1214 Vernier, Geneva, Switzerland Tel. +41 22 749 01 11 Fax +41 22 749 09 47 copyright@iso.org www.iso.org

Page

Contents

Forew	vord	iv	
Intro	duction	V	
1	Scope	1	
2	Normative references		
3	Terms and definitions	2	
4	Principle	2	
5	Chemicals and materials	2	
6	Apparatus	2	
7	Apparatus preparation 7.1 General 7.2 Location of the apparatus 7.3 Cleaning the test cup 7.4 Apparatus assembly 7.5 Apparatus verification	3 3 3 3 3 3	
8	Sampling	4	
9	Sample handling 9.1 Petroleum products 9.1.1 Sub-sampling 9.1.2 Samples containing undissolved water 9.1.3 Samples that are liquid at ambient temperature 9.1.4 Samples that are very viscous, semi-solid or solid at ambient temperature 9.2 Paints and varnishes	4 4 4 4 4 4 5	
10	Procedure 10.1 General 10.2 Procedure A 10.3 Procedure B 10.4 Procedure C	5 5 6 7	
11	Calculation11.1Conversion of barometric pressure reading11.2Correction of observed flash point to standard atmospheric pressure	8 8	
12	Expression of results	8	
13	Precision 13.1 General 13.2 Repeatability, <i>r</i> 13.3 Reproducibility, <i>R</i>	8 8 8 9	
14	Test report	10	
Annex	x A (informative) Apparatus verification	11	
Annex	x B (normative) Pensky-Martens closed cup test apparatus	14	
Annez	x C (normative) Temperature measuring device specification	20	
Biblio	ography	22	

This is a free page sample. Access the full version online. I.S. EN ISO 2719:2016&A1:2021

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: <u>www.iso.org/iso/foreword.html</u>.

The committee responsible for this document is ISO/TC 28, *Petroleum products and related products of synthetic or biological origin*, in conjunction with ISO/TC 35, *Paints and varnishes*.

This fourth edition cancels and replaces the third edition (ISO 2719:2002), which has been technically revised.

The main technical updates include:

- a) introduction of procedure C for FAME products;
- b) revision of temperature measuring device requirements, allowing alternatives for Hg-containing thermometers;
- c) removal of the original Annex D on the adaptor for the low-range thermometer, it being optional when a permanent ferrule is attached to the thermometer;
- d) revision of procedures regarding sampling and sample handling;
- e) inclusion, for automated apparatus, manufactured after 1 January 2017, of a device to automatically dispense an inert gas or vapour over the test cup in the event of a test cup fire.

Introduction

Flash point values are used in shipping, storage, handling, and safety regulations, as a classification property to define "flammable" and "combustible" materials. Precise definition of the classes is given in each particular regulation.

A flash point value can indicate the presence of highly volatile material(s) in a relatively non-volatile or non-flammable material and flash point testing can be a preliminary step to other investigations into the composition of unknown materials.

It is not appropriate for flash point determinations to be carried out on potentially unstable, decomposable, or explosive materials, unless it has been previously established that heating the specified quantity of such materials in contact with the metallic components of the flash point apparatus, within the temperature range required for the method, does not induce decomposition, explosion or other adverse effects.

Flash point values are not a constant physical-chemical property of materials tested. They are a function of the apparatus design, the condition of the apparatus used, and the operational procedure carried out. Flash point can therefore be defined only in terms of a standard test method, and no general valid correlation can be guaranteed between results obtained by different test methods or with test apparatus different from that specified.

ISO/TR 29662^[6] (CEN/TR 15138^[12]) gives useful advice in carrying out flash point tests and interpreting their results.

This is a free page sample. Access the full version online. I.S. EN ISO 2719:2016&A1:2021

Determination of flash point — Pensky-Martens closed cup method

WARNING — The use of this International Standard can involve hazardous materials, operations and equipment. This International Standard does not purport to address all of the safety problems associated with its use. It is the responsibility of users of this International Standard to take appropriate measures to ensure the safety and health of personnel prior to application of this International Standard, and fulfil statutory and regulatory requirements for this purpose.

1 Scope

This International Standard describes three procedures, A, B and C, using the Pensky-Martens closed cup tester, for determining the flash point of combustible liquids, liquids with suspended solids, liquids that tend to form a surface film under the test conditions, biodiesel and other liquids in the temperature range of 40 °C to 370 °C.

CAUTION — For certain mixtures no flash point, as defined, is observed; instead a significant enlargement of the test flame (not halo effect) and a change in colour of the test flame from blue to yellowish-orange can occur. Continued heating can result in significant burning of vapours outside the test cup, and can be a potential fire hazard.

NOTE 1 Although, technically, kerosene with a flash point above 40 °C can be tested using this International Standard, it is standard practice to test kerosene according to ISO 13736.^[5] Similarly, lubricating oils are normally tested according to ISO 2592^[2].

Procedure A is applicable to distillate fuels (diesel, biodiesel blends, heating oil and turbine fuels), new and in-use lubricating oils, paints and varnishes, and other homogeneous liquids not included in the scope of Procedures B or C.

Procedure B is applicable to residual fuel oils, cutback residua, used lubricating oils, mixtures of liquids with solids, liquids that tend to form a surface film under test conditions or are of such kinematic viscosity that they are not uniformly heated under the stirring and heating conditions of Procedure A.

Procedure C is applicable to fatty acid methyl esters (FAME) as specified in specifications such as EN 14214^[11] or ASTM D6751^[13].

This International Standard is not applicable to water-borne paints and varnishes.

NOTE 2 Water-borne paints and varnishes can be tested using ISO 3679^[3]. Liquids containing traces of highly volatile materials can be tested using ISO 1523^[1] or ISO 3679.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 1513, Paints and varnishes — Examination and preparation of test samples

ISO 3170, Petroleum liquids — Manual sampling

ISO 3171, Petroleum liquids — Automatic pipeline sampling

ISO 15528, Paints, varnishes and raw materials for paints and varnishes — Sampling

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1

flash point

lowest temperature of the test portion, corrected to a standard barometric pressure of 101,3 kPa, at which application of an ignition source causes the vapour of the test portion to ignite and the flame to propagate across the surface of the liquid under the specified conditions of test

4 Principle

The test portion is placed into the test cup of a Pensky-Martens apparatus and heated to give a constant temperature increase with continuous stirring. An ignition source is directed through an opening in the test cup lid at regular temperature intervals with simultaneous interruption of stirring. The lowest temperature at which the application of the ignition source causes the vapour of the test portion to ignite and a flame propagate over the surface of the liquid is recorded as the flash point at the absolute barometric pressure. This temperature is corrected to standard atmospheric pressure using a specified formula.

5 Chemicals and materials

5.1 Cleaning solvent, for removal of traces of sample from the test cup and cover.

The choice of solvent will depend upon the previous material tested and the tenacity of the residue. Low volatility aromatic (benzene free) solvents may be used to remove traces of oil, and mixed solvents such as toluene-acetone-methanol can be efficacious for the removal of gum-type deposits.

5.2 Liquids for verification, see <u>Annex A</u>.

6 Apparatus

6.1 Flash point apparatus, as described in <u>Annex B</u>.

If automated equipment is used, ensure that the test cup and cover assembly conform to the key dimensions specified in <u>Annex B</u> and the procedure described in <u>Clause 10</u> is followed. The user shall ensure that all of the manufacturer's instructions for adjusting and operating the instrument are followed.

NOTE Under certain circumstances, the use of electric ignition sources can give different results to those obtained when using a flame ignition source.

In cases of dispute, unless explicitly agreed otherwise, the flash point as determined using a flame ignition source shall be considered the referee test.

6.2 Temperature measuring devices, meeting the requirements for accuracy and have the response as specified in <u>Annex C</u>.

6.3 Barometer, absolute pressure reading, accuracy of ±0,5 kPa, and with a resolution of 0,1 kPa.

Barometers pre-corrected to give sea level readings, such as those used at weather stations and airports, shall not be used.

NOTE Some automated apparatus include an integral barometer that automatically measures and records the absolute barometric pressure and makes the required corrections to the detected flash point.

6.4 Heating bath or oven, capable of controlling the temperature to ±5 °C, for warming the sample if required.

The oven shall be ventilated and constructed in such a way that it will not cause ignition of any flammable vapours that can be produced when the sample is heated.

It is recommended that the oven is an explosion-protected type.

7 Apparatus preparation

7.1 General

Follow the manufacturer's instructions for the correct setup, calibration, verification (7.5) and operation of the apparatus including the integral barometer (if fitted) and temperature measuring device (see <u>Annex C</u>), especially the operation and setting of the ignition source.

7.2 Location of the apparatus

Support the flash point apparatus (<u>6.1</u>) on a level and steady surface in a draught-free position.

NOTE When draughts cannot be avoided, it is good practice to surround the apparatus with a shield.

The flash point apparatus may be located in a fume hood with individual air flow control, adjusted so that vapours can be withdrawn without causing air currents around the test cup during the test.

7.3 Cleaning the test cup

Wash the test cup, cover and its accessories with an appropriate cleaning solvent (5.1) to remove any traces of gum or residue remaining from a previous test. Dry using a stream of clean air to ensure complete removal of the solvent used.

7.4 Apparatus assembly

Examine the test cup, the cover and other parts to ensure that they are free from signs of damage and deposits. Assemble the apparatus in accordance with <u>Annex B</u>.

7.5 Apparatus verification

7.5.1 Verify the correct functioning of the apparatus at least once a year by testing a certified reference material (CRM) (5.2). The result obtained shall be equal to or less than $R / \sqrt{2}$ from the certified value of the CRM, where *R* is the reproducibility of the method as shown in Table 4.

Use the reproducibility for Procedure A (Table 4) to calculate the verification tolerance for verification materials certified for Procedures A, B or C.

It is recommended that more frequent verification checks are made using secondary working standards (SWSs) (5.2).

A recommended procedure for apparatus verification using CRMs and SWSs, and the production of SWSs, is described in Annex A.

7.5.2 The numerical values obtained during the verification check shall not be used to provide a bias statement, nor shall they be used to make any correction to the flash points subsequently determined using the apparatus.

8 Sampling

8.1 Unless otherwise specified, obtain representative samples for analysis in accordance with the procedures given in ISO 3170, ISO 3171, ISO 15528 or an equivalent National Standard.

8.2 When obtaining a sample of residual fuel oil, the sample container shall be from 85 % to 95 % full. For FAME samples, a typical one litre container filled to 85 % volume is recommended. For other types of samples, the size of the container shall be chosen such that the container is not more than 85 % full or less than 50 % full prior to any sample aliquot being taken.

8.3 Erroneously high flash points can be obtained if precautions are not taken to avoid the loss of volatile material. Do not open containers unnecessarily, to prevent loss of volatile material or possible introduction of moisture. Avoid storage of samples at temperatures in excess of 30 °C. For samples for storage, ensure that the sample container is tightly closed and leak free. Do not make a transfer unless the sample temperature is at least the equivalent of 18 °C below the expected flash point.

8.4 Do not store samples in gas-permeable containers, since volatile material may diffuse through the walls of the enclosure. Samples in leaky containers are suspect and not a source of valid results.

9 Sample handling

9.1 Petroleum products

9.1.1 Sub-sampling

9.1.1.1 Sub-sample at a temperature at least 18 °C below the expected flash point.

9.1.1.2 Successive test portions may be taken from the same sample container. Repeat tests have been shown to be within the precision of the method when the second test portion is taken with the sample container at least 50 % filled.

IMPORTANT — Results of flash point determinations can be affected if the sample volume falls below 50 % of the container capacity.

9.1.2 Samples containing undissolved water

9.1.2.1 Flash point results can be affected by the presence of water. If a sample contains undissolved water, decant a water-free aliquot prior to mixing.

9.1.2.2 For certain fuel oils and lubricants, it is not always possible to decant the sample from the free water. In such cases, separate the water from the aliquot physically, prior to mixing, or, if this is not possible, test the material in accordance with ISO 3679^[3].

9.1.3 Samples that are liquid at ambient temperature

Mix samples by gentle manual shaking prior to removal of the test portion, taking care to minimize the loss of volatile components and proceed in accordance with <u>Clause 10</u>.

9.1.4 Samples that are very viscous, semi-solid or solid at ambient temperature

Samples shall be heated in their containers, with lid/cap slightly loosened to avoid build-up of dangerous pressure, at the lowest temperature adequate to liquefy any solids, not exceeding 18 °C below the expected flash point, for 30 min. If the sample is then not completely liquefied, extend the heating period for additional 30 min periods as necessary. Then gently agitate the sample to provide

mixing, such as orbiting the container horizontally, before transferring to the test cup. No sample shall be heated and transferred unless its temperature is more than 18 °C below its expected flash point.

IMPORTANT — Volatile vapours can escape during heating when the sample container lid/cap is too loose.

9.2 Paints and varnishes

Prepare the samples in accordance with ISO 1513.

10 Procedure

10.1 General

10.1.1 The applicability of the three procedures A, B and C is explained in <u>Clause 1</u>.

Electronic flash point detection shall be used for procedure C, as it is difficult to observe a flash by visual means, and may be used for procedures A and B.

10.1.2 Care should be taken when testing samples of residual fuel oil that contain significant amounts of water, as heating these samples can cause them to foam and eject from the test cup.

10.1.3 As a safety practice, it is strongly advised to apply the ignition source to the test portion in the cup before heating the test cup and test portion, to check for the presence of unexpected volatile material.

10.1.4 As a safety practice, it is strongly advised that, for an expected flash point above 130 °C, to dip the ignition source every 10 °C throughout the test until the sample temperature reaches 28 °C below the expected flash point and then follow the prescribed dipping procedure. This practice has been shown to reduce the possibility of a fire and not to significantly affect the result.

10.1.5 At the end of a test, when the apparatus has cooled down to a safe handling temperature, remove the test cover and the test cup and clean the apparatus as recommended by the manufacturer.

10.2 Procedure A

10.2.1 Record the laboratory absolute barometric pressure reading at the time of test (see <u>6.3</u>).

It is not necessary to correct the barometric pressure to 0 $^\circ$ C, although some barometers are designed to make this correction automatically.

10.2.2 Fill the test cup (see 7.3) with the test portion to the level indicated by the filling mark. Place the lid on the test cup and put them in the heating chamber. Ensure that the locating or locking device is properly engaged and insert the temperature measuring device (<u>6.2</u>). Light the test flame and adjust to a diameter of 3,2 mm to 4,8 mm or enable the alternative ignition source. Supply heat at such a rate that the temperature of the test portion, as indicated by the temperature measuring device, increases at 5,0 °C/min to 6,0 °C/min, and maintain this heating rate throughout the test. Stir the test portion at a rotational rate of 90 r/min (1,5 Hz) to 120 r/min (2,0 Hz), stirring in a downward direction.

10.2.3 When the test portion is expected to have a flash point of 110 °C or below, make the first application of the ignition source when the temperature of the test portion is 23 °C \pm 5 °C below the expected flash point, and thereafter at 1 °C temperature intervals. Cease stirring and apply the ignition source by operating the mechanism on the cover, that controls the shutter and ignition source, so that the source is lowered into the vapour space of the test cup in 0,5 s, left in its lowered position for 1 s, and quickly raised to its high position.



This is a free preview. Purchase the entire publication at the link below:

Product Page

S Looking for additional Standards? Visit Intertek Inform Infostore

> Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation