



NSAI
Standards

Irish Standard
I.S. EN ISO 80601-2-85:2021

Medical electrical equipment - Part 2-85:
Particular requirements for the basic
safety and essential performance of
cerebral tissue oximeter equipment (ISO
80601-2-85:2021)

I.S. EN ISO 80601-2-85:2021

Incorporating amendments/corrigenda/National Annexes issued since publication:

The National Standards Authority of Ireland (NSAI) produces the following categories of formal documents:

I.S. xxx: Irish Standard — national specification based on the consensus of an expert panel and subject to public consultation.

S.R. xxx: Standard Recommendation — recommendation based on the consensus of an expert panel and subject to public consultation.

SWiFT xxx: A rapidly developed recommendatory document based on the consensus of the participants of an NSAI workshop.

This document replaces/revises/consolidates the NSAI adoption of the document(s) indicated on the CEN/CENELEC cover/Foreword and the following National document(s):

NOTE: The date of any NSAI previous adoption may not match the date of its original CEN/CENELEC document.

This document is based on:

EN ISO 80601-2-85:2021

Published:

2021-04-14

This document was published under the authority of the NSAI and comes into effect on:

2021-05-06

ICS number:

11.040.10

NOTE: If blank see CEN/CENELEC cover page

NSAI
1 Swift Square,
Northwood, Santry
Dublin 9

T +353 1 807 3800
F +353 1 807 3838
E standards@nsai.ie
W NSAI.ie

Sales:
T +353 1 857 6730
F +353 1 857 6729
W standards.ie

Údarás um Chaighdeáin Náisiúnta na hÉireann

National Foreword

I.S. EN ISO 80601-2-85:2021 is the adopted Irish version of the European Document EN ISO 80601-2-85:2021, Medical electrical equipment - Part 2-85: Particular requirements for the basic safety and essential performance of cerebral tissue oximeter equipment (ISO 80601-2-85:2021)

This document does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

For relationships with other publications refer to the NSAI web store.

Compliance with this document does not of itself confer immunity from legal obligations.

In line with international standards practice the decimal point is shown as a comma (,) throughout this document.

This page is intentionally left blank

EUROPEAN STANDARD

EN ISO 80601-2-85

NORME EUROPÉENNE

EUROPÄISCHE NORM

April 2021

ICS 11.040.10

English Version

Medical electrical equipment - Part 2-85: Particular requirements for the basic safety and essential performance of cerebral tissue oximeter equipment (ISO 80601-2-85:2021)

Appareils électromédicaux - Partie 2-85: Exigences particulières pour la sécurité de base et les performances essentielles des oxymètres pour tissu cérébral (ISO 80601-2-85:2021)

Medizinische elektrische Geräte - Teil 2-85: Besondere Anforderungen für die Sicherheit einschließlich der wesentlichen Leistungsmerkmale von Geräten für die nicht-invasive zerebrale Oxymetrie (ISO 80601-2-85:2021)

This European Standard was approved by CEN on 21 January 2021.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

EN ISO 80601-2-85:2021 (E)

Contents	Page
European foreword.....	3

European foreword

This document (EN ISO 80601-2-85:2021) has been prepared by Technical Committee ISO/TC 121 "Anaesthetic and respiratory equipment" in collaboration with Technical Committee CEN/TC 215 "Respiratory and anaesthetic equipment" the secretariat of which is held by BSI.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by October 2021, and conflicting national standards shall be withdrawn at the latest by April 2024.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 80601-2-85:2021 has been approved by CEN as EN ISO 80601-2-85:2021 without any modification.

This page is intentionally left blank

INTERNATIONAL
STANDARD

ISO
80601-2-85

First edition
2021-03

Medical electrical equipment —

Part 2-85:

**Particular requirements for the basic
safety and essential performance of
cerebral tissue oximeter equipment**

Appareils électromédicaux —

*Partie 2-85: Exigences particulières pour la sécurité de base et les
performances essentielles des oxymètres pour tissu cérébral*



Reference number
ISO 80601-2-85:2021(E)

© ISO 2021

ISO 80601-2-85:2021(E)



COPYRIGHT PROTECTED DOCUMENT

© ISO 2021

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword.....	vi
Introduction.....	vii
201.1 Scope, object and related standards.....	1
201.1.1 * Scope.....	1
201.1.2 Object.....	2
201.1.3 Collateral standards.....	2
201.1.4 Particular standards.....	3
201.2 Normative references.....	4
201.3 Terms and definitions.....	5
201.4 General requirements.....	10
201.4.3.101 * Additional requirements for <i>essential performance</i>	11
201.4.102 Additional requirements for acceptance criteria.....	11
201.4.103 Additional requirements for <i>cerebral tissue oximeter equipment, parts and accessories</i>	11
201.5 General requirements for testing of <i>ME equipment</i>	12
201.6 Classification of <i>ME equipment</i> and <i>ME systems</i>	12
201.7 <i>ME equipment</i> identification, marking and documents.....	12
201.7.1.101 <i>Information to be supplied by the manufacturer</i>	12
201.7.2.3 Consult <i>accompanying documents</i>	12
201.7.2.9.101 IP classification.....	12
201.7.2.101 Additional requirements for marking on the outside of <i>ME equipment</i> parts.....	13
201.7.4.3 Units of measurement.....	13
201.7.9.2 Instructions for use.....	13
201.7.9.2.1.101 Additional general requirements.....	13
201.7.9.2.2.101 Additional requirements for warnings and safety notices.....	15
201.7.9.2.9.101 Additional requirements for operating instructions.....	15
201.7.9.2.14.101 Additional requirements for <i>accessories, supplementary equipment, used material</i>	15
201.7.9.3.1.101 * Additional general requirements.....	16
201.8 Protection against electrical <i>hazards</i> from <i>ME equipment</i>	16
201.8.3.101 Additional requirements for classification of <i>applied parts</i>	16
201.8.5.5.1.101 Defibrillation protection.....	16
201.8.7.4.7.101 Additional requirements for measurement of the <i>patient leakage current</i>	16
201.9 Protection against mechanical <i>hazards</i> of <i>ME equipment</i> and <i>ME systems</i>	17
201.10 Protection against unwanted and excessive radiation <i>hazards</i>	17
201.10.4 Lasers.....	17
201.11 Protection against excessive temperatures and other <i>hazards</i>	17
201.11.1.2.2 <i>Applied parts</i> not intended to supply heat to a <i>patient</i>	17
201.11.6.5.101 * Additional requirements for ingress of water or particulate matter into the <i>ME equipment</i> or <i>ME system</i>	18
201.11.6.7 <i>Sterilization</i> of <i>ME equipment</i> or <i>ME system</i>	18
201.11.8.101 Additional requirements for interruption of the power supply/ <i>supply mains</i> to <i>ME equipment</i>	18

ISO 80601-2-85:2021(E)

201.11.8.101.1	<i>Technical alarm condition for power supply failure</i>	18
201.11.8.101.2	Settings and data storage following short interruptions or automatic switchover	19
201.11.8.101.3	Operation following long interruptions	19
201.12	Accuracy of controls and instruments and protection against hazardous outputs.....	19
201.12.1.101	* <i>StO₂ accuracy of cerebral tissue oximeter equipment</i>	19
201.12.1.101.1	* Specification	19
201.12.1.101.2	* Data collection for determination of <i>StO₂ accuracy</i>	21
201.12.1.101.3	* Data analysis for determination of <i>StO₂ accuracy</i>	22
201.12.1.101.4	Characteristics of the study used for determination of <i>StO₂ accuracy</i>	23
201.12.4	Protection against hazardous output	23
201.12.4.101	* <i>Data update period</i>	23
201.12.4.102	* Signal inadequacy	23
201.13	<i>Hazardous situations and fault conditions for ME equipment</i>	24
201.13.101	Detection of <i>probe faults and probe cable extender faults</i>	24
201.14	<i>Programmable electrical medical systems (PEMS)</i>	24
201.15	Construction of <i>ME equipment</i>	24
201.15.3.5.101	* Additional requirements for rough handling.....	25
201.15.3.5.101.1	* Shock and vibration (robustness).....	25
201.15.3.5.101.2	* Shock and vibration for a <i>transit-operable cerebral tissue oximeter</i> during operation	26
201.15.101	Mode of operation.....	27
201.16	<i>ME systems</i>	27
201.17	Electromagnetic compatibility of <i>ME equipment and ME systems</i>	27
201.101	* <i>Cerebral tissue oximeter probes and probe cable extenders</i>	27
201.101.1	General	27
201.101.2	Labelling	28
201.102	<i>Functional connection</i>	28
201.102.1	General	28
201.102.2	* Connection to an electronic health record or <i>integrated clinical environment</i>	28
201.102.3	Connection to a <i>distributed alarm system</i>	28
202	Electromagnetic disturbances — Requirements and tests.....	29
202.4.3.1	Configurations.....	29
202.5.2.2.1	Requirements applicable to all <i>ME equipment and ME systems</i>	29
202.8.1.101	Additional general requirements	29
202.8.2	<i>Patient physiological simulation</i>	29
206	Usability.....	30
208	General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems.....	30
208.6.1.2.101	* Additional requirements for <i>alarm condition priority</i>	30
208.6.5.4.101	* Additional requirements for <i>default alarm preset</i>	31
208.6.8.5.101	Additional requirements for <i>alarm signal inactivation states, indication and</i> access.....	31
211	Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment	31
212	Requirements for medical electrical equipment and medical electrical systems used in the emergency medical services environment.....	31

Annex C (informative) Guide to marking and labelling requirements for <i>ME</i> equipment and <i>ME</i> systems	32
Annex D (informative) Symbols on marking	36
Annex AA (informative) Particular guidance and rationale	37
Annex BB (informative) Skin temperature at the <i>cerebral tissue oximeter probe</i>	48
Annex CC (informative) Determination of <i>accuracy</i>	50
Annex DD (informative) Characteristics of a <i>tissue haemoglobin phantom</i> for the verification of the <i>accuracy</i> of <i>cerebral tissue oximeter equipment</i>	56
Annex EE (informative) Guideline for evaluating and documenting <i>StO₂ accuracy</i> in human subjects	66
Annex FF (informative) <i>Functional testers for cerebral tissue oximeter equipment</i>	72
Annex GG (informative) Concepts of <i>ME equipment</i> response time	75
Annex HH (normative) Data interface requirements	80
Annex II (informative) Comparison of methods of performance evaluation	84
Annex JJ (informative) Reference to the <i>IMDRF essential principles</i> and labelling guidances	89
Annex KK (informative) Reference to the <i>essential principles</i>	92
Annex LL (informative) Reference to the general safety and performance requirements	95
Annex MM (informative) Terminology — alphabetized index of defined terms	98
Bibliography	102

ISO 80601-2-85:2021(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared jointly by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 3, *Respiratory devices and related equipment used for patient care*, and Technical Committee IEC/TC 62, *Electrical equipment in medical practice*, Subcommittee 62D, *Electromedical equipment*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 215, *Respiratory and anaesthetic equipment*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

A list of all parts in the ISO and IEC 80601 series can be found on the ISO and IEC websites.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

The estimation of blood oxygen saturation in the brain tissue by *cerebral tissue oximetry equipment* is increasingly used in many areas of medicine. This document covers *basic safety* and *essential performance* requirements achievable within the limits of existing technology.

Annex AA contains a rationale for some of the requirements. It is included to provide additional insight into the reasoning of the committees that led to a requirement and into the *hazards* that the requirement addresses.

Annex BB is a literature review and provides recommendations relevant to determining the maximum safe temperature of the interface between a *cerebral tissue oximeter probe* and a *patient's* tissue.

Annex CC discusses both the formulae used to evaluate the *StO₂ accuracy* of *cerebral tissue oximeter equipment* measurements, and the names that are assigned to those formulae.

Annex DD presents guidance on using in-vitro methods (phantoms) for *verification* of *StO₂ accuracy* of *cerebral tissue oximeter equipment*.

Annex EE presents a guideline for an in-vivo (human subjects) *controlled desaturation study* for the *verification* of *StO₂ accuracy* of *cerebral tissue oximeter equipment*.

Annex FF is a description of *functional testers* for use with *cerebral tissue oximeter equipment*.

Annex GG describes concepts of *cerebral tissue oximeter equipment* response time.

Annex HH describes data interface requirements.

Annex II is a comparison between human desaturations (in-vivo) and *tissue haemoglobin phantom* desaturations (in-vitro) for assessing *StO₂ accuracy*.

In this document, the following print types are used:

- requirements and definitions: roman type;
- *Instructions, test specifications and terms defined in Clause 3 of the general standard, in this document or as noted: italic type;*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type; normative text of tables is also in a smaller type.

In referring to the structure of this document, the term

- “clause” means one of the numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 201.7 includes subclauses 201.7.1, 201.7.2) and
- “subclause” means a numbered subdivision of a clause (e.g. 201.7.1, 7.2 and 201.7.2.1 are all subclauses of Clause 201.7).

References to clauses within this document are preceded by the term “Clause” followed by the clause number. References to subclauses within this document are by number only.

ISO 80601-2-85:2021(E)

In this document, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

For the purposes of this document, the auxiliary verb:

- “shall” means that conformance with a requirement or a test is mandatory for conformance with this document;
- “should” means that conformance with a requirement or a test is recommended but is not mandatory for conformance with this document; and
- “may” is used to describe permission (e.g. a permissible way to achieve conformance with a requirement or test);
- “can” is used to describe a possibility or capability; and
- “must” is used to express an external constraint.

Annex C contains a guide to the marking and labelling requirements in this document.

Annex D contains a summary of the symbols referenced in this document.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in AA.

Medical electrical equipment —

Part 2-85:

Particular requirements for the basic safety and essential performance of cerebral tissue oximeter equipment

201.1 Scope, object and related standards

Clause 1 of the general standard applies, except as follows.

NOTE The general standard is IEC 60601-1:2005+AMD1:2012+AMD2:2020.

201.1.1 * Scope

Replacement:

This document applies to *basic safety* and *essential performance* of *cerebral tissue oximeter equipment*, that employs light at multiple wavelengths to derive a quantitative measure of oxygen saturation of haemoglobin within the volume of tissue sampled under the *probe* attached to the head. The *cerebral tissue oximeter equipment* can be based on continuous light, frequency domain or time domain technologies. This document applies to *ME equipment* used in a hospital environment as well as when used outside the hospital environment, such as in ambulances and air transport. Additional standards may apply to *ME equipment* for those environments of use.

NOTE 1 *Cerebral tissue oximeters* are sometimes referred to as near infrared spectroscopy equipment in medical literature.

Not included within the scope of this document are:

- invasive tissue or vascular oximeters;
- oximeters that require a blood sample from the *patient*;
- equipment measuring dissolved oxygen;
- *ME equipment*, or part thereof, that measures path-length-dependent haemoglobin change. The requirements for functional near-infrared spectroscopy equipment are found in ISO 80601-2-71^[4];
- *ME equipment*, or part thereof, that measures arterial saturation based on pulsatile changes in tissue optical properties (SpO_2). The requirements for pulse oximeter equipment are found in ISO 80601-2-61^[3];
- *ME equipment*, or any part thereof, that claims to monitor tissue in parts of the body other than the head.

This document also applies to *cerebral tissue oximeter equipment*, including *cerebral tissue oximeter monitors*, *cerebral tissue oximeter probes* and *probe cable extenders*, that have been *remanufactured*.

This is a free preview. Purchase the entire publication at the link below:

[Product Page](#)

-
- [Looking for additional Standards? Visit Intertek Inform Infostore](#)
 - [Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation](#)
-