AS 3200.2.20—1992 IEC 601-2-20;1990

Australian Standard®

Medical electrical equipment

Part 2.20: Particular requirements for safety—Baby incubators (transport)

[IEC title: Medical electrical equipment

Part 2: Particular requirements for safety of transport incubators]

This Australian Standard was prepared by Committee HT/19, Baby Incubators/Infant Care Units. It was approved on behalf of the Council of Standards Australia on 18 August 1992 and published on 21 December 1992.

The following interests are represented on Committee HT/19:

Australasian College of Physical Scientists and Engineers in Medicine

Australian Chamber of Commerce and Industry

Australian College of Paediatrics

Australian Federation for Medical and Biological Engineering

Australian Nursing Federation

Department of Health, Housing and Community Services

Department of Health, N.S.W.

Health Department, Vic.

Institute of Hospital Engineering Australia

Medical Industry Association of Australia

Monash Medical Centre

Newborn Emergency Transport Service

Queensland Health

South Australian Health Commission

Standards New Zealand

Review of Australian Standards. To keep abreast of progress in industry, Australian Standards are subject to periodic review and are kept up to date by the issue of amendments or new editions as necessary. It is important therefore that Standards users ensure that they are in possession of the latest edition, and any amendments thereto.

Full details of all Australian Standards and related publications will be found in the Standards Australia Catalogue of Publications; this information is supplemented each month by the magazine 'The Australian Standard', which subscribing members receive, and which gives details of new publications, new editions and amendments, and of withdrawn Standards.

Suggestions for improvements to Australian Standards, addressed to the head office of Standards Australia, are welcomed. Notification of any inaccuracy or ambiguity found in an Australian Standard should be made without delay in order that the matter may be investigated and appropriate action taken.

AS 3200.2.20—1992

Australian Standard®

Medical electrical equipment

Part 2.20: Particular requirements for safety—Baby incubators (transport)

First published as part of AS 3201.2—1971 Ap. Revised and redesignated in part as AS 3200.2.20—1992.

Incorporating: Amdt 1—1999

2

PREFACE

This Standard was prepared on behalf of the Standards Australia Multitechnics Standards Policy Board to supersede AS 3201.2—1971, Approval and test specification for electro-medical equipment, Part 2: Electrically heated incubators for babies. It is technically equivalent to and has been reproduced from IEC 601-2-20:1990, Medical electrical equipment, Part 2: Particular requirements for safety of transport incubators.

The Standard is one of a series of Approval and Test Specifications issued by Standards Australia for individual items of medical electrical equipment. It is supplementary to AS 3200.1—1990, Approval and test specification—Medical electrical equipment, Part 1: General requirements for safety. Only safety matters and closely aligned conditions are covered by this series. In some instances, however, these are more stringent than for most electrical appliances because of the additional requirements necessary to ensure the safety of the patient, and because of the environmental conditions, e.g. high humidity and hazardous locations, in which some equipment is likely to be used.

The International Standard IEC 601-2-20 modifies and supplements the corresponding clauses of IEC 601-1:1977, *Safety of medical electrical equipment*, Part 1: *General requirements*, first edition. This has since been revised and published as IEC 601-1:1988, *Medical electrical equipment*, Part 1: *General requirements for safety*, second edition. The second edition has been adopted as the Australian Standard AS 3200.1—1990, *Approval and test specification—Medical electrical equipment*, Part 1: *General requirements for safety*, hereinafter referred to as the General Standard.

In order that IEC 601-2-20 and the General Standard are compatible, the titles of the following sections, clauses and subclauses have been amended:

- i) Sections 2; 6; 7; 8; and 9.
- ii) Clauses 10; 24; and 55.
- iii) Subclause 56.6.

The clauses of this Particular Standard supplement or modify the corresponding clauses in AS 3200.1—1990. As stated in Clause 1.3 of AS 3200.1—1990, the requirements of a Particular Standard take priority, where appropriate over those of AS 3200.1—1990. Where the reference in the text of this Standard indicates an 'amendment;', 'addition' or 'replacement' of the relevant requirements, tests or explanatory notes of AS 3200.1—1990, these changes are made to the relevant text which then becomes part of the Standard.

Subclauses or figures which are additional to the General Standard are numbered starting from 101; additional appendices are indicated by a sequence of double capitals e.g. AA), BB), and additional items are denoted by an aa), bb), sequence.

In some parts of this series, the Australian requirements differ from the IEC specifications. To accommodate Australian conditions and well-accepted safety practices, a marginal bar is placed alongside the IEC text, and Appendix ZZ details the Australian specifications and rationale for the deviations.

In addition an Australian Appendix has been developed (Appendix TT), which lists acceptable tolerances for measurements not otherwise specified in the main text of this Standard.

In the text of this Standard, the following print types are used:

- Requirements, compliance with which can be tested and definitions: in roman type.
- Explanations, advice, introductions, general statements, exceptions and references in smaller roman type.
- Headings of subclauses and test specifications in italic type.
- An asterisk is placed before each clause for which rationale is included in Appendix AA.

This Standard requires reference to other Standards which are detailed in Appendix LL.

Under arrangements made between Australia and the international Standards bodies, ISO and IEC, as well as certain other Standards organizations, users of this Australian Standard are advised that copyright is vested in Standards Australia.

© Copyright — STANDARDS AUSTRALIA

Users of Standards are reminded that copyright subsists in all Standards Australia publications and software. Except where the Copyright Act allows and except where provided for below no publications or software produced by Standards Australia may be reproduced, stored in a retrieval system in any form or transmitted by any means without prior permission in writing from Standards Australia. Permission may be conditional on an appropriate royalty payment. Requests for permission and information on commercial software royalties should be directed to the head office of Standards Australia.

Standards Australia will permit up to 10 percent of the technical content pages of a Standard to be copied for use exclusively inhouse by purchasers of the Standard without payment of a royalty or advice to Standards Australia.

Standards Australia will also permit the inclusion of its copyright material in computer software programs for no royalty payment provided such programs are used exclusively in-house by the creators of the programs.

Care should be taken to ensure that material used is from the current edition of the Standard and that it is updated whenever the Standard is amended or revised. The number and date of the Standard should therefore be clearly identified.

The use of material in print form or in computer software programs to be used commercially, with or without payment, or in commercial contracts is subject to the payment of a royalty. This policy may be varied by Standards Australia at any time.

3

CONTENTS

SECTION ONE — GENERAL

1. 2. 3. 4. 5. 6. 7.	Scope and object Terminology and definitions General requirements General requirements for tests Classification Identification, marking and documents Power input	5 5a 6 6 7 7 9					
	SECTION TWO - ENVIRONMENTAL CONDITIONS						
10.	Environmental conditions	10					
	SECTION THREE - PROTECTION AGAINST ELECTRIC SHOCK HAZARDS						
20.	Dielectric strength	10					
	SECTION FOUR - PROTECTION AGAINST MECHANICAL HAZARDS						
22. 23.	Mechanical strength Moving parts Surfaces, corners and edges Stability in NORMAL USE	10 11 11 12					
	SECTION FIVE - PROTECTION AGAINST HAZARDS FROM UNWANTED OR EXCESSIVE RADIATION						
	SECTION SIX - PROTECTION AGAINST HAZARDS OF IGNITION OF FLAMMABLE ANAESTHETIC MIXTURES						
SECTION SEVEN - PROTECTION AGAINST EXCESSIVE TEMPERATURES, AND OTHER SAFETY HAZARDS							
43. 44. 45. 46.	Excessive temperatures Fire prevention Overflow, spillage, leakage, humidity, ingress of liquids, cleaning, sterilization and disinfection Pressure vessels and parts subject to PRESSURE Human errors Interruption of the power supply	14 14 15 16 16 17					
	Section Eight - Accuracy of operating data and protection against hazardous output						
	Accuracy of operating data	17 20					
	SECTION NINE - ABNORMAL OPERATION AND FAULT CONDITIONS; ENVIRONMENTAL TESTS						
SECTION TEN - CONSTRUCTIONAL REQUIREMENTS							
55.	General	20 21 21					



The is a new provider i arenade and chare publication at the limit below	This is a free preview.	Purchase the	entire publication	at the link below:
--	-------------------------	--------------	--------------------	--------------------

Product Page

- Dooking for additional Standards? Visit Intertek Inform Infostore
- Dearn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation