

Australian Standard™

**Hydrometers—Portable syringe-type for
lead-acid batteries**



This Australian Standard was prepared by Committee CH-001, Laboratory Glassware and Related Apparatus. It was approved on behalf of the Council of Standards Australia on 20 October 2005.

This Standard was published on 23 December 2005.

The following are represented on Committee CH-001:

National Association of Testing Authorities
National Measurement Institute
Royal Australian Chemical Institute
Royal College of Pathologists of Australasia
Science Industry Australia Inc.
University of New South Wales
University of Sydney

Keeping Standards up-to-date

Standards are living documents which reflect progress in science, technology and systems. To maintain their currency, all Standards are periodically reviewed, and new editions are published. Between editions, amendments may be issued. Standards may also be withdrawn. It is important that readers assure themselves they are using a current Standard, which should include any amendments which may have been published since the Standard was purchased.

Detailed information about Standards can be found by visiting the Standards Web Shop at www.standards.com.au and looking up the relevant Standard in the on-line catalogue.

Alternatively, the printed Catalogue provides information current at 1 January each year, and the monthly magazine, *The Global Standard*, has a full listing of revisions and amendments published each month.

Australian Standards™ and other products and services developed by Standards Australia are published and distributed under contract by SAI Global, which operates the Standards Web Shop.

We also welcome suggestions for improvement in our Standards, and especially encourage readers to notify us immediately of any apparent inaccuracies or ambiguities. Contact us via email at mail@standards.org.au, or write to the Chief Executive, Standards Australia, GPO Box 476, Sydney, NSW 2001.

This Standard was issued in draft form for comment as DR 05086.

Australian Standard™

Hydrometers—Portable syringe-type for lead-acid batteries

Originated as AS 2562—1982.
Second edition 2005.

COPYRIGHT

© Standards Australia

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher.

Published by Standards Australia, GPO Box 476, Sydney, NSW 2001, Australia

ISBN 0 7337 7070 3

PREFACE

This Standard was prepared by the Standards Australia Committee CH-001, Laboratory Glassware and Related Apparatus to supersede AS 2562—1982, *Hydrometers—Portable syringe-type for lead-acid batteries*.

The objective of this Standard is to ensure that the requirements for battery hydrometers are achieved. It was prepared at the request of manufacturers and users of such instruments.

The objective of this revision is to reconfirm the Standard with minor changes including the following:

- (a) Referenced documents have been upgraded.
- (b) The Standard has been brought into line with current editorial practices.

The terms ‘normative’ and ‘informative’ have been used in this Standard to define the application of the appendix to which they apply. A ‘normative’ appendix is an integral part of a Standard, whereas an ‘informative’ appendix is only for information and guidance.

CONTENTS

	<i>Page</i>
SECTION 1 SCOPE AND GENERAL	
1.1 SCOPE	4
1.2 APPLICATION	4
1.3 REFERENCED DOCUMENTS	4
1.4 DEFINITIONS	4
1.5 CLASSIFICATION	5
SECTION 2 DESIGN AND CONSTRUCTION	
2.1 SCOPE OF SECTION	7
2.2 GENERAL REQUIREMENTS	7
2.3 FLOAT	7
2.4 BARREL	8
2.5 ELASTOMERIC COMPONENTS	8
2.6 FLOAT SCALE.....	9
SECTION 3 PERFORMANCE REQUIREMENTS	
3.1 SCOPE OF SECTION	11
3.2 FLOAT ACCURACY.....	11
3.3 ELASTIC BULB OF BATTERY HYDROMETER.....	11
3.4 RESISTANCE OF THE BARREL TO THERMAL SHOCK	11
3.5 ELASTOMERIC AND PLASTICS COMPONENTS	11
SECTION 4 MARKING	12
APPENDICES	
A SAFETY PRECAUTIONS	13
B PROCEDURE FOR USING A BATTERY HYDROMETER.....	14
C CARE OF HYDROMETERS	15
D METHOD FOR TESTING THE ACCURACY OF A HYDROMETER FLOAT	16
E METHOD FOR TESTING THE PERFORMANCE OF THE ELASTOMERIC COMPONENTS	19
F METHOD FOR DETERMINING THE RESISTANCE OF THE BARREL TO THERMAL SHOCK.....	21
G ACCELERATED AGEING TEST FOR THE ELASTOMERIC AND PLASTICS COMPONENTS OF A BATTERY HYDROMETER.....	22

This is a free preview. Purchase the entire publication at the link below:

[Product Page](#)

-
- [Looking for additional Standards? Visit Intertek Inform Infostore](#)
 - [Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation](#)
-