

AS 3864—1997

Australian Standard[®]

**Medical refrigeration equipment—
For the storage of blood and blood
products**

This Australian Standard was prepared by Committee HT/10, Medical Refrigeration. It was approved on behalf of the Council of Standards Australia on 14 March 1997 and published on 5 August 1997.

The following interests are represented on Committee HT/10:

Australian Hospital Association
Australian Institute of Refrigeration, Air Conditioning and Heating
Australian Red Cross Society
Commercial Refrigeration Manufacturers Association of Australia
Commonwealth Department of Health and Family Services
Commonwealth Serum Laboratories
Institute of Medical and Veterinary Science
N.S.W Health Department
Queensland Health
Royal College of Pathologists of Australasia

Review of Australian Standards. To keep abreast of progress in industry, Australian Standards are subject to periodic review and are kept up to date by the issue of amendments or new editions as necessary. It is important therefore that Standards users ensure that they are in possession of the latest edition, and any amendments thereto.

Full details of all Australian Standards and related publications will be found in the Standards Australia Catalogue of Publications; this information is supplemented each month by the magazine 'The Australian Standard', which subscribing members receive, and which gives details of new publications, new editions and amendments, and of withdrawn Standards.

Suggestions for improvements to Australian Standards, addressed to the head office of Standards Australia, are welcomed. Notification of any inaccuracy or ambiguity found in an Australian Standard should be made without delay in order that the matter may be investigated and appropriate action taken.

This Standard was issued in draft form for comment as DR 95428.

AS 3864—1997

Australian Standard[®]

**Medical refrigeration equipment—
For the storage of blood and blood
products**

Originated as AS 3864—1991.
Second edition 1997.

Incorporating:
Amdt 1—1998

PUBLISHED BY STANDARDS AUSTRALIA
(STANDARDS ASSOCIATION OF AUSTRALIA)
1 THE CRESCENT, HOMEBUSH, NSW 2140

ISBN 0 7337 1151 0

PREFACE

This Standard was prepared by the Standards Australia Committee HT/10, Medical Refrigeration, to supersede AS 3864—1991, *Medical refrigeration equipment—For the storage of blood and blood products, and containers for the transport of blood and blood products*.

In the preparation of this Standard, account has been taken of testing conducted by the N.S.W. Red Cross Society Blood Transfusion Service.

Whilst this Standard specifies test methods, other methods which have been demonstrated to produce equivalent or comparable results, are acceptable.

As this Standard deals with the manufacture of new medical refrigeration equipment, existing equipment may not meet all aspects of this Standard.

The objective of this Standard is to ensure that blood and blood products can be safely stored at pre-determined temperatures prior to use in patients.

The term 'normative' has been used in this Standard to define the application of the appendix to which it applies. A 'normative' appendix is an integral part of a Standard.

The principal differences between this edition and the 1991 edition are as follows:

- (a) A temperature not warmer than -25°C is specified for the storage of plasma products, and a broader temperature range of 2°C to 6°C is specified for the storage of blood and blood products.
- (b) The alarms are to activate if the temperature of the lagged probe falls to 2.5°C or reaches 5.5°C for the storage of red cells, or rises to -27°C for the storage of plasma. In practice, it would mean that alarms would activate within the accepted temperature range.
- (c) Lagging material for temperature probes of recording systems of freezers and freezer rooms has been changed to the thermal capacity equivalent to 100 ± 10 g of ice. The use of aluminium cylinders as suitable lagging material has been incorporated.
- (d) The operation test for cabinets (Appendices D and H) may be performed at ambient temperatures of 10°C and 32°C only.

© Copyright — STANDARDS AUSTRALIA

Users of Standards are reminded that copyright subsists in all Standards Australia publications and software. Except where the Copyright Act allows and except where provided for below no publications or software produced by Standards Australia may be reproduced, stored in a retrieval system in any form or transmitted by any means without prior permission in writing from Standards Australia. Permission may be conditional on an appropriate royalty payment. Requests for permission and information on commercial software royalties should be directed to the head office of Standards Australia.

Standards Australia will permit up to 10 percent of the technical content pages of a Standard to be copied for use exclusively in-house by purchasers of the Standard without payment of a royalty or advice to Standards Australia.

Standards Australia will also permit the inclusion of its copyright material in computer software programs for no royalty payment provided such programs are used exclusively in-house by the creators of the programs.

Care should be taken to ensure that material used is from the current edition of the Standard and that it is updated whenever the Standard is amended or revised. The number and date of the Standard should therefore be clearly identified.

The use of material in print form or in computer software programs to be used commercially, with or without payment, or in commercial contracts is subject to the payment of a royalty. This policy may be varied by Standards Australia at any time.

CONTENTS

	<i>Page</i>
SECTION 1 SCOPE AND GENERAL	
1.1 SCOPE	5
1.2 REFERENCED DOCUMENTS	5
1.3 DEFINITIONS	6
1.4 ELECTRICAL SAFETY	6
1.5 CARE AND MAINTENANCE	6
1.6 MARKING	6
 SECTION 2 SELF-CONTAINED REFRIGERATION EQUIPMENT FOR THE STORAGE OF BLOOD AND BLOOD PRODUCTS IN THE TEMPERATURE RANGE 2°C TO 6°C	
2.1 SCOPE OF SECTION	7
2.2 CONSTRUCTION	7
2.3 LIGHTING	8
2.4 REFRIGERATION SYSTEM	8
2.5 GENERAL CONDITIONS OF TESTING	8
2.6 PERFORMANCE REQUIREMENTS	8
2.7 INSTRUMENTS AND TEMPERATURE MEASURING APPARATUS	9
2.8 INFORMATION TO BE SUPPLIED BY THE MANUFACTURER	10
 SECTION 3 SELF-CONTAINED REFRIGERATION EQUIPMENT FOR THE STORAGE OF BLOOD PLASMA AND PLASMA PRODUCTS AT OR BELOW –25°C	
3.1 SCOPE OF SECTION	12
3.2 CONSTRUCTION	12
3.3 REFRIGERATION SYSTEM	13
3.4 GENERAL CONDITIONS OF TESTING	13
3.5 PERFORMANCE REQUIREMENTS	14
3.6 INSTRUMENTS AND TEMPERATURE CONTROL APPARATUS	14
3.7 INFORMATION TO BE SUPPLIED BY THE MANUFACTURER	15
 SECTION 4 WALK-IN TYPE COOLROOMS FOR THE STORAGE OF BLOOD AND BLOOD PRODUCTS IN THE TEMPERATURE RANGE 2°C TO 6°C	
4.1 SCOPE OF SECTION	16
4.2 CONSTRUCTION	16
4.3 LIGHTING	18
4.4 REFRIGERATION SYSTEM	18
4.5 GENERAL CONDITIONS OF TESTING	18
4.6 PERFORMANCE REQUIREMENTS	18
4.7 INSTRUMENTS AND TEMPERATURE MONITORING APPARATUS . . .	19
4.8 INFORMATION TO BE SUPPLIED BY THE MANUFACTURER	20

This is a free preview. Purchase the entire publication at the link below:

[Product Page](#)

-
- Looking for additional Standards? Visit Intertek Inform Infostore
 - Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation
-