

Superseded by AS/NZS 2496:1995

AS 2496—1981
UDC 615.471:616.089.5-78:621.415

615.471;
616-089.5

Australian Standard 2496—1981

BREATHING ATTACHMENTS FOR ANAESTHETIC PURPOSES FOR HUMAN USE



STANDARDS ASSOCIATION OF AUSTRALIA

Incorporated by Royal Charter



THE FOLLOWING SCIENTIFIC, INDUSTRIAL AND GOVERNMENTAL ORGANIZATIONS and departments were officially represented on the committee entrusted with the preparation of this standard:

Australian and New Zealand Intensive Care Society
Australian Chamber of Commerce
Australian Society of Anaesthetists
Confederation of Australian Industry
Metal Trades Industry Association of Australia
Royal Australasian College of Surgeons
State Departments of Health

This standard, prepared by Committee MD/7, Anaesthetic Equipment and Medical Breathing Machines, was approved on behalf of the Council of the Standards Association of Australia on 3 August 1981, and was published on 12 October 1981.

To keep abreast of progress in industry, Australian standards are subject to continuous review and are kept up-to-date by the issue of amendments or new editions as necessary. It is important therefore that standards users ensure that their standards are up-to-date. Full details of all SAA publications will be found in the Annual List of Australian Standards; these details are supplemented by listings in the SAA monthly journal 'The Australian Standard'. Information on the Annual List and 'The Australian Standard' may be obtained from any sales office of the Association, where details are also available of the current status of individual standards. Suggestions for improvements to published standards, addressed to the head office of the Association, are welcomed.

This standard was issued in draft form for public review as DR 79156.

AUSTRALIAN STANDARD

**BREATHING ATTACHMENTS
FOR ANAESTHETIC PURPOSES
FOR HUMAN USE**

AS 2496—1981

First published (as AS T37)	1966 ✓
AS 2496 first published	1981

**PUBLISHED BY THE STANDARDS ASSOCIATION OF AUSTRALIA
STANDARDS HOUSE, 80 ARTHUR ST, NORTH SYDNEY, N.S.W. 2060**

ISBN 0 7262 2370 0



22 OCT 1981

PREFACE

This standard was prepared by the Association's Committee on Anaesthetic Equipment and Medical Breathing Machines. It supersedes AS T37—1966, Breathing Attachments for Anaesthetic Apparatus.

The standard relies heavily on work done in this area by BSI* and ISO/TC 121/SC 1† on specifications for breathing attachments for anaesthetic apparatus. Among a number of differences from ISO practice, this standard specifies the connection port for the reservoir bag on a circle absorber unit as female.

Although this standard specifies requirements for breathing attachments, it alone cannot ensure the satisfactory gas-tight seal required for safe anaesthesia, as the circuit is constructed of a variety of components outside the scope of this standard. It is therefore strongly recommended that all anaesthetic circuits should be pressure tested before use. A satisfactory result requires a leak of less than 1 L/min with a pressure of 4 kPa. A low pressure gauge with 22 mm conical tapers has been incorporated in circuits in Fig. B1(a) to illustrate this.

The standard describes two methods to reduce the risk of incorrect connection of circle system hoses, one by the use of male/female taper connections and the other by the use of colour coding.

The physical characteristics of plastics materials pose particular problems in that the dimensions of fittings may have to vary slightly from those specified for the metal components in order to make a satisfactory fit. A gauging test for conical fittings made of plastics materials is therefore included in this standard.

Several components of breathing systems, e.g. carbon dioxide absorbers and Ayres T-pieces, have a side arm to receive anaesthetic gases from the common gas outlet of an anaesthetic machine or other gas source. The dimensional requirements of these 'fresh gas inlets' are therefore also specified in this standard.

Provision is made for the 23 mm conical fittings for vaporizers which are *unsuitable* for use in the breathing system. Usually such vaporizers impose a high resistance.

This standard may require reference to the following standards:

AS 1721	General Purpose Metric Screw Threads
AS K185	Colours for Specific Purposes
BS 2050	Electrical Resistance of Conducting and Antistatic Products made from Flexible Polymeric Material.

*British Standards Institution.

†International Organization for Standardization.

CONTENTS

	<i>Page</i>
SPECIFICATION	
1 Scope	6
2 Application	6
3 Definitions	6
4 Materials for Rigid Components for Breathing Attachments	6
5 Electrical Conductivity	6
6 Sterilization	6
7 Conical Fittings Made of Rigid Material	7
8 Conical Fitting Made of Semi-rigid Plastics Material Excluding Elastomers	7
9 Sockets Made of Elastomeric Material	7
10 Connections for Reservoir Bag Port, Reservoir Bag and Corrugated Tubing	7
11 Size and Sequence of Conical Fittings for Adult Use	7
12 Fittings Connecting Anaesthetic Ventilator, Patient, Spirometer and Humidifier	7
13 Size and Sequence of Conical Fittings for Paediatric Use	7
14 Marking of Flow-direction-sensitive Components	8
15 Application of Screw-threaded, Weight-bearing Fittings	8
16 Connections for Vaporizers	9
17 Fresh Gas Inlet Connectors to Components of Breathing Systems	9
APPENDICES	
A Breathing Attachments for Use with Adults	16
B Some Widely Used Systems Incorporating Flow-direction-sensitive Components	17
FIGURES	
1 Rigid conical fittings of 15 mm size for anaesthetic breathing attachments	10
2 Plug and ring test gauges for conical fittings of 15 mm size made of semi-rigid plastics material	11
3 Rigid conical fittings of 22 mm size (with and without recess) for anaesthetic breathing attachments	11
4 Plug and ring test gauges for conical fittings of 22 mm size made of semi-rigid plastics material	12
5 Rigid conical fittings of 23 mm size	12
6 Plug and ring test gauges for conical fittings of 23 mm size made of semi-rigid plastics material	13
7 Rigid conical fittings of 30 mm size for anaesthetic breathing attachments	13
8 Plug and ring test gauges for conical fittings of 30 mm size made of semi-rigid plastics material	14
9 Adaptors for semi-permanent connection to reservoir bags and corrugated tubing	14
10 Fresh gas inlet nipple—plain.....	15
11 Fresh gas inlet—screw-threaded connector assembly and machine-end connection	15
B1 Schematic layouts of breathing attachments for adult use:	
(a) Schematic layout of partial rebreathing system	18
(b) Schematic layout of circle absorption system	18
(c) Schematic layout of circle absorption system incorporating a flow-direction-sensitive component	19
(d) Schematic layout of coaxial breathing system	19
(e) Schematic layout of non-rebreathing system	20
(f) Schematic layout with non-rebreathing valve	20

This is a free preview. Purchase the entire publication at the link below:

[Product Page](#)

-
- Looking for additional Standards? Visit Intertek Inform Infostore
 - Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation
-