

REFERENCE COPY
INFORMATION CENTRE
STANDARDS AUSTRALIA
under Revision see DR96275

AS/NZS 3200.2.22:1993

Australian/New Zealand Standard

**Approval and test specification—
Medical electrical equipment**

**Part 2.22: Particular requirements
for safety—Diagnostic and
therapeutic laser equipment**

STANDARDS AUSTRALIA


STANDARDS NEW ZEALAND

AS/NZS 3200.2.22:1993

This Joint Australian/New Zealand Standard was prepared by Joint Technical Committee HT/28, Lasers in Medical Procedures. It was approved on behalf of the Council of Standards Australia on 21 December 1992 and on behalf of the Council of Standards New Zealand on 14 May 1993. It was published on 17 May 1993.

The following interests are represented on Committee HT/28:

Australasian College of Dermatologists
Australasian College of Physical Scientists and Engineers in Medicine
Australian Centre for Medical Laser Technology
Australian Confederation of Operating Room Nurses
Australian Radiation Protection Association
Department of Health, Housing and Community Services
Department of Health, N.S.W.
Health Department, Vic.
Health Department, W.A.
Institute of Biomedical Engineering
Medical Industry Association of Australia
Ministry of Commerce, New Zealand, Energy & Resources Division
National Health & Medical Research Council
Royal Australasian College of Physicians
Royal Australasian College of Surgeons
Royal Australian College of Obstetricians and Gynaecologists
Royal Australian College of Ophthalmologists
Standards New Zealand

Additional interests participating in preparation of Standard:

Equipment manufacturers/suppliers

Review of Standards. To keep abreast of progress in industry, Joint Australian/ New Zealand Standards are subject to periodic review and are kept up to date by the issue of amendments or new editions as necessary. It is important therefore that Standards users ensure that they are in possession of the latest edition, and any amendments thereto.

Full details of all Joint Standards and related publications will be found in the Standards Australia and Standards New Zealand Catalogue of Publications; this information is supplemented each month by the magazines 'The Australian Standard' and 'Standards New Zealand', which subscribing members receive, and which give details of new publications, new editions and amendments, and of withdrawn Standards.

Suggestions for improvements to Joint Standards, addressed to the head office of either Standards Australia or Standards New Zealand, are welcomed. Notification of any inaccuracy or ambiguity found in a Joint Australian/New Zealand Standard should be made without delay in order that the matter may be investigated and appropriate action taken.

AS/NZS 3200.2.22:1993

Australian/New Zealand Standard

Approval and test specification— Medical electrical equipment

Part 2.22: Particular requirements for safety—Diagnostic and therapeutic laser equipment

First published as Joint Standard AS/NZS 3200.2.22:1993.

PUBLISHED JOINTLY BY:

STANDARDS AUSTRALIA
1 The Crescent,
Homebush NSW 2140 Australia

STANDARDS NEW ZEALAND
6th Floor, Wellington Trade Centre,
181-187 Victoria Street,
Wellington 1 New Zealand

ISBN 0 7262 8046 1

PREFACE

This Standard is issued as a Joint Standard under the terms of the Active Cooperation Agreement between Standards Australia and Standards New Zealand. It was prepared by the Committee on Lasers in Medical Procedures on behalf of the Standards Australia Multitechnics Standards Policy Board. It is technically equivalent to IEC 601-2-22 (1992), *Medical electrical equipment, Part 2: Particular requirements for the safety of diagnostic and therapeutic laser equipment*.

The Standard is one of a series of Approval and Test Specifications issued by Standards Australia and Standards New Zealand for individual items of medical electrical equipment. It is supplementary to AS 3200.1—1990/NZS 6150:1990, *Approval and test specification—Medical electrical equipment, Part 1: General requirements for safety*. Only safety matters and closely aligned conditions are covered by this series. In some instances, however, these are more stringent than for most electrical appliances because of the additional requirements necessary to ensure the safety of the patient, and because of the environmental conditions, e.g. high humidity and hazardous locations, in which some equipment is likely to be used.

The international Standard IEC 601-2-22 modifies and supplements the corresponding clauses of IEC 601-1 second edition (1988), *Medical electrical equipment, Part 1: General requirements for safety*. This second edition has been adopted as the Australian/New Zealand Standard AS 3200.1—1990/NZS 6150:1990, *Approval and test specification—Medical electrical equipment, Part 1: General requirements for safety*, hereinafter referred to as the General Standard.

The clauses of this Particular Standard supplement or modify the corresponding clauses in the General Standard. As stated in Clause 1.3 of AS 3200.1—1990/NZS 6150:1990, the requirements of a Particular Standard take priority, where appropriate, over those of the General Standard. Where the reference in the text of this Standard indicates an 'amendment', 'addition' or 'replacement' of the relevant requirements, tests or explanatory notes of AS 3200.1—1990/NZS 6150:1990, these changes are made to the relevant text which then becomes part of the Standard.

Sub-clauses or figures which are additional to the General Standard are numbered starting from 101; additional appendices are indicated by a sequence of double capitals e.g. AA), BB), and additional items are denoted by an aa), bb), ... sequence.

In some parts of this series Australian/New Zealand requirements differ from the IEC specifications. To accommodate Australian/New Zealand conditions and well-accepted safety practices, a marginal bar is placed alongside the IEC text, and Appendix ZZ details the Australian/New Zealand specifications and rationale for the deviations.

In the text of this Standard, the following print types are used:

- (a) Requirements, compliance with which can be tested and definitions: in roman type;
- (b) Explanations, advice, introductions, general statements, exceptions and references: in smaller roman type;
- (c) Headings of subclauses and test specifications: *in italic type*;
- (d) Terms used throughout the Standard, which have been defined in Clause 2 and which are also in the index: IN SMALL CAPITALS;
- (e) An asterisk is located prior to each clause for which rationale is included in Appendix AA.

For the purposes of this Joint Standard, the IEC text should be modified as follows:

Decimal marker—substitute a full point for a comma as a decimal marker.

This Standard requires reference to other Standards which are detailed in Appendix LL.

3

Under arrangements made between Standards Australia/Standards New Zealand and the international Standards bodies, ISO and IEC, as well as certain other Standards organizations, users of this Joint Standard are advised that in Australia, copyright is vested in Standards Australia; in New Zealand, copyright is vested in Standards New Zealand.

© Copyright – STANDARDS AUSTRALIA/STANDARDS NEW ZEALAND

Users of Standards are reminded that copyright subsists in all Standards Australia and Standards New Zealand publications and software. Except where the Copyright Act allows and except where provided for below no publications or software produced by Standards Australia or Standards New Zealand may be reproduced, stored in a retrieval system in any form or transmitted by any means without prior permission in writing from Standards Australia or Standards New Zealand. Permission may be conditional on an appropriate royalty payment. Australian requests for permission and information on commercial software royalties should be directed to the head office of Standards Australia. New Zealand requests should be directed to Standards New Zealand.

Up to 10 percent of the technical content pages of a Standard may be copied for use exclusively in-house by purchasers of the Standard without payment of a royalty or advice to Standards Australia or Standards New Zealand.

Inclusion of copyright material in computer software programs is also permitted without royalty payment provided such programs are used exclusively in-house by the creators of the programs.

Care should be taken to ensure that material used is from the current edition of the Standard and that it is updated whenever the Standard is amended or revised. The number and date of the Standard should therefore be clearly identified.

The use of material in print form or in computer software programs to be used commercially, with or without payment, or in commercial contracts is subject to the payment of a royalty. This policy may be varied by Standards Australia or Standards New Zealand at any time.

This is a free preview. Purchase the entire publication at the link below:

[Product Page](#)

-
- Looking for additional Standards? Visit Intertek Inform Infostore
 - Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation
-