



**NSAI**  
Standards

Irish Standard Recommendation  
S.R. CEN/TS 17789-1:2022

Organo-mineral fertilizers - Identification  
of chelating agents - Part 1:  
Determination of EDTA, HEEDTA and  
DTPA by ion chromatography

**S.R. CEN/TS 17789-1:2022**

*Incorporating amendments/corrigenda/National Annexes issued since publication:*

The National Standards Authority of Ireland (NSAI) produces the following categories of formal documents:

I.S. xxx: Irish Standard — national specification based on the consensus of an expert panel and subject to public consultation.

S.R. xxx: Standard Recommendation — recommendation based on the consensus of an expert panel and subject to public consultation.

SWiFT xxx: A rapidly developed recommendatory document based on the consensus of the participants of an NSAI workshop.

*This document replaces/revises/consolidates the NSAI adoption of the document(s) indicated on the CEN/CENELEC cover/Foreword and the following National document(s):*

*NOTE: The date of any NSAI previous adoption may not match the date of its original CEN/CENELEC document.*

*This document is based on:*

CEN/TS 17789-1:2022

*Published:*

2022-04-06

*This document was published under the authority of the NSAI and comes into effect on:*

2022-04-24

ICS number:

65.080

NOTE: If blank see CEN/CENELEC cover page

NSAI  
1 Swift Square,  
Northwood, Santry  
Dublin 9

T +353 1 807 3800  
F +353 1 807 3838  
E standards@nsai.ie  
W NSAI.ie

Sales:  
T +353 1 857 6730  
F +353 1 857 6729  
W standards.ie

Údarás um Chaighdeáin Náisiúnta na hÉireann

## National Foreword

S.R. CEN/TS 17789-1:2022 is the adopted Irish version of the European Document CEN/TS 17789-1:2022, Organo-mineral fertilizers - Identification of chelating agents - Part 1: Determination of EDTA, HEEDTA and DTPA by ion chromatography

This document does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

For relationships with other publications refer to the NSAI web store.

**Compliance with this document does not of itself confer immunity from legal obligations.**

*In line with international standards practice the decimal point is shown as a comma (,) throughout this document.*

This page is intentionally left blank

TECHNICAL SPECIFICATION  
SPÉCIFICATION TECHNIQUE  
TECHNISCHE SPEZIFIKATION

CEN/TS 17789-1

April 2022

ICS 65.080

English Version

Organo-mineral fertilizers - Identification of chelating  
agents - Part 1: Determination of EDTA, HEEDTA and  
DTPA by ion chromatography

Engrais organo-minéraux - Identification des agents  
chélatants - Partie 1 : Détermination de l'EDTA, de  
l'HEEDTA et du DTPA par chromatographie ionique

Organisch-mineralische Düngemittel - Identifizierung  
von Chelatbildnern - Teil 1: Bestimmung von EDTA,  
HEEDTA und DTPA mittels Ionenchromatographie

This Technical Specification (CEN/TS) was approved by CEN on 21 February 2022 for provisional application.

The period of validity of this CEN/TS is limited initially to three years. After two years the members of CEN will be requested to submit their comments, particularly on the question whether the CEN/TS can be converted into a European Standard.

CEN members are required to announce the existence of this CEN/TS in the same way as for an EN and to make the CEN/TS available promptly at national level in an appropriate form. It is permissible to keep conflicting national standards in force (in parallel to the CEN/TS) until the final decision about the possible conversion of the CEN/TS into an EN is reached.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

## CEN/TS 17789-1:2022 (E)

<b>Contents</b>		<b>Page</b>
<b>European foreword</b> .....		<b>3</b>
<b>Introduction</b> .....		<b>4</b>
<b>1</b>	<b>Scope</b> .....	<b>5</b>
<b>2</b>	<b>Normative references</b> .....	<b>5</b>
<b>3</b>	<b>Terms and definitions</b> .....	<b>5</b>
<b>4</b>	<b>Principle</b> .....	<b>5</b>
<b>5</b>	<b>Interferences</b> .....	<b>5</b>
<b>6</b>	<b>Reagents</b> .....	<b>6</b>
<b>7</b>	<b>Apparatus</b> .....	<b>7</b>
<b>8</b>	<b>Sampling and sample preparation</b> .....	<b>8</b>
<b>9</b>	<b>Procedure</b> .....	<b>8</b>
<b>9.1</b>	<b>Preparation of the test solution</b> .....	<b>8</b>
<b>9.2</b>	<b>Preparation of the standard solutions</b> .....	<b>8</b>
<b>9.3</b>	<b>Chromatographic analysis</b> .....	<b>8</b>
<b>10</b>	<b>Expression of results</b> .....	<b>9</b>
<b>11</b>	<b>Test report</b> .....	<b>9</b>
<b>Annex A (informative) Complete names of chelating agents</b> .....		<b>10</b>
<b>Annex B (informative) Chromatogram</b> .....		<b>11</b>
<b>Bibliography</b> .....		<b>12</b>

## **European foreword**

This document (CEN/TS 17789-1:2022) has been prepared by Technical Committee CEN/TC 260 “Fertilizers and liming materials”, the secretariat of which is held by DIN.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a Standardization Request given to CEN by the European Commission and the European Free Trade Association.

Any feedback and questions on this document should be directed to the users’ national standards body. A complete listing of these bodies can be found on the CEN website.

According to the CEN/CENELEC Internal Regulations, the national standards organisations of the following countries are bound to announce this Technical Specification: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

**CEN/TS 17789-1:2022 (E)****Introduction**

Micronutrients are considered to be, in plant nutrition, a number of elements known to be needed in small amounts for proper plant growth and development. The most common are Iron (Fe), Manganese (Mn), Molybdenum (Mo), Copper (Cu), Zinc (Zn) and Boron (B).

If an organo-mineral fertilizer contains a substance, or one of the substances in the mixture, which is intended to enhance the long term availability to plants of micronutrients in the EU fertilizing product, that substance is either a chelating agent or a complexing agent.

The chelating agents are divided into two groups<sup>1</sup>:

- Group 1: EDTA, DTPA, HEEDTA, IDHA and [S,S]-EDDS;
- Group 2: Chelating agents present in UVCB (unknown or variable composition, complex reaction products or biological materials) chelates including [o,o] EDDHA, [o,p] EDDHA, [o,o] EDDHMA, HBED and EDDHSA.

This document specifies the test method to be used in order to determine the content of certain chelating agents (EDTA, DTPA and HEEDTA) in organo-mineral fertilizers (product function category (PFC) 1(B) according to Regulation (EU) 2019/1009 [4]).

---

<sup>1</sup> Abbreviated terms are described in Annex A.



## 1 Scope

This document specifies a method for the determination by ion chromatography of the total amount of each of the individual chelating agents EDTA, HEEDTA, and DTPA in organo-mineral fertilizers, having an organic matrix based on vegetal residues (cocoa shells, grape residue, soybean residue, ...), algae extract, and animal meal (feather, bones, blood, ...) and containing one or more of these substances.

The method allows the identification and the determination of the total water-soluble fraction of each of these chelating agents. It does not allow to distinguish between the free form and the metal bound form of the chelating agents.

This method applies to organo-mineral fertilizers containing chelates of one or more of the following micronutrients: cobalt, copper, iron, manganese, zinc and with a mass fraction of at least 0,1 %.

## 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 12944-1, *Fertilizers and liming materials — Vocabulary — Part 1: General terms*

EN 12944-2, *Fertilizers and liming materials — Vocabulary — Part 2: Terms relating to fertilizers*

## 3 Terms and definitions

For the purposes of this document, the terms and definitions given in EN 12944-1 and EN 12944-2 apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

## 4 Principle

The micronutrients associated with the chelating agents present in an aqueous extract of the sample are replaced by iron(III). The iron chelates are separated and determined by ion chromatography. The separation is based on anion exchange, by elution with a nitrate acetate solution. The detection is based on UV photometry at 330 nm, after post-column reaction with diluted perchloric acid.

## 5 Interferences

Several substances can interfere, to a degree largely dependent on the type of column used. With the column described in 7.2, the following phenomena have been observed.

- a) Injection of solutions having high concentrations of salts can cause shifts in the retention times, mostly decreasing the retention when compared to the standard solutions. In these cases, the identity of the peaks can be confirmed by standard addition.
- b) Solutions having high concentrations of salts can also create a large signal at the void volume, poorly resolved from the HEEDTA peak.
- c) High concentrations of nitrate, carbonate, sulfate, and phosphate do not interfere. However, large amounts of chloride create a negative fronting peak poorly resolved from the DTPA peak and altering its peak shape.

This is a free preview. Purchase the entire publication at the link below:

[Product Page](#)

- 
- [Looking for additional Standards? Visit Intertek Inform Infostore](#)
  - [Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation](#)
-