



National Standards Authority of Ireland

IRISH STANDARD

I.S. 68:1955

ICS 67.100.10

National Standards
Authority of Ireland
Glasnevin, Dublin 9
Ireland

Tel: +353 1 807 3800
Fax: +353 1 807 3838
<http://www.nsai.ie>

DETERMINATION OF THE PERCENTAGE OF FAT IN CREAM

Sales
<http://www.standards.ie>

*This Irish Standard was
published under the
authority of the National
Standards Authority of
Ireland and comes into
effect on:
June 1, 1956*

**NO COPYING WITHOUT NSAI
PERMISSION EXCEPT AS
PERMITTED BY COPYRIGHT
LAW**

© NSAI 1955

Price Code D

Údarás um Chaighdeáin Náisiúnta na hÉireann

I. S. 68 : 1955

Irish Standard Specification

**DETERMINATION OF THE
PERCENTAGE OF FAT IN
CREAM**

Institute for Industrial Research and Standards

IRISH STANDARD SPECIFICATION

DETERMINATION OF THE PERCENTAGE OF FAT IN CREAM

I. S. 68 : 1955

Price ~~4/-~~ 2/6

INSTITUTE FOR INDUSTRIAL RESEARCH AND STANDARDS

Glasnevin House, Ballymun Road, Dublin

Telegrams : Research, Dublin

Telephone : Dublin 45161

SPECIFICATION

ENTITLED

STANDARD SPECIFICATION (DETERMINATION OF THE PERCENTAGE OF FAT
IN CREAM), 1955.

The Institute for Industrial Research and Standards in exercise of the power conferred by section 2 of the Industrial Research and Standards (Amendment) Act, 1954 (No. 11 of 1954), and with the consent of the Minister for Industry and Commerce (signified by his Official Seal affixed hereto), hereby declares as follows :

1. This instrument may be cited as the Standard Specification (Determination of the Percentage of Fat in Cream), 1955.

2. (1) The specification set forth in the Schedule shall be the standard specification for Determination of the Percentage of Fat in Cream.

(2) The said standard specification may be cited as Irish Standard 68 : 1955 or I.S. 68 : 1955.

SCHEDULE

Determination of the Percentage of Fat in Cream

SCOPE

1. This specification deals with the determination of the percentage of fat in cream by the Gerber method.

BUTYROMETER

2. The butyrometer shall be made of clear glass, and shall be as free as possible from striae and similar defects. It shall be well annealed so as to be resistant to the thermal shocks incidental to the method.

The butyrometer shall be a cream butyrometer of the form shown in Figure 1, shall have an overall length of 190 mm. \pm 5 mm. and shall comply with the requirements set out below.

(a) *Neck* : The butyrometer shall have a corrugated neck or a plain neck.

For butyrometers with corrugated necks the corrugations shall be of the "ring" type shown in Figure 1. The corrugations shall be 3 or 4 in number and shall be horizontal and not formed in a helix to produce a screw thread.

For butyrometers with plain necks, as shown in Figure 1, the wall thickness at the neck shall be not less than 1.5 mm. The neck may have an external rim at the top but the total thickness at the rim shall be not more than 2.5 mm.

The length of the neck shall be 14.5 mm. \pm 1.5 mm. The internal diameter shall be 11.5 mm. \pm 0.5 mm. The internal diameter of butyrometers with corrugated necks shall be measured between the crests of the corrugations.

(b) *Body* : The external diameter of the body shall be not more than 25 mm. The capacity of the butyrometer between the base of the neck and the graduation mark nearest to the body, *i.e.*, between the points A and B in Figure 1, shall be not less than 18 ml. and not more than 19 ml.

This is a free preview. Purchase the entire publication at the link below:

[Product Page](#)

-
- Looking for additional Standards? Visit Intertek Inform Infostore
 - Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation
-