



National Standards Authority of Ireland

IRISH STANDARD

I.S. EN 12955:1999

ICS 67.060
67.080.10
67.200.10

**Foodstuffs - Determination Of Aflatoxin B₁,
And The Sum Of Aflatoxins B₁, B₂, G₁ And G₂ In
Cereals, Shell-fruits And Derived Products -
High Performance Liquid Chromatographic
Method With Post Column Derivatization And
Immunoaffinity Column Clean Up**

National Standards
Authority of Ireland
Glasnevin, Dublin 9
Ireland

Tel: +353 1 807 3800
Fax: +353 1 807 3838
<http://www.nsai.ie>

Sales
<http://www.standards.ie>

*This Irish Standard was
published under the authority
of the National Standards
Authority of Ireland
and comes into effect on:*

December 17, 1999

**NO COPYING WITHOUT NSAI
PERMISSION EXCEPT AS
PERMITTED BY COPYRIGHT
LAW**

© NSAI 1999

Price Code F

Údarás um Chaighdeáin Náisiúnta na hÉireann

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN 12955

July 1999

ICS 67.060; 67.080.10; 67.200.10

English version

Foodstuffs - Determination of aflatoxin B₁, and the sum of
aflatoxins B₁, B₂, G₁ and G₂ in cereals, shell-fruits and derived
products - High performance liquid chromatographic method
with post column derivatization and immunoaffinity column clean
up

Produits alimentaires - Dosage de l'aflatoxine B₁ et de la
somme des aflatoxines B₁, B₂, G₁ et G₂ dans les céréales,
les fruits à coque et les produits dérivés - Méthode de
chromatographie en phase liquide haute performance avec
dérivation post-colonne et purification en colonne
d'immuno-affinité

Lebensmittel - Bestimmung von Aflatoxin B₁ und der
Summe von Aflatoxin B₁, B₂, G₁ und G₂ in Getreiden,
Schalenfrüchten und verwandten Produkten -
Hochleistungs-flüssigchromatographisches Verfahren mit
Nachsäulenderivatisierung und Immunoaffinitätssäulen-
Reinigung

This European Standard was approved by CEN on 7 June 1999.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Central Secretariat: rue de Stassart, 36 B-1050 Brussels

Contents

	Page
1 Scope	3
2 Normative reference(s)	3
3 Principle	3
4 Reagents	3
5 Apparatus	5
6 Procedure	6
7 Calculation of results	7
8 Precision	8
9 Test report	9
Annex A (informative) Precision data	10
Annex B (informative) Bibliography	12

Foreword

This European Standard has been prepared by Technical Committee CEN/TC 275 "Food analysis - Horizontal methods", the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by January 2000, and conflicting national standards shall be withdrawn at the latest by January 2000.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and the United Kingdom.

NOTE Existing and developing legislation (national or EU) in this area will require a method with lower levels of detection. Such a method is currently being developed as an EU SMT project.

1 Scope

This European Standard specifies a method for the determination of aflatoxin contents of greater than 8 µg/kg.

The method has been successfully validated in an interlaboratory study according to ISO 5725:1986 on maize containing 24,5 µg/kg, peanut butter containing 8,4 µg/kg and raw peanuts containing 16 µg/kg of total aflatoxins.

Some laboratory experiences have shown that this method can be used to several types of cereals, oilseed products, shell-fruits, dried fruits and derived products, after in-house validation.

2 Normative reference(s)

This European Standard incorporates by dated or undated reference, provisions from other publications. These normative references are cited at the appropriate places in the text and the publications are listed hereafter. For dated references, subsequent amendments to or revisions of any of these publications apply to this European Standard only when incorporated in it by amendment or revision. For undated references the latest edition of the publication referred to applies.

EN ISO 3696 *Water for analytical laboratory use - Specification and test methods (ISO 3696:1987)*

3 Principle

The test sample is extracted with a mixture of methanol and water. The sample extract is filtered, diluted with water, and applied to an affinity column containing antibodies specific for aflatoxins B₁, B₂, G₁ and G₂. The aflatoxins are isolated, purified and concentrated on the column then removed from the antibodies with methanol. The aflatoxins are quantified by reverse-phase high performance liquid chromatography (HPLC) with fluorescence detection and postcolumn iodine derivatization.

WARNING - The method described requires the use of solutions of aflatoxins. Aflatoxins are carcinogenic to humans. Attention is drawn to the statement made by the International Agency for Research on Cancer (WHO) [1], [2].

4 Reagents

4.1 General

During the analysis, unless otherwise stated, use only reagents of recognized analytical grade and only water according to grade 1 of EN ISO 3696.

4.2 Sodium chloride

4.3 Iodine, crystalline

4.4 Aflatoxin, in crystal form or as a film in ampoules.

WARNING: Protect the laboratory, where the analyses are done, adequately from daylight. This can be achieved effectively by using Ultraviolet (UV) absorbing foil on the windows in combination with subdued light (no direct sunlight) or curtains or blinds in combination with artificial light (fluorescent tubes are acceptable).

Protect Aflatoxin containing solutions from light as much as possible (keep in the dark, use aluminium foil or amber-coloured glassware).

4.5 Acetonitrile, for HPLC

4.6 Methanol, for analysis

4.7 Methanol, for HPLC

4.8 Toluene

4.9 Extraction solvent

Mix 7 parts per volume of methanol (4.6) with 3 parts per volume of water.

4.10 Immunoaffinity column

The immuno affinity (IA) column contains antibodies raised against aflatoxin B₁, B₂, G₁ and G₂. The column shall have a minimum binding capacity of not less than 100 ng of aflatoxin B₁ and shall give a recovery of not less than 80 % for aflatoxin B₁, B₂, G₁ and not less than 60 % for aflatoxin G₂ when a standard solution in 15 ml of a methanol/water-mixture (1 part per volume of methanol and 3,4 parts per volume of water) containing 5 ng of each toxin is applied on the IA column. The IA column should provide an appropriate solvent reservoir, e.g. a syringe with adapter.

4.11 Mobile phase

Mix 3 parts per volume of water with 1 part per volume of acetonitrile (4.5) and 1 part per volume of methanol (4.7). Degas the solution before use.

4.12 Postcolumn derivatization reagent

Dissolve 100 mg of iodine (4.3) in 2 ml of methanol (4.6). Add 200 ml of water, stir for 1 h, and filter through a 0,45 µm filter (5.8). Prepare the solution on the week of use and store the solution in the dark or in a bottle of brown glass. Before use stir the solution for 10 min.

4.13 Toluene/acetonitrile mixture

Mix 98 parts per volume of toluene (4.8) with 2 parts per volume of acetonitrile (4.5).

4.14 Aflatoxin B₁, B₂, G₁ and G₂ stock solutions

Dissolve aflatoxin B₁, B₂, G₁ and G₂ separately in the toluene/acetonitrile mixture (4.13) to give separate solutions containing 10 µg/ml.

To determine the exact concentration of aflatoxin in each stock solution, record the absorption curve between a wavelength of 330 nm and 370 nm in 1 cm quartz glass cells (5.7) in a spectrometer with toluene/acetonitrile mixture (4.13) in the reference path. Calculate the aflatoxin mass concentration of each aflatoxin, ρ_i, in micrograms per millilitre, using equation (1):

$$\rho_i = \frac{A_{\max} \times M_i \times 100}{\epsilon_i \times d} \quad (1)$$

where:

A_{\max} is the absorbance determined at the maximum of the absorption curve;

M_i is the relative molecular mass of each aflatoxin, in grams per mol;

ϵ_i is the molar absorptivity of each aflatoxin in toluene/acetonitrile (4.13), in metres squared per mol;

d is the optical path length of the cell, in centimetres.

M_i and ϵ_i are given in table 1.

Table 1 — Relative molecular mass and molar absorptivity of aflatoxins B₁, B₂, G₁ and G₂

(Mixture of toluene and acetonitrile 98 + 2)

Aflatoxin	M_i , g/mol	ϵ_i , m ² /mol
B ₁	312	1930
B ₂	314	2040
G ₁	328	1660
G ₂	330	1790

This is a free preview. Purchase the entire publication at the link below:

[Product Page](#)

-
- [Looking for additional Standards? Visit Intertek Inform Infostore](#)
 - [Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation](#)
-