



National Standards Authority of Ireland

IRISH STANDARD

I.S. EN 50147-2:1999

ICS 17.220
29.020

**ANECHOIC CHAMBERS. PART 2:
ALTERNATIVE TEST SITE SUITABILITY WITH
RESPECT TO SITE ATTENUATION**

National Standards
Authority of Ireland
Glasnevin, Dublin 9
Ireland

Tel. +353 1 807 3800
Fax: +353 1 807 3838
<http://www.nsai.ie>

Sales
<http://www.standards.ie>

*This Irish Standard was
published under the
authority of the National
Standards Authority of
Ireland and comes into
effect on:*

January 29, 1999

**NO COPYING WITHOUT NSAI
PERMISSION EXCEPT AS
PERMITTED BY COPYRIGHT
LAW**

© NSAI 1999

Price Code I

Údarás um Chaighdeán Náisiúnta na hÉireann

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN 50147-2

March 1996

ICS 17.220.00; 29.020

Descriptors: Electromagnetic compatibility, tests, anechoic chambers, enclosures, specifications

English version

Anechoic chambers
Part 2: Alternative test site suitability with respect to site attenuation

Chambres anéchoïques
Partie 2: Aptitude d'un emplacement
d'essai de substitution en ce qui
concerne l'affaiblissement de
l'emplacement

Absorberräume
Teil 2: Eignung alternativer Meßplätze
bezüglich Meßplatzdämpfung

This European Standard was approved by CENELEC on 1995-07-04. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the Central Secretariat has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom.

CENELEC

European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

Central Secretariat: rue de Stassart 35, B - 1050 Brussels

Foreword

This European Standard was prepared by working group WG 4 of CENELEC Technical Committee TC 210, EMC.

It was submitted to the CENELEC Unique Acceptance procedure (UAP) in August 1994 and was approved by CENELEC as EN 50147-2 on 1995-07-04.

The following dates were fixed:

- latest date by which the EN has to be implemented
at national level by publication of an identical
national standard or by endorsement (dop) 1996-09-01
 - latest date by which national standards
conflicting with the EN have to be withdrawn (dow) 1996-09-01
-

Contents

	Page
Foreword	2
1 Scope	4
2 Normative references	4
3 Abbreviations and definitions	4
4 Requirements	5
4.1 Site attenuation	5
4.2 Conducting ground plane	5
5 Test procedure	5
Figures	7
Annexes	
A Tables for NSA (normative)	9
B Formulas for NSA (informative)	22

1 Scope

There are many different test sites and facilities that have been constructed to make radiated emission measurements at frequencies above 30 MHz according to CISPR 16-1. Most are protected from the weather and the adverse effects of the radio frequency ambient noise. These include all weather covered open area test sites and absorber lined shielded enclosures.

Whenever construction material encloses a test site, there exists the possibility that the result of a single normalized site attenuation measurement as specified in subclause 16.6 of CISPR 16-1 is not adequate to show such alternate site suitability.

NOTE: Radiated emission sources located near dielectric interfaces have been shown to have variations in current distribution that can affect the radiated properties of the source at that location. When the EUT can be located near these interfaces, additional site attenuation measurements are required.

This standard specifies requirements for alternative test sites regarding site attenuation. As long as the document CISPR/A(CO) 63 is not published as part of CISPR 16, this standard shall be used for test site qualification.

2 Normative references

This European Standard incorporates by dated or undated reference, provisions from other publications. These normative references are cited at the appropriate places in the text and the publications listed hereafter. Dated references, subsequent amendments to or revisions of any of these publications apply to this European Standard only when incorporated in it by amendment or revision. For undated references the latest edition of the publication referred to applies.

IEC 50(161)	1990	International Electrotechnical Vocabulary (IEV) Chapter 161: Electromagnetic Compatibility
CISPR 16-1	1993	Specification for radio disturbance and immunity measuring apparatus and methods - Part 1: Radio disturbance and immunity measuring apparatus

3 Abbreviations and definitions

3.1 Abbreviations

EUT equipment under test (test sample including cabling)

NSA normalized site attenuation (according to CISPR 16-1)

3.2 Definitions

For the purposes of this standard, the following definition and the definitions contained in IEC50(161) apply.

3.2.1 alternative test site

A location other than an open area test site, such as an all weather protected site, an absorber-lined room, a salt mine, a tunnel etc.

This is a free preview. Purchase the entire publication at the link below:

[Product Page](#)

-
- Looking for additional Standards? Visit Intertek Inform Infostore
 - Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation
-