



National Standards Authority of Ireland

IRISH STANDARD

I.S. EN 867-4:2000

ICS 11.080

**NON-BIOLOGICAL SYSTEMS FOR USE IN
STERILIZERS - PART 4: SPECIFICATION
FOR INDICATORS AS AN ALTERNATIVE TO
THE BOWIE AND DICK TEST FOR THE
DETECTION OF STEAM PENETRATION**

National Standards
Authority of Ireland
Dublin 9
Ireland

Tel: (01) 807 3800
Tel: (01) 807 3838

*This Irish Standard was
published under the
authority of the National
Standards Authority of
Ireland
and comes into effect on.
April 12, 2001*

**NO COPYING WITHOUT NSAI
PERMISSION EXCEPT AS
PERMITTED BY COPYRIGHT
LAW**

© NSAI 2000

Price Code L

Údarás um Chaighdeáin Náisiúnta na hÉireann

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN 867-4

November 2000

ICS 11.080

English version

**Non-biological systems for use in sterilizers - Part 4:
Specification for indicators as an alternative to the Bowie and
Dick test for the detection of steam penetration**

Systèmes non-biologiques utilisés dans les stérilisateurs -
Partie 4: Spécifications relatives aux indicateurs utilisés en
alternative à l'essai de Bowie-Dick pour la détection de la
pénétration de vapeur d'eau

Nichtbiologische Systeme für den Gebrauch in
Sterilisatoren - Teil 4: Festlegungen für Indikatoren, die
alternativ zum Bowie-Dick-Test für den Nachweis der
Dampfdurchdringung verwendet werden

This European Standard was approved by CEN on 18 October 2000.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Management Centre has the same status as the official versions

CEN members are the national standards bodies of Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITE EUROPEEN DE NORMALISATION
EUROPAISCHES KOMITEE FÜR NORMUNG

Management Centre: rue de Stassart, 36 B-1050 Brussels

Contents

	Page
Foreword	3
Introduction	4
1 Scope	5
2 Normative references	5
3 Terms and definitions	6
4 General requirements	7
5 Indicator system format	7
6 Performance requirements	8
7 Packaging and labelling	9
8 Quality assurance	10
 Annex A (normative) Method for determining strength during and after steam sterilization	 11
Annex B (normative) Standard test cycles	13
Annex C (normative) Method for the estimation of the visual difference between the colour of the substrate and the changed or unchanged indicator system by determination of the relative reflectance density	18
Annex D (normative) Method for demonstrating uniform colour change on exposure to saturated steam	22
Annex E (normative) Method for determining the equivalence of the alternative indicator to the Bowie and Dick test	23
Annex F (normative) Method for determining the reproducibility of fail conditions created in a standard test pack by air injection, air leak and retained air systems	25
Annex G (normative) Method for evaluating indicator colour change on exposure to dry heat	29
Annex H (normative) Method for demonstrating the shelf life of the product	30
Annex J (normative) Method for accelerated ageing of test samples	31
Annex K (normative) Steam exposure apparatus and steam for test purposes	32
Annex L (normative) Standard test pack	34
Annex M (normative) Air injection system	35
Bibliography	37

Foreword

This European Standard has been prepared by Technical Committee CEN/TC 102 "Sterilizers for medical purposes", the secretariat of which is held by DIN.

This document is currently submitted to the Formal Vote.

EN 867 consists of the following Parts under the general title "Non-biological systems for use in sterilizers":

- Part 1: General requirements
- Part 2: Process indicators (Class A)
- Part 3: Specification for Class B indicators for use in the Bowie and Dick test
- Part 4: Specification for indicators as an alternative to the Bowie and Dick test for the detection of steam penetration
- Part 5: Specification for indicator systems and process challenge devices for use in performance testing for small sterilizers Type B and Type S

In addition, CEN/TC 102 Working Group 7 has prepared EN 866 consisting of the following Parts under the general title "Biological systems for testing sterilizers and sterilization processes":

- Part 1: General requirements
- Part 2: Particular systems for use in ethylene oxide sterilizers
- Part 3: Particular systems for use in moist heat sterilizers
- Part 4: Particular systems for use in irradiation sterilizers
- Part 5: Particular systems for use in low temperature steam and formaldehyde sterilizers
- Part 6: Particular systems for use in dry heat sterilizers
- Part 7: Particular requirements for self-contained biological indicator systems for use in moist heat sterilizers
- Part 8: Particular requirements for self-contained biological systems for use in ethylene oxide sterilizers

Introduction

The Bowie and Dick test was conceived as a test for successful air removal from high vacuum porous load sterilizers¹⁾. A successful Bowie and Dick test indicates rapid and even penetration of steam into the test pack. The presence of air within the pack, due either to an inefficient air removal stage or an air leak during this stage, or non-condensable gases in the steam supply are circumstances which can lead to a failure of the test. The result of the test may also be affected by other factors which inhibit steam penetration. The test does not necessarily demonstrate either achievement of the required temperature or maintenance of that temperature for the required time to achieve sterilization.

A failure of the Bowie and Dick test is not conclusive proof that the fault in the sterilizer is due to air retention, air leakage or non-condensable gases and it can be necessary to investigate other causes of failure.

The Bowie and Dick test is a performance test for steam sterilizers for wrapped goods and porous loads. As such it is performed during the demonstration of conformity of steam sterilizers to EN 285 and as a routine test of performance in EN 554. The method of carrying out the test is described in EN 285.

A test pack for the Bowie and Dick test consists of two components:

- a) a small standardised test load; and
- b) a chemical indicator system to detect the presence of steam (see EN 867-3).

The Bowie and Dick test as originally described¹⁾ utilized huckaback towels as the material for the test load. The test described in EN 285 uses cotton sheets for this purpose.

Indicators intended as an alternative to the Bowie and Dick Test use different materials for the test load and employ indicator systems specifically formulated for use with the defined test load. This standard specifies the performance of the indicator system in combination with the test load with which it is intended to be used. The test load may be presented with the indicator system already incorporated and intended for single-use or may be intended for multiple use with a new indicator system to be inserted prior to each use.

The indicator for which the performance is specified in this European standard is intended to indicate that steam penetration has been inadequate. The performance of the indicator specified in this European standard should be equivalent, but not necessarily identical, to the performance obtained in a Bowie and Dick test as described in EN 285. Equivalence should be regarded as providing a similar response to steam penetration with any differences being predictable and such that the necessary level of assurance of satisfactory steam penetration is provided. An indicator meeting this specification is not intended to identify which of the potential causes of poor steam penetration was responsible for the failure indicated by the test.

¹⁾ Bowie, J.H., Kelsey, J.C., and Thompson, G.R., Lancet, i, p. 586 (1963)

This is a free preview. Purchase the entire publication at the link below:

[Product Page](#)

-
- Looking for additional Standards? Visit Intertek Inform Infostore
 - Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation
-