



National Standards Authority of Ireland

IRISH STANDARD

I.S. EN 13150:2001

ICS 71.040.10

**WORKBENCHES FOR LABORATORIES -
DIMENSIONS, SAFETY REQUIREMENTS AND
TEST METHODS**

National Standards
Authority of Ireland
Dublin 9
Ireland

Tel. (01) 807 3800

Tel: (01) 807 3838

*This Irish Standard was
published under the
authority of the National
Standards Authority of
Ireland
and comes into effect on
October 26, 2001*

**NO COPYING WITHOUT NSAI
PERMISSION EXCEPT AS
PERMITTED BY COPYRIGHT
LAW**

© NSAI 2001

Price Code H

Údarás um Chaighdeáin Náisiúnta na hÉireann

EUROPEAN STANDARD

EN 13150

NORME EUROPÉENNE

EUROPÄISCHE NORM

May 2001

ICS 71.040.10

English version

Workbenches for laboratories - Dimensions, safety requirements and test methods

Paillasses de laboratoire - Dimensions, spécification de
sécurité et méthodes d'essai

Arbeitsische für Laboratorien - Maße,
Sicherheitsanforderungen und Prüfverfahren

This European Standard was approved by CEN on 16 April 2001.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Management Centre or to any CEN member

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Management Centre has the same status as the official versions

CEN members are the national standards bodies of Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITE EUROPEEN DE NORMALISATION
EUROPAISCHES KOMITEE FÜR NORMUNG

Management Centre: rue de Stassart, 36 B-1050 Brussels

Contents

	Page
Foreword	2
1 Scope	2
2 Normative references	3
3 Terms and definitions	3
4 Recommended dimensions	4
5 General safety requirements	6
6 Specific requirements	7
7 User information	7
8 Marking and labelling	8
9 Test report	8
Annex A (normative) Test methods for safety testing	9
Annex B (informative) Optional tests	12

Foreword

This European Standard has been prepared by Technical Committee CEN/TC 207 "Furniture", the secretariat of which is held by IBN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by November 2001, and conflicting national standards shall be withdrawn at the latest by November 2001.

This European Standard EN 13150 is applicable with date of publication (dop.) For workbenches for laboratories complying with a national standard before dop, this previous national standard may continue to apply for production and placing on the market until 2 years after dop. This standard does not apply to workbenches for laboratories, which have been put into service in the laboratory before 2 years after dop

Annex A is normative. Annex B is informative

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and the United Kingdom.

1 Scope

This European Standard specifies safety requirements and test methods for workbenches for laboratories including laboratory tables and gives recommendations for their dimensions.

This European Standard applies to workbenches, movable tables and workbench shelves designed for use in research, educational, quality control and similar laboratories.

This European Standard does not apply to workbenches for pupils in scientific class rooms of schools. It does not apply to workbenches for special purposes, e.g. for heavy diagnostic or processing machines.

It should be understood that fulfilling the requirements does not ensure that failure will not occur as a result of habitual misuse or after an excessively long period of service. The tests are designed to be applied to a stand-alone workbench that is fully assembled and ready for use.

Requirements and test methods related to the fire safety of workbenches and to the resistance of the work surface are not included in this European Standard.

2 Normative references

This European Standard incorporates by dated or undated reference, provisions from other publications. These normative references are cited at the appropriate places in the text and the publications are listed hereafter. For dated references, subsequent amendments to or revisions of any of these publications apply to this European standard only when incorporated in it by amendment or revision. For undated references the latest edition of the publication referred to applies (including amendments).

EN 527-1:2000, *Office furniture - Work tables and desks - Part 1: Dimensions*

EN 1730, *Domestic furniture - Tables - Test methods for determination of strength, durability and stability*

prEN 12600:1999, *Glass in building - Pendulum test - Impact test method for flat glass and performance requirements*

EN 61010-1, *Safety requirements for electrical equipment for measurement, control and laboratory use - Part 1: General requirements (IEC 61010-1:1990 + A1:1992, modified)*

ISO 48, *Rubber, vulcanized or thermoplastic - Determination of hardness (hardness between 10 IRHD and 100 IRHD)*

ISO 2813, *Paints and varnishes - Determination of specular gloss of non-metallic paint films at 20°, 60° and 85°*

3 Terms and definitions

For the purposes of this European Standard, the following terms and definitions apply.

3.1

work surface height, h_1

vertical distance between the floor and the top of the work surface (see Figure 1)

3.2

overworkbench shelf height, h_2

vertical distance between the floor and the top of the shelf surface (see Figure 1)

3.3

bottom rail height, h_3

vertical distance between the floor and the lower part of bottom rail (see Figure 2)

3.4

overall depth, d_1

horizontal distance between the front and the rear edge of the workbench including possible service zones (see Figure 2)

3.5

clear work surface depth, d_2

horizontal distance between the front and the rear edge of the workbench excluding edges and possible service zones (see Figure 2)

3.6

overworkbench shelf depth, d_3

horizontal distance between the front and the rear edge of the overworkbench shelf (see Figure 2)

3.7

service zone depth, d_4

horizontal distance of overall depth d_1 minus clear work surface depth d_2 (see Figure 2)

3.8

static test

test consisting of heavy loads being applied a few times to ensure that the workbench has sufficient strength to perform its function under the highest levels of loading that might reasonably be expected to occur

3.9

impact test

test to assess the strength of the workbench under the rapid rates of loading that occasionally occur

3.10

fatigue test

test simulating the repeated application of loads or movement of components occurring during long-term use and assessing the durability of the workbench under such conditions

3.11

structure

load bearing parts of the workbench including the frame, top and legs

3.12

workbench

work surface with supporting structure and services as required

3.13

movable workbench or table

unit which is not fixed to the floor or wall nor has any service connections

NOTE This definition is for the purpose of the drop test only (see A.3.7)

4 Recommended dimensions

Other dimensions than those recommended in this clause 4 may be agreed between customer and manufacturer.

4.1 Heights

4.1.1 Work surface height

The nominal heights h_1 are listed in Table 1. In every case there should be an levelling range of minimum 20 mm. For seated work the relationship between seat height and work surface height is critical. The use of adjustable height seat is recommended.

Table 1 - Work surface heights

Type of workbench	Nominal heights h_1 in mm
Sitting	720
Standing ^{a)}	900
^{a)} or sitting on high chairs or stools	

4.1.2 Leg room

Minimum leg room under the workbench should be in accordance with clause 6 of EN 527-1:2000.

This is a free preview. Purchase the entire publication at the link below:

[Product Page](#)

-
- [Looking for additional Standards? Visit Intertek Inform Infostore](#)
 - [Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation](#)
-