



National Standards Authority of Ireland

STANDARD

I.S. EN 61326:1998

ICS 25.040.40
33.100

**ELECTRICAL EQUIPMENT FOR
MEASUREMENT, CONTROL AND
LABORATORY USE EMC REQUIREMENTS
(IEC 61326:1997)**

National Standards
Authority of Ireland
Dublin 9
Ireland

Tel: (01) 807 3800
Fax: (01) 807 3838

*This Irish Standard was
published under the authority
of the National Standards
Authority of Ireland
and comes into effect on:*

April 24, 1998

**NO COPYING WITHOUT NSAI
PERMISSION EXCEPT AS
PERMITTED BY COPYRIGHT
LAW**

© NSAI 1998

Price Code J

Údarás um Chaighdeáin Náisiúnta na hÉireann

CEI 61326
(Deuxième édition – 2002)

**Matériels électriques de mesure, de
commande et de laboratoire –
Prescriptions relatives à la CEM**

IEC 61326
(Second edition – 2002)

**Electrical equipment for measurement,
control and laboratory use –
EMC requirements**

CORRIGENDUM 1

Page de couverture et page de titre:

Cover page and title page:

Remplacer "Deuxième édition" par "Première édition".

Replace "Second edition" by "First edition".

Page 6

Page 7

Remplacer le deuxième alinéa qui suit le point 6) par:

Replace the second paragraph following item 6) by:

Cette norme annule et remplace la première édition de la CEI 61326-1 parue en 1997, son amendement 1 (1998) et son amendement 2 (2000). Elle en constitue une révision technique.

This standard cancels and replaces the first edition of IEC 61326-1 published in 1997, its amendment 1 (1998) and its amendment 2 (2000), and constitutes a technical revision.

Page 20

Page 21

Remplacer le titre et le texte de la définition 3.12 par le mot "Vacant".

Replace the title and text of definition 3.12 with the word "Void".

Page 50

Page 51

Paragraphe E.3.2

Subclause E.3.2

Dans la première phrase, remplacer "et critère d'aptitude A tel que défini en 3.12" par "durant l'essai, comportement normal dans les limites de la spécification. Ceci suppose que les variations sont permises en dehors de l'erreur intrinsèque maximale indiquée dans les données techniques du manuel de l'utilisateur. Les variations se limitent à 5 fois l'erreur intrinsèque mais n'excèdent pas ± 20 % de la valeur mesurée lorsque la mesure est faite entre 50 % et 100 % de la pleine échelle."

In the first sentence, replace "and performance criterion A as defined in 3.12" by "during testing normal performance within the specification limits. This includes that variations are allowed outside the maximum intrinsic error documented in the technical data of the user's manual. The variations shall be limited to five times the intrinsic error but not more than ± 20 % of the measured value when measured at between 50 % and 100 % of full scale."

Juillet 2002

July 2002



National Standards Authority of Ireland

IRISH STANDARD

I.S. EN 61326/A3:2003

ICS 25.040.40
33.100

**ELECTRICAL EQUIPMENT FOR
MEASUREMENT, CONTROL AND
LABORATORY USE - EMC REQUIREMENTS
(ANNEXES E & F OF IEC 61326:2002 +
CORRIGENDUM 2002)**

National Standards
Authority of Ireland
Dublin 9
Ireland

Tel: (01) 807 3800
Tel: (01) 807 3838

*This Irish Standard was
published under the
authority of the National
Standards Authority of
Ireland
and comes into effect on:
January 9, 2004*

**NO COPYING WITHOUT NSAI
PERMISSION EXCEPT AS
PERMITTED BY COPYRIGHT
LAW**

© NSAI 2003

Price Code O

Údarás um Chaighdeáin Náisiúnta na hÉireann

This is a free preview. Purchase the entire publication at the link below:

[Product Page](#)

-
- [Looking for additional Standards? Visit Intertek Inform Infostore](#)
 - [Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation](#)
-