



National Standards Authority of Ireland

IRISH STANDARD

Ì.S. EN 13311-2:2001

ICS 07.080
07.100.01

**BIOTECHNOLOGY - PERFORMANCE
CRITERIA FOR VESSELS -
PART 2: PRESSURE PROTECTION DEVICES**

National Standards
Authority of Ireland
Dublin 9
Ireland

Tel: (01) 807 3800
Tel: (01) 807 3838

*This Irish Standard was
published under the
authority of the National
Standards Authority of
Ireland
and comes into effect on.
August 24, 2001*

**NO COPYING WITHOUT NSAI
PERMISSION EXCEPT AS
PERMITTED BY COPYRIGHT
LAW**

© NSAI 2001

Price Code E

Údarás um Chaighdeáin Náisiúnta na hÉireann

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN 13311-2

March 2001

ICS 07.080; 07.100.01

English version

**Biotechnology - Performance criteria for vessels - Part 2:
Pressure protection devices**

Biotechnologie - Critères de performance des récipients -
Partie 2: Dispositifs de protection vis-à-vis de la pression

Biotechnik - Leistungskriterien für Behälter - Teil 2:
Druckentlastungseinrichtung

This European Standard was approved by CEN on 4 February 2001.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPAISCHES KOMITEE FÜR NORMUNG

Management Centre: rue de Stassart, 36 B-1050 Brussels

Contents	Page
Foreword	3
Introduction	4
1 Scope	4
2 Normative references	4
3 Terms and definitions	4
4 Hazards	4
5 Performance classes	5
6 Classification and verification of performance.....	5
7 Marking and packaging	5
8 Documentation.....	5
Annex A (informative) Guidance on test methods for determining leaktightness of pressure protection devices.....	6
Annex B (informative) Information on pressure protection devices	7
Bibliography.....	8

Foreword

This European Standard has been prepared by Technical Committee CEN/TC 233 "Biotechnology", the secretariat of which is held by AFNOR.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by September 2001, and conflicting national standards shall be withdrawn at the latest by September 2001.

This European Standard has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association.

This standard is one of a series of European Standards concerned with performance criteria for vessels. These standards are :

EN 13311-1 *Biotechnology - Performance criteria for vessels - Part 1: General performance criteria.*

EN 13311-2 *Biotechnology - Performance criteria for vessels - Part 2: Pressure protection devices.*

EN 13311-3 *Biotechnology - Performance criteria for vessels - Part 3: Glass pressure vessels.*

EN 13311-4 *Biotechnology - Performance criteria for vessels - Part 4: Bioreactors.*

EN 13311-5 *Biotechnology - Performance criteria for vessels - Part 5: Kill tanks.*

EN 13311-6 *Biotechnology - Performance criteria for vessels - Part 6: Chromatography columns.*

Annexes A and B are informative.

This standard includes a bibliography.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and the United Kingdom.

Introduction

Pressure protection devices prevent equipment from being subjected to pressures beyond their nominal pressure. Main types of pressure protection devices are bursting discs and relief valves.

Use of this European Standard will aid the equipment manufacturer in the classification of pressure protection devices with regard to safe performance in biotechnological processes. The classification is easily understandable and readily utilizable for the user and the regulatory authorities.

1 Scope

This European Standard specifies performance criteria for pressure protection devices used in biotechnological processes with respect to the potential hazards to the worker and the environment from microorganisms in use.

This European Standard applies where the intended use of pressure protection devices includes hazardous or potentially hazardous microorganisms used in biotechnological processes or where exposure of the worker or the environment to such microorganisms is restricted for reasons of safety.

2 Normative references

This European Standard incorporates by dated or undated reference, provisions from other publications. These normative references are cited at the appropriate places in the text and the publications are listed hereafter. For dated references, subsequent amendments to or revisions of any of these publications apply to this European Standard only when incorporated in it by amendment or revision. For undated references the latest edition of the publication referred to applies (including amendments).

EN 13311-1:2001, *Biotechnology - Performance criteria for vessels - Part 1 : General performance criteria*.

3 Terms and definitions

For the purposes of this standard, the terms and definitions given in EN 13311-1:2001 apply.

4 Hazards

The following hazards shall be taken into account.

- a) Release of microorganisms caused by pressure release.

NOTE Following rupture of a bursting disc, the vessel will remain open and can continue to release microorganisms.

- b) Release of microorganisms caused by breaking of a bursting disc before the nominal pressure rating is reached due to mechanical and/or thermal tensions.

This is a free preview. Purchase the entire publication at the link below:

[Product Page](#)

-
- Looking for additional Standards? Visit Intertek Inform Infostore
 - Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation
-