



**NSAI**  
Standards

Irish Standard  
I.S. EN 13795-3:2006+A1:2009

# Surgical drapes, gowns and clean air suits, used as medical devices for patients, clinical staff and equipment - Part 3: Performance requirements and performance levels

## I.S. EN 13795-3:2006+A1:2009

*Incorporating amendments/corrigenda issued since publication:*

EN 13795-3:2006/A1:2009

<i>This document replaces:</i> I.S. EN 13795-3:2006	<i>This document is based on:</i> EN 13795-3:2006+A1:2009 EN 13795-3:2006	<i>Published:</i> 15 July, 2009 4 August, 2006	
This document was published under the authority of the NSAI and comes into effect on: 9 September, 2009		ICS number:	
<b>NSAI</b> 1 Swift Square, Northwood, Santry Dublin 9	T +353 1 807 3800 F +353 1 807 3838 E standards@nsai.ie W <b>NSAI.ie</b>	<b>Sales:</b> T +353 1 857 6730 F +353 1 857 6729 W standards.ie	<b>Price Code:</b> F
Údarás um Chaighdeáin Náisiúnta na hÉireann			

English Version

**Surgical drapes, gowns and clean air suits, used as medical devices for patients, clinical staff and equipment - Part 3: Performance requirements and performance levels**

Champs chirurgicaux, casaques et tenues de bloc utilisés en tant que dispositifs médicaux pour les patients, le personnel et les équipements - Partie 3: Exigences et niveaux de performance

Operationsabdecktücher, -mäntel und Rein-Luft-Kleidung zur Verwendung als Medizinprodukte für Patienten, Klinikpersonal und Geräte - Teil 3: Gebrauchsanforderungen und Leistungsstufen

This European Standard was approved by CEN on 27 April and includes Amendment 1 approved by CEN on 13 June 2009.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**Management Centre: Avenue Marnix 17, B-1000 Brussels**

## Contents

	Page
Foreword.....	3
Introduction .....	4
1 Scope .....	5
2 Normative references .....	5
3 Terms and definitions .....	5
4 Performance requirements .....	6
Annex ZA (informative) <b>ZA</b> Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC on Medical Devices <b>ZA</b> .....	10
Bibliography .....	11

## Foreword

This document (EN 13795-3:2006+A1:2009) has been prepared by Technical Committee CEN/TC 205 “Non-active medical devices”, the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by January 2010, and conflicting national standards shall be withdrawn at the latest by March 2010.

This document includes Amendment 1, approved by CEN on 2009-06-13.

This document supersedes EN 13795-3:2006.

The start and finish of text introduced or altered by amendment is indicated in the text by tags A1 A1.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For relationship with EU Directive, see informative Annex ZA, which is an integral part of this document.

EN 13795 consists of the following parts under the general title “*Surgical drapes, gowns and clean air suits, used as medical devices, for patients, clinical staff and equipment*”.

Part 1: *General requirements for manufacturers, processors and products*

Part 2: *Test methods*

Part 3: *Performance requirements and performance levels*

Attention is also drawn to the following:

*EN ISO 22610 Surgical drapes, gowns and clean air suits, used as medical devices, for patients, clinical staff and equipment — Test method to determine the resistance to wet bacterial penetration (ISO 22610:2006)*

*EN ISO 22612 Clothing for protection against infectious agents — Test method for resistance to dry microbial penetration (ISO 22612:2005)*

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.

## **Introduction**

The series of EN 13795 specifies requirements for single-use and reusable coverings (i. e. surgical gowns, surgical drapes and clean air suits) used as medical devices for patients, clinical staff and equipment and intended to prevent the transmission of infective agents between patients and clinical staff during invasive surgical procedures.

General requirements for surgical drapes, gowns and clean air suits, used as medical devices, for patients clinical staff and equipment are specified in EN 13795-1. In this respect EN 13795-1 specifies the relevant characteristics to be evaluated for products covered by this European Standard. EN 13795-2 specifies test methods for evaluating these characteristics.

NOTE For more information on products contained within the scope of this European Standard it is recommended to refer to EN 13795-1.

## 1 Scope

This part of the series of EN 13795 specifies performance requirements for surgical drapes, gowns and clean air suits.

NOTE General performance requirements are specified for various characteristics as per EN 13795-1:2002 Tables 1, 2 and 3 and should be evaluated according to EN 13795-2, EN ISO 22610 and EN ISO 22612.

## 2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 13795-1, *Surgical drapes, gowns and clean air suits, used as medical devices, for patients, clinical staff and equipment — Part 1: General requirements for manufacturers, processors and products*

EN 13795-2, *Surgical drapes, gowns and clean air suits, used as medical devices, for patients, clinical staff and equipment — Part 2: Test methods*

EN ISO 22610, *Surgical drapes, gowns and clean air suits, used as medical devices, for patients, clinical staff, and equipment - Test method to determine the resistance to wet bacterial penetration (ISO 22610:2006)*

## 3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

### 3.1

#### **less critical product area**

product area less likely to be involved in the transfer of infective agents to or from the wound

NOTE 1 For the definition of critical product area please see EN 13795-1.

NOTE 2 If the manufacturer differentiates between critical and less critical areas of the product, EN 13795-1 requires the manufacturer to identify these areas, and, if requested, to supply information on the rationale for this distinction. For more information see EN 13795-1.

### 3.2

#### **performance level**

refers to products designated as “standard” or “high performance” according to Clause 4 of this standard

NOTE With the introduction of two performance levels the EN 13795 series of standards acknowledges the fact that products are challenged to differing extents during surgical procedures, dependent upon the duration, mechanical stress and liquid challenge throughout the surgical procedure.

#### 3.2.1

##### **standard performance**

classification addressing minimum performance requirements for various characteristics of products (see Clause 4) used as medical devices in invasive surgical procedures

#### 3.2.2

##### **high performance**

classification addressing elevated performance requirements for various characteristics of products used as medical devices in invasive surgical procedures

NOTE Examples of surgical procedures where elevated performance level should be considered are those where extensive exposure to liquid, mechanical stresses or longer surgical procedures can be expected

This is a free preview. Purchase the entire publication at the link below:

[Product Page](#)

- 
- [Looking for additional Standards? Visit Intertek Inform Infostore](#)
  - [Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation](#)
-