



NSAI
Standards

Irish Standard
I.S. EN 15483:2008

Ambient air quality - Atmospheric measurements near ground with FTIR spectroscopy

I.S. EN 15483:2008

Incorporating amendments/corrigenda issued since publication:

<i>This document replaces:</i>	<i>This document is based on:</i> EN 15483:2008	<i>Published:</i> 12 November, 2008	
This document was published under the authority of the NSAI and comes into effect on: 16 January, 2009		ICS number: 13.040.20	
NSAI 1 Swift Square, Northwood, Santry Dublin 9	T +353 1 807 3800 F +353 1 807 3838 E standards@nsai.ie W NSAI.ie	Sales: T +353 1 857 6730 F +353 1 857 6729 W standards.ie	Price Code: T
Údarás um Chaighdeáin Náisiúnta na hÉireann			

ICS 13.040.20

English Version

Ambient air quality - Atmospheric measurements near ground with FTIR spectroscopy

Qualité de l'air ambiant - Mesurages de l'air ambiant à
proximité du sol par spectroscopie à transformée de
Fourier (FTIR)

Luftqualität - Messungen in der bodennahen Atmosphäre
mit FTIR-Spektroskopie

This European Standard was approved by CEN on 11 October 2008.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: rue de Stassart, 36 B-1050 Brussels

Contents

Page

Foreword.....	3
Introduction	3
1 Scope	4
2 Normative references	4
3 Terms and definitions	4
4 Symbols and abbreviations	5
5 Principle	6
6 Measurement planning	10
7 Measurement procedure	12
8 Calibration and quality assurance	15
9 Data processing	20
10 Sources of uncertainty	24
11 Servicing	27
Annex A (informative) The classical Fourier transform spectrometer	28
Annex B (informative) Monitoring configurations	33
Annex C (informative) Equipment	35
Annex D (informative) Conditions for measuring emission flux	39
Annex E (informative) Servicing	40
Annex F (normative) Performance characteristics	42
Annex G (informative) Influence of fog on the spectra	46
Annex H (informative) Sample form for a measurement record	49
Annex I (informative) Calibration by using spectral lines from databases and determination of the instrument line shape (example)	54
Annex J (informative) Example applications	56
Bibliography	67

Foreword

This document (EN 15483:2008) has been prepared by Technical Committee CEN/TC 264 "Air quality", the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by May 2009, and conflicting national standards shall be withdrawn at the latest by May 2009.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

Introduction

Fourier transform infrared spectroscopy (FTIR spectroscopy) has been successfully developed from an established laboratory analytical method to a versatile remote sensing method for atmospheric gases.

In this method, the long-path absorption of IR radiation by gaseous air pollutants is measured over an open path between an artificial IR source and an IR spectrometer and used to calculate the integrated concentration over the monitoring path. Since IR radiation is used for remote sensing, the measurements can be made without contact, that is to say without direct sampling, and can be made in various directions. These measurements include monitoring diffuse emissions from large-area sources, for example landfills, road traffic routes, sewage treatment plants, areas used for industrial or agricultural purposes, and in addition the minimization of production losses by tracing leaks in plant sections or piping systems. FTIR spectroscopy is thus suitable for a great number of analytical tasks which cannot adequately be performed using in-situ methods that make point measurements.

Generally, using a suitable measuring arrangement, an overview of the local air pollution may be obtained on site in a short time. This also includes measurements in areas to which access is difficult or impossible, or where the direct presence of staff or set-up of instruments is dangerous. FTIR spectroscopy can be used to determine different compounds at the same time.

This European Standard presents the function and performance of FTIR analytical systems. At the same time, operational notes are given, so that reproducible and valid measurements can be obtained. In addition, questions of measurement planning are discussed and the appendices give a selection of typical applications.

In some circumstances (e. g. CO) the method might be applicable for measurement of air quality as required by European legislation [1].

1 Scope

This European Standard is applicable to open-path absorption measurements of 'concentration \times path length' product using the Fourier transform infrared (FTIR) technique with an artificial radiation source. It is applicable to the continuous measurement of infrared active organic and inorganic compounds in the gaseous state in ambient air using fixed tropospheric open paths up to approximately 1 km in length and provides a spatial average.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN ISO 6142, *Gas analysis - Preparation of calibration gas mixtures - Gravimetric method (ISO 6142:2001)*

EN ISO 6144, *Gas analysis - Preparation of calibration gas mixtures - Static volumetric method (ISO 6144:2003)*

EN ISO 9169, *Air quality - Definition and determination of performance characteristics of an automatic measuring system (ISO 9169:2006)*

ISO 6145 (all parts), *Gas analysis - Preparation of calibration gas mixtures using dynamic volumetric methods*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1

absorbance

the negative logarithm of the transmission, $A(\nu) = -\lg(I(\nu)/I_0(\nu))$, where $I(\nu)$ is the spectral transmitted intensity of the radiation and $I_0(\nu)$ is the incident spectral intensity

NOTE $\lg = \log_{10}$

3.2

apodisation

application of a weighting function to interferogram data to alter the instrument's response function

3.3

background spectrum

with all other conditions being equal, that spectrum taken in the absence of the particular absorbing species of interest

3.4

instrument line shape (ILS)

mathematical function which describes the effect of the instrument's response on a monochromatic line.

3.5

intensity

radiant power per unit solid angle (non-collimated beam) or per unit area (collimated beam)

3.6

interferogram

effects of interference that are detected and recorded by a two-beam interferometer

This is a free preview. Purchase the entire publication at the link below:

[Product Page](#)

-
- [Looking for additional Standards? Visit Intertek Inform Infostore](#)
 - [Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation](#)
-