



NSAI
Standards

Standard Recommendation
S.R. CEN/TS 16115-1:2011

Ambient air quality - Measurement of bioaerosols - Part 1: Determination of moulds using filter sampling systems and culture-based analyses

S.R. CEN/TS 16115-1:2011

Incorporating amendments/corrigenda/National Annexes issued since publication:

The National Standards Authority of Ireland (NSAI) produces the following categories of formal documents:

I.S. xxx: Irish Standard – national specification based on the consensus of an expert panel and subject to public consultation.

S.R. xxx: Standard Recommendation - recommendation based on the consensus of an expert panel and subject to public consultation.

SWIFT xxx: A rapidly developed recommendatory document based on the consensus of the participants of an NSAI workshop.

This document replaces:

This document is based on:
CEN/TS 16115-1:2011

Published:
5 May, 2011

This document was published under the authority of the NSAI and comes into effect on:
5 May, 2011

ICS number:
13.040.20

NSAI
1 Swift Square,
Northwood, Santry
Dublin 9

T +353 1 807 3800
F +353 1 807 3838
E standards@nsai.ie
W NSAI.ie

Sales:
T +353 1 857 6730
F +353 1 857 6729
W standards.ie

Údarás um Chaighdeáin Náisiúnta na hÉireann

ICS 13.040.20

English Version

Ambient air quality - Measurement of bioaerosols - Part 1: Determination of moulds using filter sampling systems and culture-based analyses

Qualité de l'air ambiant - Mesurage de bioaérosols - Partie
1: Dosage des moisissures à l'aide de systèmes de
prélèvement sur filtres et d'analyses de cultures

Luftbeschaffenheit - Messen von Bioaerosolen - Teil 1:
Bestimmung von Schimmelpilzen mittels Probenahme auf
Filtern und kulturellem Nachweis

This Technical Specification (CEN/TS) was approved by CEN on 4 October 2010 for provisional application.

The period of validity of this CEN/TS is limited initially to three years. After two years the members of CEN will be requested to submit their comments, particularly on the question whether the CEN/TS can be converted into a European Standard.

CEN members are required to announce the existence of this CEN/TS in the same way as for an EN and to make the CEN/TS available promptly at national level in an appropriate form. It is permissible to keep conflicting national standards in force (in parallel to the CEN/TS) until the final decision about the possible conversion of the CEN/TS into an EN is reached.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents	Page
Foreword	3
Introduction	4
1 Scope	5
2 Normative references	5
3 Terms and definitions	5
4 Symbols and abbreviations	8
5 Basic principle of the method	8
6 Sampling	9
7 Culture-based analyses	15
8 Performance characteristics and minimum requirements	22
9 Quality assurance	23
10 Trouble shooting during sampling and analyses	23
Annex A (informative) Example for a validated sampling device	26
Annex B (informative) Recovery of spores on gelatine filters in combination with polycarbonate filters	32
Annex C (informative) Examples for sampling and analyses reports	35
Annex D (informative) Membrane filtration technique	41
Annex E (informative) Calculation by weighted mean	42
Bibliography	44

Foreword

This document (CEN/TS 16115-1:2011) has been prepared by Technical Committee CEN/TC 264 “Air quality”, the secretariat of which is held by DIN.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to announce this Technical Specification: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

Introduction

Airborne particles of biological origin are called bioaerosols. Depending on the emission source bioaerosols vary in composition; one component of ambient bioaerosols with possible ecological and health relevance can be moulds. Natural and anthropogenic sources for mould spores are widely distributed in the environment. Anthropogenic sources can for example be agriculture and construction activities or waste treatment.

Mould is a common name for filamentous fungi from different taxonomic groups (Zygomycetes, Ascomycetes, Deuteromycetes). They form a mycelium (hyphae) and spores – namely conidiospores (conidia), sporangiospores or ascospores – by which they become visible macroscopically. Most spores are in the size range of 2 µm to 10 µm, some up to 30 µm and only few up to 100 µm. Spores of some mould genera are small and become airborne very easily (e.g., *Aspergillus*, *Penicillium*) while others are bigger and/or embedded in a slime matrix (e.g., *Stachybotrys*, *Fusarium*) and less mobile.

The procedure described in this document is based on VDI 4252 Part 2 [1], VDI 4253 Part 2 [2] and is related to the ISO standards on indoor air ISO 16000-16 [3] and ISO 16000-17 [4].

1 Scope

This Technical Specification describes the measurement of moulds in ambient air in order to identify, quantify and characterize bioaerosol pollution in ambient air resulting from emissions from different sources.

The method described specifies the sampling of moulds as part of the suspended particulate matter (SPM, here particles with aerodynamic diameter up to ca. 30 µm) using a filter sampling system with gelatine/poly-carbonate filter combination followed by the culture-based analyses on DG18 agar. The sampling duration can be varied between 10 min to 24 h. The health effect of bioaerosols is not limited to any particle fraction, therefore, this document describes the sampling of moulds as part of the suspended particulate matter as a convention method.

NOTE The sampling method described in this document in principle is likely to be appropriate for the sampling of actinomycetes and other spore-forming bacteria (resistant to desiccation). For these species a special analytical procedure using different culture media should be applied, but this is not within the scope of this document.

The standard method set out in this Technical Specification is accepted by convention as reference method. The measured quantity, here the number of colony forming units per cubic meter (CFU/m³), is determined by the inlet design of the sampling head, the associated operational parameters and the analytical procedure.

Standardized methods for sampling, detection and enumeration of moulds including standards for sampling strategies are important for comparative assessment of moulds in ambient air. Before doing any measurements a plan for the measurement strategy is necessary (see CEN/TS 16115-2 [5]).

WARNING — The use of this Technical Specification may involve hazardous materials, operations and equipment. This Technical Specification does not purport to address all the safety problems associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN ISO 8199:2007, *Water quality — General guidance on the enumeration of micro-organisms by culture (ISO 8199:2005)*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1

aerodynamic diameter

diameter of a sphere of density 1 g/cm³ with the same terminal velocity due to gravitational force in calm air as the particle, under the prevailing conditions of temperature, pressure and relative humidity

[ISO 7708:1995, 2.2 [6]]

3.2

ambient air

outdoor air in the lower troposphere excluding workplace air

[EN 14907:2005, 3.1.1 [7]]

3.3

analytical blank value

value determined by a blank sample covering the analytical procedure to ensure that no significant contamination occurs during the complete analytical procedure including autoclaving, agar preparation, suspension and extraction of the filters, dilution, incubation, counting, etc.

This is a free preview. Purchase the entire publication at the link below:

[Product Page](#)

-
- [Looking for additional Standards? Visit Intertek Inform Infostore](#)
 - [Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation](#)
-