



NSAI
Standards

Irish Standard
I.S. EN ISO 18113-4:2011

In vitro diagnostic medical devices -
Information supplied by the manufacturer
(labelling) - Part 4: In vitro diagnostic
reagents for self-testing (ISO 18113
-4:2009)

I.S. EN ISO 18113-4:2011

Incorporating amendments/corrigenda/National Annexes issued since publication:

The National Standards Authority of Ireland (NSAI) produces the following categories of formal documents:

I.S. xxx: Irish Standard – national specification based on the consensus of an expert panel and subject to public consultation.

S.R. xxx: Standard Recommendation - recommendation based on the consensus of an expert panel and subject to public consultation.

SWIFT xxx: A rapidly developed recommendatory document based on the consensus of the participants of an NSAI workshop.

This document replaces:
EN ISO 18113-4:2009

This document is based on:
EN ISO 18113-4:2011
EN ISO 18113-4:2009

Published:
31 October, 2011
15 December, 2009

This document was published
under the authority of the NSAI
and comes into effect on:
31 October, 2011

ICS number:
11.100.10

NSAI
1 Swift Square,
Northwood, Santry
Dublin 9

T +353 1 807 3800
F +353 1 807 3838
E standards@nsai.ie
W NSAI.ie

Sales:
T +353 1 857 6730
F +353 1 857 6729
W standards.ie

Údarás um Chaighdeáin Náisiúnta na hÉireann

English Version

In vitro diagnostic medical devices - Information supplied by the manufacturer (labelling) - Part 4: In vitro diagnostic reagents for self-testing (ISO 18113-4:2009)

Dispositifs médicaux de diagnostic in vitro - Informations fournies par le fabricant (étiquetage) - Partie 4: Réactifs de diagnostic in vitro pour auto-tests (ISO 18113-4:2009)

In-vitro-Diagnostika - Bereitstellung von Informationen durch den Hersteller - Teil 4: Reagenzien für in-vitro-diagnostische Untersuchungen zur Eigenanwendung (ISO 18113-4:2009)

This European Standard was approved by CEN on 20 September 2011.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents

Page

Foreword.....3

**Annex ZA (informative) Relationship between this European Standard and the Essential
Requirements of the EU Directive 98/79/EC on “in vitro Diagnostic Medical Devices”4**

Foreword

This document (EN ISO 18113-4:2011) has been prepared by Technical Committee ISO/TC 212 "Clinical laboratory testing and in vitro diagnostic test systems" in collaboration with Technical Committee CEN/TC 140 "In vitro diagnostic medical devices" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by April 2012, and conflicting national standards shall be withdrawn at the latest by December 2012.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 18113-4:2009.

This new edition contains a revised Annex ZA.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive.

For relationship with EU Directive, see informative Annex ZA, which is an integral part of this document.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.

Endorsement notice

The text of ISO 18113-4:2009 has been approved by CEN as EN ISO 18113-4:2011 without any modification.

Annex ZA (informative)

Relationship between this European Standard and the Essential Requirements of the EU Directive 98/79/EC on “in vitro Diagnostic Medical Devices”

This European Standard has been prepared under a mandate given to CEN by the European Commission to provide a means of conforming to the Essential Requirements of the New Approach Directive 98/79/EC on “*in vitro* Diagnostic Medical Devices”.

Once this European Standard is cited in the Official Journal of the European Union under that Directive and has been implemented as national standard in at least one Member State, compliance with the clauses of this standard given in Table ZA.1 confers, within the limits of the scope of this European Standard, a presumption of conformity with the corresponding Essential Requirements of that Directive and associated EFTA regulations.

Table ZA.1 — Correspondence between this European Standard and European Directive 98/79/EC

Clauses of this European Standard	Essential Requirements (ERs) of Directive 98/79/EC	Qualifying comments/Notes
4.3, 6.1, 7.3, 7.11	B.7	These clauses only cover the second sentence of ER B.7, namely the labelling requirements.
4.2, 4.3, 5, 6, 7	B.8.1	Presumption of conformity with ER B.8.1 also requires compliance with clauses 4.1, 4.2.1 and 4.6 of EN ISO 18113-1.
5.8, 6.8, 7.9	B.8.3	NOTE 2
5.1, 6.2	B.8.4(a)	NOTE 1
5.2, 5.3, 6.3.1, 6.4	B.8.4(b)	
5.2.2, 6.3.2	B.8.4(d)	Full compliance to ER B.8.4(d) requires the use of EN 980, clause 5.4, symbol (LOT).
5.7, 6.7	B.8.4(e)	
5.5, 6.5	B.8.4(g)	
5.6, 6.6	B.8.4(h)	
5.8, 6.8	B.8.4(j)	NOTE 2
5.4	B.8.4(k)	
5.4, 7.3	B.8.5	
5.2.2, 6.3.2	B.8.6	
7.1, 7.2, 7.3, 7.8, 7.9	B.8.7(a)	Presumption of conformity with ER B. 8.7(a) requires also compliance with clause 4.5 of EN ISO 18113-1. NOTE 1, NOTE 3
7.5	B.8.7(b)	
7.8	B.8.7(c)	
7.6	B.8.7(e)	
7.10	B.8.7(f)	
7.11	B.8.7(g)	

Clauses of this European Standard	Essential Requirements (ERs) of Directive 98/79/EC	Qualifying comments/Notes
7.4, 7.7	B.8.7(h)	
7.12	B.8.7(k)	NOTE 3
7.6	B.8.7(m)	
7.9	B.8.7(s)	
7.13, 7.14, 7.15, 7.16, 7.17	B.8.7(t)	

NOTE 1 In the European Union, the name and address of the manufacturer's "EC Authorized representative" is required on the outer container label or in the instructions for use, if the legal manufacturer is not located within the European Union.

NOTE 2 Essential Requirement B.8.3 of Directive 98/79/EC should be consulted for a comprehensive list of the information required.

NOTE 3 Essential Requirement B.8.7 of Directive 98/79/EC should be consulted for a comprehensive list of the information required.

This page is intentionally left BLANK.

I.S. EN ISO 18113-4:2011
**INTERNATIONAL
STANDARD**

**ISO
18113-4**

First edition
2009-12-15

***In vitro* diagnostic medical devices —
Information supplied by the manufacturer
(labelling) —**

Part 4:
***In vitro* diagnostic reagents for self-
testing**

*Dispositifs médicaux de diagnostic in vitro — Informations fournies par
le fabricant (étiquetage) —*

Partie 4: Réactifs de diagnostic in vitro pour auto-tests



Reference number
ISO 18113-4:2009(E)

© ISO 2009

PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.



COPYRIGHT PROTECTED DOCUMENT

© ISO 2009

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction.....	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 General	2
4.1 Essential requirements	2
4.2 Identification of kit components	2
4.3 Presentation of the instructions for use	2
5 Content of the outer container label.....	2
5.1 Manufacturer.....	2
5.2 Identification of the IVD reagent	2
5.3 Contents	3
5.4 Intended use	3
5.5 <i>In vitro</i> diagnostic use	3
5.6 Storage and handling conditions	3
5.7 Expiry date	3
5.8 Warnings and precautions	4
6 Content of the immediate container label.....	4
6.1 General provisions	4
6.2 Manufacturer.....	4
6.3 Identification of the IVD reagent	4
6.4 Contents	4
6.5 <i>In vitro</i> diagnostic use	4
6.6 Storage and handling conditions	5
6.7 Expiry date	5
6.8 Warnings and precautions	5
7 Content of the instructions for use	5
7.1 Manufacturer.....	5
7.2 Identification of the IVD reagent	5
7.3 Intended use	5
7.4 Principles of the examination method	6
7.5 Components.....	6
7.6 Additional required equipment	6
7.7 Reagent preparation.....	6
7.8 Storage and shelf life after first opening	6
7.9 Warnings and precautions	6
7.10 Primary sample collection, handling and storage	7
7.11 Examination procedure.....	7
7.12 Control procedure	7
7.13 Reading of examination results	7
7.14 Interpretation of results	7
7.15 Performance characteristics	7
7.16 Biological reference intervals	8
7.17 Limitations of examination procedure	8
7.18 Literature references.....	8
Bibliography.....	9

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 18113-4 was prepared by Technical Committee ISO/TC 212, *Clinical laboratory testing and in vitro diagnostic test systems*

ISO 18113 consists of the following parts, under the general title *In vitro diagnostic medical devices — Information supplied by the manufacturer (labelling)*:

- *Part 1: Terms, definitions and general requirements*
- *Part 2: In vitro diagnostic reagents for professional use*
- *Part 3: In vitro diagnostic instruments for professional use*
- *Part 4: In vitro diagnostic reagents for self-testing*
- *Part 5: In vitro diagnostic instruments for self-testing*

Introduction

Manufacturers of *in vitro* diagnostic (IVD) reagents for self-testing supply users with information to enable the safe use and expected performance of their devices. The type and level of detail varies according to the intended uses and country-specific regulations.

The Global Harmonization Task Force (GHTF) encourages convergence of the evolution of regulatory systems for medical devices at the global level. Eliminating differences among regulatory jurisdictions could allow patients earlier access to new technologies and treatments. See Reference [9]. This part of ISO 18113 provides a basis for harmonization of labelling requirements for IVD reagents for self-testing.

This part of ISO 18113 is concerned solely with information supplied with IVD reagents, calibrators and control materials intended for self-testing. It is intended to be used in conjunction with ISO 18113-1, which contains the general requirements for information supplied by the manufacturer and definitions of general labelling concepts.

This part of ISO 18113 is based on EN 376:2002^[5]. The text has been modified to conform to Part 2 of the ISO/IEC Directives^[4], but the requirements, including those in ISO 18113-1, are substantially equivalent to the original European harmonized standard. This part of ISO 18113 is intended to support the essential labelling requirements of all the GHTF partners, as well as other countries that have enacted or plan to enact labelling regulations for IVD medical devices.

For IVD reagents, calibrators and/or control materials that are intended to be used as a system with an instrument provided by the same manufacturer, this part of ISO 18113 is also intended to be used together with ISO 18113-1 and ISO 18113-5^[3].

I.S. EN ISO 18113-4:2011

***In vitro* diagnostic medical devices — Information supplied by the manufacturer (labelling) —**

Part 4: *In vitro* diagnostic reagents for self-testing

1 Scope

This part of ISO 18113 specifies requirements for information supplied by the manufacturer of IVD reagents for self-testing.

This part of ISO 18113 also applies to information supplied by the manufacturer with calibrators and control materials intended for use with IVD medical devices for self-testing.

This part of ISO 18113 can also be applied to accessories.

This part of ISO 18113 applies to the labels for outer and immediate containers and to the instructions for use.

This part of ISO 18113 does not apply to:

- a) IVD instruments or equipment,
- b) IVD reagents for professional use.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 14971, *Medical devices — Application of risk management to medical devices*

ISO 15223-1, *Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied — Part 1: General requirements*

ISO 18113-1, *In vitro diagnostic medical devices — Information supplied by the manufacturer (labelling) — Part 1: Terms, definitions and general requirements*

EN 980, *Symbols for use in the labelling of medical devices*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 18113-1 apply.

4 General

4.1 Essential requirements

The requirements of ISO 18113-1 apply. For the use of symbols, the requirements of ISO 15223-1 and EN 980 apply.

ISO standards for specific IVD medical devices may also contain requirements for information supplied by the manufacturer.

EXAMPLES ISO 15197^[1]; ISO 17593^[2].

4.2 Identification of kit components

In the case of a kit, each component shall be identified by name, letter, number, symbol, colour or graphics in the same manner on all labels and in the instructions for use.

4.3 Presentation of the instructions for use

4.3.1 The instructions for use shall be written to be easily understood and applied by a lay person, and where appropriate, supplemented with drawings and diagrams.

Some devices may require separate information for the healthcare professional.

4.3.2 The information supplied shall be sufficient to enable a lay person to use the IVD reagent safely and properly, and to understand the IVD examination results.

NOTE Recommendations for developing user instruction manuals for IVD medical devices used in home health care are found in Reference [10].

5 Content of the outer container label

5.1 Manufacturer

The name and address of the manufacturer shall be given.

NOTE In the European Union, the name and address of the manufacturer's "EC Authorized Representative" is required on the outer container label or in the instructions for use, if the legal manufacturer is not located within the EU. See Reference [8].

5.2 Identification of the IVD reagent

5.2.1 IVD reagent name

The name of the IVD reagent shall be given.

When the name does not uniquely identify the IVD reagent, an additional means of identification shall also be given.

EXAMPLES Catalogue number, commodity number.

5.2.2 Batch code

A batch code shall be given.

If a kit contains different components bearing different batch codes, the batch code indicated on the outer container shall enable the individual batch code of each component to be traced from the manufacturer's production record.

5.3 Contents

The mass, volume and/or the number of examinations shall be indicated.

5.4 Intended use

If the intended use is not indicated by the name of the IVD reagent, then an abbreviated intended use statement shall be given or included in the instructions for use in terminology suitable for a lay person.

EXAMPLE Pregnancy test.

The fact that the IVD reagent is intended for self-testing shall be clearly stated.

5.5 *In vitro* diagnostic use

The *in vitro* diagnostic use of the reagent shall be indicated in terminology suitable for a lay person.

EXAMPLE Only for use outside the body.

5.6 Storage and handling conditions

The storage conditions necessary to maintain the stability of the reagents, calibrators and control materials in the unopened state shall be indicated.

EXAMPLE 1 2 °C to 8 °C or 2...8 °C or graphical symbol.
–18 °C or below or graphical symbol.

Other conditions that affect stability shall be indicated.

EXAMPLE 2 Light, humidity.

Any other conditions that affect the handling or storage of the reagents, calibrators and control materials shall be specified.

EXAMPLE 3 Fragile.

5.7 Expiry date

An expiry date based upon the stated storage instructions shall be indicated.

Expiry dates shall be expressed in a format generally familiar to the lay person.

EXAMPLES 2007-05-01, 2007-May-01, May 01, 2007.

If only the year and month are given, the expiry date shall be the last day of the month indicated.

The label on the outer container shall indicate the expiry date of the component having the earliest expiry date or an earlier date.

Local, national or regional regulations may apply.

This is a free preview. Purchase the entire publication at the link below:

[Product Page](#)

-
- Looking for additional Standards? Visit Intertek Inform Infostore
 - Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation
-