



NSAI
Standards

Irish Standard
I.S. EN ISO/IEC 17025:2005

General requirements for the competence of testing and calibration laboratories (ISO/IEC 17025:2005)

I.S. EN ISO/IEC 17025:2005

Incorporating amendments/corrigenda/National Annexes issued since publication:
EN ISO/IEC 17025:2005/AC:2006

The National Standards Authority of Ireland (NSAI) produces the following categories of formal documents:

I.S. xxx: Irish Standard – national specification based on the consensus of an expert panel and subject to public consultation.

S.R. xxx: Standard Recommendation - recommendation based on the consensus of an expert panel and subject to public consultation.

SWiFT xxx: A rapidly developed recommendatory document based on the consensus of the participants of an NSAI workshop.

<i>This document replaces:</i> EN ISO/IEC 17025:2000	<i>This document is based on:</i> EN ISO/IEC 17025:2005 EN ISO/IEC 17025:2000	<i>Published:</i> 15 May, 2005 11 May, 2000
This document was published under the authority of the NSAI and comes into effect on: 1 April, 2007		ICS number: 03.120.20
NSAI 1 Swift Square, Northwood, Santry Dublin 9	T +353 1 807 3800 F +353 1 807 3838 E standards@nsai.ie W NSAI.ie	Sales: T +353 1 857 6730 F +353 1 857 6729 W standards.ie
Údarás um Chaighdeáin Náisiúnta na hÉireann		

English version
Version Française
Deutsche Fassung

General requirements for the competence of testing and calibration
laboratories (ISO/IEC 17025:2005/Cor.1:2006)

Exigences générales concernant la
compétence des laboratoires d'étalonnages
et d'essais (ISO/IEC
17025:2005/Cor.1:2006)

Allgemeine Anforderungen an die
Kompetenz von Prüf- und
Kalibrierlaboratorien (ISO/IEC
17025:2005/Cor.1:2006)

This corrigendum becomes effective on 18 October 2006 for incorporation in the three official language versions of the EN.

Ce corrigendum prendra effet le 18 octobre 2006 pour incorporation dans les trois versions linguistiques officielles de la EN.

Die Berichtigung tritt am 18. Oktober 2006 zur Einarbeitung in die drei offiziellen Sprachfassungen der EN in Kraft.



CEN Management Centre:
rue de Stassart, 36 B-1050 Brussels

CENELEC Central Secretariat:
rue de Stassart, 35 B-1050 Brussels

**I.S. EN ISO/IEC 17025:2005
EN ISO/IEC 17025:2005/AC:2006 (E/F/D)**

English version

The text of ISO/IEC 17025:2005/Cor.1:2006 has been approved by CEN and CENELEC as a European Corrigendum without any modifications.

Version française

Le texte de l'ISO/IEC 17025:2005/Cor.1:2006 a été approuvé par le CEN et le CENELEC comme Corrigendum européen sans aucune modification.

Deutsche Fassung

Tabelle 1 ist wie folgt zu ersetzen:

Tabelle 1 — Formale Querverweisungen zu ISO 9001:2000

ISO 9001:2000	ISO/IEC 17025^a
Abschnitt 1	Abschnitt 1
Abschnitt 2	Abschnitt 2
Abschnitt 3	Abschnitt 3
4.1	4.1, 4.1.1, 4.1.2, 4.1.3, 4.1.4, 4.1.5, 4.2, 4.2.1, 4.2.2, 4.2.3, 4.2.4
4.2.1	4.2.2, 4.2.3, 4.3.1
4.2.2	4.2.2, 4.2.5
4.2.3	4.3
4.2.4	4.3.1, 4.13
5.1	4.2.2, 4.2.3
5.1 a)	4.1.2, 4.2.4
5.1 b)	4.2.2
5.1 c)	4.2.2
5.1 d)	4.15
5.1 e)	4.1.5
5.2	4.4.1
5.3	4.2.2
5.3 a)	4.2.2
5.3 b)	4.2.3
5.3 c)	4.2.2
5.3 d)	4.2.2
5.3 e)	4.2.2
5.4.1	4.2.2 c)
5.4.2	4.2.1
5.4.2 a)	4.2.1
5.4.2 b)	4.2.7
5.5.1	4.1.5 a), f), h)
5.5.2	4.1.5 i)
5.5.2 a)	4.1.5 i)

**I.S. EN ISO/IEC 17025:2005
EN ISO/IEC 17025:2005/AC:2006 (E/F/D)**

ISO 9001:2000	ISO/IEC 17025^a
5.5.2 b)	4.1.5 i)
5.5.2 c)	4.2.4
5.5.3	4.1.6
5.6.1	4.15
5.6.2	4.15
5.6.3	4.15
6.1 a)	4.10
6.1 b)	4.4.1, 4.7, 5.4.2, 5.4.3, 5.4.4, 5.10.1
6.2.1	5.2.1
6.2.2 a)	5.2.2, 5.5.3
6.2.2 b)	5.2.1, 5.2.2
6.2.2 c)	5.2.2
6.2.2 d)	4.1.5 k)
6.2.2 e)	5.2.5
6.3 a)	4.1.3, 5.3
6.3 b)	5.4.7.2, 5.5, 5.6,
6.3 c)	4.6, 5.5.6, 5.6.3.4, 5.8, 5.10
6.4	5.3.1, 5.3.2, 5.3.3, 5.3.4, 5.3.5
7.1	
7.1 a)	4.2.2
7.1 b)	4.1.5 a), 4.2.1, 4.2.3
7.1 c)	5.4, 5.9
7.1 d)	4.1, 5.4, 5.9
7.2.1	4.4.1, 4.4.2, 4.4.3, 4.4.4, 4.4.5, 5.4
7.2.2	4.4.1, 4.4.2, 4.4.3, 4.4.4, 4.4.5, 5.4
7.2.3	4.4.2, 4.4.4, 4.5, 4.7, 4.8
7.3	5.4, 5.9
7.4.1	4.6.1, 4.6.2, 4.6.4
7.4.2	4.6.3
7.4.3	4.6.2
7.5.1	5.1, 5.2, 5.4, 5.5, 5.6, 5.7, 5.8, 5.9
7.5.1 a)	4.3.1
7.5.1 b)	4.2.1
7.5.1 c)	5.3, 5.5
7.5.1 d)	5.5
7.5.1 e)	5.3
7.5.1 f)	4.7, 5.8, 5.9, 5.10
7.5.2	5.2.5, 5.4.2, 5.4.5
7.5.2 a)	5.4.1
7.5.2 b)	5.2.5, 5.5.2
7.5.2 c)	5.4.1
7.5.2 d)	4.13
7.5.2 e)	5.9
7.5.3	5.8.2
7.5.4	4.1.5 (c), 5.8
7.5.5	4.6.1, 4.12, 5.8, 5.10

**I.S. EN ISO/IEC 17025:2005
EN ISO/IEC 17025:2005/AC:2006 (E/F/D)**

ISO 9001:2000	ISO/IEC 17025^a
7.6	5.4, 5.5
8.1	4.10, 5.4, 5.9
8.1 a)	5.4, 5.9
8.1 b)	4.14
8.1 c)	4.10
8.2.1	4.7.2
8.2.2	4.11.5, 4.14
8.2.3	4.11.5, 4.14, 5.9
8.2.4	4.5, 4.6, 4.9, 5.5.2, 5.5.9, 5.8
8.3	4.9
8.4	4.10, 5.9
8.4 a)	4.7.2
8.4 b)	4.4, 5.4
8.4 c)	5.9
8.4 d)	4.6.4
8.5.1	4.10, 4.12
8.5.2	4.11, 4.12
8.5.3	4.9, 4.11, 4.12
^a ISO/IEC 17025 enthält einige Anforderungen an die technische Kompetenz, die nicht durch ISO 9001:2000 abgedeckt sind.	



I.S. EN ISO/IEC 17025:2005
INTERNATIONAL STANDARD ISO/IEC 17025:2005
TECHNICAL CORRIGENDUM 1

Published 2006-08-15

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION
INTERNATIONAL ELECTROTECHNICAL COMMISSION • МЕЖДУНАРОДНАЯ ЭЛЕКТРОТЕХНИЧЕСКАЯ КОМИССИЯ • COMMISSION ÉLECTROTECHNIQUE INTERNATIONALE

General requirements for the competence of testing and calibration laboratories

TECHNICAL CORRIGENDUM 1

Exigences générales concernant la compétence des laboratoires d'étalonnages et d'essais

RECTIFICATIF TECHNIQUE 1

Technical Corrigendum 1 to ISO/IEC 17025:2005 was prepared by the *ISO Committee on conformity assessment (CASCO)*.

Pages 24 and 25, Table A.1

Replace the existing Table A.1 with the following, as some of the cross-references were incorrect.

Table A.1 — Nominal cross-references to ISO 9001:2000

ISO 9001:2000	ISO/IEC 17025 ^a
Clause 1	Clause 1
Clause 2	Clause 2
Clause 3	Clause 3
4.1	4.1, 4.1.1, 4.1.2, 4.1.3, 4.1.4, 4.1.5, 4.2, 4.2.1, 4.2.2, 4.2.3, 4.2.4
4.2.1	4.2.2, 4.2.3, 4.3.1
4.2.2	4.2.2, 4.2.5
4.2.3	4.3
4.2.4	4.3.1, 4.13
5.1	4.2.2, 4.2.3
5.1 a)	4.1.2, 4.2.4
5.1 b)	4.2.2
5.1 c)	4.2.2
5.1 d)	4.15
5.1 e)	4.1.5
5.2	4.4.1
5.3	4.2.2
5.3 a)	4.2.2
5.3 b)	4.2.3
5.3 c)	4.2.2
5.3 d)	4.2.2
5.3 e)	4.2.2
5.4.1	4.2.2
5.4.2	4.2.1
5.4.2 a)	4.2.1
5.4.2 b)	4.2.7
5.5.1	4.1.5 a), f), h)
5.5.2	4.1.5 i)
5.5.2 a)	4.1.5 i)
5.5.2 b)	4.1.5 i)
5.5.2 c)	4.2.4
5.5.3	4.1.6
5.6.1	4.15
5.6.2	4.15
5.6.3	4.15
6.1 a)	4.10
6.1 b)	4.4.1, 4.7, 5.4.2, 5.4.3, 5.4.4, 5.10.1
6.2.1	5.2.1
6.2.2 a)	5.2.2, 5.5.3
6.2.2 b)	5.2.1, 5.2.2
6.2.2 c)	5.2.2
6.2.2 d)	4.1.5 k)
6.2.2 e)	5.2.5
6.3 a)	4.1.3, 5.3
6.3 b)	5.4.7.2, 5.5, 5.6
6.3 c)	4.6, 5.5.6, 5.6.3.4, 5.8, 5.10
6.4	5.3.1, 5.3.2, 5.3.3, 5.3.4, 5.3.5
7.1	
7.1 a)	4.2.2
7.1 b)	4.1.5 a), 4.2.1, 4.2.3
7.1 c)	5.4, 5.9
7.1 d)	4.1, 5.4, 5.9

7.2.1	4.4.1, 4.4.2, 4.4.3, 4.4.4, 4.4.5, 5.4
7.2.2	4.4.1, 4.4.2, 4.4.3, 4.4.4, 4.4.5, 5.4
7.2.3	4.4.2, 4.4.4, 4.5, 4.7, 4.8
7.3	5.4, 5.9
7.4.1	4.6.1, 4.6.2, 4.6.4
7.4.2	4.6.3
7.4.3	4.6.2
7.5.1	5.1, 5.2, 5.4, 5.5, 5.6, 5.7, 5.8, 5.9
7.5.1 a)	4.3.1
7.5.1 b)	4.2.1
7.5.1 c)	5.3, 5.5
7.5.1 d)	5.5
7.5.1 e)	5.3
7.5.1 f)	4.7, 5.8, 5.9, 5.10
7.5.2	5.2.5, 5.4.2, 5.4.5
7.5.2 a)	5.4.1
7.5.2 b)	5.2.5, 5.5.2
7.5.2 c)	5.4.1
7.5.2 d)	4.13
7.5.2 e)	5.9
7.5.3	5.8.2
7.5.4	4.1.5 c), 5.8
7.5.5	4.6.1, 4.12, 5.8, 5.10
7.6	5.5, 5.6
8.1	4.10, 5.4, 5.9
8.1 a)	5.4, 5.9
8.1 b)	4.14
8.1 c)	4.10
8.2.1	4.7.2
8.2.2	4.11.5, 4.14
8.2.3	4.11.5, 4.14, 5.9
8.2.4	4.5, 4.6, 4.9, 5.5.2, 5.5.9, 5.8
8.3	4.9
8.4	5.9
8.4 a)	4.7.2
8.4 b)	4.4, 5.4
8.4 c)	5.9
8.4 d)	4.6.4
8.5.1	4.10
8.5.2	4.11
8.5.3	4.12
^a ISO/IEC 17025 covers several technical competence requirements that are not covered by ISO 9001:2000.	

This page is intentionally left BLANK.

I.S. EN ISO/IEC 17025:2005

EUROPEAN STANDARD

EN ISO/IEC 17025

NORME EUROPÉENNE

EUROPÄISCHE NORM

May 2005

ICS 03.120.20

Supersedes EN ISO/IEC 17025:2000

English version

General requirements for the competence of testing and calibration laboratories (ISO/IEC 17025:2005)

Exigences générales concernant la compétence des laboratoires d'étalonnages et d'essais (ISO/IEC 17025:2005)

Allgemeine Anforderungen an die Kompetenz von Prüf- und Kalibrierlaboratorien (ISO/IEC 17025:2005)

This European Standard was approved by CEN and CENELEC on 15 March 2005.

CEN and CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN Management Centre or to any CEN or CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN or CENELEC member into its own language and notified to the CEN Management Centre has the same status as the official versions.

CEN and CENELEC members are the national standards bodies and national electrotechnical committees, respectively, of Austria, Belgium, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.



CEN Management Centre:
rue de Stassart, 36 B-1050 Brussels

CENELEC Central Secretariat:
rue de Stassart, 35 B-1050 Brussels

I.S. EN ISO/IEC 17025:2005

EN ISO/IEC 17025:2005 (E)

Foreword

This document (EN ISO/IEC 17025:2005) has been prepared by Technical Committee ISO/CASCO "Committee on conformity assessment" in collaboration with Technical Committee CEN/CLC/TC 1 "Criteria for conformity assessment bodies", the secretariat of which is held by SN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by November 2005, and conflicting national standards shall be withdrawn at the latest by November 2005.

This document supersedes EN ISO/IEC 17025:2000.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.

Endorsement notice

The text of ISO/IEC 17025:2005 has been approved by CEN and CENELEC as EN ISO/IEC 17025:2005 without any modifications.

I.S. EN ISO/IEC 17025:2005
**INTERNATIONAL
STANDARD**

**ISO/IEC
17025**

Second edition
2005-05-15

**General requirements for the competence
of testing and calibration laboratories**

*Exigences générales concernant la compétence des laboratoires
d'étalonnages et d'essais*

Reference number
ISO/IEC 17025:2005(E)



PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

© ISO 2005

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword.....	v
Introduction	vi
1 Scope.....	1
2 Normative references	2
3 Terms and definitions.....	2
4 Management requirements	2
4.1 Organization	2
4.2 Management system.....	3
4.3 Document control	4
4.3.1 General	4
4.3.2 Document approval and issue.....	4
4.3.3 Document changes	5
4.4 Review of requests, tenders and contracts	5
4.5 Subcontracting of tests and calibrations	6
4.6 Purchasing services and supplies	6
4.7 Service to the customer	6
4.8 Complaints.....	7
4.9 Control of nonconforming testing and/or calibration work.....	7
4.10 Improvement.....	7
4.11 Corrective action.....	8
4.11.1 General	8
4.11.2 Cause analysis	8
4.11.3 Selection and implementation of corrective actions.....	8
4.11.4 Monitoring of corrective actions	8
4.11.5 Additional audits	8
4.12 Preventive action	8
4.13 Control of records.....	9
4.13.1 General	9
4.13.2 Technical records	9
4.14 Internal audits.....	9
4.15 Management reviews	10
5 Technical requirements.....	10
5.1 General	10
5.2 Personnel.....	11
5.3 Accommodation and environmental conditions	12
5.4 Test and calibration methods and method validation.....	12
5.4.1 General	12
5.4.2 Selection of methods.....	13
5.4.3 Laboratory-developed methods	13
5.4.4 Non-standard methods.....	13
5.4.5 Validation of methods.....	14
5.4.6 Estimation of uncertainty of measurement.....	14
5.4.7 Control of data.....	15
5.5 Equipment.....	15
5.6 Measurement traceability	17
5.6.1 General	17
5.6.2 Specific requirements.....	17
5.6.3 Reference standards and reference materials.....	18
5.7 Sampling	19

5.8	Handling of test and calibration items	19
5.9	Assuring the quality of test and calibration results	20
5.10	Reporting the results	20
5.10.1	General	20
5.10.2	Test reports and calibration certificates	20
5.10.3	Test reports	21
5.10.4	Calibration certificates	22
5.10.5	Opinions and interpretations	22
5.10.6	Testing and calibration results obtained from subcontractors	23
5.10.7	Electronic transmission of results	23
5.10.8	Format of reports and certificates	23
5.10.9	Amendments to test reports and calibration certificates	23
Annex A	(informative) Nominal cross-references to ISO 9001:2000	24
Annex B	(informative) Guidelines for establishing applications for specific fields	26
Bibliography	27

Foreword

ISO (the International Organization for Standardization) and IEC (the International Electrotechnical Commission) form the specialized system for worldwide standardization. National bodies that are members of ISO or IEC participate in the development of International Standards through technical committees established by the respective organization to deal with particular fields of technical activity. ISO and IEC technical committees collaborate in fields of mutual interest. Other international organizations, governmental and non-governmental, in liaison with ISO and IEC, also take part in the work. In the field of conformity assessment, the ISO Committee on conformity assessment (CASCO) is responsible for the development of International Standards and Guides.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

Draft International Standards are circulated to the national bodies for voting. Publication as an International Standard requires approval by at least 75 % of the national bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO/IEC 17025 was prepared by the *ISO Committee on conformity assessment (CASCO)*.

It was circulated for voting to the national bodies of both ISO and IEC, and was approved by both organizations.

This second edition cancels and replaces the first edition (ISO/IEC 17025:1999), which has been technically revised.

Introduction

The first edition (1999) of this International Standard was produced as the result of extensive experience in the implementation of ISO/IEC Guide 25 and EN 45001, both of which it replaced. It contained all of the requirements that testing and calibration laboratories have to meet if they wish to demonstrate that they operate a management system, are technically competent, and are able to generate technically valid results.

The first edition referred to ISO 9001:1994 and ISO 9002:1994. These standards have been superseded by ISO 9001:2000, which made an alignment of ISO/IEC 17025 necessary. In this second edition, clauses have been amended or added only when considered necessary in the light of ISO 9001:2000.

Accreditation bodies that recognize the competence of testing and calibration laboratories should use this International Standard as the basis for their accreditation. Clause 4 specifies the requirements for sound management. Clause 5 specifies the requirements for technical competence for the type of tests and/or calibrations the laboratory undertakes.

Growth in the use of management systems generally has increased the need to ensure that laboratories which form part of larger organizations or offer other services can operate to a quality management system that is seen as compliant with ISO 9001 as well as with this International Standard. Care has been taken, therefore, to incorporate all those requirements of ISO 9001 that are relevant to the scope of testing and calibration services that are covered by the laboratory's management system.

Testing and calibration laboratories that comply with this International Standard will therefore also operate in accordance with ISO 9001.

Conformity of the quality management system within which the laboratory operates to the requirements of ISO 9001 does not of itself demonstrate the competence of the laboratory to produce technically valid data and results. Nor does demonstrated conformity to this International Standard imply conformity of the quality management system within which the laboratory operates to all the requirements of ISO 9001.

The acceptance of testing and calibration results between countries should be facilitated if laboratories comply with this International Standard and if they obtain accreditation from bodies which have entered into mutual recognition agreements with equivalent bodies in other countries using this International Standard.

The use of this International Standard will facilitate cooperation between laboratories and other bodies, and assist in the exchange of information and experience, and in the harmonization of standards and procedures.

General requirements for the competence of testing and calibration laboratories

1 Scope

1.1 This International Standard specifies the general requirements for the competence to carry out tests and/or calibrations, including sampling. It covers testing and calibration performed using standard methods, non-standard methods, and laboratory-developed methods.

1.2 This International Standard is applicable to all organizations performing tests and/or calibrations. These include, for example, first-, second- and third-party laboratories, and laboratories where testing and/or calibration forms part of inspection and product certification.

This International Standard is applicable to all laboratories regardless of the number of personnel or the extent of the scope of testing and/or calibration activities. When a laboratory does not undertake one or more of the activities covered by this International Standard, such as sampling and the design/development of new methods, the requirements of those clauses do not apply.

1.3 The notes given provide clarification of the text, examples and guidance. They do not contain requirements and do not form an integral part of this International Standard.

1.4 This International Standard is for use by laboratories in developing their management system for quality, administrative and technical operations. Laboratory customers, regulatory authorities and accreditation bodies may also use it in confirming or recognizing the competence of laboratories. This International Standard is not intended to be used as the basis for certification of laboratories.

NOTE 1 The term 'management system' in this International Standard means the quality, administrative and technical systems that govern the operations of a laboratory.

NOTE 2 Certification of a management system is sometimes also called registration.

1.5 Compliance with regulatory and safety requirements on the operation of laboratories is not covered by this International Standard.

1.6 If testing and calibration laboratories comply with the requirements of this International Standard, they will operate a quality management system for their testing and calibration activities that also meets the principles of ISO 9001. Annex A provides nominal cross-references between this International Standard and ISO 9001. This International Standard covers technical competence requirements that are not covered by ISO 9001.

NOTE 1 It might be necessary to explain or interpret certain requirements in this International Standard to ensure that the requirements are applied in a consistent manner. Guidance for establishing applications for specific fields, especially for accreditation bodies (see ISO/IEC 17011) is given in Annex B.

NOTE 2 If a laboratory wishes accreditation for part or all of its testing and calibration activities, it should select an accreditation body that operates in accordance with ISO/IEC 17011.

This is a free preview. Purchase the entire publication at the link below:

[Product Page](#)

-
- [Looking for additional Standards? Visit Intertek Inform Infostore](#)
 - [Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation](#)
-