



NSAI
Standards

Standard Recommendation
S.R. CEN/TS 16170:2012

Sludge, treated biowaste and soil - Determination of elements using inductively coupled plasma optical emission spectrometry (ICP-OES)

© CEN 2012

No copying without NSAI permission except as permitted by copyright law.

S.R. CEN/TS 16170:2012

Incorporating amendments/corrigenda/National Annexes issued since publication:

The National Standards Authority of Ireland (NSAI) produces the following categories of formal documents:

I.S. xxx: Irish Standard – national specification based on the consensus of an expert panel and subject to public consultation.

S.R. xxx: Standard Recommendation - recommendation based on the consensus of an expert panel and subject to public consultation.

SWIFT xxx: A rapidly developed recommendatory document based on the consensus of the participants of an NSAI workshop.

This document replaces:

This document is based on:
CEN/TS 16170:2012

Published:
10 December, 2012

This document was published
under the authority of the NSAI
and comes into effect on:
10 December, 2012

ICS number:

13.030.01
13.080.10

NSAI
1 Swift Square,
Northwood, Santry
Dublin 9

T +353 1 807 3800
F +353 1 807 3838
E standards@nsai.ie
W NSAI.ie

Sales:
T +353 1 857 6730
F +353 1 857 6729
W standards.ie

Údarás um Chaighdeáin Náisiúnta na hÉireann

ICS 13.030.01; 13.080.10

English Version

**Sludge, treated biowaste and soil - Determination of elements
using inductively coupled plasma optical emission spectrometry
(ICP-OES)**

Boues, bio-déchets et sols - Détermination des éléments
en traces par spectrométrie d'émission optique avec
plasma induit par haute fréquence (ICP-OES)

Schlamm, behandelter Bioabfall und Boden - Bestimmung
von Elementen mittels optischer Emissionsspektrometrie
mit induktiv gekoppeltem Plasma (ICP-OES)

This Technical Specification (CEN/TS) was approved by CEN on 16 July 2012 for provisional application.

The period of validity of this CEN/TS is limited initially to three years. After two years the members of CEN will be requested to submit their comments, particularly on the question whether the CEN/TS can be converted into a European Standard.

CEN members are required to announce the existence of this CEN/TS in the same way as for an EN and to make the CEN/TS available promptly at national level in an appropriate form. It is permissible to keep conflicting national standards in force (in parallel to the CEN/TS) until the final decision about the possible conversion of the CEN/TS into an EN is reached.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents

Page

Foreword.....	3
Introduction	4
1 Scope	5
2 Normative references	5
3 Principle	5
4 Interferences	5
5 Reagents	6
6 Apparatus	8
7 Procedure	9
7.1 Test sample solution	9
7.2 Test solution	9
7.3 Instrument set-up	9
7.4 Calibration	10
7.5 Sample measurement	10
8 Calculation	11
9 Expression of results	11
10 Performance characteristics	11
10.1 Calibration check	11
10.2 Interference	11
10.3 Recovery	12
10.4 Precision	12
11 Test report	12
Annex A (informative) Wavelengths, spectral interferences and estimated instrumental detection limits	13
Annex B (informative) Precision data	16
Annex C (informative) Inter-element correction	19
Bibliography	21

Foreword

This document (CEN/TS 16170:2012) has been prepared by Technical Committee CEN/TC 400 "Project Committee - Horizontal standards in the fields of sludge, biowaste and soil", the secretariat of which is held by DIN.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association.

The preparation of this document by CEN is based on a mandate by the European Commission (Mandate M/330), which assigned the development of standards on sampling and analytical methods for hygienic and biological parameters as well as inorganic and organic determinants, aiming to make these standards applicable to sludge, treated biowaste and soil as far as this is technically feasible.

According to the CEN/CENELEC Internal Regulations, the national standards organisations of the following countries are bound to announce this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Introduction

This Technical Specification is applicable for several types of matrices as indicated in Table 1.

Table 1 — Matrices for which this Technical Specification is applicable

Matrix	Materials tested
Sludge	Municipal sludge
Biowaste	Compost
Soil	Sludge amended soils

WARNING — Persons using this Technical Specification should be familiar with usual laboratory practice. This Technical Specification does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of the user to establish appropriate safety and health practices and to ensure compliance with any national regulatory conditions.

IMPORTANT — It is absolutely essential that tests conducted according to this Technical Specification be carried out by suitably trained staff.

1 Scope

This Technical Specification specifies a method for the determination of the following elements in *aqua regia*, nitric acid digest solutions of sludge, treated biowaste and soil:

Aluminium (Al), antimony (Sb), arsenic (As), barium (Ba), beryllium (Be), bismuth (Bi), boron (B), cadmium (Cd), calcium (Ca), chromium (Cr), cobalt (Co), copper (Cu), gallium (Ga), indium (In), iron (Fe), lead (Pb), lithium (Li), magnesium (Mg), manganese (Mn), mercury (Hg), molybdenum (Mo), nickel (Ni), phosphorus (P), potassium (K), selenium (Se), silicon (Si), silver (Ag), sodium (Na), strontium (Sr), sulfur (S), thallium (Tl), tin (Sn), titanium (Ti), tungsten (W), uranium (U), vanadium (V), zinc (Zn) and zirconium (Zr).

Table A.1 lists the elements for which this method is applicable along with the recommended wavelength and typical instrumental detection limits for clean matrices.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 15934, *Sludge, treated biowaste, soil and waste — Calculation of dry matter fraction after determination of dry residue or water content*

EN 16173, *Sludge, treated biowaste and soil — Digestion of nitric acid soluble fractions of elements*

EN 16174, *Sludge, treated biowaste and soil — Digestion of aqua regia soluble fractions of elements*

EN ISO 3696, *Water for analytical laboratory use — Specification and test methods (ISO 3696)*

3 Principle

Digests of sludge, treated biowaste or soil with nitric acid or *aqua regia* (see EN 16173 and EN 16174) are analysed by inductively coupled plasma optical emission spectrometry (ICP-OES) using sequential or simultaneous optical systems and axial or radial viewing of the plasma.

The instrument measures characteristic emission spectra by optical spectrometry. Analyte species originating in the digest solution are nebulised and the resulting aerosol is transported to the plasma torch. Element-specific emission spectra are produced by a radio-frequency inductively coupled plasma. The spectra are dispersed by a grating spectrometer, and the intensities of the emission lines are monitored by photosensitive devices.

4 Interferences

Background correction is required for trace element determination. Background correction is not required in cases of line broadening where a background correction measurement would actually degrade the analytical result. Additional interferences and matrix effects shall be recognised and appropriate corrections made. Tests for their presence are described below.

Spectral interferences are caused by background emission from continuous or recombination phenomena, stray light which causes background increase or overlap of a spectral line from another element, or unresolved overlap of molecular band spectra.

This is a free preview. Purchase the entire publication at the link below:

[Product Page](#)

-
- Looking for additional Standards? Visit Intertek Inform Infostore
 - Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation
-