This is a free page sample. Access the full version online.



Irish Standard I.S. EN 60601-2-4:2011

Medical electrical equipment -- Part 2 -4: Particular requirements for the basic safety and essential performance of cardiac defibrillators (IEC 60601-2 -4:2010 (EQV))

 $\ensuremath{\mathbb{C}}$ NSAI 2011 No copying without NSAI permission except as permitted by copyright law.

Incorporating amendments/corrigenda issued since publication:

The National Standards Authority of Ireland (NSAI) produces the following categories of formal documents:

I.S. xxx: Irish Standard – national specification based on the consensus of an expert panel and subject to public consultation.

S.R. xxx: Standard Recommendation - recommendation based on the consensus of an expert panel and subject to public consultation.

SWiFT xxx: A rapidly developed recommendatory document based on the consensus of the participants of an NSAI workshop.

<i>This document replaces:</i> EN 60601-2-4:2003	<i>This document is based on:</i> EN 60601-2-4:2011 EN 60601-2-4:2003	Publisi 19 Aug 31 Jan	<i>hed:</i> ğust, 2011 uary, 2003
This document was published under the authority of the NSAI and c 23 August, 2011	omes into effect on:		ICS number: 11.040.10
NSAI T +353 1 807 3800 Sales: 1 Swift Square, F +353 1 807 3838 T +353 1 857 6730 Northwood, Santry E standards@nsai.ie F +353 1 857 6729 Dublin 9 W NSAI.ie W standards.ie			
Údarás um Chaighdeáin Náisiúnta na hÉireann			

EUROPEAN STANDARD

EN 60601-2-4

NORME EUROPÉENNE EUROPÄISCHE NORM

August 2011

ICS 11.040.10

Supersedes EN 60601-2-4:2003

English version

Medical electrical equipment -Part 2-4: Particular requirements for the basic safety and essential performance of cardiac defibrillators

(IEC 60601-2-4:2010)

Appareils électromédicaux -Partie 2-4: Exigences particulières pour la sécurité de base et les performances essentielles des défibrillateurs cardiaques (CEI 60601-2-4:2010) Medizinische elektrische Geräte -Teil 2-4: Besondere Festlegungen für die Sicherheit einschließlich der wesentlichen Leistungsmerkmale von Defibrillatoren (IEC 60601-2-4:2010)

This European Standard was approved by CENELEC on 2011-01-12. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the Central Secretariat has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

CENELEC

European Committee for Electrotechnical Standardization Comité Européen de Normalisation Electrotechnique Europäisches Komitee für Elektrotechnische Normung

Management Centre: Avenue Marnix 17, B - 1000 Brussels

© 2011 CENELEC - All rights of exploitation in any form and by any means reserved worldwide for CENELEC members.

EN 60601-2-4:2011

- 2 -

Foreword

The text of document 62D/857/FDIS, future edition 3 of IEC 60601-2-4, prepared by SC 62D, Electromedical equipment, of IEC TC 62, Electrical equipment in medical practice, was submitted to the IEC-CENELEC parallel vote and was approved by CENELEC as EN 60601-2-4 on 2011-01-12.

This European Standard supersedes EN 60601-2-4:2003.

EN 60601-2-4:2011 constitutes a technical revision, revised to structurally align it with EN 60601-1:2006 and to implement the decision of IEC SC 62A that the clause numbering structure of particular standards written to EN 60601-1:2006 would adhere to the form specified in ISO/IEC Directives, Part 2:2004. The aim of this third edition is to bring this particular standard up to date with reference to the third edition of the general standard through reformatting and technical changes.

The principle technical changes are as follows:

- 201.8.8.3, test 4: added additional test options;
- Figure 201.105: provided example of stainless steel plates. Added note for 10 Hz generator or shockable rhythm generator;
- Figure 201.101: Changed orientation of the lower diode at the oscilloscope connection;
- 202.6.1, .2, .4: "Additions" and "Replacements" corrected to be as originally intended;
- 201.101.1: Clarified preconditioning of a non-rechargeable battery;
- 201.3.207: Clarified definition of DUMMY COMPONENT;
- 201.15.4.101: In paragraph b), added reduced flex requirements for sterilizable internal paddles with specified limit on sterilization cycles;
- 201.15.4.3.103: Added an option for devices having non-changeable pre-programmed energy-setting sequences;
- 201.102.3.1, 2: Changed from specified defibrillation cycles to use of pre-programmed defibrillation sequence;
- 202.6.2.2.1: Changed ESD discharge sequence to match EN 60601-1-2:2007.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN and CENELEC shall not be held responsible for identifying any or all such patent rights.

The following dates were fixed:

-	latest date by which the EN has to be implemented at national level by publication of an identical national standard or by endorsement	(dop)	2012-02-12
_	latest date by which the national standards conflicting with the EN have to be withdrawn	(dow)	2014-01-12

In this standard, the following print types are used:

- Requirements and definitions: roman type.

- Test specifications: italic type.

- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.

– TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- "clause" means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);

- "subclause" means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term "Clause" followed by the clause number. References to subclauses within this particular standard are by number only.

In this standard, the conjunctive "or" is used as an "inclusive or" so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- "shall" means that compliance with a requirement or a test is mandatory for compliance with this standard;

- "should" means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;

- "may" is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

This European Standard has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association and covers essential requirements of EC Directive MDD (93/42/EEC). See Annex ZZ.

Annexes ZA and ZZ have been added by CENELEC.

Endorsement notice

The text of the International Standard IEC 60601-2-4:2010 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following note has to be added for the standard indicated:

[2] IEC 60601-2-27 NOTE Harmonized as EN 60601-2-27.

EN 60601-2-4:2011

- 4 -

Annex ZA

(normative)

Normative references to international publications with their corresponding European publications

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE When an international publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

Annex ZA of EN 60601-1:2006 a	applies except	as follows:
-------------------------------	----------------	-------------

Publication	Year	Title	<u>EN/HD</u>	Year
Replace IEC 6060	1-1-2 by:			
IEC 60601-1-2 (mod)	2007	Medical electrical equipment - Part 1-2: General requirements for basic safety and essential performance - Collateral standard: Electromagnetic compatibility - Requirements and tests	EN 60601-1-2 + corr. March	2007 2010
Add:				
IEC 61000-4-2	-	Electromagnetic compatibility (EMC) - Part 4-2: Testing and measurement techniques - Electrostatic discharge immunity test	EN 61000-4-2	-
ISO 15223-1	2007	Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied - Part 1: General requirements	-	-

- 5 -

EN 60601-2-4:2011

Annex ZZ

(informative)

Coverage of Essential Requirements of EC Directives

This European Standard has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association and within its scope the standard covers all relevant essential requirements as given in Annex I of the EC Directive 93/42/EEC

Compliance with this standard provides one means of conformity with the specified essential requirements of the Directive concerned.

WARNING: Other requirements and other EC Directives may be applicable to the products falling within the scope of this standard.

This is a free page sample. Access the full version online.

I.S. EN 60601-2-4:2011

This page is intentionally left BLANK.

60601-2-4 © IEC:2010

CONTENTS

- 2 -

FOREWOR	D	4
201.1	Scope, object and related standards	7
201.2	Normative references	9
201.3	Terms and definitions	9
201.4	General requirements	11
201.5	General requirements for testing of ME EQUIPMENT	12
201.6	Classification of ME EQUIPMENT and ME SYSTEMS	12
201.7	ME EQUIPMENT identification, marking and documents	13
201.8	Protection against electrical HAZARDS from ME EQUIPMENT	17
201.9	Protection against MECAHNICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS	23
201.10	Protection against unwanted and excessive radiation HAZARDS	23
201.11	Protection against excessive temperatures and other HAZARDS	23
201.12	* Accuracy of controls and instruments and protection against hazardous outputs	25
201.13	HAZARDOUS SITUATIONS and fault conditions	27
201.14	PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS)	27
201.15	Construction of ME EQUIPMENT	27
201.16	ME SYSTEMS	32
201.17	Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	32
201.101	* Charging time	32
201.102	Internal electrical power source	35
201.103	* Endurance	36
201.104	* Synchronizer	37
201.105	* Recovery of the MONITOR and/or ECG input after defibrillation	37
201.106	* Disturbance to the MONITOR from charging or internal discharging	41
201.107	* Requirements for RHYTHM RECOGNITION DETECTOR	42
201.108	DEFIBRILLATOR ELECTRODES	43
201.109	* External pacing (U.S.)	45
202	* Electromagnetic compatibility – Requirements and tests	49
Annexes		52
Annex C (ir	formative) Guide to marking and labelling requirements for ME EQUIPMENT	53
Annex AA (informative) Particular guidance and rationale	55
Annex BB (IEC 60601-	informative) Mapping between the elements of the second edition of 2-4 and IEC 60601-2-4:2010	68
Bibliograph	V	73
Index of de	fined terms used in this particular standard	74
		-
Figure 201. ME EQUIPME	101 – Dynamic test for limitation of energy from different parts of the	18
Figure 201.	102 – Allowed current versus applied test voltage	22
Figure 201.	103 – Examples of cord anchorages that require testing	31
Figure 201.	104 – Test apparatus for flexible cords and their anchorages	32

60601-2-4 © IEC:2010	- 3 -
----------------------	-------

Figure 201.105 – Arrangement for test of recovery after defibrillation	39
Figure 201.106 – Arrangement of monitoring electrodes on sponge	40
Figure 201.107 – Arrangement for recovery test after defibrillation	40
Figure 201.108 – Arrangement for test of disturbance from charging and internal discharging	42
Figure 201.109 – Test circuit for offset instability/internal noise determination	49
Figure 201.110 – Test circuit for DEFIBRILLATOR overload test of pacing output circuitry	49
Table 201.101 – Distributed ESSENTIAL PERFORMANCE requirements	12
Table 201.102 – Rhythm recognition detector categories	42
Table 201.C.101 – Marking on the outside of a CARDIAC DEFIBRILLATOR or its parts	53
Table 201.C.102 – Marking of controls and instruments of a CARDIAC DEFIBRILLATOR	53
Table 201.C.103 – Accompanying documents, general	53
Table 201.C.104 – Accompanying documents, instructions for use	54
Table 201.C.105 – Accompanying documents, technical description	54
Table BB.1 – Mapping between the elements of the second edition of IEC 60601-2-4 and IEC 60601-2-4:2010	68

- 4 -

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT -

Part 2-4: Particular requirements for the basic safety and essential performance of cardiac defibrillators

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International standard IEC 60601-2-4 has been prepared by IEC subcommittee 62D: Electromedical equipment, of IEC technical committee 62: Electrical equipment in medical practice.

This third edition cancels and replaces the second edition published in 2002. This edition constitutes a technical revision, revised to structurally align it with IEC 60601-1:2005 and to implement the decision of IEC SC 62A that the clause numbering structure of particular standards written to IEC 60601-1:2005 would adhere to the form specified in ISO/IEC Directives, Part 2:2004. The aim of this third edition is to bring this particular standard up to date with reference to the third edition of the general standard through reformatting and technical changes.

The principle technical changes are as follows:

201.8.8.3, test 4: added additional test options;

60601-2-4 © IEC:2010

- 5 -

- Figure 201.105: provided example of stainless steel plates. Added note for 10 Hz generator or shockable rhythm generator;
- Figure 201.101: Changed orientation of the lower diode at the oscilloscope connection;
- 202.6.1, .2, .4: "Additions" and "Replacements" corrected to be as originally intended;
- 201.101.1: Clarified preconditioning of a non-rechargeable battery;
- 201.3.207: Clarified definition of DUMMY COMPONENT;
- 201.15.4.101: In paragraph b), added reduced flex requirements for sterilizable internal paddles with specified limit on sterilization cycles;
- 201.15.4.3.103: Added an option for devices having non-changeable pre-programmed energy-setting sequences;
- 201.102.3.1, 2: Changed from specified defibrillation cycles to use of pre-programmed defibrillation sequence;
- 202.6.2.2.1: Changed ESD discharge sequence to match IEC 60601-1-2, third edition.

The text of this particular standard is based on the following documents:

FDIS	Report on voting
62D/857/FDIS	62D/878/RVD

Full information on the voting for the approval of this particular standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this standard, the following print types are used:

- Requirements and definitions: roman type.
- Test specifications: italic type.
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type.
 Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- "clause" means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- "subclause" means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term "Clause" followed by the clause number. References to subclauses within this particular standard are by number only.

In this standard, the conjunctive "or" is used as an "inclusive or" so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- "shall" means that compliance with a requirement or a test is mandatory for compliance with this standard;
- "should" means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;

- 6 -

 "may" is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 series, published under the general title *Medical electrical equipment,* can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC web site under "http://webstore.iec.ch" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

60601-2-4 © IEC:2010

- 7 -

MEDICAL ELECTRICAL EQUIPMENT -

Part 2-4: Particular requirements for the basic safety and essential performance of cardiac defibrillators

201.1 Scope, object and related standards

Clause 1 of the general standard¹ applies, except as follows:

201.1.1 * Scope

Replacement:

This International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of CARDIAC DEFIBRILLATORS, hereafter referred to as ME EQUIPMENT.

If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

HAZARDS inherent in the intended physiological function of ME EQUIPMENT or ME SYSTEMS within the scope of this standard are not covered by specific requirements in this standard except in 7.2.13 and 8.4.1 of the general standard.

NOTE See also 4.2 of the general standard.

This particular standard does not apply to implantable defibrillators, remote control DEFIBRILLATORS, external transcutaneous pacemakers, or separate stand-alone cardiac monitors (which are standardized by IEC 60601-2-27 [2]²). Cardiac monitors which use separate ECG monitoring electrodes are not within the scope of this standard unless they are used as the sole basis for AED rhythm recognition detection or beat detection for synchronized cardioversion.

Defibrillation waveform technology is evolving rapidly. Published studies indicate that the effectiveness of waveforms varies. The choice of a particular waveform including waveshape, delivered energy, efficacy, and safety has been specifically excluded from the scope of this standard.

However, due to the critical importance of the therapeutic waveform, comments have been added to the rationale which addresses considerations in waveform selection.

201.1.2 Object

Replacement:

The object of this particular standard is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for cardiac defibrillators as defined in 201.3.202.

¹ The general standard is IEC 60601-1:2005, Medical electrical equipment – Part 1: General requirements for basic safety and essential performance.

² Numbers in square brackets refer to the bibliography.



This is a free preview. Purchase the entire publication at the link below:

Product Page

S Looking for additional Standards? Visit Intertek Inform Infostore

> Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation