



**NSAI**  
Standards

Irish Standard  
I.S. EN ISO 6974-5:2014

Natural gas - Determination of composition and associated uncertainty by gas chromatography - Part 5: Isothermal method for nitrogen, carbon dioxide, C1 to C5 hydrocarbons and C6+ hydrocarbons (ISO 6974-5:2014)

**I.S. EN ISO 6974-5:2014**

*Incorporating amendments/corrigenda/National Annexes issued since publication:*

The National Standards Authority of Ireland (NSAI) produces the following categories of formal documents:

I.S. xxx: Irish Standard — national specification based on the consensus of an expert panel and subject to public consultation.

S.R. xxx: Standard Recommendation — recommendation based on the consensus of an expert panel and subject to public consultation.

SWiFT xxx: A rapidly developed recommendatory document based on the consensus of the participants of an NSAI workshop.

*This document replaces/revises/consolidates the NSAI adoption of the document(s) indicated on the CEN/CENELEC cover/Foreword and the following National document(s):*

*NOTE: The date of any NSAI previous adoption may not match the date of its original CEN/CENELEC document.*

*This document is based on:*

EN ISO 6974-5:2014

*Published:*

2014-07-16

*This document was published  
under the authority of the NSAI  
and comes into effect on:*

2014-08-02

ICS number:

75.060

NOTE: If blank see CEN/CENELEC cover page

NSAI  
1 Swift Square,  
Northwood, Santry  
Dublin 9

T +353 1 807 3800  
F +353 1 807 3838  
E standards@nsai.ie  
W NSAI.ie

Sales:  
T +353 1 857 6730  
F +353 1 857 6729  
W standards.ie

Údarás um Chaighdeáin Náisiúnta na hÉireann

EUROPEAN STANDARD

**EN ISO 6974-5**

NORME EUROPÉENNE

EUROPÄISCHE NORM

July 2014

ICS 75.060

Supersedes EN ISO 6974-5:2001

English Version

**Natural gas - Determination of composition and associated uncertainty by gas chromatography - Part 5: Isothermal method for nitrogen, carbon dioxide, C1 to C5 hydrocarbons and C6+ hydrocarbons (ISO 6974-5:2014)**

Gaz naturel - Détermination de la composition et de l'incertitude associée par chromatographie en phase gazeuse - Partie 5: Méthode isotherme pour l'azote, le dioxyde de carbone, les hydrocarbures C1 à C5 et C6+ (ISO 6974-5:2014)

Erdgas - Bestimmung der Zusammensetzung und der zugehörigen Unsicherheit durch Gaschromatographie - Teil 5: Isothermes Verfahren für Stickstoff, Kohlenstoffdioxid, C1- bis C5-Kohlenwasserstoffe und C6+-Kohlenwasserstoffe (ISO 6974-5:2014)

This European Standard was approved by CEN on 28 June 2014.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels**

## **Contents**

Page

<b>Foreword.....</b>	<b>3</b>
----------------------	----------

## **Foreword**

This document (EN ISO 6974-5:2014) has been prepared by Technical Committee ISO/TC 193 "Natural gas".

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by January 2015, and conflicting national standards shall be withdrawn at the latest by January 2015.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 6974-5:2001.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

### **Endorsement notice**

The text of ISO 6974-5:2014 has been approved by CEN as EN ISO 6974-5:2014 without any modification.

This page is intentionally left blank

# INTERNATIONAL STANDARD

ISO  
6974-5

Second edition  
2014-07-15

---

---

## Natural gas — Determination of composition and associated uncertainty by gas chromatography —

Part 5:

### Isothermal method for nitrogen, carbon dioxide, C<sub>1</sub> to C<sub>5</sub> hydrocarbons and C<sub>6+</sub> hydrocarbons

*Gaz naturel — Détermination de la composition et de l'incertitude  
associée par chromatographie en phase gazeuse —*

*Partie 5: Méthode isotherme pour l'azote, le dioxyde de carbone, les  
hydrocarbures C<sub>1</sub> à C<sub>5</sub> et C<sub>6+</sub>*



Reference number  
ISO 6974-5:2014(E)

© ISO 2014

**ISO 6974-5:2014(E)**



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2014

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.org](mailto:copyright@iso.org)  
Web [www.iso.org](http://www.iso.org)

Published in Switzerland



# Contents

	Page
<b>Foreword</b> .....	<b>iv</b>
<b>Introduction</b> .....	<b>v</b>
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>2</b>
<b>3 Principle</b> .....	<b>2</b>
<b>4 Materials</b> .....	<b>4</b>
<b>5 Apparatus</b> .....	<b>4</b>
<b>6 Scheme of the configuration</b> .....	<b>6</b>
<b>7 Procedure</b> .....	<b>7</b>
7.1 Control of the apparatus.....	7
7.2 Operation of the apparatus.....	8
<b>8 Expression of results</b> .....	<b>11</b>
8.1 Uncertainty.....	11
8.2 Test report.....	11
<b>Annex A (informative) Example of application</b> .....	<b>12</b>
<b>Annex B (informative) Procedure for Setting Valve timings and Restrictor Setting</b> .....	<b>22</b>
<b>Bibliography</b> .....	<b>24</b>

## ISO 6974-5:2014(E)

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 [www.iso.org/directives](http://www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received [www.iso.org/patents](http://www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT), see the following URL: Foreword - Supplementary information

The committee responsible for this document is ISO/TC 193, *Natural Gas*, Subcommittee SC 1, *Analysis of Natural Gas*.

This second edition cancels and replaces the first edition (ISO 6974-5:2000).

ISO 6974 consists of the following parts, under the general title *Natural gas — Determination of composition and associated uncertainty by gas chromatography*:

- *Part 1: General guidelines and calculation of composition*
- *Part 2: Uncertainty calculations*
- *Part 3: Determination of hydrogen, helium, oxygen, nitrogen, carbon dioxide and hydrocarbons up to C<sub>8</sub> using two capillary columns and one packed column*
- *Part 4: Determination of nitrogen, carbon dioxide and C<sub>1</sub> to C<sub>5</sub> and C<sub>6</sub>+ hydrocarbons for a laboratory and on-line measuring system using two columns*
- *Part 5: Isothermal method for nitrogen, carbon dioxide, C<sub>1</sub> to C<sub>5</sub> hydrocarbons and C<sub>6</sub>+ hydrocarbons*
- *Part 6: Determination of helium, oxygen, nitrogen, carbon dioxide and C<sub>1</sub> to C<sub>8</sub> hydrocarbons using three capillary columns*

## Introduction

This part of ISO 6974 describes a method for the analysis of natural gas that is commonly used for online process applications, but can be applied to laboratory instruments. The compositional data obtained are used for the calculation of calorific value, density and Wobbe index.

It is assumed that the natural gas does not contain any oxygen at source and that any oxygen which may be present is due to contamination during sampling.

The primary use of this chromatographic method is the calculation of calorific value (CV) according to ISO 6976. It is based on a column switching technique in which multiple columns, chosen for their separating ability for particular groups of components, are switched under automatic control.

Only one injection is necessary and the first phase of the method involves accelerated backflush of C<sub>6</sub>+ (which is measured as a recombined “pseudo component” rather than by the summation of individual component measurements). Lighter components (nitrogen, methane, carbon dioxide and ethane) are stored on the appropriate separating column while the heavier, C<sub>3</sub> to C<sub>5</sub> hydrocarbons are eluted. The lighter components are then separated by redirecting carrier gas on to the appropriate column.

A Thermal Conductivity Detector (TCD) is used for measurement of the above components.

When the method is first set up, the repeatability of measurement is established by repetitive analysis of a cylinder of test gas, commonly a typical natural gas. For each component, a control chart showing the mean value, and the bounds representing 2 and 3 standard deviations, is drawn up. Subsequently, this test gas is analysed after each calibration of the analyser, and the results are compared with the data in the control charts. The performance of the analyser is assessed by this procedure.

Any change in the method setup can give rise to differences in component responses and hence (where applied) to calculated uncertainties. In these circumstances fitting data to an existing control chart is not a suitable procedure, and the operations that were undertaken when the method was first set up shall be repeated.

This part of ISO 6974 provides one of the methods that may be used for determining the compositions of natural gas in accordance with ISO 6974-1 and ISO 6974-2.



# Natural gas — Determination of composition and associated uncertainty by gas chromatography —

## Part 5:

## Isothermal method for nitrogen, carbon dioxide, C<sub>1</sub> to C<sub>5</sub> hydrocarbons and C<sub>6+</sub> hydrocarbons

### 1 Scope

This part of this International Standard describes a gas chromatographic method for the quantitative determination of the content of nitrogen, carbon dioxide and C<sub>1</sub> to C<sub>5</sub> hydrocarbons individually and a composite C<sub>6+</sub> measurement, which represents all hydrocarbons of carbon number 6 and above in natural gas samples. It is applicable to the analysis of gases containing constituents within the working ranges given in [Table 1](#).

**Table 1 — Component working ranges**

Component		Mole fraction	
		%	
		Min.	Max.
Nitrogen	N <sub>2</sub>	0,1	22
Carbon dioxide	CO <sub>2</sub>	0,05	15
Methane	CH <sub>4</sub>	34	100
Ethane	C <sub>2</sub> H <sub>6</sub>	0,1	23
Propane	C <sub>3</sub> H <sub>8</sub>	0,05	10
iso-Butane	i-C <sub>4</sub> H <sub>10</sub>	0,01	2,0
n-Butane	n-C <sub>4</sub> H <sub>10</sub>	0,01	2,0
neo-Pentane	neo-C <sub>5</sub> H <sub>12</sub>	0,005	0,35
iso-Pentane	i-C <sub>5</sub> H <sub>12</sub>	0,005	0,35
n-Pentane	n-C <sub>5</sub> H <sub>12</sub>	0,005	0,35
Hexanes +	C <sub>6+</sub>	0,005	0,35
NOTE 1 The working ranges in <a href="#">Table 1</a> are those for which the method has been shown to be satisfactory, and are offered for guidance. However, there is no reason why wider ranges should not be used, provided that the successful measurement of such wider ranges has been demonstrated.			
NOTE 2 Hydrocarbons above n-pentane are expressed as the “pseudo-component” C <sub>6+</sub> which is measured as one composite peak and calibrated as such. The properties of C <sub>6+</sub> are calculated from an extended analysis of the individual C <sub>6</sub> and higher hydrocarbons.			
NOTE 3 Oxygen is not a normal constituent of natural gas and would not be expected to be present in gas sampled to an online instrument. If any oxygen is present as a result of air contamination, it will be measured with the nitrogen. The resulting measured (nitrogen + oxygen) value will be in error to a small extent because of the slight difference between the detector responses of oxygen and nitrogen.			
NOTE 4 The helium and argon contents are assumed to be sufficiently small and unvarying that they need not be analysed for.			
NOTE 5 The gas sample shall not contain any hydrocarbon condensate and/or water.			

This is a free preview. Purchase the entire publication at the link below:

[Product Page](#)

- 
- Looking for additional Standards? Visit Intertek Inform Infostore
  - Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation
-