



NSAI
Standards

Standard Recommendation
S.R. CEN/TS 16023:2013

Characterization of waste - Determination of gross calorific value and calculation of net calorific value

S.R. CEN/TS 16023:2013

Incorporating amendments/corrigenda/National Annexes issued since publication:

The National Standards Authority of Ireland (NSAI) produces the following categories of formal documents:

I.S. xxx: Irish Standard – national specification based on the consensus of an expert panel and subject to public consultation.

S.R. xxx: Standard Recommendation - recommendation based on the consensus of an expert panel and subject to public consultation.

SWIFT xxx: A rapidly developed recommendatory document based on the consensus of the participants of an NSAI workshop.

This document replaces:

This document is based on:
CEN/TS 16023:2013

Published:
27 November, 2013

This document was published
under the authority of the NSAI
and comes into effect on:
27 November, 2013

ICS number:
13.030.01

NSAI
1 Swift Square,
Northwood, Santry
Dublin 9

T +353 1 807 3800
F +353 1 807 3838
E standards@nsai.ie
W NSAI.ie

Sales:
T +353 1 857 6730
F +353 1 857 6729
W standards.ie

Údarás um Chaighdeáin Náisiúnta na hÉireann

ICS 13.030.01

English Version

Characterization of waste - Determination of gross calorific value and calculation of net calorific value

Caractérisation des déchets - Détermination du pouvoir
calorifique supérieur et calcul du pouvoir calorifique
inférieur

Charakterisierung von Abfällen - Bestimmung des
Brennwertes und Berechnung des Heizwertes

This Technical Specification (CEN/TS) was approved by CEN on 6 August 2013 for provisional application.

The period of validity of this CEN/TS is limited initially to three years. After two years the members of CEN will be requested to submit their comments, particularly on the question whether the CEN/TS can be converted into a European Standard.

CEN members are required to announce the existence of this CEN/TS in the same way as for an EN and to make the CEN/TS available promptly at national level in an appropriate form. It is permissible to keep conflicting national standards in force (in parallel to the CEN/TS) until the final decision about the possible conversion of the CEN/TS into an EN is reached.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents

Page

Foreword.....	3
Introduction	4
1 Scope	5
2 Normative references	5
3 Terms and definitions	5
4 Principle.....	6
4.1 Gross calorific value	6
4.2 Net calorific value	6
5 Reagents.....	6
6 Apparatus	7
7 Sample storage	8
8 Sample preparation	8
9 Procedure	9
9.1 General.....	9
9.2 General preparations, measurements and temperature corrections	10
9.3 Calibration	13
9.4 Samples	13
10 Calculation of effective heat capacity.....	14
11 Gross calorific value	15
12 Calculation of net calorific value.....	16
12.1 General.....	16
12.2 Net calorific value at constant pressure	16
13 Expression of results	16
14 Quality control.....	17
15 Test report	17
Annex A (informative) Example of a calorimeter	18
Annex B (informative) Temperature evolution	19
Annex C (informative) Calculation of the gross calorific value – Correct calculation versus the simplified calculation	20
C.1 Correct calculation	20
C.2 Sulfur correction	21
C.3 Nitrogen correction.....	21
C.4 Halogens correction	21
C.5 Influence of the thermo-chemical corrections.....	22
Annex D (informative) Typical hydrogen contents in waste products	24
Annex E (informative) Summary of general requirements and recommendations	25
Bibliography	26

Foreword

This document (CEN/TS 16023:2013) has been prepared by Technical Committee CEN/TC 292 "Characterization of waste", the secretariat of which is held by NEN.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to announce this Technical Specification: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Introduction

This method is a quick and easy way to evaluate the suitability of waste to be treated by thermal processes.

The determination of gross calorific value is carried out without thermo-chemical corrections. These corrections typically result in minor changes of the result. The influence of these corrections is shown in Annex C.

The result obtained is the gross calorific value at constant volume with both the water of the combustion products and the moisture of the waste as liquid water.

The net calorific value is obtained by calculation from the gross calorific value. For the purpose of this Technical Specification, the value of the net calorific value does not need to be determined exactly. The correction for hydrogen is simplified by the use of typical hydrogen contents derived from table values of hydrogen contents in common types of waste.

This Technical Specification specifies a quick method to determine calorific value; a more comprehensive analysis is described in ISO 1928.

Waste can contain water and (unburnable) solids in large amounts. Therefore their calorific value – especially on the “as received” basis – can be quite low. For some purposes it may be sufficient to determine the gross calorific value only, and not the net calorific value.

WARNING — Strict adherence to all of the provisions prescribed in this Technical Specification should ensure against explosive rupture of the bomb, or a blow-out, provided that the bomb is of proper design and construction and in good mechanical condition. Anyone dealing with waste and sludge analysis is required to be aware of the typical risks of this kind of material irrespective of the parameter to be determined. Waste and sludge samples may contain hazardous (e.g. toxic, reactive, flammable, infectious) substances, which can be liable to biological and/or chemical reaction. Consequently, it is recommended that these samples be handled with special care. The gases that may be produced by microbiological or chemical activity are potentially flammable and will pressurize sealed bottles. Bursting bottles are likely to result in hazardous shrapnel, dust and/or aerosol. National regulations should be followed with respect to all hazards associated with this method.

1 Scope

This Technical Specification specifies a simplified method for the determination of the gross calorific value of waste at constant volume and at the reference temperature of 25 °C in a bomb calorimeter calibrated by combustion of certified benzoic acid. This Technical Specification does not include thermo-chemical corrections.

This Technical Specification also specifies a simplified calculation of the net calorific value from the gross calorific value.

This Technical Specification is applicable for the evaluation of suitability of waste to be treated by thermal processes and for the energy to be recovered.

This Technical Specification is applicable to all kinds of waste.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 13965-2:2010, *Characterization of waste - Terminology - Part 2: Management related terms and definitions*

EN 14346, *Characterization of waste - Calculation of dry matter by determination of dry residue or water content*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in EN 13965-2:2010 and the following apply.

3.1

corrected temperature rise

change in calorimeter temperature caused solely by the processes taking place within the combustion bomb

Note 1 to entry: It is the total observed temperature rise corrected for heat exchange, stirring power etc.

3.2

gross calorific value at constant volume

absolute value of the specific energy of combustion, in Joules, for unit mass of waste burned in oxygen in a calorimetric bomb under the conditions specified

Note 1 to entry: The products of combustion are assumed to consist of gaseous oxygen, nitrogen, carbon dioxide and sulfur dioxide, of liquid water (in equilibrium with its vapour) saturated with carbon dioxide under the conditions of the bomb reaction, and of solid ash, all at the reference temperature.

3.3

net calorific value at constant pressure

absolute value of the specific energy of combustion, in Joules, for unit mass of waste burned in oxygen at constant pressure under such conditions that all the water of the reaction products remains as water vapour (in a hypothetical state at 0,1 MPa)

Note 1 to entry: The other products are assumed to remain at the reference temperature.

This is a free preview. Purchase the entire publication at the link below:

[Product Page](#)

-
- Looking for additional Standards? Visit Intertek Inform Infostore
 - Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation
-