



NSAI
Standards

Irish Standard
I.S. EN 61672-1:2013

Electroacoustics - Sound level meters -- Part 1: Specifications

I.S. EN 61672-1:2013

Incorporating amendments/corrigenda/National Annexes issued since publication:

The National Standards Authority of Ireland (NSAI) produces the following categories of formal documents:

I.S. xxx: Irish Standard — national specification based on the consensus of an expert panel and subject to public consultation.

S.R. xxx: Standard Recommendation — recommendation based on the consensus of an expert panel and subject to public consultation.

SWiFT xxx: A rapidly developed recommendatory document based on the consensus of the participants of an NSAI workshop.

This document replaces/revises/consolidates the NSAI adoption of the document(s) indicated on the CEN/CENELEC cover/Foreword and the following National document(s):

NOTE: The date of any NSAI previous adoption may not match the date of its original CEN/CENELEC document.

This document is based on:

EN 61672-1:2013

Published:

2013-12-13

*This document was published
under the authority of the NSAI
and comes into effect on:*

2013-12-24

ICS number:

17.140.50

NOTE: If blank see CEN/CENELEC cover page

NSAI
1 Swift Square,
Northwood, Santry
Dublin 9

T +353 1 807 3800
F +353 1 807 3838
E standards@nsai.ie
W NSAI.ie

Sales:
T +353 1 857 6730
F +353 1 857 6729
W standards.ie

Údarás um Chaighdeáin Náisiúnta na hÉireann

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN 61672-1

December 2013

ICS 17.140.50

Supersedes EN 61672-1:2003

English version

**Electroacoustics -
Sound level meters -
Part 1: Specifications
(IEC 61672-1:2013)**

Electroacoustique -
Sonomètres -
Partie 1: Spécifications
(CEI 61672-1:2013)

Elektroakustik -
Schallpegelmesser -
Teil 1: Anforderungen
(IEC 61672-1:2013)

This European Standard was approved by CENELEC on 2013-11-04. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

CENELEC

European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Avenue Marnix 17, B - 1000 Brussels

Foreword

The text of document 29/812/FDIS, future edition 2 of IEC 61672-1, prepared by IEC/TC 29 "Electroacoustics" in cooperation with the International Organization of Legal Metrology (OIML), was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN 61672-1:2013.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2013-08-04
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2016-11-04

This document supersedes EN 61672-1:2003.

EN 61672-1:2013 includes the following significant technical changes with respect to EN 61672-1:2003.

In this second edition, conformance to specifications is demonstrated when

- a) measured deviations from design goals do not exceed the applicable acceptance limits, and
- b) the uncertainty of measurement does not exceed the corresponding maximum-permitted uncertainty, with both uncertainties determined for a coverage probability of 95 %.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC [and/or CEN] shall not be held responsible for identifying any or all such patent rights.

Endorsement notice

The text of the International Standard IEC 61672-1:2013 was approved by CENELEC as a European Standard without any modification.

Annex ZA

(normative)

Normative references to international publications with their corresponding European publications

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE When an international publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 60942	-	Electroacoustics - Sound calibrators	EN 60942	-
IEC 61000-4-2	2008	Electromagnetic compatibility (EMC) Part 4-2: Testing and measurement techniques - Electrostatic discharge immunity test	EN 61000-4-2	2009
IEC 61000-6-2	2005	Electromagnetic compatibility (EMC) Part 6-2: Generic standards - Immunity for industrial environments	EN 61000-6-2 + corr. September	2005 2005
IEC 61094-6	-	Measurement microphones Part 6: Electrostatic actuators for determination of frequency response	EN 61094-6	-
IEC 61183	-	Electroacoustics - Random-incidence and diffuse-field calibration of sound level meters	EN 61183	-
IEC 62585	-	Electroacoustics - Methods to determine corrections to obtain the free-field response of a sound level meter	EN 62585	-
ISO/IEC Guide 98-4	2012	Uncertainty of measurement Part 4: Role of measurement uncertainty in conformity assessment	-	-
ISO/IEC Guide 99		International vocabulary of metrology - Basic and general concepts and associated terms (VIM)	-	-
CISPR 16-1-1	2010	Specification for radio disturbance and	EN 55016-1-1	2010
+ corr. October	2010	immunity measuring apparatus and	-	-
+ corr. October	2011	methods	-	-
+ A1	2010	Part 1-1: Radio disturbance and immunity measuring apparatus - Measuring apparatus	+ A1	2010

This page is intentionally left blank



IEC 61672-1

Edition 2.0 2013-09

INTERNATIONAL STANDARD

NORME INTERNATIONALE

**Electroacoustics – Sound level meters –
Part 1: Specifications**

**Electroacoustique – Sonomètres –
Partie 1: Spécifications**



THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2013 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester.

If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de la CEI ou du Comité national de la CEI du pays du demandeur.

Si vous avez des questions sur le copyright de la CEI ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de la CEI de votre pays de résidence.

IEC Central Office
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
Fax: +41 22 919 03 00
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigenda or an amendment might have been published.

Useful links:

IEC publications search - www.iec.ch/searchpub

The advanced search enables you to find IEC publications by a variety of criteria (reference number, text, technical committee,...).

It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available on-line and also once a month by email.

Electropedia - www.electropedia.org

The world's leading online dictionary of electronic and electrical terms containing more than 30 000 terms and definitions in English and French, with equivalent terms in additional languages. Also known as the International Electrotechnical Vocabulary (IEV) on-line.

Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: csc@iec.ch.

A propos de la CEI

La Commission Electrotechnique Internationale (CEI) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

A propos des publications CEI

Le contenu technique des publications de la CEI est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

Liens utiles:

Recherche de publications CEI - www.iec.ch/searchpub

La recherche avancée vous permet de trouver des publications CEI en utilisant différents critères (numéro de référence, texte, comité d'études,...).

Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

Just Published CEI - webstore.iec.ch/justpublished

Restez informé sur les nouvelles publications de la CEI. Just Published détaille les nouvelles publications parues. Disponible en ligne et aussi une fois par mois par email.

Electropedia - www.electropedia.org

Le premier dictionnaire en ligne au monde de termes électroniques et électriques. Il contient plus de 30 000 termes et définitions en anglais et en français, ainsi que les termes équivalents dans les langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (VEI) en ligne.

Service Clients - webstore.iec.ch/csc

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: csc@iec.ch.



IEC 61672-1

Edition 2.0 2013-09

INTERNATIONAL STANDARD

NORME INTERNATIONALE

**Electroacoustics – Sound level meters –
Part 1: Specifications**

**Electroacoustique – Sonomètres –
Partie 1: Spécifications**

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

PRICE CODE
CODE PRIX

XA

ICS 17.140.50

ISBN 978-2-8322-1087-1

**Warning! Make sure that you obtained this publication from an authorized distributor.
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

CONTENTS

FOREWORD.....	4
INTRODUCTION.....	6
1 Scope.....	7
2 Normative references	8
3 Terms and definitions	8
4 Reference environmental conditions	14
5 Performance specifications.....	14
5.1 General.....	14
5.2 Adjustments at the calibration check frequency	17
5.3 Corrections to indicated levels.....	17
5.3.1 General	17
5.3.2 Reflections and diffraction	17
5.3.3 Windscreens.....	18
5.3.4 Format for correction data	18
5.3.5 Corrections for use during periodic testing.....	19
5.4 Directional response.....	19
5.5 Frequency weightings.....	20
5.6 Level linearity.....	23
5.7 Self-generated noise	24
5.8 Time-weightings F and S	24
5.9 Toneburst response.....	24
5.10 Response to repeated tonebursts	26
5.11 Overload indication	27
5.12 Under-range indication	27
5.13 C-weighted peak sound level.....	27
5.14 Stability during continuous operation	28
5.15 High-level stability	28
5.16 Reset	29
5.17 Thresholds	29
5.18 Display.....	29
5.19 Analogue or digital output.....	29
5.20 Timing facilities	30
5.21 Radio frequency emissions and disturbances to a public power supply.....	30
5.22 Crosstalk.....	31
5.23 Power supply.....	31
6 Environmental, electrostatic, and radio-frequency requirements	32
6.1 General.....	32
6.2 Static pressure	32
6.3 Air temperature	32
6.4 Humidity.....	33
6.5 Electrostatic discharge	33
6.6 A.C. power-frequency and radio-frequency fields.....	33
6.7 Mechanical vibration.....	34
7 Provision for use with auxiliary devices	35
8 Marking	35
9 Instruction Manual	35

9.1	General	35
9.2	Information for operation	36
9.2.1	General	36
9.2.2	Design features	36
9.2.3	Power supply	37
9.2.4	Adjustments at the calibration check frequency	37
9.2.5	Corrections to indicated levels	37
9.2.6	Operating the sound level meter	37
9.2.7	Accessories	38
9.2.8	Influence of variations in environmental conditions	38
9.3	Information for testing	39
Annex A (informative)	Relationship between tolerance interval, corresponding acceptance interval and the maximum-permitted uncertainty of measurement	41
Annex B (normative)	Maximum-permitted uncertainties of measurement	42
Annex C (informative)	Example assessments of conformance to specifications of this standard	44
Annex D (normative)	Frequencies at fractional-octave intervals	47
Annex E (normative)	Analytical expressions for frequency-weightings C, A, and Z	49
Figure 1	– Principal steps involved in forming a time-weighted sound level	10
Figure A.1	– Relationship between tolerance interval, corresponding acceptance interval and the maximum-permitted uncertainty of measurement	41
Figure C.1	– Examples of assessment of conformance	46
Table 1	– Acceptance limits for the difference between a measured windscreen correction and the corresponding correction given in the Instruction Manual	18
Table 2	– Acceptance limits for deviations of directional response from the design goal	20
Table 3	– Frequency weightings and acceptance limits	22
Table 4	– Reference 4 kHz toneburst responses and acceptance limits	25
Table 5	– Reference differences for C-weighted peak sound levels and acceptance limits	28
Table 6	– Limits for conducted disturbance to the voltage of a public supply of electric power	31
Table B.1	– Maximum-permitted uncertainties of measurement for a coverage probability of 95 %	42
Table C.1	– Examples of assessment of conformance	45
Table D.1	– Frequencies at one-third-octave intervals	47
Table D.2	– Frequencies at one-sixth-octave intervals	48
Table D.3	– Frequencies at one-twelfth-octave intervals	48

INTERNATIONAL ELECTROTECHNICAL COMMISSION

ELECTROACOUSTICS – SOUND LEVEL METERS –

Part 1: Specifications

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as “IEC Publication(s)”). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 61672-1 has been prepared by IEC technical committee 29, Electroacoustics, in cooperation with the International Organization of Legal Metrology (OIML).

This second edition cancels and replaces the first edition published in 2002. This second edition constitutes a technical revision.

The main technical changes with respect to the previous edition are as follows:

In this second edition, conformance to specifications is demonstrated when:

- a) measured deviations from design goals do not exceed the applicable acceptance limits, and
- b) the uncertainty of measurement does not exceed the corresponding maximum-permitted uncertainty, with both uncertainties determined for a coverage probability of 95 %.

The text of this second edition is based on that of the first edition and the following documents:

FDIS	Report on voting
29/812/FDIS	29/823/RVD

Full information on the voting for the approval of this standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

A list of all parts of the IEC 61672 series, published under the general title *Electroacoustics – Sound level meters*, can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

INTRODUCTION

For assessments of conformance to performance specifications, this second edition of IEC 61672-1 uses different criteria than were used for the 2002 first edition.

In the period from 1961 to 1985, International Standards for sound level meters did not provide any requirements or recommendations to account for the uncertainty of measurement in assessments of conformance to specifications.

This absence of requirements or recommendations to account for uncertainty of measurement created ambiguity in determinations of conformance to specifications for situations where a measured deviation from a design goal was close to a limit of the allowed deviation. If conformance was determined based on whether a measured deviation did or did not exceed the limits, the end-user of the sound level meter incurred the risk that the true deviation from a design goal exceeded the limits.

To remove this ambiguity, IEC Technical Committee 29, at its meeting in 1996, adopted a policy to account for measurement uncertainty in assessments of conformance in International Standards that it prepares.

The first edition (2002) of IEC 61672-1 accounted for measurement uncertainty by giving two explicit criteria for determining conformance to the specifications. The two criteria were (a) that measured deviations from design goals, extended by the expanded uncertainty of measurement, do not exceed the applicable tolerance limits and (b) that the expanded uncertainty of measurement does not exceed agreed-upon maximum values. For most performance specifications, the tolerance limits were calculated essentially by extending the allowances for design and manufacturing from the 1979 and 1985 International Standards for sound level meters by the applicable maximum-permitted expanded uncertainties of measurement. Tolerance limits were intended to represent the limits for true deviations from design goals with a coverage probability of 95 %.

This second edition of IEC 61672-1 uses an amended criterion for assessing conformance to a specification. Conformance is demonstrated when (a) measured deviations from design goals do not exceed the applicable *acceptance limits* and (b) the uncertainty of measurement does not exceed the corresponding maximum-permitted uncertainty. Acceptance limits are analogous to the allowances for design and manufacturing implied in the first edition (2002) of IEC 61672-1. Actual and maximum-permitted uncertainties are determined for a coverage probability of 95 %. The amended criterion for assessing conformance does not necessitate any change to the design of a sound level meter in order to conform to the specifications of this International Standard.

The maximum-permitted uncertainties of measurement are not equivalent to the uncertainties associated with the measurement of a sound level. The uncertainty of a measured sound level is evaluated from the anticipated deviations of the electroacoustical performance of the sound level meter from the relevant design goals as well as estimates of the uncertainties associated with the specific measurement situation. Unless more-specific information is available, the evaluation of the contribution of a specific sound level meter to a total measurement uncertainty can be based on the acceptance limits and maximum-permitted uncertainties specified in this standard.

ELECTROACOUSTICS – SOUND LEVEL METERS –

Part 1: Specifications

1 Scope

This part of IEC 61672 gives electroacoustical performance specifications for three kinds of sound measuring instruments:

- a time-weighting sound level meter that measures exponential-time-weighted, frequency-weighted sound levels;
- an integrating-averaging sound level meter that measures time-averaged, frequency-weighted sound levels; and
- an integrating sound level meter that measures frequency-weighted sound exposure levels.

Sound level meters conforming to the requirements of this standard have a specified frequency response for sound incident on the microphone from one principal direction in an acoustic free field or successively from random directions.

Sound level meters specified in this standard are intended to measure sounds generally in the range of human hearing.

NOTE The AU frequency weighting specified in IEC 61012 can be applied for measurements of A-weighted sound levels of audible sound in the presence of a source that contains spectral components at frequencies greater than 20 kHz.¹

Two performance categories, class 1 and class 2, are specified in this standard. In general, specifications for class 1 and class 2 sound level meters have the same design goals and differ mainly in the acceptance limits and the range of operational temperature. Acceptance limits for class 2 are greater than, or equal to, those for class 1.

This standard is applicable to a range of designs for sound level meters. A sound level meter may be a self-contained hand-held instrument with an attached microphone and a built-in display device. A sound level meter may be comprised of separate components in one or more enclosures and may be capable of displaying a variety of acoustical signal levels. Sound level meters may include extensive analogue or digital signal processing, separately or in combination, with multiple analogue and digital outputs. Sound level meters may include general-purpose computers, recorders, printers, and other devices that form a necessary part of the complete instrument.

Sound level meters may be designed for use with an operator present or for automatic and continuous measurements of sound level without an operator present. Specifications in this standard for the response to sound waves apply without an operator present in the sound field.

¹ IEC 61012, *Filters for the measurement of audible sound in the presence of ultrasound*.

This is a free preview. Purchase the entire publication at the link below:

[Product Page](#)

-
- Looking for additional Standards? Visit Intertek Inform Infostore
 - Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation
-