



NSAI
Standards

Irish Standard
I.S. EN 62585:2012

Electroacoustics - Methods to determine corrections to obtain the free-field response of a sound level meter (IEC 62585:2012 (EQV))

I.S. EN 62585:2012

Incorporating amendments/corrigenda issued since publication:

The National Standards Authority of Ireland (NSAI) produces the following categories of formal documents:

I.S. xxx: Irish Standard – national specification based on the consensus of an expert panel and subject to public consultation.

S.R. xxx: Standard Recommendation - recommendation based on the consensus of an expert panel and subject to public consultation.

SWiFT xxx: A rapidly developed recommendatory document based on the consensus of the participants of an NSAI workshop.

<i>This document replaces:</i>	<i>This document is based on:</i> EN 62585:2012	<i>Published:</i> 5 October, 2012
This document was published under the authority of the NSAI and comes into effect on: 10 October, 2012		ICS number: 17.140.50
NSAI 1 Swift Square, Northwood, Santry Dublin 9	T +353 1 807 3800 F +353 1 807 3838 E standards@nsai.ie W NSAI.ie	Sales: T +353 1 857 6730 F +353 1 857 6729 W standards.ie
Údarás um Chaighdeáin Náisiúnta na hÉireann		

I.S. EN 62585:2012

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN 62585

October 2012

ICS 17.140.50

English version

**Electroacoustics -
Methods to determine corrections to obtain the free-field response
of a sound level meter
(IEC 62585:2012)**

Électroacoustique -
Méthode de détermination de corrections
pour obtenir la réponse en champ libre
d'un sonomètre
(CEI 62585:2012)

Elektroakustik -
Verfahren zur Ermittlung
von Korrekturwerten für die Bestimmung
des Freifeld-Frequenzgangs eines
Schallpegelmessers
(IEC 62585:2012)

This European Standard was approved by CENELEC on 2012-08-29. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

CENELEC

European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

Management Centre: Avenue Marnix 17, B - 1000 Brussels

Foreword

The text of document 29/770/FDIS, future edition 1 of IEC 62585, prepared by IEC/TC 29 "Electroacoustics" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN 62585:2012.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2013-05-29
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2015-08-29

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC [and/or CEN] shall not be held responsible for identifying any or all such patent rights.

Endorsement notice

The text of the International Standard IEC 62585:2012 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following notes have to be added for the standards indicated:

IEC 61094-8	NOTE	Harmonised as EN 61094-8.
IEC 61260	NOTE	Harmonised as EN 61260.
IEC 61094-2	NOTE	Harmonised as EN 61094-2.
IEC 61094-3	NOTE	Harmonised as EN 61094-3.

Annex ZA (normative)

Normative references to international publications with their corresponding European publications

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE When an international publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 60942	-	Electroacoustics - Sound calibrators	EN 60942	-
IEC 61094-1	-	Measurement microphones - Part 1: Specifications for laboratory standard microphones	EN 61094-1	-
IEC 61094-5	-	Measurement microphones - Part 5: Methods for pressure calibration of working standard microphones by comparison	EN 61094-5	-
IEC 61094-6	-	Measurement microphones - Part 6: Electrostatic actuators for determination of frequency response	EN 61094-6	-
IEC/TS 61094-7	-	Measurement microphones - Part 7: Values for the difference between free-field and pressure sensitivity levels of laboratory standard microphones	-	-
IEC 61183	-	Electroacoustics - Random-incidence and diffuse-field calibration of sound level meters	EN 61183	-
IEC 61672-1	-	Electroacoustics - Sound level meters - Part 1: Specifications	EN 61672-1	-
IEC 61672-2	-	Electroacoustics - Sound level meters - Part 2: Pattern evaluation tests	EN 61672-2	-
IEC 61672-3	-	Electroacoustics - Sound level meters - Part 3: Periodic tests	EN 61672-3	-
ISO/IEC Guide 98-3	-	Uncertainty of measurement - Part 3: Guide to the expression of uncertainty in measurement (GUM)	-	-
ISO/IEC Guide 99	-	International vocabulary of metrology - Basic - and general concepts and associated terms (VIM)		-

This page is intentionally left BLANK.

CONTENTS

FOREWORD.....	4
INTRODUCTION.....	6
1 Scope.....	7
2 Normative references	8
3 Terms and definitions	8
4 Reference environmental conditions	9
5 Uncertainties of measurement	9
6 General requirements for measurement of adjustment value and corrections.....	10
7 Submission and inspection	12
8 Determination of the adjustment value at the calibration check frequency.....	12
9 Determination of corrections for the typical effects of reflections from the case of the sound level meter and diffraction of sound around the microphone	13
10 Determination of corrections for the deviation of the microphone free-field frequency response from a uniform frequency response	13
11 Determination of corrections for the use of windscreens and similar accessories.....	14
12 Determination of corrections over a range of frequencies when a sound calibrator is recommended for use by the sound level meter manufacturer.....	14
13 Determination of corrections over a range of frequencies when a comparison coupler is recommended for use by the sound level meter manufacturer	15
14 Determination of corrections over a range of frequencies when an electrostatic actuator is recommended for use by the sound level meter manufacturer	16
15 Documentation	16
Annex A (normative) Calibration check frequency – Determination of the adjustment value	18
Annex B (normative) Determination of the effects of reflections from the case of the sound level meter and diffraction of sound around the microphone	20
Annex C (normative) Determination of the effects of the use of windscreens and similar accessories	21
Annex D (informative) Determination of corrections for use with sound calibrators to determine equivalent free-field response at all frequencies of interest	22
Annex E (informative) Determination of corrections for use with comparison couplers to determine equivalent free-field frequency response at all frequencies of interest	25
Annex F (informative) Determination of corrections for use with electrostatic actuators to determine equivalent free-field frequency response at all frequencies of interest	28
Annex G (informative) Comparison techniques.....	31
Annex H (informative) Exact one-twelfth-octave frequencies	34
Annex I (informative) Example calculations of expanded uncertainty of measurement.....	36
Bibliography.....	40
Figure 1 – Mounting methods for the sound level meter	11
Figure A.1 – Diagrammatic representation of factors for consideration in optimising adjustment value at the calibration check frequency	19
Figure D.1 – Measurement steps using a sound calibrator	22
Figure E.1 – Measurement steps using a comparison coupler	25

I.S. EN 62585:2012

62585 © IEC:2012

– 3 –

Figure F.1 – Measurement steps using an electrostatic actuator	28
Table H.1 – Exact frequencies for one-twelfth-octave steps over one decade	35
Table I.1 – Description of likely uncertainty components	37
Table I.2 – Uncertainty example for a frequency of 1 kHz	38
Table I.3 – Uncertainty example for a frequency of 8 kHz	39

INTERNATIONAL ELECTROTECHNICAL COMMISSION

**ELECTROACOUSTICS –
METHODS TO DETERMINE CORRECTIONS TO OBTAIN
THE FREE-FIELD RESPONSE OF A SOUND LEVEL METER**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 62585 has been prepared by IEC technical committee 29: Electroacoustics.

The text of this International Standard is based on the following documents:

FDIS	Report on voting
29/770/FDIS	29/782/RVD

Full information on the voting for the approval of this International Standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

I.S. EN 62585:2012

62585 © IEC:2012

– 5 –

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

INTRODUCTION

The performance specification International Standard for sound level meters, IEC 61672-1, requires that at least one model of sound calibrator, conforming to the requirements of IEC 60942 be specified in the instruction manual for checking and maintaining the correct indication on the display of the sound level meter at the calibration check frequency. The sound level meter manufacturer specifies the adjustment value to be applied to obtain the required indication on the display in response to the sound pressure level generated by the sound calibrator, in order to optimize performance over the complete frequency range.

In addition, various corrections need to be available over a range of frequencies in order for a periodic test of a sound level meter to be performed according to IEC 61672-3. For example, corrections are needed for any effects of the sound level meter case or of accessories such as windscreens on the equivalent free-field sound level. Information on these corrections is also required by users of sound level meters and sound calibrators on a regular basis.

Also, a manufacturer producing a sound level meter to the specifications of IEC 61672-1, may recommend, in the instruction manual, the use of a sound calibrator, comparison coupler or electrostatic actuator to determine the acoustical response of a sound level meter at various frequencies. In this case the manufacturer is required to provide corrections to obtain equivalent sound levels that would be displayed under reference environmental conditions in response to plane progressive sinusoidal waves that are incident from the reference direction at each frequency used for periodic testing. These corrections will either be given in the instruction manual, or the instruction manual will state where they can be found.

ELECTROACOUSTICS – METHODS TO DETERMINE CORRECTIONS TO OBTAIN THE FREE-FIELD RESPONSE OF A SOUND LEVEL METER

1 Scope

This International Standard provides information on the corrections required over a range of frequencies in order for a periodic test of a sound level meter to be performed according to IEC 61672-3. These corrections include:

- corrections for the typical effects of reflections from the case of the sound level meter and diffraction of sound around the microphone;
- corrections for the deviation of the typical microphone frequency response from a uniform frequency response, where the actual microphone response cannot be measured;
- corrections for the influence on the frequency response of a typical microphone of a specified windscreen and any other accessory that is part of the configuration for normal use of the particular sound level meter submitted for testing.

This International Standard includes discussion about uncertainties of measurement of the required corrections. In some instances a maximum permitted expanded uncertainty for the manufacturer or testing laboratory is given. This maximum permitted expanded uncertainty excludes any component due to the variability of different samples of artefact (for example, microphone or windscreen). It should be noted that if large uncertainties of measurement are quoted for each of the individual corrections, when they are combined to account for the configuration of sound level meter under test, the large individual uncertainties may result in a failure to conform to the maximum permitted expanded uncertainties of measurement given in Table A.1 of IEC 61672-1:—¹ and hence a failure of the sound level meter to conform to IEC 61672-1

In addition, this International Standard describes methods for determining these corrections, over the frequency range of interest, and explains the adjustment value at the calibration check frequency to be quoted by the manufacturer of the sound level meter (also required by IEC 61672-3).

When the sound level meter manufacturer recommends the use of a sound calibrator, comparison coupler, or an electrostatic actuator for periodic testing of the acoustical response of a sound level meter at various frequencies, this International Standard describes methods of measurement of the corrections required to adjust the indication on the sound level meter to an equivalent free-field level, over the frequency range of interest. These corrections relate to a specific model of sound calibrator, comparison coupler or electrostatic actuator, microphone and sound level meter (also required by IEC 61672-3).

The aim of this International Standard is to ensure that the adjustment value at the calibration check frequency and all corrections are determined using consistent and appropriate methods.

It is intended that this International Standard will be used by manufacturers to determine adjustment values and corrections, by laboratories performing pattern evaluation tests according to IEC 61672-2, and by laboratories performing periodic tests according to IEC 61672-3. Laboratories performing periodic tests according to IEC 61672-3 will also need to consult this International Standard to ensure that the expanded uncertainties of

¹ Second edition to be published. (A revision of 61672-1:2002.)

measurement for the corrections quoted by the manufacturer do not exceed the maximum permitted values.

The corrections obtained by use of the methods given in this International Standard are the result of measurements made using samples of the devices. It is possible that these corrections may not be totally representative either for all batches produced or over time. Repeating the measurements at regular intervals is recommended to ensure that no changes are required to the corrections stated in the instruction manual.

This International Standard does not specifically cover the case where the sound level meter is fitted with a microphone intended for use in random-incidence sound fields, as information is given in IEC 61183.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60942, *Electroacoustics – Sound calibrators*

IEC 61094-1, *Measurement microphones – Part 1: Specifications for laboratory standard microphones*

IEC 61094-5, *Measurement microphones – Part 5: Methods for pressure calibration of working standard microphones by comparison*

IEC 61094-6, *Measurement microphones – Part 6: Electrostatic actuators for determination of frequency response*

IEC/TS 61094-7, *Measurement microphones – Part 7: Values for the difference between free-field and pressure sensitivity levels of laboratory standard microphones*

IEC 61183, *Electroacoustics-Random-incidence and diffuse-field calibration of sound level meters*

IEC 61672-1: —², *Electroacoustics – Sound level meters – Part 1: Specifications*

IEC 61672-2, *Electroacoustics – Sound level meters – Part 2: Pattern evaluation tests*

IEC 61672-3, *Electroacoustics – Sound level meters – Part 3: Periodic tests*

ISO/IEC Guide 98-3, *Uncertainty of measurement – Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)*

ISO/IEC Guide 99, *International vocabulary of metrology – Basic and general concepts and associated terms (VIM)*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in the ISO/IEC Guide 99, as well as the following apply.

² Second edition to be published. (A revision of 61672-1:2002.)

This is a free preview. Purchase the entire publication at the link below:

[Product Page](#)

-
- Looking for additional Standards? Visit Intertek Inform Infostore
 - Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation
-