

Irish Standard I.S. EN 14180:2014

Sterilizers for medical purposes - Low temperature steam and formaldehyde sterilizers - Requirements and testing

© CEN 2014 No copying without NSAI permission except as permitted by copyright law.

I.S. EN 14180:2014

Incorporating amendments/corrigenda/National Annexes issued since publication:

The National Standards Authority of Ireland (NSAI) produces the following categories of formal documents:

I.S. xxx: Irish Standard – national specification based on the consensus of an expert panel and subject to public consultation.

S.R. xxx: Standard Recommendation — recommendation based on the consensus of an expert panel and subject to public consultation.

SWIFT xxx: A rapidly developed recommendatory document based on the consensus of the participants of an NSAI workshop.

This document replaces/revises/consolidates the NSAI adoption of the document(s) indicated on the CEN/CENELEC cover/Foreword and the following National document(s):

NOTE: The date of any NSAI previous adoption may not match the date of its original CEN/CENELEC document.

This document is based on: EN 14180:2014

Published: 2014-05-21

This document was published		ICS number:
and comes into effect on:		11.080.10
2014-06-07		
		NOTE: If blank see CEN/CENELEC cover page
NSAI	T +353 1	. 807 3800 Sales:
1 Swift Square,	F +353 1	807 3838 T +353 1 857 6730
Northwood, Santry	E standa	ards@nsai.ie F +353 1 857 6729
Dublin 9	W NSAI.i	ie W standards.ie
Úda	rás um Cha	ighdeáin Náisiúnta na hÉireann

This is a free page sample. Access the full version online. I.S. EN 14180:2014

EUROPEAN STANDARD

EN 14180

NORME EUROPÉENNE

EUROPÄISCHE NORM

May 2014

ICS 11.080.10

Supersedes EN 14180:2003+A2:2009

English Version

Sterilizers for medical purposes - Low temperature steam and formaldehyde sterilizers - Requirements and testing

Stérilisateurs à usage médical - Stérilisateurs à la vapeur et au formaldéhyde à basse température - Exigences et essais Sterilisatoren für medizinische Zwecke - Niedertemperatur-Dampf-Formaldehyd-Sterilisatoren - Anforderungen und Prüfung

This European Standard was approved by CEN on 10 April 2014.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION COMITÉ EUROPÉEN DE NORMALISATION EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

Ref. No. EN 14180:2014 E

EN 14180:2014 (E)

Contents

Page

Forewo	ord	. 3
Introdu	iction	. 4
1	Scope	. 5
2	Normative references	. 5
3	Terms and definitions	. 6
4	Technical requirements	11
4.1	Sterilizer chamber	11
4.2	Design and construction	14
4.3	Indicating, measuring, operating and recording devices	16
5	Process control	22
5.1	General	22
5.2	Software verification and validation	22
5.3	Operating cycle and automatic control	23
5.4	Override of automatic control	24
5.5	Fault	24
6	Performance requirements	25
6.1	Sterilizing performance	25
6.2	Desorption efficacy	27
6.3	Drying	27
7	Sound power and vibration	27
8	Packaging, marking and labelling	28
9	Information to be supplied by the manufacturer	29
10	Service and local environment	31
10.1	General	31
10.2	Electricity	31
10.3	Sterilant	31
10.4	Steam	32
10.5	Water	32
10.6	Compressed air	32
10.7	Drainage and discharges	32
10.8	Ventilation and environment	33
10.9	Lighting	33
Annex	A (normative) Test methids	34
Annex	B (normative) Sterilizer classification and testing	40
Annex	C (normative) Test equipment	43
Annex	D (normative) Determination of formaldehyde residuals in a filter indicator	46
Annex	E (informative) Formaldehyde residues on medical devices	49
Annex	F (informative) Environmental aspects	51
Annex	ZA (informative) Relationship between this European Standard and the Essential Requirement	S
	of EU Directive 93/42/EEC on medical devices	54
Bibliog	jraphy	60

Foreword

This document (EN 14180:2014) has been prepared by Technical Committee CEN/TC 102 "Sterilizers for medical purposes", the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by November 2014 and conflicting national standards shall be withdrawn at the latest by November 2014.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN 14180:2003+A2:2009.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For relationship with EU Directive(s), see informative Annex ZA, which is an integral part of this document.

The standard is a full technical revision of the previous version. The following amendments have been made in comparison to EN 14180:2003+A2:2009:

- normative references were updated;
- terms risk assessment, risk analysis and software validation were added;
- align biological testing with method from EN ISO 25424;
- requirements for heat isolation were updated;
- safety requirements, mainly as a consequence of compliance with the machinery directive were added;
- requirements and testing for sound power, also including vibration, were updated;
- Annex ZA including Tables ZA1 and ZA2 were updated.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Introduction

This European Standard specifies minimum requirements and test methods for sterilizers working below ambient atmospheric pressure performing a low temperature steam and formaldehyde (LTSF) process.

LTSF sterilizers are primarily used for the sterilization of medical devices in health care facilities, but can also be used during the commercial production of medical devices.

LTSF processes are specified by physical parameters and verified using physical, chemical and microbiological means [8]. The sterilizers operate automatically using pre-set cycles.

The test methods and test equipment given could also be applicable to validation and routine control.

Validation and routine control of sterilization processes are essential to ensure their efficacy. This standard does not cover validation and routine control of a LTSF process. Criteria for validation and routine control of LTSF sterilization processes are given in EN ISO 25424.

At the present state of knowledge, LTSF sterilizers should not be assumed to deliver processes effectively inactivating the causative agents of spongiform encephalopathies such as scrapie, Bovine Spongiform Encephalopathy and Creutzfeld-Jakob Disease. Specific recommendations have been produced in particular countries for the processing of materials potentially contaminated with these agents. See also EN ISO 25424:2011, 1.2.1.

Planning and design of products applying to this standard should consider not only technical issues but also the environmental impact from the product during its life-cycle. Environmental aspects are addressed in Annex F of this standard.

NOTE Specifications on operator safety are addressed in EN 61010–1, EN 61010–2–040 and are not repeated in this standard. EN 60204–1 can also give valuable guidelines.

1 Scope

This European Standard specifies requirements and tests for LTSF sterilizers, which use a mixture of low temperature steam and formaldehyde as sterilizing agent, and which are working below ambient pressure only.

These sterilizers are primarily used for the sterilization of heat labile medical devices in health care facilities.

This European Standard specifies minimum requirements:

- for the performance and design of sterilizers to ensure that the process is capable of sterilizing medical devices;
- for the equipment and controls of these sterilizers necessary for the validation and routine control of the sterilization processes.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 764-7, Pressure equipment - Part 7: Safety systems for unfired pressure equipment

EN 867–5, Non-biological systems for use in sterilizers - Part 5: Specification for indicator systems and process challenge devices for use in performance testing for small sterilizers Type B and Type S

EN 868–5, Packaging for terminally sterilized medical devices - Part 5: Sealable pouches and reels of porous materials and plastic film construction - Requirements and test methods

EN 14222:2003, Stainless steel shell boilers

EN 60584-2, Thermocouples - Part 2: Tolerances

EN 60751:2008, Industrial platinum resistance thermometers and platinum temperature sensors (IEC 60751:2008)

EN 61010–1:2010, Safety requirements for electrical equipment for measurement, control and laboratory use — Part 1: General requirements (IEC 61010-1:2010)

EN 61010-2-040:2005, Safety requirements for electrical equipment for measurement, control, and laboratory use — Part 2-040: Particular requirements for sterilizers and washer-disinfectors used to treat medical materials (IEC 61010-2-040:2005)

EN 61326-1:2013, Electrical equipment for measurement, control and laboratory use — EMC requirements — Part 1: General requirements (IEC 61326-1:2012)

EN ISO 228-1:2003, Pipe threads where pressure-tight joints are not made on the threads - Part 1: Dimensions, tolerances and designation (ISO 228-1:2000)

EN ISO 1874-1, Plastics - Polyamide (PA) moulding and extrusion materials - Part 1: Designation system and basis for specification (ISO 1874-1)

EN ISO 3746:2010, Acoustics - Determination of sound power levels and sound energy levels of noise sources using sound pressure - Survey method using an enveloping measurement surface over a reflecting plane (ISO 3746:2010)



This is a free preview. Purchase the entire publication at the link below:

Product Page

S Looking for additional Standards? Visit Intertek Inform Infostore

> Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation