



**NSAI**  
Standards

Irish Standard  
I.S. EN ISO 5840-1:2015

# Cardiovascular implants - Cardiac valve prostheses - Part 1: General requirements (ISO 5840-1:2015)

© CEN 2015 No copying without NSAI permission except as permitted by copyright law.

## I.S. EN ISO 5840-1:2015

*Incorporating amendments/corrigenda/National Annexes issued since publication:*

The National Standards Authority of Ireland (NSAI) produces the following categories of formal documents:

I.S. xxx: Irish Standard — national specification based on the consensus of an expert panel and subject to public consultation.

S.R. xxx: Standard Recommendation — recommendation based on the consensus of an expert panel and subject to public consultation.

SWiFT xxx: A rapidly developed recommendatory document based on the consensus of the participants of an NSAI workshop.

*This document replaces/revises/consolidates the NSAI adoption of the document(s) indicated on the CEN/CENELEC cover/Foreword and the following National document(s):*

*NOTE: The date of any NSAI previous adoption may not match the date of its original CEN/CENELEC document.*

*This document is based on:*

EN ISO 5840-1:2015

*Published:*

2015-09-30

*This document was published  
under the authority of the NSAI  
and comes into effect on:*

2015-10-22

ICS number:

11.040.40

NOTE: If blank see CEN/CENELEC cover page

NSAI  
1 Swift Square,  
Northwood, Santry  
Dublin 9

T +353 1 807 3800  
F +353 1 807 3838  
E standards@nsai.ie  
W NSAI.ie

Sales:  
T +353 1 857 6730  
F +353 1 857 6729  
W standards.ie

Údarás um Chaighdeáin Náisiúnta na hÉireann

## National Foreword

I.S. EN ISO 5840-1:2015 is the adopted Irish version of the European Document EN ISO 5840-1:2015, Cardiovascular implants - Cardiac valve prostheses - Part 1: General requirements (ISO 5840-1:2015)

This document does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

**Compliance with this document does not of itself confer immunity from legal obligations.**

*In line with international standards practice the decimal point is shown as a comma (,) throughout this document.*

This page is intentionally left blank

**EUROPEAN STANDARD**

**EN ISO 5840-1**

**NORME EUROPÉENNE**

**EUROPÄISCHE NORM**

September 2015

ICS 11.040.40

English Version

## **Cardiovascular implants - Cardiac valve prostheses - Part 1: General requirements (ISO 5840-1:2015)**

Implants cardiovasculaires - Prothèses valvulaires -  
Partie 1: Exigences générales (ISO 5840-1:2015)

Herz- und Gefäßimplantate - Herzklappenprothesen -  
Teil 1: Allgemeine Anforderungen (ISO 5840-1:2015)

This European Standard was approved by CEN on 10 July 2015.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels**

**EN ISO 5840-1:2015 (E)**

**Contents**

Page

<b>European foreword.....</b>	<b>3</b>
<b>Annex ZA (informative) Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC on medical devices .....</b>	<b>4</b>

## **European foreword**

This document (EN ISO 5840-1:2015) has been prepared by Technical Committee ISO/TC 150 “Implants for surgery” in collaboration with Technical Committee CEN/TC 285 “Non-active surgical implants” the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by March 2016, and conflicting national standards shall be withdrawn at the latest by March 2016.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For relationship with EU Directive(s), see informative Annex ZA, which is an integral part of this document.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

### **Endorsement notice**

The text of ISO 5840-1:2015 has been approved by CEN as EN ISO 5840-1:2015 without any modification.

## Annex ZA (informative)

### Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC on medical devices

This European Standard has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association to provide a means of conforming to Essential Requirements of the New Approach Directive 93/42/EEC on medical devices.

Once this standard is cited in the Official Journal of the European Union under that Directive and has been implemented as a national standard in at least one Member State, compliance with the clauses of this standard given in Table ZA.1 confers, within the limits of the scope of this standard, a presumption of conformity with the corresponding Essential Requirements of that Directive and associated EFTA regulations.

**Table ZA.1 — Correspondence between this European Standard  
and Directive 93/42/EEC on medical devices**

Clause(s)/sub-clause(s) of this European Standard	Essential Requirements (ERs) of Directive 93/42/EEC	Qualifying remarks/Notes
5, 6, 7	7.1	
5, 6.2.4, 6.4, 6.5	7.2	
6.2.2, 6.2.3, 7.3	7.3	
6.5	7.5	
6.2.2, 6.5	7.6	
5, 6.4, 6.5	8.1	
6.2.4, 6.4	8.3	
6.2.4	8.4	
6.4	8.5	
6.2.4	8.6	
6.2.4	8.7	
6.2.1, 6.3, 7	9.1	
6.2.1, 6.3, 6.4, 6.5, 7	9.2, 1. indent	
6.4, 6.5, 7	9.2, 2. indent	
6.4, 6.5, 7	9.2, 3. indent	
6.2.1, 6.3, 6.4, 6.5, 7	9.2, 4. indent	
6.2.4	13	

**WARNING** — Other requirements and other EU Directives may be applicable to the product(s) falling within the scope of this standard.



# INTERNATIONAL STANDARD

**ISO  
5840-1**

First edition  
2015-09-15

---

---

## **Cardiovascular implants — Cardiac valve prostheses —**

### **Part 1: General requirements**

*Implants cardiovasculaires — Prothèses valvulaires —*

*Partie 1: Exigences générales*



Reference number  
ISO 5840-1:2015(E)

© ISO 2015

**ISO 5840-1:2015(E)**



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2015, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Ch. de Blandonnet 8 • CP 401  
CH-1214 Vernier, Geneva, Switzerland  
Tel. +41 22 749 01 11  
Fax +41 22 749 09 47  
copyright@iso.org  
www.iso.org

# Contents

Page

Foreword .....	iv
Introduction .....	v
<b>1 Scope .....</b>	<b>1</b>
<b>2 Normative references .....</b>	<b>1</b>
<b>3 Terms and definitions .....</b>	<b>2</b>
<b>4 Abbreviations .....</b>	<b>11</b>
<b>5 Fundamental requirements .....</b>	<b>12</b>
<b>6 Device description .....</b>	<b>12</b>
6.1 Intended use .....	12
6.2 Design inputs .....	12
6.2.1 Operational specifications .....	12
6.2.2 Performance specifications .....	12
6.2.3 Implant procedure .....	12
6.2.4 Packaging, labelling, and sterilization .....	13
6.3 Design outputs .....	13
6.4 Design transfer (manufacturing verification/validation) .....	14
6.5 Risk management .....	14
<b>7 Design verification testing and analysis/design validation .....</b>	<b>15</b>
7.1 General requirements .....	15
7.2 <i>In vitro</i> assessment .....	15
7.3 Preclinical <i>in vivo</i> evaluation .....	15
7.4 Clinical investigations .....	15
<b>Annex A (informative) Rationale for the provisions of this part of ISO 5840 .....</b>	<b>16</b>
<b>Annex B (normative) Packaging .....</b>	<b>19</b>
<b>Annex C (normative) Product labels, instructions for use, and training .....</b>	<b>20</b>
<b>Annex D (normative) Sterilization .....</b>	<b>23</b>
<b>Annex E (informative) <i>In vitro</i> test guidelines for paediatric devices .....</b>	<b>24</b>
<b>Annex F (informative) Statistical procedures when using <i>in vitro</i> performance criteria .....</b>	<b>28</b>
<b>Annex G (informative) Examples and definitions of some physical and material properties of heart valve systems .....</b>	<b>29</b>
<b>Annex H (informative) Examples of standards applicable to testing of materials and components of heart valve systems .....</b>	<b>40</b>
<b>Annex I (informative) Raw and post-conditioning mechanical properties for support structure materials .....</b>	<b>46</b>
<b>Annex J (informative) Corrosion assessment .....</b>	<b>48</b>
<b>Annex K (informative) Echocardiographic protocol .....</b>	<b>51</b>
<b>Bibliography .....</b>	<b>54</b>

## ISO 5840-1:2015(E)

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: [Foreword - Supplementary Information](#)

The committee responsible for this document is ISO/TC 150, *Implants for surgery*, Subcommittee SC 2, *Cardiovascular implants and extracorporeal systems*.

This first edition of ISO 5840-1, together with ISO 5840-2 and ISO 5840-3, cancels and replaces ISO 5840:2005, which has been technically revised.

ISO 5840 consists of the following parts, under the general title *Cardiovascular implants — Cardiac valve prostheses*:

- *Part 1: General requirements*
- *Part 2: Surgically implanted heart valve substitutes*
- *Part 3: Heart valve substitutes implanted by transcatheter techniques*

## Introduction

There is, as yet, no heart valve substitute which can be regarded as ideal.

The ISO 5840-series has been prepared by a group well aware of the issues associated with heart valve substitutes and their development. In several areas, the provisions of the ISO 5840-series deliberately have not been specified to encourage development and innovation. It does specify the types of tests, test methods, and/or requirements for test apparatus and requires documentation of test methods and results. The areas with which the ISO 5840-series are concerned are those which will ensure that associated risks to the patient and other users of the device have been adequately mitigated, facilitate quality assurance, aid the clinician in choosing a heart valve substitute, and ensure that the device will be presented at the operating table in convenient form. Emphasis has been placed on specifying types of *in vitro* testing, on preclinical *in vivo* and clinical evaluations, on reporting of all *in vitro*, preclinical *in vivo*, and clinical evaluations, and on the labelling and packaging of the device. Such a process involving *in vitro*, preclinical *in vivo*, and clinical evaluations is intended to clarify the required procedures prior to market release and to enable prompt identification and management of any subsequent problems.

With regard to *in vitro* testing and reporting, apart from basic material testing for mechanical, physical, chemical, and biocompatibility characteristics, the ISO 5840-series also covers important hydrodynamic and durability characteristics of heart valve substitutes. The ISO 5840-series does not specify exact test methods for hydrodynamic and durability testing, but it offers guidelines for the test apparatus.

The ISO 5840-series is incomplete in several areas. It is intended to be revised, updated, and/or amended as knowledge and techniques in heart valve substitute technology improve.



# Cardiovascular implants — Cardiac valve prostheses —

## Part 1: General requirements

### 1 Scope

This part of ISO 5840 is applicable to heart valve substitutes intended for human implantation and provides general requirements. Subsequent parts of the ISO 5840-series provide specific requirements.

This part of ISO 5840 is applicable to both newly developed and modified heart valve substitutes and to the accessories, packaging, and labelling required for their implantation and for determining the appropriate size of the heart valve substitute to be implanted.

This part of ISO 5840 outlines an approach for qualifying the design and manufacture of a heart valve substitute through risk management. The selection of appropriate qualification tests and methods are derived from the risk assessment. The tests may include those to assess the physical, chemical, biological, and mechanical properties of heart valve substitutes and of their materials and components. The tests may also include those for preclinical *in vivo* evaluation and clinical evaluation of the finished heart valve substitute.

This part of ISO 5840 defines operational conditions for heart valve substitutes.

This part of ISO 5840 excludes homografts.

NOTE A rationale for the provisions of this part of ISO 5840 is given in [Annex A](#).

### 2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 5840-2, *Cardiovascular implants — Cardiac valve prostheses — Part 2: Surgically implanted heart valve substitutes*

ISO 5840-3, *Cardiovascular implants — Cardiac valve prostheses — Part 3: Heart valve substitutes implanted by transcatheter techniques*

ISO 11135, *Sterilization of health-care products — Ethylene oxide — Requirements for the development, validation and routine control of a sterilization process for medical devices*

ISO 11137 (all parts), *Sterilization of health care products — Radiation*

ISO 11607 (all parts), *Packaging for terminally sterilized medical devices*

ISO 14155, *Clinical investigation of medical devices for human subjects — Good clinical practice*

ISO 14160, *Sterilization of health care products — Liquid chemical sterilizing agents for single-use medical devices utilizing animal tissues and their derivatives — Requirements for characterization, development, validation and routine control of a sterilization process for medical devices*

ISO 14630:2012, *Non-active surgical implants — General requirements*

This is a free preview. Purchase the entire publication at the link below:

[Product Page](#)

- 
- Looking for additional Standards? Visit Intertek Inform Infostore
  - Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation
-