



NSAI
Standards

Irish Standard
I.S. EN ISO 23145-2:2016

Fine ceramics (advanced ceramics, advanced technical ceramics) - Determination of bulk density of ceramic powders - Part 2:
Untapped density (ISO 23145-2:2012)

I.S. EN ISO 23145-2:2016

Incorporating amendments/corrigenda/National Annexes issued since publication:

The National Standards Authority of Ireland (NSAI) produces the following categories of formal documents:

I.S. xxx: Irish Standard — national specification based on the consensus of an expert panel and subject to public consultation.

S.R. xxx: Standard Recommendation — recommendation based on the consensus of an expert panel and subject to public consultation.

SWiFT xxx: A rapidly developed recommendatory document based on the consensus of the participants of an NSAI workshop.

This document replaces/revises/consolidates the NSAI adoption of the document(s) indicated on the CEN/CENELEC cover/Foreword and the following National document(s):

NOTE: The date of any NSAI previous adoption may not match the date of its original CEN/CENELEC document.

This document is based on:

EN ISO 23145-2:2016

Published:

2016-04-20

*This document was published
under the authority of the NSAI
and comes into effect on:*

2016-05-08

ICS number:

81.060.30

NOTE: If blank see CEN/CENELEC cover page

NSAI
1 Swift Square,
Northwood, Santry
Dublin 9

T +353 1 807 3800
F +353 1 807 3838
E standards@nsai.ie
W NSAI.ie

Sales:
T +353 1 857 6730
F +353 1 857 6729
W standards.ie

Údarás um Chaighdeáin Náisiúnta na hÉireann

National Foreword

I.S. EN ISO 23145-2:2016 is the adopted Irish version of the European Document EN ISO 23145-2:2016, Fine ceramics (advanced ceramics, advanced technical ceramics) - Determination of bulk density of ceramic powders - Part 2: Untapped density (ISO 23145-2:2012)

This document does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

Compliance with this document does not of itself confer immunity from legal obligations.

In line with international standards practice the decimal point is shown as a comma (,) throughout this document.

This page is intentionally left blank

EUROPEAN STANDARD

EN ISO 23145-2

NORME EUROPÉENNE

EUROPÄISCHE NORM

April 2016

ICS 81.060.30

Supersedes EN 725-9:2006

English Version

Fine ceramics (advanced ceramics, advanced technical ceramics) - Determination of bulk density of ceramic powders - Part 2: Untapped density (ISO 23145-2:2012)

Céramiques techniques - Détermination de la masse volumique des poudres céramiques - Partie 2: Masse volumique sans tassement (ISO 23145-2:2012)

Hochleistungskeramik - Bestimmung der Dichte von keramischen Pulvern - Teil 2: Schüttdichte (ISO 23145-2:2012)

This European Standard was approved by CEN on 25 March 2016.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

EN ISO 23145-2:2016 (E)

Contents	Page
European foreword.....	3

European foreword

The text of ISO 23145-2:2012 has been prepared by Technical Committee ISO/TC 206 “Fine ceramics” of the International Organization for Standardization (ISO) and has been taken over as EN ISO 23145-2:2016 by Technical Committee CEN/TC 184 “Advanced technical ceramics” the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by October 2016, and conflicting national standards shall be withdrawn at the latest by October 2016.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN 725-9:2006.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 23145-2:2012 has been approved by CEN as EN ISO 23145-2:2016 without any modification.

This page is intentionally left blank

INTERNATIONAL STANDARD

**ISO
23145-2**

First edition
2012-06-01

Fine ceramics (advanced ceramics, advanced technical ceramics) — Determination of bulk density of ceramic powders —

Part 2: Untapped density

*Céramiques techniques — Détermination de la masse volumique des
poudres céramiques —*

Partie 2: Masse volumique sans tassement



Reference number
ISO 23145-2:2012(E)

© ISO 2012

ISO 23145-2:2012(E)



COPYRIGHT PROTECTED DOCUMENT

© ISO 2012

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
1 Scope	1
2 Normative references	1
3 Principle	1
4 Apparatus	1
5 Sampling	4
6 Procedure	5
7 Calculation	5
8 Test report	5
Bibliography	7

ISO 23145-2:2012(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 23145-2 was prepared by Technical Committee ISO/TC 206, *Fine ceramics*.

ISO 23145 consists of the following parts, under the general title *Fine ceramics (advanced ceramics, advanced technical ceramics)* — *Determination of bulk density of ceramic powders*:

- *Part 1: Tap density*
- *Part 2: Untapped density*

Fine ceramics (advanced ceramics, advanced technical ceramics) — Determination of bulk density of ceramic powders —

Part 2: Untapped density

1 Scope

This part of ISO 23145 specifies the test method to determine the untapped density of granulated or ungranulated ceramic powders by a constant-volume measuring method.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 565:1990, *Test sieves — Metal wire cloth, perforated metal plate and electroformed sheet — Nominal sizes of openings*

ISO 80000-1:2009, *Quantities and units — Part 1: General*

ISO/IEC 17025:2005, *General requirements for the competence of testing and calibration laboratories*

3 Principle

The mass of a known volume of the powder is determined after allowing it to fall freely into a stationary container, avoiding vibration. The mass of the powder divided by its volume after the test gives its untapped density.

4 Apparatus

4.1 Funnel (Figure 1), of stainless steel, with an orifice of diameter 2,5 mm and the other with an orifice of diameter 5,0 mm. The funnel shall be made of a non-magnetic, corrosion-resistant metallic material such as stainless steel (for example SUS 304) having sufficient wall thickness and hardness to withstand distortion and excessive wear.

4.2 Cylindrical container (Figure 2), of stainless steel, with a volume of 100 cm³ and a diameter-to-height ratio of approximately 1.

4.3 Sieve, as specified in ISO 565, with an aperture size of 0,71 mm.

4.4 Balance, with a precision of 0,1 g or 0,01 g.

A balance with a precision of 0,01 g should be used for very fluffy powders such as aerosil (fumed silica).

4.5 Straight edge (Figure 3), to remove the cone of surplus powder by gently drawing it.

4.6 Stand and horizontal vibration-free base. A stand to support the funnel concentric with the cylindrical container so that the bottom of the funnel orifice is approximately 50 mm above the top of the cylindrical container when the apparatus is assembled as shown in Figure 4.

This is a free preview. Purchase the entire publication at the link below:

[Product Page](#)

-
- Looking for additional Standards? Visit Intertek Inform Infostore
 - Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation
-