



NSAI
Standards

Irish Standard
I.S. EN ISO 6874:2015

Dentistry - Polymer-based pit and fissure sealants (ISO 6874:2015)

I.S. EN ISO 6874:2015

Incorporating amendments/corrigenda/National Annexes issued since publication:

The National Standards Authority of Ireland (NSAI) produces the following categories of formal documents:

I.S. xxx: Irish Standard — national specification based on the consensus of an expert panel and subject to public consultation.

S.R. xxx: Standard Recommendation — recommendation based on the consensus of an expert panel and subject to public consultation.

SWiFT xxx: A rapidly developed recommendatory document based on the consensus of the participants of an NSAI workshop.

This document replaces/revises/consolidates the NSAI adoption of the document(s) indicated on the CEN/CENELEC cover/Foreword and the following National document(s):

NOTE: The date of any NSAI previous adoption may not match the date of its original CEN/CENELEC document.

This document is based on:

EN ISO 6874:2015

Published:

2015-09-16

This document was published under the authority of the NSAI and comes into effect on:

2015-10-05

ICS number:

11.060.10

NOTE: If blank see CEN/CENELEC cover page

NSAI
1 Swift Square,
Northwood, Santry
Dublin 9

T +353 1 807 3800
F +353 1 807 3838
E standards@nsai.ie
W NSAI.ie

Sales:
T +353 1 857 6730
F +353 1 857 6729
W standards.ie

Údarás um Chaighdeáin Náisiúnta na hÉireann

National Foreword

I.S. EN ISO 6874:2015 is the adopted Irish version of the European Document EN ISO 6874:2015, Dentistry - Polymer-based pit and fissure sealants (ISO 6874:2015)

This document does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

Compliance with this document does not of itself confer immunity from legal obligations.

In line with international standards practice the decimal point is shown as a comma (,) throughout this document.

This page is intentionally left blank

EUROPEAN STANDARD

EN ISO 6874

NORME EUROPÉENNE

EUROPÄISCHE NORM

September 2015

ICS 11.060.10

Supersedes EN ISO 6874:2005

English Version

Dentistry - Polymer-based pit and fissure sealants (ISO 6874:2015)

Médecine bucco-dentaire - Produits dentaires à base de polymères pour comblement des puits et fissures (ISO 6874:2015)

Zahnheilkunde - Versiegelungskunststoffe für Grübchen und Fissuren (ISO 6874:2015)

This European Standard was approved by CEN on 10 July 2015.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

EN ISO 6874:2015 (E)

Contents	Page
European foreword.....	3

European foreword

This document (EN ISO 6874:2015) has been prepared by Technical Committee ISO/TC 106 "Dentistry" in collaboration with Technical Committee CEN/TC 55 "Dentistry" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by March 2016, and conflicting national standards shall be withdrawn at the latest by March 2016.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 6874:2005.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 6874:2015 has been approved by CEN as EN ISO 6874:2015 without any modification.

This page is intentionally left blank

INTERNATIONAL STANDARD

**ISO
6874**

Third edition
2015-09-01

Dentistry — Polymer-based pit and fissure sealants

*Médecine bucco-dentaire — Produits dentaires à base de polymères
pour comblement des puits et fissures*



Reference number
ISO 6874:2015(E)

© ISO 2015

ISO 6874:2015(E)



COPYRIGHT PROTECTED DOCUMENT

© ISO 2015, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Contents

	Page
Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Classification	1
4 Requirements	1
4.1 Biocompatibility.....	1
4.2 Physical properties.....	1
4.2.1 Working time, Class 1 sealant.....	1
4.2.2 Setting time, Class 1 sealant.....	1
4.2.3 Depth of cure, Class 2 sealant.....	1
5 Sampling	2
6 Test methods	2
6.1 Inspection.....	2
6.2 Test conditions.....	2
6.3 Preparation of test specimens.....	2
6.4 Working time, Class 1 sealant.....	2
6.4.1 Apparatus.....	2
6.4.2 Procedure.....	4
6.4.3 Treatment of results.....	4
6.5 Setting time, Class 1 sealant.....	5
6.5.1 Apparatus.....	5
6.5.2 Procedure.....	5
6.5.3 Treatment of results.....	5
6.6 Depth of cure, Class 2 sealant.....	6
6.6.1 Apparatus.....	6
6.6.2 Procedure.....	6
6.6.3 Treatment of results.....	6
7 Packaging, marking and instructions and information to be supplied by the manufacturer	6
7.1 Packaging.....	7
7.2 Marking.....	7
7.2.1 Capsule or single dose container.....	7
7.2.2 Outer pack.....	7
7.3 Manufacturer's instructions and information for the user.....	7
Bibliography	9

ISO 6874:2015(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: [Foreword - Supplementary information](#)

The committee responsible for this document is ISO/TC 106, *Dentistry*, Subcommittee SC1, *Filling and restorative materials*.

This third edition cancels and replaces the second edition (ISO 6874:2005), of which it constitutes a minor revision.

Introduction

The efficacy of pit and fissure sealants for the prevention of dental caries is widely accepted. The polymer-based materials intended for this purpose and covered by this International Standard harden by a free-radical polymerisation reaction that is either initiated by mixing components or by external energy, e.g. visible light.

Specific qualitative and quantitative requirements for freedom from biological hazard are not included in this International Standard but, when assessing possible biological hazards, reference can be made to ISO 10993 (all parts) and ISO 7405.

Dentistry — Polymer-based pit and fissure sealants

1 Scope

This International Standard specifies requirements and test methods for polymer-based materials intended for sealing pits and fissures in teeth.

This International Standard covers both self-curing and external-energy-activated materials.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 1942, *Dentistry — Vocabulary*

ISO 8601, *Data elements and interchange formats — Information interchange — Representation of dates and times*

3 Classification

For the purposes of this International Standard, polymer-based pit and fissure sealants are classified, according to the method of curing, as follows:

Class 1: Materials whose setting is effected by mixing an initiator and activator (“self-curing” materials).

Class 2: Materials whose setting is effected by the application of energy from an external source, such as visible light (“external-energy-activated” materials).

4 Requirements

4.1 Biocompatibility

See the Introduction for guidance on biocompatibility, ISO 7405 and ISO 10993-1.

4.2 Physical properties

4.2.1 Working time, Class 1 sealant

The working time for Class 1 sealants, determined in accordance with [6.4](#), shall not be less than 40 s.

4.2.2 Setting time, Class 1 sealant

The setting time for Class 1 sealants, determined in accordance with [6.5](#), shall not be greater than 5 min.

4.2.3 Depth of cure, Class 2 sealant

The depth of cure for Class 2 sealants, determined in accordance with [6.6](#), shall be not less than 1,5 mm. If the material is supplied in more than one shade, each shade shall comply with this requirement.

This is a free preview. Purchase the entire publication at the link below:

[Product Page](#)

-
- [Looking for additional Standards? Visit Intertek Inform Infostore](#)
 - [Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation](#)
-