



NSAI
Standards

Irish Standard
I.S. EN 60601-2-29:2008

Medical electrical equipment -- Part 2
-29: Particular requirements for the
basic safety and essential performance
of radiotherapy simulators (IEC 60601-2
-29:2008 (EQV))

I.S. EN 60601-2-29:2008

Incorporating amendments/corrigenda issued since publication:

EN 60601-2-29:2008/A11:2011

The National Standards Authority of Ireland (NSAI) produces the following categories of formal documents:

I.S. xxx: Irish Standard – national specification based on the consensus of an expert panel and subject to public consultation.

S.R. xxx: Standard Recommendation - recommendation based on the consensus of an expert panel and subject to public consultation.

SWiFT xxx: A rapidly developed recommendatory document based on the consensus of the participants of an NSAI workshop.

<p><i>This document replaces:</i> EN 60601-2-29:1999</p>	<p><i>This document is based on:</i> EN 60601-2-29:2008 EN 60601-2-29:1999</p>	<p><i>Published:</i> 28 November, 2008 1 April, 1999</p>
<p>This document was published under the authority of the NSAI and comes into effect on:</p> <p>26 January, 2009</p>		<p>ICS number: 11.040.60</p>
<p>NSAI 1 Swift Square, Northwood, Santry Dublin 9</p>	<p>T +353 1 807 3800 F +353 1 807 3838 E standards@nsai.ie W NSAI.ie</p>	<p>Sales: T +353 1 857 6730 F +353 1 857 6729 W standards.ie</p>
<p>Údarás um Chaighdeáin Náisiúnta na hÉireann</p>		

EUROPEAN STANDARD

EN 60601-2-29/A11

NORME EUROPÉENNE

EUROPÄISCHE NORM

October 2011

ICS 11.040.60

English version

**Medical electrical equipment -
Part 2-29: Particular requirements for the basic safety and essential
performance of radiotherapy simulators**

Appareils électromédicaux -
Partie 2-29: Exigences particulières pour
la sécurité de base et les performances
essentielle des simulateurs de
radiothérapie

Medizinische elektrische Geräte -
Teil 2-29: Besondere Festlegungen für die
Sicherheit einschließlich der wesentlichen
Leistungsmerkmale von
Strahlentherapiesimulatoren

This amendment A11 modifies the European Standard EN 60601-2-29:2008; it was approved by CENELEC on 2011-10-01. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this amendment the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This amendment exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

CENELEC

European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

Management Centre: Avenue Marnix 17, B - 1000 Brussels

Foreword

This document (EN 60601-2-29:2008/A11:2011) has been prepared by CLC/TC 62 "Electrical equipment in medical practice".

The following dates are fixed:

- latest date by which this document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2012-10-01
- latest date by which the national standards conflicting with this document have to be withdrawn (dow) 2014-10-01

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC [and/or CEN] shall not be held responsible for identifying any or all such patent rights.

Replace Annex ZZ of EN 60601-2-29:2008 by:

Annex ZZ
(informative)

Coverage of Essential Requirements of EC Directives

This European Standard has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association and within its scope the standard covers all relevant essential requirements as given in Annex I of the EC Directive 93/42/EEC except as follows:

- Essential Requirement 6a
- Essential Requirement 7.4
- Essential Requirement 7.5 paragraph 2 & 3
- Essential Requirement 13.6 (q)

Compliance with this standard provides one means of conformity with the specified essential requirements of the Directive concerned.

WARNING: Other requirements and other EC Directives may be applicable to the products falling within the scope of this standard.

This page is intentionally left BLANK.

EUROPEAN STANDARD

EN 60601-2-29

NORME EUROPÉENNE

EUROPÄISCHE NORM

November 2008

ICS 11.040.60

Supersedes EN 60601-2-29:1999

English version

**Medical electrical equipment -
Part 2-29: Particular requirements for the basic safety and essential
performance of radiotherapy simulators
(IEC 60601-2-29:2008)**

Appareils électromédicaux -
Partie 2-29: Exigences particulières
pour la sécurité de base
et les performances essentielles
des simulateurs de radiothérapie
(CEI 60601-2-29:2008)

Medizinische elektrische Geräte -
Teil 2-29: Besondere Festlegungen
für die Sicherheit einschließlich
der wesentlichen Leistungsmerkmale
von Strahlentherapiesimulatoren
(IEC 60601-2-29:2008)

This European Standard was approved by CENELEC on 2008-11-01. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the Central Secretariat has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

CENELEC

European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

Central Secretariat: rue de Stassart 35, B - 1050 Brussels

Foreword

The text of document 62C/423/CDV, future edition 3 of IEC 60601-2-29, prepared by SC 62C, Equipment for radiotherapy, nuclear medicine and radiation dosimetry, of IEC TC 62, Electrical equipment in medical practice, was submitted to the IEC-CENELEC parallel Unique Acceptance Procedure and was approved by CENELEC as EN 60601-2-29 on 2008-11-01.

This European Standard supersedes EN 60601-2-29:1999.

EN 60601-2-29:2008 constitutes a technical revision, which brings EN 60601-2-29 in line with EN 60601-1:2006 and its collateral standards.

The following dates were fixed:

- latest date by which the EN has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2009-08-01
- latest date by which the national standards conflicting with the EN have to be withdrawn (dow) 2011-11-01

This European Standard has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association and covers essential requirements of EC Directive 93/42/EEC. See Annex ZZ.

In this standard, the following print types are used:

- Requirements and definitions: roman type.
- *Test specifications: italic type.*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes Subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this particular standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

Annexes ZA and ZZ have been added by CENELEC.

Endorsement notice

The text of the International Standard IEC 60601-2-29:2008 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following notes have to be added for the standards indicated:

IEC 60601-1-3	NOTE	Harmonized as EN 60601-1-3:2008 (not modified).
IEC 60601-1-8	NOTE	Harmonized as EN 60601-1-8:2007 (not modified).
IEC 60601-2-1	NOTE	Harmonized as EN 60601-2-1:1998 (not modified).

Annex ZA
(normative)

**Normative references to international publications
with their corresponding European publications**

Addition to Annex ZA of EN 60601-1:2006:

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC/TR 60788	2004	Medical electrical equipment - Glossary of defined terms	-	-
IEC 61217	- ¹⁾	Radiotherapy equipment - Coordinates, movements and scales	EN 61217	1996 ²⁾

¹⁾ Undated reference.

²⁾ Valid edition at date of issue.

Annex ZZ
(informative)

Coverage of Essential Requirements of EC Directives

This European Standard has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association and within its scope the standard covers all relevant essential requirements as given in Annex I of the EC Directive 93/42/EEC.

Compliance with this standard provides one means of conformity with the specified essential requirements of the Directive concerned.

WARNING: Other requirements and other EC Directives may be applicable to the products falling within the scope of this standard.

This page is intentionally left BLANK.

CONTENTS

FOREWORD.....	3
INTRODUCTION.....	5
201.1 Scope, object and related standards	6
201.2 Normative references	7
201.3 Terms and definitions.....	8
201.4 General requirements.....	8
201.5 General requirements for testing of ME EQUIPMENT.....	8
201.6 Classification of ME EQUIPMENT and ME SYSTEMS	8
201.7 ME EQUIPMENT identification, marking and documents.....	9
201.8 Protection against electrical HAZARDS from ME EQUIPMENT.....	11
201.9 Protection against mechanical HAZARDS of ME EQUIPMENT and ME SYSTEMS	11
201.10 Protection against unwanted and excessive radiation HAZARDS.....	18
201.11 Protection against excessive temperatures and other HAZARDS.....	18
201.12 Accuracy of controls and instruments and protection against hazardous outputs.....	19
201.13 HAZARDOUS SITUATIONS and fault conditions.....	19
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS)	19
201.15 Construction of ME EQUIPMENT	19
201.16 ME SYSTEMS	19
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	19
Annex C (informative) Guide to marking and labelling requirements for ME EQUIPMENT and ME SYSTEMS.....	20
Annex AA (informative) Particular guidance and rationale	21
Bibliography.....	22
Index of defined terms used in this particular standard.....	23
Figure 101 – Equipment movements and scales – Rotary GANTRY with identification of axes 1 to 8, directions 9 to 13, and dimensions 14 and 15 (see accompanying table)	14
Figure 102 – Equipment movements and scales – ISOCENTRIC RADIOTHERAPY SIMULATOR or TELERADIOTHERAPY ME EQUIPMENT, with identification of axes 1; 4 to 6; 19, of directions 9 to 12; 16 to 18 and of dimensions 14; 15 (see accompanying table).....	15
Figure 103 – Equipment movements and scales – View from RADIATION SOURCE of TELERADIOTHERAPY RADIATION FIELD or RADIOTHERAPY SIMULATOR DELINEATED RADIATION FIELD (see accompanying table)	16
Table 201.101 – Description of equipment movements	17
Table 201.C.101 – ACCOMPANYING DOCUMENTS, General	20
Table 201.C.102 – ACCOMPANYING DOCUMENTS, Instructions for use	20
Table 201.C.103 – ACCOMPANYING DOCUMENTS, technical description.....	20

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –**Part 2-29: Particular requirements for the basic safety and essential performance of radiotherapy simulators**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC provides no marking procedure to indicate its approval and cannot be rendered responsible for any equipment declared to be in conformity with an IEC Publication.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International standard IEC 60601-2-29 has been prepared by IEC subcommittee 62C: Equipment for radiotherapy, nuclear medicine and radiation dosimetry, of IEC technical committee 62: Electrical equipment in medical practice.

This third edition cancels and replaces the second edition published in 1999. This edition constitutes a technical revision, which brings this standard in line with the third edition of IEC 60601-1 and its collateral standards.

The text of this particular standard is based on the following documents:

CDV	Report on voting
62C/423/CDV	62C/434/RVC

Full information on the voting for the approval of this particular standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this standard, the following print types are used:

- Requirements and definitions: roman type.
- *Test specifications: italic type.*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this particular standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex A.

A list of all parts of the IEC 60601 series, published under the general title: *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the maintenance result date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

INTRODUCTION

This particular standard establishes requirements to be complied with by MANUFACTURERS in the design and construction of RADIOTHERAPY SIMULATORS; it does not attempt to define their optimum performance requirements. Its purpose is to identify those features of design that are regarded, at the present time, as essential for the safe operation of such ME EQUIPMENT. It places limits on the degradation of ME EQUIPMENT performance beyond which it can be presumed that a fault condition exists, for example a component failure, and where an INTERLOCK then operates to prevent continued operation of the ME EQUIPMENT.

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-29: Particular requirements for the basic safety and essential performance of radiotherapy simulators

201.1 Scope, object and related standards

Clause 1 of the general standard¹⁾ applies, except as follows:

201.1.1 Scope

Replacement:

This International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of RADIOTHERAPY SIMULATORS, hereafter referred to as ME EQUIPMENT.

If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

HAZARDS inherent in the intended physiological function of ME EQUIPMENT or ME SYSTEMS within the scope of this standard are not covered by specific requirements in this standard except in 7.2.13 and 8.4.1 of the general standard.

NOTE See also 4.2 of the general standard.

201.1.2 Object

Replacement:

The object of this particular standard is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for RADIOTHERAPY SIMULATORS [as defined in 201.3.204].

201.1.3 *Collateral standards

Addition:

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of the general standard and Clause 2 of this particular standard.

The following collateral standard does not apply:

- IEC 60601-1-10.

201.1.4 Particular standards

Replacement:

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in the general standard and collateral standards as appropriate for the particular

¹⁾ The general standard is IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*

ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

A requirement of a particular standard takes priority over the general standard.

For brevity, IEC 60601-1 is referred to in this particular standard as the general standard. Collateral standards are referred to by their document number.

The numbering of sections, clauses and subclauses of this particular standard corresponds to that of the general standard with the prefix "201" (e.g. 201.1 in this standard addresses the content of Clause 1 of the general standard) or applicable collateral standard with the prefix "20x" where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this particular standard addresses the content of Clause 4 of the 60601-1-2 collateral standard, 203.4 in this particular standard addresses the content of Clause 4 of the 60601-1-3 collateral standard, etc.). The changes to the text of the general standard are specified by the use of the following words:

"Replacement" means that the clause or subclause of the general standard or applicable collateral standard is replaced completely by the text of this particular standard.

"Addition" means that the text of this particular standard is additional to the requirements of the general standard or applicable collateral standard.

"Amendment" means that the clause or subclause of the general standard or applicable collateral standard is amended as indicated by the text of this particular standard.

Subclauses or figures which are additional to those of the general standard are numbered starting from 201.101. However due to the fact that definitions in the general standard are numbered 3.1 through 3.139, additional definitions in this standard are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses or figures which are additional to those of a collateral standard are numbered starting from 20x, where "x" is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

The term "this standard" is used to make reference to the general standard, any applicable collateral standards and this particular standard taken together.

Where there is no corresponding section, clause or subclause in this particular standard, the section, clause or subclause of the general standard or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of the general standard or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this particular standard.

201.2 Normative references

Clause 2 of the general standard applies, except as follows:

Addition:

IEC TR 60788:2004, *Medical electrical equipment – Glossary of defined terms*

IEC 61217, *Radiotherapy equipment – Coordinates, movements and scales*

This is a free preview. Purchase the entire publication at the link below:

[Product Page](#)

-
- [Looking for additional Standards? Visit Intertek Inform Infostore](#)
 - [Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation](#)
-