



**NSAI**  
Standards

Irish Standard  
I.S. EN ISO 9680:2014

## Dentistry - Operating lights (ISO 9680:2014)

## I.S. EN ISO 9680:2014

*Incorporating amendments/corrigenda/National Annexes issued since publication:*

The National Standards Authority of Ireland (NSAI) produces the following categories of formal documents:

I.S. xxx: Irish Standard — national specification based on the consensus of an expert panel and subject to public consultation.

S.R. xxx: Standard Recommendation — recommendation based on the consensus of an expert panel and subject to public consultation.

SWiFT xxx: A rapidly developed recommendatory document based on the consensus of the participants of an NSAI workshop.

*This document replaces/revises/consolidates the NSAI adoption of the document(s) indicated on the CEN/CENELEC cover/Foreword and the following National document(s):*

*NOTE: The date of any NSAI previous adoption may not match the date of its original CEN/CENELEC document.*

*This document is based on:*

EN ISO 9680:2014

*Published:*

2014-10-29

*This document was published under the authority of the NSAI and comes into effect on:*

2014-11-26

ICS number:

11.060.20

NOTE: If blank see CEN/CENELEC cover page

NSAI  
1 Swift Square,  
Northwood, Santry  
Dublin 9

T +353 1 807 3800  
F +353 1 807 3838  
E standards@nsai.ie  
W NSAI.ie

Sales:  
T +353 1 857 6730  
F +353 1 857 6729  
W standards.ie

Údarás um Chaighdeáin Náisiúnta na hÉireann

EUROPEAN STANDARD

EN ISO 9680

NORME EUROPÉENNE

EUROPÄISCHE NORM

October 2014

ICS 11.060.20

Supersedes EN ISO 9680:2007

English Version

## Dentistry - Operating lights (ISO 9680:2014)

Médecine bucco-dentaire - Appareils d'éclairage (ISO 9680:2014)

Zahnheilkunde - Behandlungsleuchten (ISO 9680:2014)

This European Standard was approved by CEN on 9 August 2014.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels**

**Contents**

Page

**Foreword.....3**

## **Foreword**

This document (EN ISO 9680:2014) has been prepared by Technical Committee ISO/TC 106 "Dentistry" in collaboration with Technical Committee CEN/TC 55 "Dentistry" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by April 2015, and conflicting national standards shall be withdrawn at the latest by April 2015.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 9680:2007.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

### **Endorsement notice**

The text of ISO 9680:2014 has been approved by CEN as EN ISO 9680:2014 without any modification.

This page is intentionally left blank

# INTERNATIONAL STANDARD

**ISO  
9680**

Third edition  
2014-10-15

---

---

## **Dentistry — Operating lights**

*Médecine bucco-dentaire — Appareils d'éclairage*



Reference number  
ISO 9680:2014(E)

© ISO 2014

**ISO 9680:2014(E)**



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2014

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.org](mailto:copyright@iso.org)  
Web [www.iso.org](http://www.iso.org)

Published in Switzerland



# Contents

	Page
<b>Foreword</b> .....	<b>iv</b>
<b>Introduction</b> .....	<b>vi</b>
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>1</b>
<b>3 Terms and definitions</b> .....	<b>1</b>
<b>4 Classification</b> .....	<b>2</b>
4.1 According to type of protection against electric shock.....	2
4.2 According to mode of operation.....	2
<b>5 Requirements and recommendations</b> .....	<b>2</b>
5.1 General requirements.....	2
5.2 Optical requirements.....	3
5.3 Mechanical requirements.....	6
5.4 Cleaning and disinfection.....	7
5.5 Electrical requirements.....	7
5.6 Usability.....	7
<b>6 Sampling</b> .....	<b>7</b>
<b>7 Testing</b> .....	<b>7</b>
7.1 General.....	7
7.2 Visual inspection.....	7
7.3 Optical tests.....	8
7.4 Mechanical tests.....	11
7.5 Cleaning and disinfection.....	11
<b>8 Manufacturer's instructions</b> .....	<b>11</b>
8.1 Documents.....	11
8.2 General.....	12
8.3 Instructions for use.....	12
8.4 Technical description.....	12
8.5 Check.....	12
<b>9 Packaging</b> .....	<b>12</b>
<b>10 Marking</b> .....	<b>13</b>
10.1 Marking on the outside of mains-operated operating lights.....	13
10.2 Marking on the inside of operating lights.....	13
10.3 Graphical symbols.....	13
10.4 Colours of the insulation of conductors.....	13
10.5 Indicator lights and push-buttons.....	13
<b>Annex A (informative) Transformation formulas</b> .....	<b>14</b>
<b>Annex B (normative) Normalized absorbance of camphorquinone</b> .....	<b>15</b>
<b>Bibliography</b> .....	<b>17</b>

## ISO 9680:2014(E)

### Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2. [www.iso.org/directives](http://www.iso.org/directives)

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received. [www.iso.org/patents](http://www.iso.org/patents)

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT), see the following URL: [Foreword - Supplementary information](#)

The committee responsible for this document is ISO/TC 106, *Dentistry*, Subcommittee SC 6, *Dental equipment*.

This third edition cancels and replaces the second edition (ISO 9680:2007), which has been technically revised via the following changes:

- a) The scope was expanded to consider any light source technology, including light emitting diodes (LEDs);
- b) Normative references have been updated;
- c) The requirement for the adjustable level of illuminance has been changed to eliminate an upper limit on the minimum illuminance level and to reduce the lower limit on maximum illuminance to 15 000 lx;
- d) The requirement for the Illumination pattern has been revised to specify a minimum size and shape of the outer area of illumination, area B;
- e) A requirement has been added to measure, plot and report the 10 %, 50 % and 75 % of maximum illuminance isolux lines;
- f) The CIE chromaticity coordinates for corner point 1 in [Table 1](#) have been changed to set this corner point within 0,02 of the Planckian locus in the CIE 1960 Uniform Chromaticity Space [i.e. (u,v) chromaticity space];
- g) The colour rendering index requirement was revised to exclude LED operating lights since current LED operating lights may not meet the requirement and an accepted method for measuring the colour rendering properties of white LEDs is not yet established;
- h) In the requirement for ultraviolet light irradiance the lower limit of the wavelength range was changed from 180 nm to 200 nm in order to reflect the measurement range of available radiometers;
- i) A requirement for compatibility with light-activated restorative materials has been added;
- j) The requirement for operating forces has been simplified;
- k) The requirement for expelled parts has been revised;

- l) References in electrical requirements were updated and simplified;
- m) A requirement on usability has been added;
- n) Requirements for test conditions have been simplified;
- o) Electrical tests have been deleted due to reference to IEC 60601-1:2005+A1:2012;
- p) Optical tests have been clarified and a test for compatibility with light-activated restorative materials has been added;
- q) The requirements for instructions for use and technical description have been revised;
- r) The requirements for marking on the outside of mains-operated operating lights have been revised;
- s) The requirement for marking of operating controls has been eliminated in favour of the broader requirement for graphical symbols;
- t) The Bibliography has been expanded.

## **ISO 9680:2014(E)**

### **Introduction**

The aim of this International Standard is to provide the dentist and his staff with means to enable them to work with optimum visual ease and comfort, i.e. a visual acuity of 90 % to 100 % according to zone, without adversely affecting their perception of colour or causing excessive fatigue.

In this International Standard, the safety of an operating light is assessed in combination with its power supply. Such power supplies may be incorporated in dental units or dental patient chairs.

Any item of equipment recommended by the manufacturer for use in conjunction with an operating light should not render the equipment unsafe nor affect its qualities adversely.

In preparing this International Standard account has been taken of IEC 60598-1.

This International Standard refers to IEC 60601-1, the basic standard on safety of medical electrical equipment, wherever relevant, by stating the respective clause numbers of IEC 60601-1.

This International Standard takes priority over IEC 60601-1 as specified in the individual Clauses of this International Standard.

Only the specifications laid down in this International Standard are applicable.

# Dentistry — Operating lights

## 1 Scope

This International Standard specifies requirements and test methods for operating lights used in the dental office and intended for illuminating the oral cavity of patients. It also contains specifications on manufacturers' instructions for use, marking and packaging.

This International Standard applies to operating lights, irrespective of the technology of the light source.

This International Standard excludes auxiliary light sources, e.g. from dental handpieces and dental headlamps and also operating lights which are specifically designed for use in oral surgery.

## 2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 1942, *Dentistry — Vocabulary*

ISO 4073, *Dentistry — Information system on the location of dental equipment in the working area of the oral health care provider*

ISO 9687, *Dentistry — Graphical symbols for dental equipment*

ISO 15223-1, *Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied — Part 1: General requirements*

ISO 14971, *Medical devices — Application of risk management to medical devices*

ISO 21530, *Dentistry — Materials used for dental equipment surfaces — Determination of resistance to chemical disinfectants*

IEC 60050-845, *International Electrotechnical Vocabulary, Lighting*

IEC 60598-1, *Luminaires — Part 1: General requirements and tests*

IEC 60601-1:2005, *Medical electrical equipment — Part 1: General requirements for basic safety and essential performance* + A1:2012

IEC 62366, *Medical devices — Application of usability engineering to medical devices*

IEC 62471:2006, *Photobiological safety of lamps and lamp systems*

IEC 80601-2-60, *Medical electrical equipment — Part 2-60: Particular requirements for basic safety and essential performance of dental equipment*

CIE 013.3, *Method of measuring and specifying colour rendering properties of light sources*

CIE S 017, *ILV: International Lighting Vocabulary*

## 3 Terms and definitions

For the purposes of this document, the terms and definitions given in CIE S 017, IEC 60050-845, IEC 60598-1, IEC 60601-1, ISO 1942, ISO 4073 and the following apply.

This is a free preview. Purchase the entire publication at the link below:

[Product Page](#)

- 
- [Looking for additional Standards? Visit Intertek Inform Infostore](#)
  - [Learn about LexConnect, All Jurisdictions, Standards referenced in Australian legislation](#)
-